

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051123\
 Data File : BP015068.D
 Acq On : 11 May 2023 17:50
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020708

Quant Time: May 11 23:22:10 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 11 23:06:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.140	152	352962	20.000	ng/u1	0.00
20) Naphthalene-d8	10.975	136	1591541	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.781	164	1059997	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.539	188	2373230	20.000	ng/u1	0.00
79) Chrysene-d12	21.616	240	2264211	20.000	ng/u1	0.00
88) Perylene-d12	24.192	264	2697886	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.440	96	66652	7.337	ng/uL	0.00
4) Pyridine-d5	3.875	84	488591	18.974	ng/u1	0.00
7) Phenol-d5	7.287	99	612873	18.701	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.457	67	364030	19.233	ng/u1	0.00
11) 2-Chlorophenol-d4	7.663	132	473729	19.660	ng/u1	0.00
15) 4-Methylphenol-d8	8.852	113	513087	19.648	ng/u1	0.00
21) Nitrobenzene-d5	9.316	128	250473	20.028	ng/u1	0.00
24) 2-Nitrophenol-d4	10.046	143	286504	20.589	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.587	165	521683	20.776	ng/u1	0.00
31) 4-Chloroaniline-d4	11.110	131	779484	20.981	ng/u1	0.00
46) Dimethylphthalate-d6	14.187	166	1631866	20.373	ng/u1	0.00
49) Acenaphthylene-d8	14.481	160	1872448	20.439	ng/u1	0.00
54) 4-Nitrophenol-d4	14.975	143	300900	19.346	ng/u1	0.00
60) Fluorene-d10	15.775	176	1413206	20.909	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.892	200	293198	19.604	ng/u1	0.00
73) Anthracene-d10	17.633	188	2230708	20.527	ng/u1	0.00
81) Pyrene-d10	19.851	212	2553830	19.447	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.021	264	2636643	19.687	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.475	88	73864	7.567	ng/uL	97
5) Pyridine	3.899	79	500913	18.704	ng/u1	99
6) Benzaldehyde	7.269	77	159463	17.703	ng/u1	94
8) Phenol	7.316	94	638253	18.993	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.552	93	527322	19.613	ng/u1	99
12) 2-Chlorophenol	7.699	128	501094	19.661	ng/u1	97
13) 2-Methylphenol	8.587	108	494707	19.048	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.669	45	650338	19.307	ng/u1	99
16) Acetophenone	8.975	105	845864	20.754	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.957	70	425587	19.910	ng/u1	97
18) 4-Methylphenol	8.916	108	564071	19.732	ng/u1	98
19) Hexachloroethane	9.234	117	205502	19.747	ng/u1	86
22) Nitrobenzene	9.357	77	597275	19.337	ng/u1	96
23) Isophorone	9.887	82	1252287	19.725	ng/u1	98
25) 2-Nitrophenol	10.075	139	308255	20.201	ng/u1	100
26) 2,4-Dimethylphenol	10.134	107	613419	20.145	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.369	93	766198	19.835	ng/u1	98
29) 2,4-Dichlorophenol	10.616	162	533250	21.024	ng/u1	97
30) Naphthalene	11.022	128	1751717	20.245	ng/u1	100
32) 4-Chloroaniline	11.134	127	810084	21.215	ng/u1	100
33) Hexachlorobutadiene	11.304	225	334244	21.865	ng/u1	98
34) Caprolactam	11.904	113	128457	13.433	ng/u1	98
35) 4-Chloro-3-methylphenol	12.245	107	577816	20.175	ng/u1	95
36) 2-Methylnaphthalene	12.622	142	1205299	20.674	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051123\
 Data File : BP015068.D
 Acq On : 11 May 2023 17:50
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 11 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD020708


Quant Time: May 11 23:22:10 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 11 23:06:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.840	142	1217950	20.803	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.987	216	676288	21.333	ng/ul	99
40) Hexachlorocyclopentadiene	12.957	237	413719	20.142	ng/ul	97
41) 2,4,6-Trichlorophenol	13.222	196	456292	20.383	ng/ul	97
42) 2,4,5-Trichlorophenol	13.292	196	504682	20.793	ng/ul	100
43) 1,1'-Biphenyl	13.616	154	1675496	20.373	ng/ul	98
44) 2-Chloronaphthalene	13.663	162	1310260	20.400	ng/ul	98
45) 2-Nitroaniline	13.863	65	349548	19.021	ng/ul	99
47) Dimethylphthalate	14.234	163	1708731	20.484	ng/ul	99
48) 2,6-Dinitrotoluene	14.357	165	340991	20.633	ng/ul	92
50) Acenaphthylene	14.504	152	2088778	20.435	ng/ul	99
51) 3-Nitroaniline	14.687	138	254365	17.093	ng/ul	98
52) Acenaphthene	14.845	153	1419726	20.336	ng/ul	98
53) 2,4-Dinitrophenol	14.892	184	197825	18.028	ng/ul	98
55) 4-Nitrophenol	14.986	109	222350	19.436	ng/ul	98
56) Dibenzofuran	15.181	168	1996616	20.562	ng/ul	97
57) 2,4-Dinitrotoluene	15.139	165	495941	20.738	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.404	232	443800	21.665	ng/ul	94
59) Diethylphthalate	15.586	149	1672104	20.097	ng/ul	100
61) Fluorene	15.828	166	1630407	20.893	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.816	204	810588	21.259	ng/ul	98
63) 4-Nitroaniline	15.845	138	227189	17.679	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.910	198	309824	20.072	ng/ul#	90
67) N-Nitrosodiphenylamine	16.033	169	1399280	20.052	ng/ul	99
68) 4-Bromophenyl-phenylether	16.716	248	526777	21.550	ng/ul	93
69) Hexachlorobenzene	16.839	284	642641	22.246	ng/ul#	91
70) Atrazine	16.980	200	534358	20.717	ng/ul	99
71) Pentachlorophenol	17.180	266	381804	21.130	ng/ul	96
72) Phenanthrene	17.580	178	2598902	20.110	ng/ul	100
74) Anthracene	17.669	178	2676066	20.478	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.587	216	698500	20.759	ng/ul	98
76) Pentachlorobenzene	15.098	250	730676	21.627	ng/ul	98
77) Carbazole	17.933	167	2385615	20.819	ng/ul	100
78) Di-n-butylphthalate	18.463	149	2895040	19.618	ng/ul	99
80) Fluoranthene	19.527	202	3106334	19.408	ng/ul	98
82) Pyrene	19.880	202	3236731	19.582	ng/ul	98
83) Butylbenzylphthalate	20.733	149	1316654	17.968	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.521	252	1073449	21.587	ng/ul	98
85) Benzo(a)anthracene	21.598	228	3139828	20.055	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.492	149	1905504	18.063	ng/ul	99
87) Chrysene	21.651	228	2998419	20.035	ng/ul	98
89) Di-n-octyl phthalate	22.474	149	3265788	17.675	ng/ul	100
90) Benzo(b)fluoranthene	23.392	252	3396107	19.857	ng/ul	99
91) Benzo(k)fluoranthene	23.445	252	3262943	19.868	ng/ul	99
93) Benzo(a)pyrene	24.074	252	2978605	19.703	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.915	276	3883150	19.838	ng/ul	96
95) Dibenzo(a,h)anthracene	26.939	278	3202004	19.928	ng/ul	98
96) Benzo(g,h,i)perylene	27.762	276	3132363	19.624	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 11 23:22:10 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 11 23:06:15 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.567 ng/uL
RT: 3.475 min Scan# 15
Delta R.T. 0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion: 88 Resp: 73864
Ion Ratio Lower Upper
88 100
43 25.5 20.4 30.6
58 66.3 55.4 83.2



#5
Pyridine
Concen: 18.704 ng/uL
RT: 3.899 min Scan# 87
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion: 79 Resp: 500913
Ion Ratio Lower Upper
79 100
52 68.4 55.4 83.2
51 32.3 25.4 38.2









#6
Benzaldehyde
Concen: 17.703 ng/ul
RT: 7.269 min Scan# 660
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion: 77 Resp: 159463
Ion Ratio Lower Upper
77 100
105 104.0 76.9 115.3
106 97.2 75.0 112.6



#8
Phenol
Concen: 18.993 ng/ul
RT: 7.316 min Scan# 668
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion: 94 Resp: 638253
Ion Ratio Lower Upper
94 100
65 25.3 20.6 31.0
66 32.7 26.1 39.1









#10
Bis(2-Chloroethyl)ether
Concen: 19.613 ng/ul
RT: 7.552 min Scan# 708
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion: 93 Resp: 527322
Ion Ratio Lower Upper
93 100
63 71.0 58.0 87.0
95 32.8 26.4 39.6



#12
2-Chlorophenol
Concen: 19.661 ng/ul
RT: 7.699 min Scan# 733
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:128 Resp: 501094
Ion Ratio Lower Upper
128 100
64 42.3 36.2 54.2
130 31.9 26.3 39.5





#13

2-Methylphenol

Concen: 19.048 ng/ul

RT: 8.587 min Scan# 884

Delta R.T. -0.000 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Instrument :

BNA_P

ClientSampleId :





SSTD020708

Tgt Ion:108 Resp: 494707

Ion Ratio Lower Upper

108 100

107 89.0 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.307 ng/ul

RT: 8.669 min Scan# 898

Delta R.T. -0.012 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50




Tgt Ion: 45 Resp: 650338


Ion Ratio Lower Upper

45 100

77 15.6 12.8 19.2

79 12.4 9.5 14.3









#16
Acetophenone
Concen: 20.754 ng/ul
RT: 8.975 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:105 Resp: 845864
Ion Ratio Lower Upper
105 100
77 72.7 59.4 89.2
51 22.6 19.2 28.8



#17
N-Nitroso-di-n-propylamine
Concen: 19.910 ng/ul
RT: 8.957 min Scan# 947
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion: 70 Resp: 425587
Ion Ratio Lower Upper
70 100
42 53.1 41.0 61.6
101 10.5 7.9 11.9
130 22.0 16.1 24.1









#18
4-Methylphenol
Concen: 19.732 ng/ul
RT: 8.916 min Scan# 941
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:108 Resp: 564071
Ion Ratio Lower Upper
108 100
107 113.2 88.5 132.7



#19
Hexachloroethane
Concen: 19.747 ng/ul
RT: 9.234 min Scan# 994
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:117 Resp: 205502
Ion Ratio Lower Upper
117 100
201 107.7 75.8 113.8
199 69.0 46.5 69.7









#22
Nitrobenzene
Concen: 19.337 ng/ul
RT: 9.357 min Scan# 1105
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion: 77 Resp: 597275
Ion Ratio Lower Upper
77 100
123 48.3 36.2 54.2
65 12.9 10.1 15.1



#23
Isophorone
Concen: 19.725 ng/ul
RT: 9.887 min Scan# 1105
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion: 82 Resp: 1252287
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.0
138 17.9 13.6 20.4









#25
2-Nitrophenol
Concen: 20.201 ng/ul
RT: 10.075 min Scan# 1138
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:139 Resp: 308255
Ion Ratio Lower Upper
139 100
65 48.3 38.4 57.6
109 27.8 22.2 33.2



#26
2,4-Dimethylphenol
Concen: 20.145 ng/ul
RT: 10.134 min Scan# 1147
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:107 Resp: 613419
Ion Ratio Lower Upper
107 100
121 51.4 41.0 61.6
122 88.4 69.5 104.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.835 ng/ul
 RT: 10.369 min Scan# 1188
 Delta R.T. -0.006 min
 Lab File: BP015068.D
 Acq: 11 May 2023 17:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020708


Tgt Ion: 93 Resp: 766198
 Ion Ratio Lower Upper
 93 100
 95 32.4 26.7 40.1
 123 11.1 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 21.024 ng/ul
 RT: 10.616 min Scan# 1229
 Delta R.T. -0.000 min
 Lab File: BP015068.D
 Acq: 11 May 2023 17:50

Tgt Ion: 162 Resp: 533250
 Ion Ratio Lower Upper
 162 100
 164 64.2 52.0 78.0
 98 33.0 28.9 43.3









#30
Naphthalene
Concen: 20.245 ng/ul
RT: 11.022 min Scan# 1299
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:128 Resp: 1751717
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 13.1 10.6 15.8



#32
4-Chloroaniline
Concen: 21.215 ng/ul
RT: 11.134 min Scan# 1317
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:127 Resp: 810084
Ion Ratio Lower Upper
127 100
129 32.1 25.8 38.8









#33
Hexachlorobutadiene
Concen: 21.865 ng/ul
RT: 11.304 min Scan# 1346
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:225 Resp: 334244
Ion Ratio Lower Upper
225 100
223 65.0 51.0 76.6
227 66.6 52.2 78.4



#34
Caprolactam
Concen: 13.433 ng/ul
RT: 11.904 min Scan# 1448
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:113 Resp: 128457
Ion Ratio Lower Upper
113 100
55 155.4 121.2 181.8
56 113.1 90.2 135.4









#35
4-Chloro-3-methylphenol
Concen: 20.175 ng/ul
RT: 12.245 min Scan# 1506
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:107 Resp: 577816
Ion Ratio Lower Upper
107 100
144 26.0 20.3 30.5
142 84.1 63.2 94.8



#36
2-Methylnaphthalene
Concen: 20.674 ng/ul
RT: 12.622 min Scan# 1570
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:142 Resp: 1205299
Ion Ratio Lower Upper
142 100
141 88.5 71.0 106.6









#37
1-Methylnaphthalene
Concen: 20.803 ng/ul
RT: 12.840 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:142 Resp: 1217950
Ion Ratio Lower Upper
142 100
141 93.1 74.6 112.0
116 3.7 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.333 ng/ul
RT: 12.987 min Scan# 1632
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:216 Resp: 676288
Ion Ratio Lower Upper
216 100
214 76.4 61.4 92.0
179 18.8 15.9 23.9
108 15.4 13.7 20.5









#40
Hexachlorocyclopentadiene
Concen: 20.142 ng/ul
RT: 12.957 min Scan# 1627
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:237 Resp: 413719
Ion Ratio Lower Upper
237 100
235 63.9 49.4 74.0
272 13.6 10.4 15.6



#41
2,4,6-Trichlorophenol
Concen: 20.383 ng/ul
RT: 13.222 min Scan# 1672
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:196 Resp: 456292
Ion Ratio Lower Upper
196 100
198 97.1 75.5 113.3
200 31.2 23.6 35.4





#42

2,4,5-Trichlorophenol

Concen: 20.793 ng/ul

RT: 13.292 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Instrument :

BNA_P

ClientSampleId :

SSTD020708





Tgt Ion:196 Resp: 504682

Ion Ratio Lower Upper

196 100

198 96.4 77.0 115.4

200 30.4 24.3 36.5



#43

1,1'-Biphenyl

Concen: 20.373 ng/ul

RT: 13.616 min Scan# 1739

Delta R.T. -0.006 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50




Tgt Ion:154 Resp: 1675496


Ion Ratio Lower Upper

154 100

153 39.2 32.7 49.1

76 13.3 10.7 16.1





#44

2-Chloronaphthalene

Concen: 20.400 ng/ul

RT: 13.663 min Scan# 1747

Delta R.T. -0.000 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Instrument :

BNA_P

ClientSampleId :

SSTD020708





Tgt Ion:162 Resp: 1310260

Ion Ratio Lower Upper

162 100

164 33.1 25.9 38.9

127 36.9 30.8 46.2



#45

2-Nitroaniline

Concen: 19.021 ng/ul

RT: 13.863 min Scan# 1781

Delta R.T. -0.006 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50




Tgt Ion: 65 Resp: 349548


Ion Ratio Lower Upper

65 100

92 72.8 59.8 89.8

138 115.1 92.3 138.5









#47
Dimethylphthalate
Concen: 20.484 ng/ul
RT: 14.234 min Scan# 1844
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:163 Resp: 1708731
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.0 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 20.633 ng/ul
RT: 14.357 min Scan# 1865
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:165 Resp: 340991
Ion Ratio Lower Upper
165 100
89 53.0 48.0 72.0
121 20.2 17.5 26.3









#50
Acenaphthylene
Concen: 20.435 ng/ul
RT: 14.504 min Scan# 1890
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:152 Resp: 2088778
Ion Ratio Lower Upper
152 100
151 20.3 16.1 24.1
153 13.3 10.3 15.5



#51
3-Nitroaniline
Concen: 17.093 ng/ul
RT: 14.687 min Scan# 1921
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:138 Resp: 254365
Ion Ratio Lower Upper
138 100
108 10.3 8.8 13.2
92 114.4 93.4 140.0









#52
Acenaphthene
Concen: 20.336 ng/ul
RT: 14.845 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:153 Resp: 1419726
Ion Ratio Lower Upper
153 100
152 48.9 38.0 57.0
154 90.0 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 18.028 ng/ul
RT: 14.892 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:184 Resp: 197825
Ion Ratio Lower Upper
184 100
63 58.6 49.7 74.5
154 59.2 47.5 71.3









#55
4-Nitrophenol
Concen: 19.436 ng/ul
RT: 14.986 min Scan# 1973
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:109 Resp: 222350
Ion Ratio Lower Upper
109 100
139 146.9 115.3 172.9
65 132.6 107.9 161.9



#56
Dibenzofuran
Concen: 20.562 ng/ul
RT: 15.181 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:168 Resp: 1996616
Ion Ratio Lower Upper
168 100
139 35.0 29.4 44.0









#57
2,4-Dinitrotoluene
Concen: 20.738 ng/ul
RT: 15.139 min Scan# 1998
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:165 Resp: 495941
Ion Ratio Lower Upper
165 100
63 39.1 33.1 49.7
182 2.4 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.665 ng/ul
RT: 15.404 min Scan# 2043
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:232 Resp: 443800
Ion Ratio Lower Upper
232 100
131 40.5 36.8 55.2
130 2.1 1.9 2.9
166 29.2 24.6 37.0









#59
Diethylphthalate
Concen: 20.097 ng/ul
RT: 15.586 min Scan# 2074
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:149 Resp: 1672104
Ion Ratio Lower Upper
149 100
177 22.1 17.5 26.3
150 12.4 9.8 14.8



#61
Fluorene
Concen: 20.893 ng/ul
RT: 15.828 min Scan# 2115
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:166 Resp: 1630407
Ion Ratio Lower Upper
166 100
165 97.0 77.8 116.6
167 13.3 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 21.259 ng/ul
RT: 15.816 min Scan# 2113
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:204 Resp: 810588
Ion Ratio Lower Upper
204 100
206 33.7 26.1 39.1
141 53.9 44.5 66.7



#63
4-Nitroaniline
Concen: 17.679 ng/ul
RT: 15.845 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:138 Resp: 227189
Ion Ratio Lower Upper
138 100
92 50.7 41.2 61.8
108 71.0 60.1 90.1









#66
4,6-Dinitro-2-methylphenol
Concen: 20.072 ng/ul
RT: 15.910 min Scan# 2129
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:198 Resp: 309824
Ion Ratio Lower Upper
198 100
182 4.2 3.7 5.5
77 20.5 21.2 31.8#



#67
N-Nitrosodiphenylamine
Concen: 20.052 ng/ul
RT: 16.033 min Scan# 2150
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:169 Resp: 1399280
Ion Ratio Lower Upper
169 100
168 67.0 53.5 80.3
167 36.8 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 21.550 ng/ul
RT: 16.716 min Scan# 2266
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:248 Resp: 526777
Ion Ratio Lower Upper
248 100
250 97.3 79.0 118.4
141 61.5 58.6 87.8



#69
Hexachlorobenzene
Concen: 22.246 ng/ul
RT: 16.839 min Scan# 2287
Delta R.T. -0.000 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:284 Resp: 642641
Ion Ratio Lower Upper
284 100
142 26.0 27.3 40.9#
249 29.1 24.6 36.8





#70

Atrazine

Concen: 20.717 ng/ul

RT: 16.980 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Instrument :

BNA_P

ClientSampleId :

SSTD020708





Tgt Ion:200 Resp: 534358

Ion Ratio Lower Upper

200 100

173 26.5 21.4 32.2

215 48.0 38.1 57.1



#71

Pentachlorophenol

Concen: 21.130 ng/ul

RT: 17.180 min Scan# 2345

Delta R.T. -0.006 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50




Tgt Ion:266 Resp: 381804


Ion Ratio Lower Upper

266 100

264 66.9 51.0 76.4

268 68.5 51.8 77.6





#72

Phenanthrene

Concen: 20.110 ng/ul

RT: 17.580 min Scan# 2413

Delta R.T. -0.000 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Instrument :

BNA_P

ClientSampleId :

SSTD020708





Tgt Ion:178 Resp: 2598902

Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.2

176 19.5 15.6 23.4



#74

Anthracene

Concen: 20.478 ng/ul

RT: 17.669 min Scan# 2428

Delta R.T. -0.006 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50




Tgt Ion:178 Resp: 2676066


Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 18.9 15.2 22.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.759 ng/uL

RT: 13.587 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Instrument :

BNA_P

ClientSampleId :

SSTD020708





Tgt Ion:216 Resp: 698500

Ion Ratio Lower Upper

216 100

214 79.2 62.2 93.4

218 48.7 38.2 57.2



#76

Pentachlorobenzene

Concen: 21.627 ng/uL

RT: 15.098 min Scan# 1991

Delta R.T. -0.006 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50




Tgt Ion:250 Resp: 730676


Ion Ratio Lower Upper

250 100

248 58.7 49.8 74.6

252 64.4 51.6 77.4









#77
 Carbazole
 Concen: 20.819 ng/ul
 RT: 17.933 min Scan# 2474
 Delta R.T. -0.006 min
 Lab File: BP015068.D
 Acq: 11 May 2023 17:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020708


Tgt Ion:167 Resp: 2385615
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.2
 139 12.6 10.2 15.4



#78
 Di-n-butylphthalate
 Concen: 19.618 ng/ul
 RT: 18.463 min Scan# 2563
 Delta R.T. -0.006 min
 Lab File: BP015068.D
 Acq: 11 May 2023 17:50

Tgt Ion:149 Resp: 2895040
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.3 10.9
 104 4.7 3.9 5.9





#80

Fluoranthene

Concen: 19.408 ng/ul

RT: 19.527 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Instrument :

BNA_P

ClientSampleId :

SSTD020708





Tgt Ion:202 Resp: 3106334

Ion Ratio Lower Upper

202 100

101 11.2 9.5 14.3

100 8.3 7.3 10.9



#82

Pyrene

Concen: 19.582 ng/ul

RT: 19.880 min Scan# 2804

Delta R.T. -0.006 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50




Tgt Ion:202 Resp: 3236731


Ion Ratio Lower Upper

202 100

101 13.1 11.2 16.8

100 10.5 9.0 13.6





#83

Butylbenzylphthalate

Concen: 17.968 ng/ul

RT: 20.733 min Scan# 2950

Delta R.T. -0.006 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Instrument :

BNA_P

ClientSampleId :

SSTD020708





Tgt Ion:149 Resp: 1316654

Ion Ratio Lower Upper

149 100

91 69.7 57.4 86.2

206 21.3 16.6 25.0



#84

3,3'-Dichlorobenzidine

Concen: 21.587 ng/ul

RT: 21.521 min Scan# 3083

Delta R.T. -0.000 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50




Tgt Ion:252 Resp: 1073449


Ion Ratio Lower Upper

252 100

254 64.2 50.8 76.2

126 9.9 9.8 14.8









#85
Benzo(a)anthracene
Concen: 20.055 ng/ul
RT: 21.598 min Scan# 3097
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:228 Resp: 3139828
Ion Ratio Lower Upper
228 100
229 19.8 16.1 24.1
226 27.2 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 18.063 ng/ul
RT: 21.492 min Scan# 3078
Delta R.T. -0.006 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:149 Resp: 1905504
Ion Ratio Lower Upper
149 100
167 27.7 21.8 32.6
279 4.4 3.2 4.8





#87

Chrysene

Concen: 20.035 ng/ul

RT: 21.651 min Scan# 3106

Delta R.T. -0.006 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Instrument :

BNA_P

ClientSampleId :

SSTD020708





Tgt Ion:228 Resp: 2998419

Ion Ratio Lower Upper

228 100

226 29.9 25.0 37.6

229 19.7 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 17.675 ng/ul




RT: 22.474 min Scan# 3245


Delta R.T. -0.006 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Tgt Ion:149 Resp: 3265788





#90

Benzo(b)fluoranthene

Concen: 19.857 ng/ul

RT: 23.392 min Scan# 3401

Delta R.T. -0.012 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Instrument :

BNA_P

ClientSampleId :

SSTD020708





Tgt Ion:252 Resp: 3396107

Ion Ratio Lower Upper

252 100

253 22.3 17.4 26.2

125 9.4 8.3 12.5



#91

Benzo(k)fluoranthene

Concen: 19.868 ng/ul

RT: 23.445 min Scan# 3410

Delta R.T. -0.006 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50




Tgt Ion:252 Resp: 3262943


Ion Ratio Lower Upper

252 100

253 22.0 17.5 26.3

125 9.6 8.5 12.7









#93
Benzo(a)pyrene
Concen: 19.703 ng/ul
RT: 24.074 min Scan# 31
Delta R.T. -0.012 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50




Instrument :
BNA_P
ClientSampleId :
SSTD020708


Tgt Ion:252 Resp: 2978605
Ion Ratio Lower Upper
252 100
253 21.4 17.2 25.8
125 10.1 9.1 13.7



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.838 ng/ul
RT: 26.915 min Scan# 4000
Delta R.T. -0.018 min
Lab File: BP015068.D
Acq: 11 May 2023 17:50

Tgt Ion:276 Resp: 3883150
Ion Ratio Lower Upper
276 100
138 21.0 20.1 30.1
277 25.1 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 19.928 ng/ul

RT: 26.939 min Scan# 4006

Delta R.T. -0.012 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Instrument :

BNA_P

ClientSampleId :

SSTD020708





Tgt Ion:278 Resp: 3202004

Ion Ratio Lower Upper

278 100

139 14.7 13.3 19.9

279 24.1 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 19.624 ng/ul

RT: 27.762 min Scan# 4144

Delta R.T. -0.018 min

Lab File: BP015068.D

Acq: 11 May 2023 17:50

Tgt Ion:276 Resp: 3132363

Ion Ratio Lower Upper

276 100

138 19.6 17.7 26.5

277 24.0 18.8 28.2

