


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051221\
 Data File : BP005501.D
 Acq On : 12 May 2021 13:19
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue May 11 11:20:59 2021






AutoFind: Scans 2457, 2458, 2459; Background Corrected with Scan 2448

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.6	8506	PASS
68	69	0.00	2	1.9	163	PASS
69	198	0.00	100	37.5	8713	PASS
70	69	0.00	2	0.3	25	PASS
127	198	10	80	34.6	8024	PASS
197	198	0.00	2	0.4	87	PASS
198	198	100	100	100.0	23219	PASS
199	198	5	9	6.9	1612	PASS
275	198	10	60	36.6	8490	PASS
365	198	1	100	7.3	1687	PASS
441	198	0.01	100	13.3	3089	PASS
442	442	100	100	100.0	20531	PASS
443	442	15	24	21.1	4330	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051221\
Data File : BP005501.D
Acq On : 12 May 2021 13:19
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: May 12 13:58:39 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 11 11:20:59 2021
Response via : Initial Calibration



TIC: BP005501.D\data.ms

(77) Benzidine

19.316min (+ 0.006) 251745.45 ng




response 109108

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.90	13.73
183.00	12.30	11.76
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051221\
 Data File : BP005501.D
 Acq On : 12 May 2021 13:19
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: May 12 13:58:39 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 11 11:20:59 2021
 Response via : Initial Calibration



TIC: BP005501.D\data.ms

(70) Pentachlorophenol (C)

16.763min (+ 0.006) 188967.87 ng

response 23818

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.70	56.61
264.00	67.10	55.37
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/12/2021	BNA_P	<u>BP005501.D</u>
Compound Name	Response	Retention Time
DDT	90735	20.5
DDD	1660	20.08
DDE	374	19.58
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
2034	92769	2.19