

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051322\
 Data File : BP010346.D
 Acq On : 13 May 2022 11:01
 Operator : CG/JU
 Sample : SSTD00538
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005638

Quant Time: May 13 13:34:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 13:33:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.905	152	116528	20.000	ng/ul	0.00
20) Naphthalene-d8	10.704	136	506902	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.534	164	366628	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.286	188	830938	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.369	240	891060	20.000	ng/ul	0.00
88) Perylene-d12	23.798	264	813513	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	4911	1.790	ng/uL	0.01
4) Pyridine-d5	3.775	84	34219	4.483	ng/ul	0.01
7) Phenol-d5	7.069	99	40133	3.708	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.158	67	117	0.017	ng/ul	-0.07
11) 2-Chlorophenol-d4	7.434	132	32325	4.055	ng/ul	0.00
15) 4-Methylphenol-d8	8.610	113	33906	3.811	ng/ul	0.00
21) Nitrobenzene-d5	9.063	128	16962	4.173	ng/ul	0.00
24) 2-Nitrophenol-d4	9.787	143	16815	3.629	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	34478	4.088	ng/ul	0.00
31) 4-Chloroaniline-d4	10.846	131	50414	4.049	ng/ul	0.00
46) Dimethylphthalate-d6	13.945	166	127419	4.362	ng/ul	0.00
49) Acenaphthylene-d8	14.228	160	140321	4.029	ng/ul	0.00
54) 4-Nitrophenol-d4	14.734	143	13082	2.335	ng/ul	0.00
60) Fluorene-d10	15.528	176	112143	4.548	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	13770	2.420	ng/ul	0.00
73) Anthracene-d10	17.386	188	171906	4.225	ng/ul	0.00
81) Pyrene-d10	19.616	212	216190	4.390	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.639	264	176706	4.073	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.387	88	4253	1.517	ng/ul#	21
5) Pyridine	3.799	79	35178	4.397	ng/ul#	26
6) Benzaldehyde	7.046	77	27363	4.575	ng/ul#	84
8) Phenol	7.099	94	45681	4.187	ng/ul#	89
10) Bis(2-Chloroethyl)ether	7.328	93	37856	4.434	ng/ul#	76
12) 2-Chlorophenol	7.469	128	35202	4.397	ng/ul	95
13) 2-Methylphenol	8.352	108	33550	4.051	ng/ul	95
14) 2,2'-oxybis(1-Chloropr...	8.434	45	67052	4.992	ng/ul#	79
16) Acetophenone	8.728	105	60103	4.321	ng/ul#	69
17) N-Nitroso-di-n-propyla...	8.710	70	31930	4.395	ng/ul#	69
18) 4-Methylphenol	8.675	108	36160	3.991	ng/ul	95
19) Hexachloroethane	8.987	117	16551	4.849	ng/ul#	81
22) Nitrobenzene	9.105	77	47460	4.600	ng/ul#	79
23) Isophorone	9.634	82	84433	4.179	ng/ul	98
25) 2-Nitrophenol	9.822	139	19221	4.038	ng/ul#	85
26) 2,4-Dimethylphenol	9.881	107	44848	4.412	ng/ul#	79
27) Bis(2-Chloroethoxy)met...	10.116	93	51334	4.329	ng/ul#	96
29) 2,4-Dichlorophenol	10.357	162	36220	4.396	ng/ul#	85
30) Naphthalene	10.757	128	124309	4.596	ng/ul	96
32) 4-Chloroaniline	10.869	127	49715	4.001	ng/ul	95
33) Hexachlorobutadiene	11.046	225	27599	4.891	ng/ul	95
34) Caprolactam	11.646	113	10292	3.279	ng/ul#	1
35) 4-Chloro-3-methylphenol	11.993	107	40455	4.080	ng/ul#	72
36) 2-Methylnaphthalene	12.363	142	86649	4.525	ng/ul	92

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051322\
 Data File : BP010346.D
 Acq On : 13 May 2022 11:01
 Operator : CG/JU
 Sample : SST00538
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005638

Quant Time: May 13 13:34:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 13:33:22 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.581	142	89925	4.619	ng/ul#	92
39) 1,2,4,5-Tetrachloroben...	12.734	216	54219	4.811	ng/ul#	94
40) Hexachlorocyclopentadiene	12.716	237	17914	2.851	ng/ul	92
41) 2,4,6-Trichlorophenol	12.975	196	29644	4.011	ng/ul#	77
42) 2,4,5-Trichlorophenol	13.051	196	32632	4.051	ng/ul#	92
43) 1,1'-Biphenyl	13.369	154	128070	4.676	ng/ul#	93
44) 2-Chloronaphthalene	13.410	162	98060	4.623	ng/ul	95
45) 2-Nitroaniline	13.622	65	27651	3.850	ng/ul#	68
47) Dimethylphthalate	13.993	163	127104	4.485	ng/ul#	92
48) 2,6-Dinitrotoluene	14.110	165	22645	3.916	ng/ul	94
50) Acenaphthylene	14.257	152	156262	4.486	ng/ul	95
51) 3-Nitroaniline	14.440	138	19657	3.886	ng/ul#	84
52) Acenaphthene	14.598	153	105885	4.695	ng/ul	96
53) 2,4-Dinitrophenol	14.651	184	6317	1.670	ng/ul#	75
55) 4-Nitrophenol	14.751	109	14564	3.065	ng/ul#	37
56) Dibenzofuran	14.934	168	158386	4.800	ng/ul	98
57) 2,4-Dinitrotoluene	14.898	165	32276	3.815	ng/ul#	66
58) 2,3,4,6-Tetrachlorophenol	15.163	232	28484	3.984	ng/ul#	87
59) Diethylphthalate	15.351	149	128369	4.388	ng/ul#	90
61) Fluorene	15.581	166	126714	4.636	ng/ul	94
62) 4-Chlorophenyl-phenyle...	15.575	204	66832	4.730	ng/ul	91
63) 4-Nitroaniline	15.604	138	17808	3.562	ng/ul#	81
66) 4,6-Dinitro-2-methylph...	15.669	198	13186	2.379	ng/ul#	82
67) N-Nitrosodiphenylamine	15.792	169	107471	4.400	ng/ul	94
68) 4-Bromophenyl-phenylether	16.475	248	40086	4.567	ng/ul	98
69) Hexachlorobenzene	16.598	284	46835	4.613	ng/ul#	88
70) Atrazine	16.745	200	38601	3.912	ng/ul#	90
71) Pentachlorophenol	16.945	266	20361	3.022	ng/ul#	77
72) Phenanthrene	17.328	178	210935	4.604	ng/ul	98
74) Anthracene	17.422	178	209501	4.509	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.340	216	65854	5.417	ng/ul	90
76) Pentachlorobenzene	14.857	250	54021	4.647	ng/ul	92
77) Carbazole	17.816	167	250	0.006	ng/ul	94
78) Di-n-butylphthalate	18.316	149	389	0.007	ng/ul	98
80) Fluoranthene	19.292	202	248225	4.434	ng/ul	96
82) Pyrene	19.645	202	265646	4.584	ng/ul#	93
83) Butylbenzylphthalate	20.516	149	93114	3.773	ng/ul#	79
84) 3,3'-Dichlorobenzidine	21.286	252	72904	3.828	ng/ul#	97
85) Benzo(a)anthracene	21.351	228	258233	4.433	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.274	149	135735	3.645	ng/ul#	83
87) Chrysene	21.404	228	253357	4.389	ng/ul	98
89) Di-n-octyl phthalate	22.204	149	238654	4.358	ng/ul	100
90) Benzo(b)fluoranthene	23.051	252	235314	4.404	ng/ul	98
91) Benzo(k)fluoranthene	23.104	252	227270	4.508	ng/ul#	97
93) Benzo(a)pyrene	23.686	252	226064	4.368	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.321	276	209741	3.666	ng/ul#	92
95) Dibenzo(a,h)anthracene	26.498	278	280	0.006	ng/ul	94
96) Benzo(g,h,i)perylene	27.204	276	251	0.005	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051322\
 Data File : BP010346.D
 Acq On : 13 May 2022 11:01
 Operator : CG/JU
 Sample : SSTD00538
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005638

Quant Time: May 13 13:34:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 13:33:22 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.517 ng/uL
RT: 3.387 min Scan# 51
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion: 88 Resp: 4253
Ion Ratio Lower Upper
88 100
43 39.1 6.8 10.2#
58 80.7 28.4 42.6#



#5
Pyridine
Concen: 4.397 ng/uL
RT: 3.799 min Scan# 121
Delta R.T. 0.018 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion: 79 Resp: 35178
Ion Ratio Lower Upper
79 100
52 99.3 35.5 53.3#
51 53.4 20.2 30.4#









#6
Benzaldehyde
Concen: 4.575 ng/ul
RT: 7.046 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion: 77 Resp: 27363
Ion Ratio Lower Upper
77 100
105 83.8 77.0 115.4
106 76.7 77.0 115.4#



#8
Phenol
Concen: 4.187 ng/ul
RT: 7.099 min Scan# 682
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion: 94 Resp: 45681
Ion Ratio Lower Upper
94 100
65 33.1 19.0 28.6#
66 47.6 34.6 51.8









#10
Bis(2-Chloroethyl)ether
Concen: 4.434 ng/ul
RT: 7.328 min Scan# 721
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion: 93 Resp: 37856
Ion Ratio Lower Upper
93 100
63 82.4 45.0 67.4#
95 30.7 26.3 39.5



#12
2-Chlorophenol
Concen: 4.397 ng/ul
RT: 7.469 min Scan# 745
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:128 Resp: 35202
Ion Ratio Lower Upper
128 100
64 58.9 45.3 67.9
130 31.8 29.0 43.6





#13

2-Methylphenol

Concen: 4.051 ng/ul

RT: 8.352 min Scan# 894

Delta R.T. 0.006 min

Lab File: BP010346.D

Acq: 13 May 2022 11:01

Instrument :

BNA_P

ClientSampleId :





SSTD005638

Tgt Ion:108 Resp: 33550

Ion Ratio Lower Upper

108 100

107 93.1 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 4.992 ng/ul

RT: 8.434 min Scan# 909

Delta R.T. 0.006 min

Lab File: BP010346.D

Acq: 13 May 2022 11:01




Tgt Ion: 45 Resp: 67052


Ion Ratio Lower Upper

45 100

77 9.8 17.1 25.7#

79 9.2 13.2 19.8#









#16
Acetophenone
Concen: 4.321 ng/ul
RT: 8.728 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:105 Resp: 60103
Ion Ratio Lower Upper
105 100
77 88.0 53.4 80.2#
51 38.4 13.8 20.8#



#17
N-Nitroso-di-n-propylamine
Concen: 4.395 ng/ul
RT: 8.710 min Scan# 956
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion: 70 Resp: 31930
Ion Ratio Lower Upper
70 100
42 68.4 31.8 47.8#
101 8.0 15.6 23.4#
130 16.1 13.7 20.5









#18
4-Methylphenol
Concen: 3.991 ng/ul
RT: 8.675 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:108 Resp: 36160
Ion Ratio Lower Upper
108 100
107 109.7 83.6 125.4



#19
Hexachloroethane
Concen: 4.849 ng/ul
RT: 8.987 min Scan# 1003
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:117 Resp: 16551
Ion Ratio Lower Upper
117 100
201 106.3 106.5 159.7#
199 64.5 59.0 88.6









#22
Nitrobenzene
Concen: 4.600 ng/ul
RT: 9.105 min Scan# 1022
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion: 77 Resp: 47460
Ion Ratio Lower Upper
77 100
123 38.1 42.3 63.5#
65 14.8 5.8 8.8#



#23
Isophorone
Concen: 4.179 ng/ul
RT: 9.634 min Scan# 1113
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion: 82 Resp: 84433
Ion Ratio Lower Upper
82 100
95 8.4 7.8 11.8
138 16.8 12.9 19.3









#25
2-Nitrophenol
Concen: 4.038 ng/ul
RT: 9.822 min Scan# 1145
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:139 Resp: 19221
Ion Ratio Lower Upper
139 100
65 64.8 49.4 74.2
109 35.8 44.7 67.1#



#26
2,4-Dimethylphenol
Concen: 4.412 ng/ul
RT: 9.881 min Scan# 1155
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:107 Resp: 44848
Ion Ratio Lower Upper
107 100
121 47.2 34.1 51.1
122 80.4 46.3 69.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 4.329 ng/ul
 RT: 10.116 min Scan# 1194
 Delta R.T. 0.006 min
 Lab File: BP010346.D
 Acq: 13 May 2022 11:01




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005638


Tgt Ion: 93 Resp: 51334
 Ion Ratio Lower Upper
 93 100
 95 34.9 26.2 39.4
 123 9.4 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 4.396 ng/ul
 RT: 10.357 min Scan# 1236
 Delta R.T. 0.006 min
 Lab File: BP010346.D
 Acq: 13 May 2022 11:01

Tgt Ion: 162 Resp: 36220
 Ion Ratio Lower Upper
 162 100
 164 63.2 37.0 55.6#
 98 39.8 31.0 46.6









#30
Naphthalene
Concen: 4.596 ng/ul
RT: 10.757 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:128 Resp: 124309
Ion Ratio Lower Upper
128 100
129 10.6 8.6 13.0
127 12.6 12.2 18.4



#32
4-Chloroaniline
Concen: 4.001 ng/ul
RT: 10.869 min Scan# 1323
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:127 Resp: 49715
Ion Ratio Lower Upper
127 100
129 33.4 24.6 37.0









#33
Hexachlorobutadiene
Concen: 4.891 ng/ul
RT: 11.046 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:225 Resp: 27599
Ion Ratio Lower Upper
225 100
223 65.9 53.4 80.0
227 59.7 53.8 80.8



#34
Caprolactam
Concen: 3.279 ng/ul
RT: 11.646 min Scan# 1455
Delta R.T. 0.012 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:113 Resp: 10292
Ion Ratio Lower Upper
113 100
55 193.7 58.2 87.2#
56 162.9 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 4.080 ng/ul
RT: 11.993 min Scan# 1514
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:107 Resp: 40455
Ion Ratio Lower Upper
107 100
144 26.9 15.8 23.8#
142 71.1 38.9 58.3#



#36
2-Methylnaphthalene
Concen: 4.525 ng/ul
RT: 12.363 min Scan# 1577
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:142 Resp: 86649
Ion Ratio Lower Upper
142 100
141 87.6 64.3 96.5









#37
1-Methylnaphthalene
Concen: 4.619 ng/ul
RT: 12.581 min Scan# 1614
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:142 Resp: 89925
Ion Ratio Lower Upper
142 100
141 95.3 69.8 104.8
116 4.9 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.811 ng/ul
RT: 12.734 min Scan# 1640
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:216 Resp: 54219
Ion Ratio Lower Upper
216 100
214 77.5 58.2 87.2
179 22.4 14.2 21.2#
108 18.3 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 2.851 ng/ul
RT: 12.716 min Scan# 1636
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:237 Resp: 17914
Ion Ratio Lower Upper
237 100
235 59.4 53.2 79.8
272 12.1 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 4.011 ng/ul
RT: 12.975 min Scan# 1681
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:196 Resp: 29644
Ion Ratio Lower Upper
196 100
198 103.9 64.1 96.1#
200 30.2 19.9 29.9#









#42
2,4,5-Trichlorophenol
Concen: 4.051 ng/ul
RT: 13.051 min Scan# 1694
Delta R.T. 0.012 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:196 Resp: 32632
Ion Ratio Lower Upper
196 100
198 86.8 68.6 103.0
200 27.5 32.8 49.2#



#43
1,1'-Biphenyl
Concen: 4.676 ng/ul
RT: 13.369 min Scan# 1748
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:154 Resp: 128070
Ion Ratio Lower Upper
154 100
153 40.3 28.6 43.0
76 12.8 8.5 12.7#









#44
2-Chloronaphthalene
Concen: 4.623 ng/ul
RT: 13.410 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:	162	Resp:	98060
Ion	Ratio	Lower	Upper
162	100		
164	32.5	27.5	41.3
127	42.0	37.1	55.7



#45
2-Nitroaniline
Concen: 3.850 ng/ul
RT: 13.622 min Scan# 1791
Delta R.T. 0.012 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:	65	Resp:	27651
Ion	Ratio	Lower	Upper
65	100		
92	56.2	77.9	116.9#
138	79.3	81.5	122.3#









#47
Dimethylphthalate
Concen: 4.485 ng/ul
RT: 13.993 min Scan# 1854
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:163 Resp: 127104
Ion Ratio Lower Upper
163 100
194 4.7 4.1 6.1
164 10.1 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 3.916 ng/ul
RT: 14.110 min Scan# 1874
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:165 Resp: 22645
Ion Ratio Lower Upper
165 100
89 69.9 51.7 77.5
121 25.2 18.3 27.5









#50
Acenaphthylene
Concen: 4.486 ng/ul
RT: 14.257 min Scan# 1898
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:152 Resp: 156262
Ion Ratio Lower Upper
152 100
151 19.7 18.3 27.5
153 13.4 11.0 16.6



#51
3-Nitroaniline
Concen: 3.886 ng/ul
RT: 14.440 min Scan# 1930
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:138 Resp: 19657
Ion Ratio Lower Upper
138 100
108 9.7 15.8 23.8#
92 127.1 117.1 175.7









#52
Acenaphthene
Concen: 4.695 ng/ul
RT: 14.598 min Scan# 1957
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:153 Resp: 105885
Ion Ratio Lower Upper
153 100
152 48.8 34.3 51.5
154 85.7 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 1.670 ng/ul
RT: 14.651 min Scan# 1966
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:184 Resp: 6317
Ion Ratio Lower Upper
184 100
63 68.9 34.6 52.0#
154 59.2 57.7 86.5









#55
4-Nitrophenol
Concen: 3.065 ng/ul
RT: 14.751 min Scan# 1981
Delta R.T. 0.012 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:109 Resp: 14564
Ion Ratio Lower Upper
109 100
139 122.2 61.4 92.2#
65 146.2 65.3 97.9#



#56
Dibenzofuran
Concen: 4.800 ng/ul
RT: 14.934 min Scan# 2014
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:168 Resp: 158386
Ion Ratio Lower Upper
168 100
139 38.9 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 3.815 ng/ul
RT: 14.898 min Scan# 2007
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:165 Resp: 32276
Ion Ratio Lower Upper
165 100
63 59.0 29.5 44.3#
182 2.6 2.8 4.2#



#58
2,3,4,6-Tetrachlorophenol
Concen: 3.984 ng/ul
RT: 15.163 min Scan# 2053
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:232 Resp: 28484
Ion Ratio Lower Upper
232 100
131 47.2 25.4 38.2#
130 2.4 4.2 6.2#
166 35.0 28.1 42.1









#59
Diethylphthalate
Concen: 4.388 ng/ul
RT: 15.351 min Scan# 2085
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:149 Resp: 128369
Ion Ratio Lower Upper
149 100
177 20.1 22.0 33.0#
150 13.3 10.2 15.2



#61
Fluorene
Concen: 4.636 ng/ul
RT: 15.581 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:166 Resp: 126714
Ion Ratio Lower Upper
166 100
165 93.4 70.3 105.5
167 14.0 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 4.730 ng/ul
RT: 15.575 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:	204	Resp:	66832
Ion Ratio	Lower	Upper	
204	100		
206	31.1	27.6	41.4
141	65.7	46.6	69.8



#63
4-Nitroaniline
Concen: 3.562 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:	138	Resp:	17808
Ion Ratio	Lower	Upper	
138	100		
92	57.9	41.6	62.4
108	94.9	97.6	146.4









#66
4,6-Dinitro-2-methylphenol
Concen: 2.379 ng/ul
RT: 15.669 min Scan# 2138
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:198 Resp: 13186
Ion Ratio Lower Upper
198 100
182 8.1 6.0 9.0
77 33.6 17.7 26.5#



#67
N-Nitrosodiphenylamine
Concen: 4.400 ng/ul
RT: 15.792 min Scan# 2160
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:169 Resp: 107471
Ion Ratio Lower Upper
169 100
168 66.2 48.3 72.5
167 35.4 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 4.567 ng/ul
RT: 16.475 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:248 Resp: 40086
Ion Ratio Lower Upper
248 100
250 98.5 79.4 119.0
141 87.2 67.2 100.8



#69
Hexachlorobenzene
Concen: 4.613 ng/ul
RT: 16.598 min Scan# 2297
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:284 Resp: 46835
Ion Ratio Lower Upper
284 100
142 34.2 19.2 28.8#
249 31.3 27.4 41.2





#70

Atrazine

Concen: 3.912 ng/ul

RT: 16.745 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP010346.D

Acq: 13 May 2022 11:01

Instrument :

BNA_P

ClientSampleId :

SSTD005638





Tgt Ion:200 Resp: 38601

Ion Ratio Lower Upper

200 100

173 26.0 19.2 28.8

215 49.2 32.2 48.4#



#71

Pentachlorophenol

Concen: 3.022 ng/ul

RT: 16.945 min Scan# 2356

Delta R.T. 0.012 min

Lab File: BP010346.D

Acq: 13 May 2022 11:01




Tgt Ion:266 Resp: 20361


Ion Ratio Lower Upper

266 100

264 57.9 57.3 85.9

268 56.8 65.7 98.5#





#72

Phenanthrene

Concen: 4.604 ng/u1

RT: 17.328 min Scan# 2421

Delta R.T. 0.000 min

Lab File: BP010346.D

Acq: 13 May 2022 11:01

Instrument :

BNA_P

ClientSampleId :

SSTD005638





Tgt Ion:178 Resp: 210935

Ion Ratio Lower Upper

178 100

179 15.7 11.7 17.5

176 19.2 14.5 21.7



#74

Anthracene

Concen: 4.509 ng/u1

RT: 17.422 min Scan# 2437

Delta R.T. 0.006 min

Lab File: BP010346.D

Acq: 13 May 2022 11:01




Tgt Ion:178 Resp: 209501


Ion Ratio Lower Upper

178 100

179 15.3 12.3 18.5

176 18.8 13.4 20.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 5.417 ng/uL

RT: 13.340 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP010346.D

Acq: 13 May 2022 11:01

Instrument :

BNA_P

ClientSampleId :

SSTD005638





Tgt Ion:216 Resp: 65854

Ion Ratio Lower Upper

216 100

214 73.7 50.4 75.6

218 47.0 35.2 52.8



#76

Pentachlorobenzene

Concen: 4.647 ng/uL

RT: 14.857 min Scan# 2001

Delta R.T. 0.000 min

Lab File: BP010346.D

Acq: 13 May 2022 11:01




Tgt Ion:250 Resp: 54021


Ion Ratio Lower Upper

250 100

248 61.3 43.5 65.3

252 62.6 46.5 69.7









#77
 Carbazole
 Concen: 0.006 ng/ul
 RT: 17.816 min Scan# 21
 Delta R.T. 0.130 min
 Lab File: BP010346.D
 Acq: 13 May 2022 11:01




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005638


Tgt Ion:	167	Resp:	250
Ion Ratio	Lower	Upper	
167	100		
166	16.1	15.8	23.8
139	11.3	8.3	12.5



#78
 Di-n-butylphthalate
 Concen: 0.007 ng/ul
 RT: 18.316 min Scan# 2589
 Delta R.T. 0.077 min
 Lab File: BP010346.D
 Acq: 13 May 2022 11:01

Tgt Ion:	149	Resp:	389
Ion Ratio	Lower	Upper	
149	100		
150	12.2	9.6	14.4
104	9.3	6.2	9.4









#80
Fluoranthene
Concen: 4.434 ng/ul
RT: 19.292 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:202 Resp: 248225
Ion Ratio Lower Upper
202 100
101 9.8 9.6 14.4
100 7.7 6.3 9.5



#82
Pyrene
Concen: 4.584 ng/ul
RT: 19.645 min Scan# 2815
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:202 Resp: 265646
Ion Ratio Lower Upper
202 100
101 11.0 10.6 15.8
100 8.8 9.8 14.6#





#83

Butylbenzylphthalate

Concen: 3.773 ng/ul

RT: 20.516 min Scan# 2962

Delta R.T. 0.006 min

Lab File: BP010346.D

Acq: 13 May 2022 11:01

Instrument :

BNA_P

ClientSampleId :

SSTD005638





Tgt Ion:149 Resp: 93114

Ion Ratio Lower Upper

149 100

91 70.5 71.1 106.7#

206 22.6 28.8 43.2#



#84

3,3'-Dichlorobenzidine

Concen: 3.828 ng/ul

RT: 21.286 min Scan# 3094

Delta R.T. 0.006 min

Lab File: BP010346.D

Acq: 13 May 2022 11:01




Tgt Ion:252 Resp: 72904


Ion Ratio Lower Upper

252 100

254 61.1 49.8 74.6

126 11.5 4.3 6.5#









#85
Benzo(a)anthracene
Concen: 4.433 ng/ul
RT: 21.351 min Scan# 3105
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:228 Resp: 258233
Ion Ratio Lower Upper
228 100
229 20.3 18.2 27.2
226 25.8 22.8 34.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 3.645 ng/ul
RT: 21.274 min Scan# 3092
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:149 Resp: 135735
Ion Ratio Lower Upper
149 100
167 29.4 32.8 49.2#
279 4.7 5.6 8.4#









#87
Chrysene
Concen: 4.389 ng/ul
RT: 21.404 min Scan# 3114
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:228 Resp: 253357
Ion Ratio Lower Upper
228 100
226 30.0 22.7 34.1
229 19.5 15.2 22.8



#89
Di-n-octyl phthalate
Concen: 4.358 ng/ul
RT: 22.204 min Scan# 3250
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:149 Resp: 238654









#90
Benzo(b)fluoranthene
Concen: 4.404 ng/ul
RT: 23.051 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:252 Resp: 235314
Ion Ratio Lower Upper
252 100
253 23.2 17.4 26.0
125 9.4 7.3 10.9



#91
Benzo(k)fluoranthene
Concen: 4.508 ng/ul
RT: 23.104 min Scan# 3403
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:252 Resp: 227270
Ion Ratio Lower Upper
252 100
253 21.9 17.9 26.9
125 9.3 5.1 7.7#









#93
Benzo(a)pyrene
Concen: 4.368 ng/ul
RT: 23.686 min Scan# 3501
Delta R.T. 0.006 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01




Instrument :
BNA_P
ClientSampleId :
SSTD005638


Tgt Ion:252 Resp: 226064
Ion Ratio Lower Upper
252 100
253 22.3 17.5 26.3
125 8.7 6.4 9.6



#94
Indeno(1,2,3-cd)pyrene
Concen: 3.666 ng/ul
RT: 26.321 min Scan# 3950
Delta R.T. 0.000 min
Lab File: BP010346.D
Acq: 13 May 2022 11:01

Tgt Ion:276 Resp: 209741
Ion Ratio Lower Upper
276 100
138 21.4 12.3 18.5#
277 26.2 19.3 28.9









#95
 Dibenzo(a,h)anthracene
 Concen: 0.006 ng/ul
 RT: 26.498 min Scan# 3952
 Delta R.T. 0.165 min
 Lab File: BP010346.D
 Acq: 13 May 2022 11:01

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005638

Tgt Ion: 278 Resp: 280

Ion	Ratio	Lower	Upper
278	100		
139	9.9	6.6	10.0
279	26.4	18.4	27.6



#96
 Benzo(g,h,i)perylene
 Concen: 0.005 ng/ul
 RT: 27.204 min Scan# 4100
 Delta R.T. 0.112 min
 Lab File: BP010346.D
 Acq: 13 May 2022 11:01

Tgt Ion: 276 Resp: 251

Ion	Ratio	Lower	Upper
276	100		
138	17.2	11.6	17.4
277	18.9	15.9	23.9

