

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051322\
 Data File : BP010347.D
 Acq On : 13 May 2022 11:35
 Operator : CG/JU
 Sample : SSTD01039
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010639

Quant Time: May 13 13:35:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 13:33:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	133798	20.000	ng/ul	0.00
20) Naphthalene-d8	10.705	136	597787	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.534	164	426041	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.281	188	949840	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.363	240	981916	20.000	ng/ul	0.00
88) Perylene-d12	23.792	264	908913	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	12249	3.889	ng/uL	0.00
4) Pyridine-d5	3.770	84	74061	8.450	ng/ul	0.00
7) Phenol-d5	7.064	99	94642	7.616	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	68513	8.912	ng/ul	0.00
11) 2-Chlorophenol-d4	7.434	132	76592	8.369	ng/ul	0.00
15) 4-Methylphenol-d8	8.611	113	81490	7.976	ng/ul	0.00
21) Nitrobenzene-d5	9.058	128	40861	8.525	ng/ul	0.00
24) 2-Nitrophenol-d4	9.781	143	41828	7.654	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.322	165	80945	8.138	ng/ul	0.00
31) 4-Chloroaniline-d4	10.840	131	118168	8.047	ng/ul	0.00
46) Dimethylphthalate-d6	13.940	166	283556	8.354	ng/ul	0.00
49) Acenaphthylene-d8	14.222	160	324911	8.029	ng/ul	0.00
54) 4-Nitrophenol-d4	14.728	143	38726	5.948	ng/ul	0.00
60) Fluorene-d10	15.528	176	248853	8.686	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	37391	5.749	ng/ul	0.00
73) Anthracene-d10	17.381	188	393400	8.458	ng/ul	0.00
81) Pyrene-d10	19.610	212	472876	8.715	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.633	264	401483	8.283	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.381	88	11887	3.692	ng/uL#	1
5) Pyridine	3.787	79	81771	8.902	ng/ul#	27
6) Benzaldehyde	7.040	77	63436	9.237	ng/ul	91
8) Phenol	7.093	94	104752	8.361	ng/ul#	91
10) Bis(2-Chloroethyl)ether	7.322	93	86345	8.807	ng/ul#	72
12) 2-Chlorophenol	7.464	128	80859	8.797	ng/ul	96
13) 2-Methylphenol	8.346	108	75733	7.964	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.428	45	149312	9.682	ng/ul#	81
16) Acetophenone	8.722	105	136196	8.529	ng/ul#	75
17) N-Nitroso-di-n-propyla...	8.711	70	74031	8.874	ng/ul#	66
18) 4-Methylphenol	8.669	108	88424	8.499	ng/ul	96
19) Hexachloroethane	8.981	117	36849	9.403	ng/ul	82
22) Nitrobenzene	9.105	77	110857	9.111	ng/ul#	83
23) Isophorone	9.628	82	199493	8.372	ng/ul	98
25) 2-Nitrophenol	9.816	139	46578	8.297	ng/ul#	85
26) 2,4-Dimethylphenol	9.875	107	104033	8.678	ng/ul#	76
27) Bis(2-Chloroethoxy)met...	10.110	93	122674	8.772	ng/ul#	98
29) 2,4-Dichlorophenol	10.352	162	85116	8.760	ng/ul#	81
30) Naphthalene	10.752	128	287442	9.012	ng/ul	97
32) 4-Chloroaniline	10.863	127	118267	8.070	ng/ul	96
33) Hexachlorobutadiene	11.046	225	62387	9.375	ng/ul	92
34) Caprolactam	11.634	113	25449	6.874	ng/ul#	1
35) 4-Chloro-3-methylphenol	11.987	107	100122	8.563	ng/ul#	72
36) 2-Methylnaphthalene	12.357	142	205505	9.100	ng/ul	90

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051322\
 Data File : BP010347.D
 Acq On : 13 May 2022 11:35
 Operator : CG/JU
 Sample : SSTD01039
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD010639


Quant Time: May 13 13:35:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 13:33:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.575	142	204789	8.920	ng/ul#	96
39) 1,2,4,5-Tetrachloroben...	12.728	216	117686	8.986	ng/ul#	94
40) Hexachlorocyclopentadiene	12.710	237	47938	6.566	ng/ul	97
41) 2,4,6-Trichlorophenol	12.969	196	69973	8.146	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.040	196	78776	8.415	ng/ul#	84
43) 1,1'-Biphenyl	13.369	154	285627	8.975	ng/ul	96
44) 2-Chloronaphthalene	13.410	162	221358	8.980	ng/ul	93
45) 2-Nitroaniline	13.610	65	66370	7.953	ng/ul#	77
47) Dimethylphthalate	13.987	163	287893	8.742	ng/ul#	91
48) 2,6-Dinitrotoluene	14.104	165	53933	8.026	ng/ul	98
50) Acenaphthylene	14.251	152	353813	8.740	ng/ul	95
51) 3-Nitroaniline	14.434	138	47236	8.035	ng/ul#	88
52) Acenaphthene	14.598	153	234110	8.933	ng/ul	97
53) 2,4-Dinitrophenol	14.646	184	17818	4.052	ng/ul#	76
55) 4-Nitrophenol	14.740	109	38217	6.921	ng/ul#	43
56) Dibenzofuran	14.928	168	341429	8.904	ng/ul	96
57) 2,4-Dinitrotoluene	14.893	165	78719	8.007	ng/ul#	75
58) 2,3,4,6-Tetrachlorophenol	15.157	232	68895	8.292	ng/ul#	85
59) Diethylphthalate	15.351	149	290044	8.532	ng/ul#	92
61) Fluorene	15.581	166	280935	8.845	ng/ul	90
62) 4-Chlorophenyl-phenyle...	15.575	204	149091	9.080	ng/ul	97
63) 4-Nitroaniline	15.722	138	227	0.039	ng/ul	83
66) 4,6-Dinitro-2-methylph...	15.657	198	39230	6.191	ng/ul#	82
67) N-Nitrosodiphenylamine	15.787	169	242012	8.668	ng/ul	94
68) 4-Bromophenyl-phenylether	16.469	248	87486	8.720	ng/ul	96
69) Hexachlorobenzene	16.592	284	104306	8.988	ng/ul#	88
70) Atrazine	16.840	200	136	0.012	ng/ul	96
71) Pentachlorophenol	16.940	266	52194	6.776	ng/ul	86
72) Phenanthrene	17.322	178	465264	8.884	ng/ul	99
74) Anthracene	17.416	178	459589	8.653	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.334	216	129631	9.328	ng/uL#	85
76) Pentachlorobenzene	14.851	250	119373	8.984	ng/uL	93
77) Carbazole	17.798	167	837	0.017	ng/ul	96
78) Di-n-butylphthalate	18.239	149	474094	7.828	ng/ul#	93
80) Fluoranthene	19.286	202	546004	8.850	ng/ul	97
82) Pyrene	19.639	202	588822	9.221	ng/ul	95
83) Butylbenzylphthalate	20.510	149	215581	7.927	ng/ul#	82
84) 3,3'-Dichlorobenzidine	21.280	252	169084	8.057	ng/ul#	96
85) Benzo(a)anthracene	21.345	228	563221	8.774	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.275	149	319497	7.787	ng/ul#	82
87) Chrysene	21.398	228	553568	8.701	ng/ul	99
89) Di-n-octyl phthalate	22.204	149	546998	8.940	ng/ul	100
90) Benzo(b)fluoranthene	23.045	252	546511	9.155	ng/ul	99
91) Benzo(k)fluoranthene	23.098	252	510711	9.067	ng/ul#	96
93) Benzo(a)pyrene	23.680	252	499163	8.632	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.316	276	452263	7.076	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.327	278	397639	7.276	ng/ul#	94
96) Benzo(g,h,i)perylene	27.092	276	386611	7.162	ng/ul#	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 13 13:35:07 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 13 13:33:22 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.692 ng/uL
RT: 3.381 min Scan# 50
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:	88	Resp:	11887
Ion Ratio	Lower	Upper	
88	100		
43	36.1	6.8	10.2#
58	97.9	28.4	42.6#



#5
Pyridine
Concen: 8.902 ng/uL
RT: 3.787 min Scan# 119
Delta R.T. 0.006 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:	79	Resp:	81771
Ion Ratio	Lower	Upper	
79	100		
52	102.6	35.5	53.3#
51	47.4	20.2	30.4#









#6
Benzaldehyde
Concen: 9.237 ng/ul
RT: 7.040 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion: 77 Resp: 63436
Ion Ratio Lower Upper
77 100
105 87.6 77.0 115.4
106 86.7 77.0 115.4



#8
Phenol
Concen: 8.361 ng/ul
RT: 7.093 min Scan# 681
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion: 94 Resp: 104752
Ion Ratio Lower Upper
94 100
65 32.6 19.0 28.6#
66 45.8 34.6 51.8









#10
Bis(2-Chloroethyl)ether
Concen: 8.807 ng/ul
RT: 7.322 min Scan# 720
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion: 93 Resp: 86345
Ion Ratio Lower Upper
93 100
63 86.6 45.0 67.4#
95 30.7 26.3 39.5



#12
2-Chlorophenol
Concen: 8.797 ng/ul
RT: 7.464 min Scan# 744
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion: 128 Resp: 80859
Ion Ratio Lower Upper
128 100
64 54.7 45.3 67.9
130 33.3 29.0 43.6





#13

2-Methylphenol

Concen: 7.964 ng/ul

RT: 8.346 min Scan# 894

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35

Instrument :

BNA_P

ClientSampleId :





SSTD010639

Tgt Ion:108 Resp: 75733

Ion Ratio Lower Upper

108 100

107 91.6 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 9.682 ng/ul

RT: 8.428 min Scan# 908

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35




Tgt Ion: 45 Resp: 149312


Ion Ratio Lower Upper

45 100

77 11.9 17.1 25.7#

79 8.7 13.2 19.8#









#16
Acetophenone
Concen: 8.529 ng/ul
RT: 8.722 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:105 Resp: 136196
Ion Ratio Lower Upper
105 100
77 83.4 53.4 80.2#
51 34.9 13.8 20.8#



#17
N-Nitroso-di-n-propylamine
Concen: 8.874 ng/ul
RT: 8.711 min Scan# 956
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion: 70 Resp: 74031
Ion Ratio Lower Upper
70 100
42 71.4 31.8 47.8#
101 9.6 15.6 23.4#
130 20.3 13.7 20.5









#18
4-Methylphenol
Concen: 8.499 ng/ul
RT: 8.669 min Scan# 949
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:108 Resp: 88424
Ion Ratio Lower Upper
108 100
107 108.2 83.6 125.4



#19
Hexachloroethane
Concen: 9.403 ng/ul
RT: 8.981 min Scan# 1002
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:117 Resp: 36849
Ion Ratio Lower Upper
117 100
201 109.3 106.5 159.7
199 62.2 59.0 88.6









#22
Nitrobenzene
Concen: 9.111 ng/ul
RT: 9.105 min Scan# 1108
Delta R.T. 0.006 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion: 77 Resp: 110857
Ion Ratio Lower Upper
77 100
123 41.2 42.3 63.5#
65 13.7 5.8 8.8#



#23
Isophorone
Concen: 8.372 ng/ul
RT: 9.628 min Scan# 1112
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion: 82 Resp: 199493
Ion Ratio Lower Upper
82 100
95 9.0 7.8 11.8
138 15.4 12.9 19.3









#25
2-Nitrophenol
Concen: 8.297 ng/ul
RT: 9.816 min Scan# 1144
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:139 Resp: 46578
Ion Ratio Lower Upper
139 100
65 58.4 49.4 74.2
109 36.5 44.7 67.1#



#26
2,4-Dimethylphenol
Concen: 8.678 ng/ul
RT: 9.875 min Scan# 1154
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:107 Resp: 104033
Ion Ratio Lower Upper
107 100
121 48.0 34.1 51.1
122 84.1 46.3 69.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 8.772 ng/ul
 RT: 10.110 min Scan# 1194
 Delta R.T. 0.000 min
 Lab File: BP010347.D
 Acq: 13 May 2022 11:35




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010639


Tgt Ion: 93 Resp: 122674
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.2 39.4
 123 11.3 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 8.760 ng/ul
 RT: 10.352 min Scan# 1235
 Delta R.T. 0.000 min
 Lab File: BP010347.D
 Acq: 13 May 2022 11:35

Tgt Ion: 162 Resp: 85116
 Ion Ratio Lower Upper
 162 100
 164 68.9 37.0 55.6#
 98 39.4 31.0 46.6









#30
Naphthalene
Concen: 9.012 ng/ul
RT: 10.752 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:128 Resp: 287442
Ion Ratio Lower Upper
128 100
129 11.1 8.6 13.0
127 13.2 12.2 18.4



#32
4-Chloroaniline
Concen: 8.070 ng/ul
RT: 10.863 min Scan# 1322
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:127 Resp: 118267
Ion Ratio Lower Upper
127 100
129 33.0 24.6 37.0





#33

Hexachlorobutadiene

Concen: 9.375 ng/ul

RT: 11.046 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD010639





Tgt Ion:225 Resp: 62387

Ion Ratio Lower Upper

225 100

223 62.6 53.4 80.0

227 57.9 53.8 80.8



#34

Caprolactam

Concen: 6.874 ng/ul

RT: 11.634 min Scan# 1453

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35




Tgt Ion:113 Resp: 25449


Ion Ratio Lower Upper

113 100

55 216.9 58.2 87.2#

56 154.1 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 8.563 ng/ul
RT: 11.987 min Scan# 1513
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:107 Resp: 100122
Ion Ratio Lower Upper
107 100
144 23.2 15.8 23.8
142 73.7 38.9 58.3#



#36
2-Methylnaphthalene
Concen: 9.100 ng/ul
RT: 12.357 min Scan# 1576
Delta R.T. -0.006 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:142 Resp: 205505
Ion Ratio Lower Upper
142 100
141 89.3 64.3 96.5









#37
1-Methylnaphthalene
Concen: 8.920 ng/ul
RT: 12.575 min Scan# 1613
Delta R.T. -0.006 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:142 Resp: 204789
Ion Ratio Lower Upper
142 100
141 91.1 69.8 104.8
116 4.8 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 8.986 ng/ul
RT: 12.728 min Scan# 1639
Delta R.T. -0.006 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:216 Resp: 117686
Ion Ratio Lower Upper
216 100
214 77.6 58.2 87.2
179 22.9 14.2 21.2#
108 17.3 15.3 22.9





#40

Hexachlorocyclopentadiene

Concen: 6.566 ng/ul

RT: 12.710 min Scan# 1636

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD010639





Tgt Ion:237 Resp: 47938

Ion Ratio Lower Upper

237 100

235 63.9 53.2 79.8

272 12.6 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 8.146 ng/ul

RT: 12.969 min Scan# 1680

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35




Tgt Ion:196 Resp: 69973


Ion Ratio Lower Upper

196 100

198 96.9 64.1 96.1#

200 29.8 19.9 29.9





#42

2,4,5-Trichlorophenol

Concen: 8.415 ng/ul

RT: 13.040 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD010639





Tgt Ion:196 Resp: 78776

Ion Ratio Lower Upper

196 100

198 99.1 68.6 103.0

200 29.8 32.8 49.2#



#43

1,1'-Biphenyl

Concen: 8.975 ng/ul

RT: 13.369 min Scan# 1748

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35




Tgt Ion:154 Resp: 285627


Ion Ratio Lower Upper

154 100

153 38.3 28.6 43.0

76 12.5 8.5 12.7









#44
2-Chloronaphthalene
Concen: 8.980 ng/ul
RT: 13.410 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:162 Resp: 221358
Ion Ratio Lower Upper
162 100
164 32.4 27.5 41.3
127 40.2 37.1 55.7



#45
2-Nitroaniline
Concen: 7.953 ng/ul
RT: 13.610 min Scan# 1789
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion: 65 Resp: 66370
Ion Ratio Lower Upper
65 100
92 65.7 77.9 116.9#
138 87.8 81.5 122.3









#47
Dimethylphthalate
Concen: 8.742 ng/ul
RT: 13.987 min Scan# 1853
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:163 Resp: 287893
Ion Ratio Lower Upper
163 100
194 4.2 4.1 6.1
164 9.8 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 8.026 ng/ul
RT: 14.104 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:165 Resp: 53933
Ion Ratio Lower Upper
165 100
89 66.8 51.7 77.5
121 22.6 18.3 27.5









#50
Acenaphthylene
Concen: 8.740 ng/ul
RT: 14.251 min Scan# 1898
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:152 Resp: 353813
Ion Ratio Lower Upper
152 100
151 19.9 18.3 27.5
153 12.8 11.0 16.6



#51
3-Nitroaniline
Concen: 8.035 ng/ul
RT: 14.434 min Scan# 1929
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:138 Resp: 47236
Ion Ratio Lower Upper
138 100
108 13.3 15.8 23.8#
92 131.2 117.1 175.7









#52
Acenaphthene
Concen: 8.933 ng/ul
RT: 14.598 min Scan# 1957
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:153 Resp: 234110
Ion Ratio Lower Upper
153 100
152 46.0 34.3 51.5
154 89.5 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 4.052 ng/ul
RT: 14.646 min Scan# 1965
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:184 Resp: 17818
Ion Ratio Lower Upper
184 100
63 78.5 34.6 52.0#
154 67.3 57.7 86.5









#55
4-Nitrophenol
Concen: 6.921 ng/ul
RT: 14.740 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:109 Resp: 38217
Ion Ratio Lower Upper
109 100
139 114.6 61.4 92.2#
65 143.0 65.3 97.9#



#56
Dibenzofuran
Concen: 8.904 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. -0.006 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:168 Resp: 341429
Ion Ratio Lower Upper
168 100
139 40.3 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 8.007 ng/ul
RT: 14.893 min Scan# 2007
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:165 Resp: 78719
Ion Ratio Lower Upper
165 100
63 53.2 29.5 44.3#
182 2.9 2.8 4.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 8.292 ng/ul
RT: 15.157 min Scan# 2052
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:232 Resp: 68895
Ion Ratio Lower Upper
232 100
131 46.5 25.4 38.2#
130 2.5 4.2 6.2#
166 31.3 28.1 42.1









#59
Diethylphthalate
Concen: 8.532 ng/ul
RT: 15.351 min Scan# 2085
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:149 Resp: 290044
Ion Ratio Lower Upper
149 100
177 21.3 22.0 33.0#
150 13.1 10.2 15.2



#61
Fluorene
Concen: 8.845 ng/ul
RT: 15.581 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:166 Resp: 280935
Ion Ratio Lower Upper
166 100
165 98.0 70.3 105.5
167 13.3 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 9.080 ng/ul
RT: 15.575 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:	204	Resp:	149091
Ion	Ratio	Lower	Upper
204	100		
206	32.5	27.6	41.4
141	60.1	46.6	69.8



#63
4-Nitroaniline
Concen: 0.039 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. 0.124 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:	138	Resp:	227
Ion	Ratio	Lower	Upper
138	100		
92	47.4	41.6	62.4
108	97.6	97.6	146.4









#66
4,6-Dinitro-2-methylphenol
Concen: 6.191 ng/ul
RT: 15.657 min Scan# 2138
Delta R.T. -0.006 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:198 Resp: 39230
Ion Ratio Lower Upper
198 100
182 5.1 6.0 9.0#
77 32.2 17.7 26.5#



#67
N-Nitrosodiphenylamine
Concen: 8.668 ng/ul
RT: 15.787 min Scan# 2159
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:169 Resp: 242012
Ion Ratio Lower Upper
169 100
168 65.4 48.3 72.5
167 34.6 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 8.720 ng/ul
RT: 16.469 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:248 Resp: 87486
Ion Ratio Lower Upper
248 100
250 97.4 79.4 119.0
141 78.9 67.2 100.8



#69
Hexachlorobenzene
Concen: 8.988 ng/ul
RT: 16.592 min Scan# 2296
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:284 Resp: 104306
Ion Ratio Lower Upper
284 100
142 34.4 19.2 28.8#
249 31.5 27.4 41.2





#70

Atrazine

Concen: 0.012 ng/ul

RT: 16.840 min Scan# 21

Delta R.T. 0.094 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD010639





Tgt Ion:200 Resp: 136

Ion Ratio Lower Upper

200 100

173 24.6 19.2 28.8

215 43.8 32.2 48.4



#71

Pentachlorophenol

Concen: 6.776 ng/ul

RT: 16.940 min Scan# 2355

Delta R.T. 0.006 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35




Tgt Ion:266 Resp: 52194


Ion Ratio Lower Upper

266 100

264 63.0 57.3 85.9

268 67.1 65.7 98.5





#72

Phenanthrene

Concen: 8.884 ng/ul

RT: 17.322 min Scan# 2421

Delta R.T. -0.006 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD010639





Tgt Ion:178 Resp: 465264

Ion Ratio Lower Upper

178 100

179 15.4 11.7 17.5

176 17.7 14.5 21.7



#74

Anthracene

Concen: 8.653 ng/ul

RT: 17.416 min Scan# 2436

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35




Tgt Ion:178 Resp: 459589


Ion Ratio Lower Upper

178 100

179 15.9 12.3 18.5

176 17.8 13.4 20.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 9.328 ng/uL

RT: 13.334 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD010639





Tgt Ion:216 Resp: 129631

Ion Ratio Lower Upper

216 100

214 78.3 50.4 75.6#

218 49.4 35.2 52.8



#76

Pentachlorobenzene

Concen: 8.984 ng/uL

RT: 14.851 min Scan# 2000

Delta R.T. -0.006 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35




Tgt Ion:250 Resp: 119373


Ion Ratio Lower Upper

250 100

248 61.8 43.5 65.3

252 60.8 46.5 69.7









#77
 Carbazole
 Concen: 0.017 ng/ul
 RT: 17.798 min Scan# 2482
 Delta R.T. 0.112 min
 Lab File: BP010347.D
 Acq: 13 May 2022 11:35




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010639


Tgt Ion:167 Resp: 837
 Ion Ratio Lower Upper
 167 100
 166 17.5 15.8 23.8
 139 10.1 8.3 12.5



#78
 Di-n-butylphthalate
 Concen: 7.828 ng/ul
 RT: 18.239 min Scan# 2576
 Delta R.T. 0.000 min
 Lab File: BP010347.D
 Acq: 13 May 2022 11:35

Tgt Ion:149 Resp: 474094
 Ion Ratio Lower Upper
 149 100
 150 9.0 9.6 14.4#
 104 5.7 6.2 9.4#









#80
Fluoranthene
Concen: 8.850 ng/ul
RT: 19.286 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:202 Resp: 546004
Ion Ratio Lower Upper
202 100
101 10.5 9.6 14.4
100 8.3 6.3 9.5



#82
Pyrene
Concen: 9.221 ng/ul
RT: 19.639 min Scan# 2814
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:202 Resp: 588822
Ion Ratio Lower Upper
202 100
101 11.5 10.6 15.8
100 10.1 9.8 14.6





#83

Butylbenzylphthalate

Concen: 7.927 ng/ul

RT: 20.510 min Scan# 2962

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD010639





Tgt Ion:149 Resp: 215581

Ion Ratio Lower Upper

149 100

91 75.1 71.1 106.7

206 21.6 28.8 43.2#



#84

3,3'-Dichlorobenzidine

Concen: 8.057 ng/ul

RT: 21.280 min Scan# 3093

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35




Tgt Ion:252 Resp: 169084


Ion Ratio Lower Upper

252 100

254 64.3 49.8 74.6

126 10.2 4.3 6.5#





#85

Benzo(a)anthracene

Concen: 8.774 ng/ul

RT: 21.345 min Scan# 3104

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD010639





Tgt Ion:228 Resp: 563221

Ion Ratio Lower Upper

228 100

229 19.3 18.2 27.2

226 26.2 22.8 34.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 7.787 ng/ul

RT: 21.275 min Scan# 3092

Delta R.T. 0.006 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35




Tgt Ion:149 Resp: 319497


Ion Ratio Lower Upper

149 100

167 28.4 32.8 49.2#

279 4.6 5.6 8.4#





#87

Chrysene

Concen: 8.701 ng/ul

RT: 21.398 min Scan# 3113

Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35

Instrument :

BNA_P

ClientSampleId :

SSTD010639





Tgt Ion:228 Resp: 553568

Ion Ratio Lower Upper

228 100

226 28.6 22.7 34.1

229 19.6 15.2 22.8



#89

Di-n-octyl phthalate

Concen: 8.940 ng/ul




RT: 22.204 min Scan# 3250


Delta R.T. 0.000 min

Lab File: BP010347.D

Acq: 13 May 2022 11:35

Tgt Ion:149 Resp: 546998









#90
Benzo(b)fluoranthene
Concen: 9.155 ng/ul
RT: 23.045 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:252 Resp: 546511
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.0
125 9.2 7.3 10.9



#91
Benzo(k)fluoranthene
Concen: 9.067 ng/ul
RT: 23.098 min Scan# 3402
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:252 Resp: 510711
Ion Ratio Lower Upper
252 100
253 21.2 17.9 26.9
125 9.0 5.1 7.7#









#93
Benzo(a)pyrene
Concen: 8.632 ng/ul
RT: 23.680 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35




Instrument :
BNA_P
ClientSampleId :
SSTD010639


Tgt Ion:252 Resp: 499163
Ion Ratio Lower Upper
252 100
253 21.3 17.5 26.3
125 10.2 6.4 9.6#



#94
Indeno(1,2,3-cd)pyrene
Concen: 7.076 ng/ul
RT: 26.316 min Scan# 3949
Delta R.T. -0.006 min
Lab File: BP010347.D
Acq: 13 May 2022 11:35

Tgt Ion:276 Resp: 452263
Ion Ratio Lower Upper
276 100
138 21.4 12.3 18.5#
277 25.6 19.3 28.9









#95
 Dibenzo(a,h)anthracene
 Concen: 7.276 ng/ul
 RT: 26.327 min Scan# 3951
 Delta R.T. -0.006 min
 Lab File: BP010347.D
 Acq: 13 May 2022 11:35

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010639

Tgt Ion:278 Resp: 397639
 Ion Ratio Lower Upper
 278 100
 139 15.2 6.6 10.0#
 279 23.5 18.4 27.6



#96
 Benzo(g,h,i)perylene
 Concen: 7.162 ng/ul
 RT: 27.092 min Scan# 4081
 Delta R.T. 0.000 min
 Lab File: BP010347.D
 Acq: 13 May 2022 11:35

Tgt Ion:276 Resp: 386611
 Ion Ratio Lower Upper
 276 100
 138 20.5 11.6 17.4#
 277 23.3 15.9 23.9

