

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
 Data File : BP024615.D
 Acq On : 13 May 2025 11:22
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Quant Time: May 13 16:02:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 13 15:55:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.664	152	90088	20.00	ng	0.00
21) Naphthalene-d8	10.434	136	378055	20.00	ng	0.00
39) Acenaphthene-d10	14.293	164	254561	20.00	ng	-0.01
64) Phenanthrene-d10	17.092	188	531548	20.00	ng	0.00
76) Chrysene-d12	21.522	240	614288	20.00	ng	0.00
86) Perylene-d12	24.810	264	724260	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.293	112	51093	9.70	ng	0.00
7) Phenol-d6	6.858	99	58619	8.72	ng	0.00
23) Nitrobenzene-d5	8.811	82	72056	9.63	ng	0.00
42) 2,4,6-Tribromophenol	15.810	330	41219	9.55	ng	0.00
45) 2-Fluorobiphenyl	12.904	172	203719	10.48	ng	0.00
79) Terphenyl-d14	19.822	244	366245	10.22	ng	0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.228	88	14435	5.84	ng	95
3) Pyridine	3.646	79	23737	4.33	ng	# 91
4) n-Nitrosodimethylamine	3.546	42	12080	4.99	ng	# 96
6) Aniline	7.011	93	38664	4.45	ng	97
8) 2-Chlorophenol	7.240	128	27022	4.52	ng	93
9) Benzaldehyde	6.828	77	22328m	5.14	ng	
10) Phenol	6.887	94	30912	4.46	ng	98
11) bis(2-Chloroethyl)ether	7.099	93	29668	5.20	ng	97
12) 1,3-Dichlorobenzene	7.552	146	35946	5.23	ng	96
13) 1,4-Dichlorobenzene	7.699	146	35091	5.07	ng	95
14) 1,2-Dichlorobenzene	8.011	146	34458	5.11	ng	99
15) Benzyl Alcohol	7.917	79	19888	3.89	ng	97
16) 2,2'-oxybis(1-Chloropr...	8.193	45	34754	5.09	ng	95
17) 2-Methylphenol	8.116	107	21754	4.37	ng	97
18) Hexachloroethane	8.728	117	13650	5.15	ng	96
19) n-Nitroso-di-n-propyla...	8.464	70	22015	4.73	ng	98
20) 3+4-Methylphenols	8.452	107	27500	4.06	ng	97
22) Acetophenone	8.487	105	45968	4.87	ng	# 99
24) Nitrobenzene	8.852	77	31500	4.78	ng	97
25) Isophorone	9.375	82	60855	4.69	ng	98
26) 2-Nitrophenol	9.569	139	12786	3.96	ng	95
27) 2,4-Dimethylphenol	9.628	122	25830	4.54	ng	92
28) bis(2-Chloroethoxy)met...	9.858	93	38071	4.91	ng	97
29) 2,4-Dichlorophenol	10.122	162	23623	4.29	ng	97
30) 1,2,4-Trichlorobenzene	10.299	180	32699	5.26	ng	99
31) Naphthalene	10.481	128	100564	5.13	ng	100
33) 4-Chloroaniline	10.616	127	32908	4.16	ng	99
34) Hexachlorobutadiene	10.763	225	21834	5.43	ng	95
35) Caprolactam	11.399	113	7742m	3.88	ng	
36) 4-Chloro-3-methylphenol	11.746	107	29849	4.44	ng	95
37) 2-Methylnaphthalene	12.099	142	63562	5.01	ng	99
38) 1-Methylnaphthalene	12.316	142	70043	5.16	ng	100
40) 1,2,4,5-Tetrachloroben...	12.469	216	37738	5.11	ng	99
41) Hexachlorocyclopentadiene	12.446	237	15939	3.56	ng	93
43) 2,4,6-Trichlorophenol	12.722	196	19429	4.06	ng	92
44) 2,4,5-Trichlorophenol	12.816	196	23918	4.49	ng	93

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
 Data File : BP024615.D
 Acq On : 13 May 2025 11:22
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Quant Time: May 13 16:02:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 13 15:55:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) 1,1'-Biphenyl	13.116	154	96956	5.15	ng	98
47) 2-Chloronaphthalene	13.157	162	72977	5.09	ng	99
48) 2-Nitroaniline	13.381	65	17778	4.19	ng	98
49) Acenaphthylene	14.010	152	117448	4.89	ng	99
50) Dimethylphthalate	13.751	163	99940	5.15	ng	100
51) 2,6-Dinitrotoluene	13.869	165	19052	4.65	ng	# 85
52) Acenaphthene	14.357	154	70302	5.09	ng	96
53) 3-Nitroaniline	14.222	138	16400	3.92	ng	97
55) Dibenzofuran	14.698	168	115180	5.23	ng	98
57) 2,4-Dinitrotoluene	14.675	165	25481	4.41	ng	92
58) Fluorene	15.363	166	93303	5.20	ng	98
59) 2,3,4,6-Tetrachlorophenol	14.946	232	23382	4.73	ng	98
60) Diethylphthalate	15.134	149	101366	5.17	ng	99
61) 4-Chlorophenyl-phenyle...	15.357	204	47886	5.35	ng	98
62) 4-Nitroaniline	15.410	138	14482m	3.55	ng	
63) Azobenzene	15.651	77	84462	5.03	ng	99
66) n-Nitrosodiphenylamine	15.575	169	80078	4.90	ng	99
67) 4-Bromophenyl-phenylether	16.275	248	30702	4.90	ng	97
68) Hexachlorobenzene	16.387	284	40234	5.06	ng	99
69) Atrazine	16.551	200	27909	4.66	ng	98
70) Pentachlorophenol	16.757	266	16456	3.68	ng	95
71) Phenanthrene	17.134	178	152030	5.14	ng	99
72) Anthracene	17.234	178	145809	4.90	ng	99
73) Carbazole	17.522	167	122666	4.56	ng	98
74) Di-n-butylphthalate	18.104	149	157014	4.54	ng	99
75) Fluoranthene	19.228	202	168865	4.89	ng	99
78) Pyrene	19.604	202	181817	4.94	ng	99
80) Butylbenzylphthalate	20.551	149	70341	4.30	ng	100
81) Benzo(a)anthracene	21.498	228	192110	4.98	ng	99
82) 3,3'-Dichlorobenzidine	21.445	252	54793	3.82	ng	99
83) Chrysene	21.575	228	186449	5.10	ng	99
84) Bis(2-ethylhexyl)phtha...	21.445	149	103286	4.29	ng	97
85) Di-n-octyl phthalate	22.692	149	160883	4.09	ng	99
87) Indeno(1,2,3-cd)pyrene	28.504	276	242814	4.59	ng	# 87
88) Benzo(b)fluoranthene	23.763	252	186058	4.49	ng	97
89) Benzo(k)fluoranthene	23.839	252	207645	4.86	ng	100
90) Benzo(a)pyrene	24.657	252	180157	4.52	ng	99
91) Dibenzo(a,h)anthracene	28.586	278	193373	4.50	ng	98
92) Benzo(g,h,i)perylene	29.651	276	199361	4.66	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

