

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051421\
 Data File : BP005559.D
 Acq On : 14 May 2021 20:57
 Operator : CG/JU
 Sample : M2322-03DL 5X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 MW-3-20210505DL

Quant Time: May 14 22:57:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 12:01:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.675	152	13153	20.00	ng	0.00
21) Naphthalene-d8	10.469	136	41805	20.00	ng	# 0.00
39) Acenaphthene-d10	14.328	164	31571	20.00	ng	0.00
64) Phenanthrene-d10	17.086	188	70390	20.00	ng	# 0.00
76) Chrysene-d12	21.186	240	107195	20.00	ng	# 0.00
86) Perylene-d12	23.492	264	142124	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.275	112	6809	9.93	ng	0.00
7) Phenol-d6	6.887	99	3783	4.43	ng	0.00
23) Nitrobenzene-d5	8.846	82	14747	13.88	ng	0.00
42) 2,4,6-Tribromophenol	15.833	330	14804	16.52	ng	0.00
45) 2-Fluorobiphenyl	12.951	172	36431	13.02	ng	0.00
79) Terphenyl-d14	19.657	244	72373	11.47	ng	0.00
Target Compounds						Qvalue
60) Diethylphthalate	15.163	149	62342	23.62	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051421\
Data File : BP005559.D
Acq On : 14 May 2021 20:57
Operator : CG/JU
Sample : M2322-03DL 5X
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
MW-3-20210505DL

Quant Time: May 14 22:57:13 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 12:01:17 2021
Response via : Initial Calibration

