

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051523\
 Data File : BP015096.D
 Acq On : 16 May 2023 03:31
 Operator : CG/JU
 Sample : 02592-12MSD
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 JREF6MSD

Quant Time: May 16 04:25:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 16 01:39:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.134	152	343579	20.000	ng/u1	0.00
20) Naphthalene-d8	10.963	136	1574127	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.775	164	985930	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.528	188	2130827	20.000	ng/u1	0.00
79) Chrysene-d12	21.604	240	1774064	20.000	ng/u1	0.00
88) Perylene-d12	24.174	264	1734356	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.423	96	93	0.011	ng/uL	-0.01
4) Pyridine-d5	3.881	84	16503	0.658	ng/u1	0.00
7) Phenol-d5	7.281	99	511324	16.029	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.452	67	300594	16.315	ng/u1	0.00
11) 2-Chlorophenol-d4	7.658	132	394377	16.814	ng/u1	0.00
15) 4-Methylphenol-d8	8.846	113	434617	17.098	ng/u1	0.00
21) Nitrobenzene-d5	9.310	128	198760	16.069	ng/u1	0.00
24) 2-Nitrophenol-d4	10.040	143	218474	15.874	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.581	165	436414	17.572	ng/u1	0.00
31) 4-Chloroaniline-d4	11.099	131	582047	15.840	ng/u1	0.00
46) Dimethylphthalate-d6	14.175	166	1277563	17.148	ng/u1	0.00
49) Acenaphthylene-d8	14.469	160	1484635	17.423	ng/u1	0.00
54) 4-Nitrophenol-d4	14.969	143	243866	16.857	ng/u1	0.00
60) Fluorene-d10	15.763	176	1087746	17.303	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.887	200	217972	16.232	ng/u1	0.00
73) Anthracene-d10	17.628	188	1666097	17.076	ng/u1	0.00
81) Pyrene-d10	19.845	212	1844754	17.929	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.004	264	1483514	17.231	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.470	88	125166	13.173	ng/uL	97
5) Pyridine	3.893	79	815554	31.284	ng/u1	95
6) Benzaldehyde	7.263	77	578072	65.929	ng/u1	96
8) Phenol	7.311	94	1185807	36.250	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.546	93	919764	35.144	ng/u1	100
12) 2-Chlorophenol	7.693	128	887239	35.763	ng/u1	99
13) 2-Methylphenol	8.581	108	905997	35.836	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.669	45	1176852	35.892	ng/u1	99
16) Acetophenone	8.969	105	1470255	37.059	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.952	70	724153	34.804	ng/u1	98
18) 4-Methylphenol	8.910	108	996117	35.797	ng/u1	97
19) Hexachloroethane	9.228	117	355001	35.045	ng/u1	96
22) Nitrobenzene	9.352	77	1054374	34.513	ng/u1	100
23) Isophorone	9.887	82	2119215	33.749	ng/u1	99
25) 2-Nitrophenol	10.069	139	522055	34.590	ng/u1	98
26) 2,4-Dimethylphenol	10.128	107	1038077	34.467	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.363	93	1312494	34.353	ng/u1	99
29) 2,4-Dichlorophenol	10.610	162	898677	35.823	ng/u1	99
30) Naphthalene	11.016	128	2924431	34.172	ng/u1	100
32) 4-Chloroaniline	11.122	127	1019875	27.004	ng/u1	100
33) Hexachlorobutadiene	11.299	225	518058	34.264	ng/u1	99
34) Caprolactam	11.910	113	313194	33.114	ng/u1	92
35) 4-Chloro-3-methylphenol	12.240	107	1005237	35.488	ng/u1	99
36) 2-Methylnaphthalene	12.616	142	1970890	34.180	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051523\
 Data File : BP015096.D
 Acq On : 16 May 2023 03:31
 Operator : CG/JU
 Sample : 02592-12MSD
 Misc :
 ALS Vial : 31 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 JREF6MSD


Quant Time: May 16 04:25:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 16 01:39:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.828	142	2042555	35.274	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.975	216	1038915	35.233	ng/ul	97
40) Hexachlorocyclopentadiene	12.951	237	614951	32.188	ng/ul	99
41) 2,4,6-Trichlorophenol	13.216	196	716590	34.416	ng/ul	98
42) 2,4,5-Trichlorophenol	13.287	196	801688	35.511	ng/ul	99
43) 1,1'-Biphenyl	13.610	154	2684440	35.093	ng/ul	99
44) 2-Chloronaphthalene	13.657	162	2092055	35.020	ng/ul	99
45) 2-Nitroaniline	13.857	65	628710	36.782	ng/ul	94
47) Dimethylphthalate	14.228	163	2654293	34.209	ng/ul	99
48) 2,6-Dinitrotoluene	14.351	165	555741	36.153	ng/ul	97
50) Acenaphthylene	14.498	152	3312683	34.843	ng/ul	100
51) 3-Nitroaniline	14.681	138	583778	42.176	ng/ul	99
52) Acenaphthene	14.840	153	2267106	34.914	ng/ul	100
53) 2,4-Dinitrophenol	14.887	184	330917	32.423	ng/ul	99
55) 4-Nitrophenol	14.981	109	373694	35.120	ng/ul	99
56) Dibenzofuran	15.175	168	3126453	34.616	ng/ul	98
57) 2,4-Dinitrotoluene	15.134	165	794229	35.706	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.398	232	658763	34.575	ng/ul	98
59) Diethylphthalate	15.587	149	2656901	34.333	ng/ul	99
61) Fluorene	15.822	166	2471292	34.048	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.810	204	1210317	34.128	ng/ul	100
63) 4-Nitroaniline	15.845	138	613514	51.327	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.898	198	460775	33.247	ng/ul	97
67) N-Nitrosodiphenylamine	16.028	169	2152176	34.350	ng/ul	100
68) 4-Bromophenyl-phenylether	16.710	248	769256	35.049	ng/ul	97
69) Hexachlorobenzene	16.828	284	896815	34.577	ng/ul	99
70) Atrazine	16.975	200	494222	21.341	ng/ul	100
71) Pentachlorophenol	17.175	266	538833	33.212	ng/ul	100
72) Phenanthrene	17.569	178	3991268	34.398	ng/ul	100
74) Anthracene	17.663	178	4017822	34.243	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.581	216	1064443	35.233	ng/ul	99
76) Pentachlorobenzene	15.092	250	1033773	34.078	ng/ul	99
77) Carbazole	17.928	167	3625957	35.242	ng/ul	100
78) Di-n-butylphthalate	18.457	149	4510201	34.040	ng/ul	100
80) Fluoranthene	19.522	202	4514191	35.996	ng/ul	100
82) Pyrene	19.875	202	4664124	36.015	ng/ul	100
83) Butylbenzylphthalate	20.727	149	1996433	34.773	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.510	252	1469691	37.720	ng/ul	99
85) Benzo(a)anthracene	21.586	228	4265476	34.773	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.486	149	2851821	34.502	ng/ul	100
87) Chrysene	21.645	228	4033796	34.400	ng/ul	99
89) Di-n-octyl phthalate	22.463	149	4735125	39.864	ng/ul	100
90) Benzo(b)fluoranthene	23.386	252	4105114	37.336	ng/ul	99
91) Benzo(k)fluoranthene	23.433	252	3923617	37.164	ng/ul	99
93) Benzo(a)pyrene	24.063	252	3376614	34.745	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.898	276	3711751	29.497	ng/ul	99
95) Dibenzo(a,h)anthracene	26.915	278	3064507	29.667	ng/ul	100
96) Benzo(g,h,i)perylene	27.739	276	2929469	28.549	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 16 04:25:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 16 01:39:21 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 13.173 ng/uL
RT: 3.470 min Scan# 14
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion: 88 Resp: 125166
Ion Ratio Lower Upper
88 100
43 28.2 20.4 30.6
58 70.8 55.4 83.2



#5
Pyridine
Concen: 31.284 ng/uL
RT: 3.893 min Scan# 86
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion: 79 Resp: 815554
Ion Ratio Lower Upper
79 100
52 63.9 55.4 83.2
51 32.2 25.4 38.2









#6
Benzaldehyde
Concen: 65.929 ng/ul
RT: 7.263 min Scan# 659
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion: 77 Resp: 578072
Ion Ratio Lower Upper
77 100
105 101.9 76.9 115.3
106 95.4 75.0 112.6



#8
Phenol
Concen: 36.250 ng/ul
RT: 7.311 min Scan# 667
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31


Tgt Ion: 94 Resp: 1185807
Ion Ratio Lower Upper
94 100
65 26.1 20.6 31.0
66 33.4 26.1 39.1








#10
Bis(2-Chloroethyl)ether
Concen: 35.144 ng/ul
RT: 7.546 min Scan# 708
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Instrument :
BNA_P
ClientSampleId :
JREF6MSD







Tgt Ion: 93 Resp: 919764
Ion Ratio Lower Upper
93 100
63 72.5 58.0 87.0
95 32.2 26.4 39.6



#12
2-Chlorophenol
Concen: 35.763 ng/ul
RT: 7.693 min Scan# 732
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion: 128 Resp: 887239
Ion Ratio Lower Upper
128 100
64 45.2 36.2 54.2
130 32.1 26.3 39.5









#13
2-Methylphenol
Concen: 35.836 ng/ul
RT: 8.581 min Scan# 883
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:108 Resp: 905997
Ion Ratio Lower Upper
108 100
107 88.2 71.8 107.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 35.892 ng/ul
RT: 8.669 min Scan# 898
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion: 45 Resp: 1176852
Ion Ratio Lower Upper
45 100
77 15.7 12.8 19.2
79 11.8 9.5 14.3









#16
Acetophenone
Concen: 37.059 ng/ul
RT: 8.969 min Scan# 949
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:105 Resp: 1470255
Ion Ratio Lower Upper
105 100
77 73.4 59.4 89.2
51 23.8 19.2 28.8



#17
N-Nitroso-di-n-propylamine
Concen: 34.804 ng/ul
RT: 8.952 min Scan# 946
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion: 70 Resp: 724153
Ion Ratio Lower Upper
70 100
42 53.0 41.0 61.6
101 9.7 7.9 11.9
130 20.5 16.1 24.1









#18
4-Methylphenol
Concen: 35.797 ng/ul
RT: 8.910 min Scan# 911
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:108 Resp: 996117
Ion Ratio Lower Upper
108 100
107 113.9 88.5 132.7



#19
Hexachloroethane
Concen: 35.045 ng/ul
RT: 9.228 min Scan# 993
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31


Tgt Ion:117 Resp: 355001
Ion Ratio Lower Upper
117 100
201 98.8 75.8 113.8
199 61.2 46.5 69.7








#22
Nitrobenzene
Concen: 34.513 ng/ul
RT: 9.352 min Scan# 1014
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Instrument :
BNA_P
ClientSampleId :
JREF6MSD







Tgt Ion: 77 Resp: 1054374
Ion Ratio Lower Upper
77 100
123 45.0 36.2 54.2
65 12.4 10.1 15.1



#23
Isophorone
Concen: 33.749 ng/ul
RT: 9.887 min Scan# 1105
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion: 82 Resp: 2119215
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.0
138 17.5 13.6 20.4









#25
2-Nitrophenol
Concen: 34.590 ng/ul
RT: 10.069 min Scan# 1136
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:139 Resp: 522055
Ion Ratio Lower Upper
139 100
65 50.0 38.4 57.6
109 28.0 22.2 33.2



#26
2,4-Dimethylphenol
Concen: 34.467 ng/ul
RT: 10.128 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:107 Resp: 1038077
Ion Ratio Lower Upper
107 100
121 51.5 41.0 61.6
122 88.0 69.5 104.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 34.353 ng/ul
 RT: 10.363 min Scan# 1187
 Delta R.T. -0.006 min
 Lab File: BP015096.D
 Acq: 16 May 2023 03:31




Instrument :
 BNA_P
 ClientSampleId :
 JREF6MSD


Tgt Ion: 93 Resp: 1312494
 Ion Ratio Lower Upper
 93 100
 95 33.1 26.7 40.1
 123 10.8 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 35.823 ng/ul
 RT: 10.610 min Scan# 1228
 Delta R.T. 0.000 min
 Lab File: BP015096.D
 Acq: 16 May 2023 03:31

Tgt Ion:162 Resp: 898677
 Ion Ratio Lower Upper
 162 100
 164 66.1 52.0 78.0
 98 36.1 28.9 43.3









#30
Naphthalene
Concen: 34.172 ng/ul
RT: 11.016 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:128 Resp: 2924431
Ion Ratio Lower Upper
128 100
129 11.2 8.9 13.3
127 13.1 10.6 15.8



#32
4-Chloroaniline
Concen: 27.004 ng/ul
RT: 11.122 min Scan# 1315
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:127 Resp: 1019875
Ion Ratio Lower Upper
127 100
129 32.0 25.8 38.8





#33

Hexachlorobutadiene

Concen: 34.264 ng/ul

RT: 11.299 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31

Instrument :

BNA_P

ClientSampleId :

JREF6MSD





Tgt Ion:225 Resp: 518058

Ion Ratio Lower Upper

225 100

223 62.9 51.0 76.6

227 65.6 52.2 78.4



#34

Caprolactam

Concen: 33.114 ng/ul

RT: 11.910 min Scan# 1449

Delta R.T. 0.006 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31




Tgt Ion:113 Resp: 313194


Ion Ratio Lower Upper

113 100

55 162.2 121.2 181.8

56 119.6 90.2 135.4









#35
4-Chloro-3-methylphenol
Concen: 35.488 ng/ul
RT: 12.240 min Scan# 1505
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:107 Resp: 1005237
Ion Ratio Lower Upper
107 100
144 26.2 20.3 30.5
142 79.9 63.2 94.8



#36
2-Methylnaphthalene
Concen: 34.180 ng/ul
RT: 12.616 min Scan# 1569
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:142 Resp: 1970890
Ion Ratio Lower Upper
142 100
141 89.3 71.0 106.6









#37
1-Methylnaphthalene
Concen: 35.274 ng/ul
RT: 12.828 min Scan# 1605
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:142 Resp: 2042555
Ion Ratio Lower Upper
142 100
141 92.2 74.6 112.0
116 3.8 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 35.233 ng/ul
RT: 12.975 min Scan# 1630
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:216 Resp: 1038915
Ion Ratio Lower Upper
216 100
214 79.2 61.4 92.0
179 20.5 15.9 23.9
108 18.0 13.7 20.5









#40
Hexachlorocyclopentadiene
Concen: 32.188 ng/ul
RT: 12.951 min Scan# 1626
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:237 Resp: 614951
Ion Ratio Lower Upper
237 100
235 61.3 49.4 74.0
272 12.4 10.4 15.6



#41
2,4,6-Trichlorophenol
Concen: 34.416 ng/ul
RT: 13.216 min Scan# 1671
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:196 Resp: 716590
Ion Ratio Lower Upper
196 100
198 96.3 75.5 113.3
200 31.4 23.6 35.4





#42

2,4,5-Trichlorophenol

Concen: 35.511 ng/ul

RT: 13.287 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31

Instrument :

BNA_P

ClientSampleId :

JREF6MSD





Tgt Ion:196 Resp: 801688

Ion Ratio Lower Upper

196 100

198 97.1 77.0 115.4

200 31.6 24.3 36.5



#43

1,1'-Biphenyl

Concen: 35.093 ng/ul

RT: 13.610 min Scan# 1738

Delta R.T. -0.006 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31




Tgt Ion:154 Resp: 2684440


Ion Ratio Lower Upper

154 100

153 40.1 32.7 49.1

76 13.6 10.7 16.1









#44
2-Chloronaphthalene
Concen: 35.020 ng/ul
RT: 13.657 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:162 Resp: 2092055
Ion Ratio Lower Upper
162 100
164 32.4 25.9 38.9
127 37.6 30.8 46.2



#45
2-Nitroaniline
Concen: 36.782 ng/ul
RT: 13.857 min Scan# 1780
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion: 65 Resp: 628710
Ion Ratio Lower Upper
65 100
92 72.1 59.8 89.8
138 106.3 92.3 138.5









#47
Dimethylphthalate
Concen: 34.209 ng/ul
RT: 14.228 min Scan# 1843
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:163 Resp: 2654293
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.0 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 36.153 ng/ul
RT: 14.351 min Scan# 1864
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:165 Resp: 555741
Ion Ratio Lower Upper
165 100
89 57.6 48.0 72.0
121 20.4 17.5 26.3









#50
Acenaphthylene
Concen: 34.843 ng/ul
RT: 14.498 min Scan# 1890
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:152 Resp: 3312683
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1
153 13.2 10.3 15.5



#51
3-Nitroaniline
Concen: 42.176 ng/ul
RT: 14.681 min Scan# 1920
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:138 Resp: 583778
Ion Ratio Lower Upper
138 100
108 10.5 8.8 13.2
92 117.6 93.4 140.0





#52

Acenaphthene

Concen: 34.914 ng/ul

RT: 14.840 min Scan# 1948

Delta R.T. -0.006 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31

Instrument :

BNA_P

ClientSampleId :

JREF6MSD





Tgt Ion:153 Resp: 2267106

Ion Ratio Lower Upper

153 100

152 47.9 38.0 57.0

154 88.8 71.0 106.4



#53

2,4-Dinitrophenol

Concen: 32.423 ng/ul

RT: 14.887 min Scan# 1955

Delta R.T. -0.006 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31




Tgt Ion:184 Resp: 330917


Ion Ratio Lower Upper

184 100

63 63.6 49.7 74.5

154 58.9 47.5 71.3









#55
4-Nitrophenol
Concen: 35.120 ng/ul
RT: 14.981 min Scan# 1971
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:109 Resp: 373694
Ion Ratio Lower Upper
109 100
139 144.6 115.3 172.9
65 133.6 107.9 161.9



#56
Dibenzofuran
Concen: 34.616 ng/ul
RT: 15.175 min Scan# 2004
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:168 Resp: 3126453
Ion Ratio Lower Upper
168 100
139 35.6 29.4 44.0









#57
2,4-Dinitrotoluene
Concen: 35.706 ng/ul
RT: 15.134 min Scan# 1997
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:165 Resp: 794229
Ion Ratio Lower Upper
165 100
63 42.2 33.1 49.7
182 2.6 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 34.575 ng/ul
RT: 15.398 min Scan# 2042
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:232 Resp: 658763
Ion Ratio Lower Upper
232 100
131 43.6 36.8 55.2
130 2.2 1.9 2.9
166 30.5 24.6 37.0









#59
Diethylphthalate
Concen: 34.333 ng/ul
RT: 15.587 min Scan# 2074
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:149 Resp: 2656901
Ion Ratio Lower Upper
149 100
177 22.5 17.5 26.3
150 12.4 9.8 14.8



#61
Fluorene
Concen: 34.048 ng/ul
RT: 15.822 min Scan# 2114
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:166 Resp: 2471292
Ion Ratio Lower Upper
166 100
165 97.5 77.8 116.6
167 13.1 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 34.128 ng/ul
RT: 15.810 min Scan# 2112
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:204 Resp: 1210317
Ion Ratio Lower Upper
204 100
206 33.0 26.1 39.1
141 55.5 44.5 66.7



#63
4-Nitroaniline
Concen: 51.327 ng/ul
RT: 15.845 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:138 Resp: 613514
Ion Ratio Lower Upper
138 100
92 51.7 41.2 61.8
108 74.0 60.1 90.1









#66
4,6-Dinitro-2-methylphenol
Concen: 33.247 ng/ul
RT: 15.898 min Scan# 2128
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:198 Resp: 460775
Ion Ratio Lower Upper
198 100
182 3.8 3.7 5.5
77 25.0 21.2 31.8



#67
N-Nitrosodiphenylamine
Concen: 34.350 ng/ul
RT: 16.028 min Scan# 2149
Delta R.T. 0.000 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:169 Resp: 2152176
Ion Ratio Lower Upper
169 100
168 67.1 53.5 80.3
167 36.6 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 35.049 ng/ul
RT: 16.710 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:248 Resp: 769256
Ion Ratio Lower Upper
248 100
250 97.6 79.0 118.4
141 68.2 58.6 87.8



#69
Hexachlorobenzene
Concen: 34.577 ng/ul
RT: 16.828 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:284 Resp: 896815
Ion Ratio Lower Upper
284 100
142 33.5 27.3 40.9
249 31.1 24.6 36.8









#70
Atrazine
Concen: 21.341 ng/ul
RT: 16.975 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:200 Resp: 494222
Ion Ratio Lower Upper
200 100
173 26.7 21.4 32.2
215 48.0 38.1 57.1



#71
Pentachlorophenol
Concen: 33.212 ng/ul
RT: 17.175 min Scan# 2344
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:266 Resp: 538833
Ion Ratio Lower Upper
266 100
264 63.7 51.0 76.4
268 64.5 51.8 77.6





#72

Phenanthrene

Concen: 34.398 ng/ul

RT: 17.569 min Scan# 2411

Delta R.T. -0.006 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31

Instrument :

BNA_P

ClientSampleId :

JREF6MSD





Tgt Ion:178 Resp: 3991268

Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.2

176 19.5 15.6 23.4



#74

Anthracene

Concen: 34.243 ng/ul

RT: 17.663 min Scan# 2427

Delta R.T. -0.006 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31




Tgt Ion:178 Resp: 4017822


Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.2

176 19.0 15.2 22.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 35.233 ng/uL

RT: 13.581 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31

Instrument :

BNA_P

ClientSampleId :

JREF6MSD





Tgt Ion:216 Resp: 1064443

Ion Ratio Lower Upper

216 100

214 76.6 62.2 93.4

218 48.0 38.2 57.2



#76

Pentachlorobenzene

Concen: 34.078 ng/uL

RT: 15.092 min Scan# 1990

Delta R.T. -0.006 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31




Tgt Ion:250 Resp: 1033773


Ion Ratio Lower Upper

250 100

248 63.3 49.8 74.6

252 64.8 51.6 77.4









#77
 Carbazole
 Concen: 35.242 ng/ul
 RT: 17.928 min Scan# 2473
 Delta R.T. -0.006 min
 Lab File: BP015096.D
 Acq: 16 May 2023 03:31




Instrument :
 BNA_P
 ClientSampleId :
 JREF6MSD


Tgt Ion:167 Resp: 3625957
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.2
 139 12.7 10.2 15.4



#78
 Di-n-butylphthalate
 Concen: 34.040 ng/ul
 RT: 18.457 min Scan# 2562
 Delta R.T. -0.006 min
 Lab File: BP015096.D
 Acq: 16 May 2023 03:31

Tgt Ion:149 Resp: 4510201
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.3 10.9
 104 5.0 3.9 5.9









#80
Fluoranthene
Concen: 35.996 ng/ul
RT: 19.522 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:202 Resp: 4514191
Ion Ratio Lower Upper
202 100
101 11.9 9.5 14.3
100 9.3 7.3 10.9



#82
Pyrene
Concen: 36.015 ng/ul
RT: 19.875 min Scan# 2803
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:202 Resp: 4664124
Ion Ratio Lower Upper
202 100
101 14.0 11.2 16.8
100 11.6 9.0 13.6





#83

Butylbenzylphthalate

Concen: 34.773 ng/ul

RT: 20.727 min Scan# 2949

Delta R.T. -0.006 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31

Instrument :

BNA_P

ClientSampleId :

JREF6MSD





Tgt Ion:149 Resp: 1996433

Ion Ratio Lower Upper

149 100

91 70.3 57.4 86.2

206 20.8 16.6 25.0



#84

3,3'-Dichlorobenzidine

Concen: 37.720 ng/ul

RT: 21.510 min Scan# 3081

Delta R.T. -0.011 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31




Tgt Ion:252 Resp: 1469691


Ion Ratio Lower Upper

252 100

254 64.3 50.8 76.2

126 11.6 9.8 14.8





#85

Benzo(a)anthracene

Concen: 34.773 ng/ul

RT: 21.586 min Scan# 3094

Delta R.T. -0.011 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31

Instrument :

BNA_P

ClientSampleId :

JREF6MSD





Tgt Ion:228 Resp: 4265476

Ion Ratio Lower Upper

228 100

229 19.9 16.1 24.1

226 27.3 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 34.502 ng/ul

RT: 21.486 min Scan# 3077

Delta R.T. -0.006 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31




Tgt Ion:149 Resp: 2851821


Ion Ratio Lower Upper

149 100

167 27.2 21.8 32.6

279 4.2 3.2 4.8





#87

Chrysene

Concen: 34.400 ng/ul

RT: 21.645 min Scan# 3105

Delta R.T. -0.006 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31

Instrument :

BNA_P

ClientSampleId :

JREF6MSD





Tgt Ion:228 Resp: 4033796

Ion Ratio Lower Upper

228 100

226 30.1 25.0 37.6

229 20.0 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 39.864 ng/ul




RT: 22.463 min Scan# 3243


Delta R.T. -0.011 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31

Tgt Ion:149 Resp: 4735125









#90
Benzo(b)fluoranthene
Concen: 37.336 ng/ul
RT: 23.386 min Scan# 3400
Delta R.T. -0.006 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31




Instrument :
BNA_P
ClientSampleId :
JREF6MSD


Tgt Ion:252 Resp: 4105114
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.2
125 9.7 8.3 12.5



#91
Benzo(k)fluoranthene
Concen: 37.164 ng/ul
RT: 23.433 min Scan# 3408
Delta R.T. -0.011 min
Lab File: BP015096.D
Acq: 16 May 2023 03:31

Tgt Ion:252 Resp: 3923617
Ion Ratio Lower Upper
252 100
253 21.5 17.5 26.3
125 10.7 8.5 12.7





#93

Benzo(a)pyrene

Concen: 34.745 ng/ul

RT: 24.063 min Scan# 31

Delta R.T. -0.011 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31

Instrument :

BNA_P

ClientSampleId :

JREF6MSD





Tgt Ion:252 Resp: 3376614

Ion Ratio Lower Upper

252 100

253 21.5 17.2 25.8

125 11.6 9.1 13.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 29.497 ng/ul

RT: 26.898 min Scan# 3997

Delta R.T. -0.017 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31




Tgt Ion:276 Resp: 3711751


Ion Ratio Lower Upper

276 100

138 24.6 20.1 30.1

277 24.4 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 29.667 ng/ul

RT: 26.915 min Scan# 4000

Delta R.T. -0.017 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31

Instrument :

BNA_P

ClientSampleId :

JREF6MSD





Tgt Ion:278 Resp: 3064507

Ion Ratio Lower Upper

278 100

139 16.6 13.3 19.9

279 23.7 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 28.549 ng/ul

RT: 27.739 min Scan# 4140

Delta R.T. -0.017 min

Lab File: BP015096.D

Acq: 16 May 2023 03:31

Tgt Ion:276 Resp: 2929469

Ion Ratio Lower Upper

276 100

138 21.6 17.7 26.5

277 23.4 18.8 28.2

