

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051623\
 Data File : BP015110.D
 Acq On : 16 May 2023 15:45
 Operator : CG/JU
 Sample : 02768-11
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CONN8

Quant Time: May 17 00:35:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 16 01:39:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.134	152	360044	20.000	ng/u1	0.00
20) Naphthalene-d8	10.963	136	1626193	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.775	164	1044492	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.528	188	2275877	20.000	ng/u1	0.00
79) Chrysene-d12	21.604	240	1701615	20.000	ng/u1	0.00
88) Perylene-d12	24.174	264	1491511	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	47187	5.092	ng/uL	0.00
4) Pyridine-d5	3.875	84	310640	11.826	ng/u1	0.00
7) Phenol-d5	7.281	99	253070	7.570	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.452	67	680295	35.236	ng/u1	0.00
11) 2-Chlorophenol-d4	7.658	132	668071	27.180	ng/u1	0.00
15) 4-Methylphenol-d8	8.840	113	494194	18.552	ng/u1	-0.01
21) Nitrobenzene-d5	9.310	128	439988	34.432	ng/u1	0.00
24) 2-Nitrophenol-d4	10.040	143	463439	32.594	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.575	165	779318	30.375	ng/u1	-0.01
31) 4-Chloroaniline-d4	11.099	131	1012759	26.679	ng/u1	0.00
46) Dimethylphthalate-d6	14.175	166	2936509	37.205	ng/u1	0.00
49) Acenaphthylene-d8	14.469	160	3342494	37.028	ng/u1	0.00
54) 4-Nitrophenol-d4	14.957	143	102703	6.701	ng/u1	-0.01
60) Fluorene-d10	15.763	176	2493427	37.439	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.881	200	455893	31.786	ng/u1	0.00
73) Anthracene-d10	17.628	188	4032018	38.690	ng/u1	0.00
81) Pyrene-d10	19.845	212	4463638	45.229	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.010	264	3058044	41.303	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051623\
Data File : BP015110.D
Acq On : 16 May 2023 15:45
Operator : CG/JU
Sample : 02768-11
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CONN8

Quant Time: May 17 00:35:38 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 16 01:39:21 2023
Response via : Initial Calibration

