

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051623\
 Data File : BP051114.D
 Acq On : 16 May 2023 18:01
 Operator : CG/JU
 Sample : 02768-15
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CONP5

Quant Time: May 17 00:36:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 16 01:39:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.134	152	323158	20.000	ng/u1	0.00
20) Naphthalene-d8	10.963	136	1478379	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.775	164	961522	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.528	188	2085455	20.000	ng/u1	0.00
79) Chrysene-d12	21.604	240	1730682	20.000	ng/u1	0.00
88) Perylene-d12	24.168	264	1714049	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	39869	4.793	ng/uL	0.00
4) Pyridine-d5	3.881	84	171992	7.295	ng/u1	0.00
7) Phenol-d5	7.281	99	224234	7.473	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.452	67	620155	35.787	ng/u1	0.00
11) 2-Chlorophenol-d4	7.658	132	598532	27.131	ng/u1	0.00
15) 4-Methylphenol-d8	8.840	113	440016	18.404	ng/u1	-0.01
21) Nitrobenzene-d5	9.310	128	402089	34.612	ng/u1	0.00
24) 2-Nitrophenol-d4	10.034	143	431819	33.407	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.575	165	716578	30.722	ng/u1	-0.01
31) 4-Chloroaniline-d4	11.099	131	961772	27.869	ng/u1	0.00
46) Dimethylphthalate-d6	14.175	166	2729790	37.570	ng/u1	0.00
49) Acenaphthylene-d8	14.469	160	3091264	37.200	ng/u1	0.00
54) 4-Nitrophenol-d4	14.957	143	88363	6.263	ng/u1	-0.01
60) Fluorene-d10	15.763	176	2345912	38.263	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.881	200	447053	34.015	ng/u1	0.00
73) Anthracene-d10	17.628	188	3967108	41.543	ng/u1	0.00
81) Pyrene-d10	19.845	212	4492004	44.752	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.010	264	3728874	43.824	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051623\
Data File : BP015114.D
Acq On : 16 May 2023 18:01
Operator : CG/JU
Sample : 02768-15
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CONP5

Quant Time: May 17 00:36:18 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 16 01:39:21 2023
Response via : Initial Calibration

