

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051624\
 Data File : BP020407.D
 Acq On : 16 May 2024 10:37
 Operator : MA/JU
 Sample : PB160893BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK893

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/16/2024
 Supervised By :mohammad ahmed 05/18/2024

Quant Time: May 16 23:38:18 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 16 10:48:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.593	152	63751	20.000	ng/u1	0.00
20) Naphthalene-d8	10.363	136	260821	20.000	ng/u1	0.01
38) Acenaphthene-d10	14.269	164	178976	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.122	188	400707	20.000	ng/u1	0.00
79) Chrysene-d12	21.616	240	450196	20.000	ng/u1	-0.01
88) Perylene-d12	24.963	264	523121	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	7991	5.294	ng/uL	0.00
4) Pyridine-d5	3.599	84	95523	21.779	ng/u1	0.00
7) Phenol-d5	6.787	99	134186	24.509	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.946	67	82495	24.805	ng/u1	0.00
11) 2-Chlorophenol-d4	7.134	132	111349	28.450	ng/u1	0.00
15) 4-Methylphenol-d8	8.305	113	107784	24.351	ng/u1	0.00
21) Nitrobenzene-d5	8.758	128	51300	26.266	ng/u1	0.01
24) 2-Nitrophenol-d4	9.469	143	56025	25.049	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	10.016	165	99808m	24.454	ng/u1	0.02
31) 4-Chloroaniline-d4	10.540	131	135078	23.647	ng/u1	0.02
46) Dimethylphthalate-d6	13.693	166	364862	27.916	ng/u1	0.00
49) Acenaphthylene-d8	13.957	160	387699	26.853	ng/u1	0.00
54) 4-Nitrophenol-d4	14.587	143	39720m	19.774	ng/u1	0.04
60) Fluorene-d10	15.298	176	311536	28.578	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.475	200	48433	19.047	ng/u1	0.00
73) Anthracene-d10	17.222	188	518569	29.915	ng/u1	-0.01
81) Pyrene-d10	19.628	212	671757	25.526	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.733	264	739675	29.244	ng/u1	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051624\
Data File : BP020407.D
Acq On : 16 May 2024 10:37
Operator : MA/JU
Sample : PB160893BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK893

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/16/2024
Supervised By :mohammad ahmed 05/18/2024

Quant Time: May 16 23:38:18 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 16 10:48:40 2024
Response via : Initial Calibration

