

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051723\
 Data File : BP015131.D
 Acq On : 17 May 2023 18:06
 Operator : CG/JU
 Sample : 02768-01RX
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CONM7

Quant Time: May 18 01:12:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 18 01:09:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.128	152	258124	20.000	ng/ul	0.00
20) Naphthalene-d8	10.957	136	1266527	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.769	164	829059	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.522	188	1788752	20.000	ng/ul	0.00
79) Chrysene-d12	21.598	240	1568351	20.000	ng/ul	0.00
88) Perylene-d12	24.163	264	1468393	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	35540	5.350	ng/uL	0.00
4) Pyridine-d5	3.881	84	124842	6.629	ng/ul	0.01
7) Phenol-d5	7.275	99	148848	6.211	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.446	67	482553	34.862	ng/ul	0.00
11) 2-Chlorophenol-d4	7.652	132	437230	24.813	ng/ul	0.00
15) 4-Methylphenol-d8	8.840	113	294699	15.432	ng/ul	0.00
21) Nitrobenzene-d5	9.305	128	302527	30.398	ng/ul	0.00
24) 2-Nitrophenol-d4	10.034	143	309922	27.987	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.575	165	516253	25.835	ng/ul	0.00
31) 4-Chloroaniline-d4	11.093	131	668787	22.621	ng/ul	0.00
46) Dimethylphthalate-d6	14.175	166	2099725	33.516	ng/ul	0.00
49) Acenaphthylene-d8	14.463	160	2378035	33.189	ng/ul	0.00
54) 4-Nitrophenol-d4	14.957	143	44436	3.653	ng/ul	0.00
60) Fluorene-d10	15.763	176	1795078	33.957	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.875	200	271844	24.115	ng/ul	0.00
73) Anthracene-d10	17.622	188	2941910	35.917	ng/ul	0.00
81) Pyrene-d10	19.839	212	3326494	36.570	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.998	264	2671876	36.655	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051723\
Data File : BP015131.D
Acq On : 17 May 2023 18:06
Operator : CG/JU
Sample : 02768-01RX
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CONM7

Quant Time: May 18 01:12:30 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 18 01:09:56 2023
Response via : Initial Calibration

