

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051823\
 Data File : BP051823.D
 Acq On : 18 May 2023 09:04
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020717

Quant Time: May 19 00:40:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 18 01:09:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.128	152	322088	20.000	ng/u1	0.00
20) Naphthalene-d8	10.963	136	1460379	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.775	164	904851	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.528	188	1927092	20.000	ng/u1	0.00
79) Chrysene-d12	21.604	240	1681851	20.000	ng/u1	0.00
88) Perylene-d12	24.168	264	2170396	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	68206	8.228	ng/uL	0.00
4) Pyridine-d5	3.870	84	511364	21.762	ng/u1	0.00
7) Phenol-d5	7.275	99	602019	20.131	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.446	67	384859	22.283	ng/u1	0.00
11) 2-Chlorophenol-d4	7.652	132	455795	20.729	ng/u1	0.00
15) 4-Methylphenol-d8	8.840	113	503793	21.142	ng/u1	0.00
21) Nitrobenzene-d5	9.305	128	234627	20.446	ng/u1	0.00
24) 2-Nitrophenol-d4	10.034	143	256341	20.076	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.575	165	478439	20.765	ng/u1	0.00
31) 4-Chloroaniline-d4	11.099	131	740704	21.728	ng/u1	0.00
46) Dimethylphthalate-d6	14.175	166	1455161	21.282	ng/u1	0.00
49) Acenaphthylene-d8	14.469	160	1705066	21.803	ng/u1	0.00
54) 4-Nitrophenol-d4	14.963	143	271396	20.441	ng/u1	0.00
60) Fluorene-d10	15.763	176	1256577	21.779	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.881	200	231470	19.059	ng/u1	0.00
73) Anthracene-d10	17.628	188	1932444	21.899	ng/u1	0.00
81) Pyrene-d10	19.845	212	2114417	21.676	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.004	264	2268256	21.053	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.470	88	75476	8.473	ng/uL	98
5) Pyridine	3.893	79	533347	21.824	ng/u1	97
6) Benzaldehyde	7.258	77	158706	19.308	ng/u1	99
8) Phenol	7.305	94	644002	21.001	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.540	93	537448	21.906	ng/u1	98
12) 2-Chlorophenol	7.687	128	497765	21.403	ng/u1	98
13) 2-Methylphenol	8.575	108	497189	20.978	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.658	45	442702	14.402	ng/u1	97
16) Acetophenone	8.963	105	859076	23.099	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.946	70	437392	22.424	ng/u1	96
18) 4-Methylphenol	8.905	108	565784	21.689	ng/u1	100
19) Hexachloroethane	9.222	117	204033	21.486	ng/u1	98
22) Nitrobenzene	9.352	77	638117	22.515	ng/u1	99
23) Isophorone	9.881	82	1239464	21.276	ng/u1	99
25) 2-Nitrophenol	10.069	139	290804	20.769	ng/u1	94
26) 2,4-Dimethylphenol	10.122	107	615491	22.028	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.363	93	769198	21.701	ng/u1	98
29) 2,4-Dichlorophenol	10.604	162	491708	21.127	ng/u1	99
30) Naphthalene	11.010	128	1736394	21.870	ng/u1	99
32) 4-Chloroaniline	11.122	127	778748	22.226	ng/u1	100
33) Hexachlorobutadiene	11.293	225	288305	20.554	ng/u1	97
34) Caprolactam	11.899	113	104997	11.966	ng/u1	87
35) 4-Chloro-3-methylphenol	12.234	107	548859	20.886	ng/u1	99
36) 2-Methylnaphthalene	12.610	142	1159453	21.674	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051823\
 Data File : BP015136.D
 Acq On : 18 May 2023 09:04
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020717


Quant Time: May 19 00:40:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 18 01:09:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.828	142	1171817	21.813	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.975	216	587060	21.693	ng/ul	98
40) Hexachlorocyclopentadiene	12.951	237	317414	18.103	ng/ul	98
41) 2,4,6-Trichlorophenol	13.210	196	395192	20.680	ng/ul	99
42) 2,4,5-Trichlorophenol	13.281	196	429614	20.735	ng/ul	99
43) 1,1'-Biphenyl	13.610	154	1565664	22.301	ng/ul	99
44) 2-Chloronaphthalene	13.651	162	1219925	22.251	ng/ul	99
45) 2-Nitroaniline	13.857	65	355568	22.666	ng/ul	91
47) Dimethylphthalate	14.222	163	1531485	21.507	ng/ul	100
48) 2,6-Dinitrotoluene	14.345	165	293158	20.780	ng/ul	97
50) Acenaphthylene	14.498	152	1942804	22.266	ng/ul	100
51) 3-Nitroaniline	14.675	138	224721	17.690	ng/ul	90
52) Acenaphthene	14.840	153	1331511	22.343	ng/ul	100
53) 2,4-Dinitrophenol	14.887	184	145704	15.555	ng/ul	93
55) 4-Nitrophenol	14.975	109	216331	22.153	ng/ul	94
56) Dibenzofuran	15.169	168	1841487	22.216	ng/ul	100
57) 2,4-Dinitrotoluene	15.134	165	431987	21.161	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.392	232	357357	20.436	ng/ul	94
59) Diethylphthalate	15.587	149	1595212	22.461	ng/ul	98
61) Fluorene	15.822	166	1499019	22.503	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.810	204	713218	21.913	ng/ul	98
63) 4-Nitroaniline	15.839	138	190952	17.407	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.898	198	256531	20.467	ng/ul	97
67) N-Nitrosodiphenylamine	16.022	169	1250188	22.063	ng/ul	99
68) 4-Bromophenyl-phenylether	16.710	248	433969	21.863	ng/ul	99
69) Hexachlorobenzene	16.828	284	504806	21.521	ng/ul	99
70) Atrazine	16.975	200	445168	21.255	ng/ul	98
71) Pentachlorophenol	17.175	266	298728	20.359	ng/ul	99
72) Phenanthrene	17.569	178	2336576	22.266	ng/ul	100
74) Anthracene	17.663	178	2388449	22.508	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.575	216	600693	21.985	ng/ul	99
76) Pentachlorobenzene	15.092	250	589442	21.485	ng/ul	99
77) Carbazole	17.928	167	2099072	22.559	ng/ul	99
78) Di-n-butylphthalate	18.457	149	2505601	20.910	ng/ul	100
80) Fluoranthene	19.522	202	2604087	21.903	ng/ul	99
82) Pyrene	19.875	202	2755583	22.444	ng/ul	99
83) Butylbenzylphthalate	20.727	149	1081289	19.866	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.510	252	737013	19.953	ng/ul	99
85) Benzo(a)anthracene	21.586	228	2517815	21.651	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.486	149	1540645	19.661	ng/ul	99
87) Chrysene	21.645	228	2398387	21.575	ng/ul	100
89) Di-n-octyl phthalate	22.463	149	2589954	17.424	ng/ul	100
90) Benzo(b)fluoranthene	23.380	252	2428711	17.652	ng/ul	99
91) Benzo(k)fluoranthene	23.433	252	2420907	18.324	ng/ul	100
93) Benzo(a)pyrene	24.057	252	2543148	20.911	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.892	276	2293701	14.566	ng/ul	100
95) Dibenzo(a,h)anthracene	26.910	278	1909817	14.774	ng/ul	100
96) Benzo(g,h,i)perylene	27.727	276	1795029	13.979	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 19 00:40:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 18 01:09:56 2023
Response via : Initial Calibration





#2

1,4-Dioxane

Concen: 8.473 ng/uL

RT: 3.470 min Scan# 14

Delta R.T. 0.000 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion: 88 Resp: 75476

Ion Ratio Lower Upper

88 100

43 28.5 20.4 30.6

58 70.2 55.4 83.2



#5

Pyridine

Concen: 21.824 ng/uL

RT: 3.893 min Scan# 86

Delta R.T. 0.000 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04




Tgt Ion: 79 Resp: 533347


Ion Ratio Lower Upper

79 100

52 71.4 55.4 83.2

51 34.5 25.4 38.2









#6
Benzaldehyde
Concen: 19.308 ng/ul
RT: 7.258 min Scan# 658
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion: 77 Resp: 158706
Ion Ratio Lower Upper
77 100
105 97.0 76.9 115.3
106 94.9 75.0 112.6



#8
Phenol
Concen: 21.001 ng/ul
RT: 7.305 min Scan# 666
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion: 94 Resp: 644002
Ion Ratio Lower Upper
94 100
65 26.5 20.6 31.0
66 33.1 26.1 39.1









#10
Bis(2-Chloroethyl)ether
Concen: 21.906 ng/ul
RT: 7.540 min Scan# 706
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion: 93 Resp: 537448
Ion Ratio Lower Upper
93 100
63 73.3 58.0 87.0
95 30.7 26.4 39.6



#12
2-Chlorophenol
Concen: 21.403 ng/ul
RT: 7.687 min Scan# 731
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:128 Resp: 497765
Ion Ratio Lower Upper
128 100
64 46.8 36.2 54.2
130 32.3 26.3 39.5





#13

2-Methylphenol

Concen: 20.978 ng/ul

RT: 8.575 min Scan# 882

Delta R.T. 0.000 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :





SSTD020717

Tgt Ion:108 Resp: 497189

Ion Ratio Lower Upper

108 100

107 89.4 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 14.402 ng/ul

RT: 8.658 min Scan# 896

Delta R.T. 0.000 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04




Tgt Ion: 45 Resp: 442702


Ion Ratio Lower Upper

45 100

77 14.6 12.8 19.2

79 11.0 9.5 14.3









#16
Acetophenone
Concen: 23.099 ng/ul
RT: 8.963 min Scan# 948
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:105 Resp: 859076
Ion Ratio Lower Upper
105 100
77 77.8 59.4 89.2
51 26.0 19.2 28.8



#17
N-Nitroso-di-n-propylamine
Concen: 22.424 ng/ul
RT: 8.946 min Scan# 945
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion: 70 Resp: 437392
Ion Ratio Lower Upper
70 100
42 55.1 41.0 61.6
101 9.3 7.9 11.9
130 19.8 16.1 24.1









#18
4-Methylphenol
Concen: 21.689 ng/ul
RT: 8.905 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:108 Resp: 565784
Ion Ratio Lower Upper
108 100
107 110.6 88.5 132.7



#19
Hexachloroethane
Concen: 21.486 ng/ul
RT: 9.222 min Scan# 992
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:117 Resp: 204033
Ion Ratio Lower Upper
117 100
201 91.7 75.8 113.8
199 57.3 46.5 69.7









#22
Nitrobenzene
Concen: 22.515 ng/ul
RT: 9.352 min Scan# 1014
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion: 77 Resp: 638117
Ion Ratio Lower Upper
77 100
123 44.0 36.2 54.2
65 12.7 10.1 15.1



#23
Isophorone
Concen: 21.276 ng/ul
RT: 9.881 min Scan# 1104
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion: 82 Resp: 1239464
Ion Ratio Lower Upper
82 100
95 6.3 5.4 8.0
138 16.7 13.6 20.4









#25
2-Nitrophenol
Concen: 20.769 ng/ul
RT: 10.069 min Scan# 1136
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:139 Resp: 290804
Ion Ratio Lower Upper
139 100
65 52.8 38.4 57.6
109 29.7 22.2 33.2



#26
2,4-Dimethylphenol
Concen: 22.028 ng/ul
RT: 10.122 min Scan# 1145
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:107 Resp: 615491
Ion Ratio Lower Upper
107 100
121 50.2 41.0 61.6
122 86.1 69.5 104.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.701 ng/ul
 RT: 10.363 min Scan# 1186
 Delta R.T. 0.006 min
 Lab File: BP015136.D
 Acq: 18 May 2023 09:04




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020717


Tgt Ion: 93 Resp: 769198
 Ion Ratio Lower Upper
 93 100
 95 31.9 26.7 40.1
 123 10.8 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 21.127 ng/ul
 RT: 10.604 min Scan# 1227
 Delta R.T. 0.000 min
 Lab File: BP015136.D
 Acq: 18 May 2023 09:04

Tgt Ion:162 Resp: 491708
 Ion Ratio Lower Upper
 162 100
 164 64.4 52.0 78.0
 98 37.0 28.9 43.3









#30
Naphthalene
Concen: 21.870 ng/ul
RT: 11.010 min Scan# 1296
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:128 Resp: 1736394
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 13.0 10.6 15.8



#32
4-Chloroaniline
Concen: 22.226 ng/ul
RT: 11.122 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:127 Resp: 778748
Ion Ratio Lower Upper
127 100
129 32.5 25.8 38.8





#33

Hexachlorobutadiene

Concen: 20.554 ng/ul

RT: 11.293 min Scan# 1344

Delta R.T. 0.000 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion:225 Resp: 288305

Ion Ratio Lower Upper

225 100

223 60.6 51.0 76.6

227 63.7 52.2 78.4



#34

Caprolactam

Concen: 11.966 ng/ul

RT: 11.899 min Scan# 1447

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04




Tgt Ion:113 Resp: 104997


Ion Ratio Lower Upper

113 100

55 169.8 121.2 181.8

56 125.3 90.2 135.4









#35
4-Chloro-3-methylphenol
Concen: 20.886 ng/ul
RT: 12.234 min Scan# 1504
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:107 Resp: 548859
Ion Ratio Lower Upper
107 100
144 25.2 20.3 30.5
142 77.4 63.2 94.8



#36
2-Methylnaphthalene
Concen: 21.674 ng/ul
RT: 12.610 min Scan# 1568
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:142 Resp: 1159453
Ion Ratio Lower Upper
142 100
141 89.4 71.0 106.6





#37

1-Methylnaphthalene

Concen: 21.813 ng/ul

RT: 12.828 min Scan# 1605

Delta R.T. 0.000 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion:142 Resp: 1171817

Ion Ratio Lower Upper

142 100

141 93.7 74.6 112.0

116 3.7 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.693 ng/ul

RT: 12.975 min Scan# 1630

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Tgt Ion:216 Resp: 587060




Ion Ratio Lower Upper


216 100

214 78.8 61.4 92.0

179 20.6 15.9 23.9

108 18.5 13.7 20.5









#40
Hexachlorocyclopentadiene
Concen: 18.103 ng/ul
RT: 12.951 min Scan# 1626
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:237 Resp: 317414
Ion Ratio Lower Upper
237 100
235 63.1 49.4 74.0
272 13.5 10.4 15.6



#41
2,4,6-Trichlorophenol
Concen: 20.680 ng/ul
RT: 13.210 min Scan# 1670
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:196 Resp: 395192
Ion Ratio Lower Upper
196 100
198 93.7 75.5 113.3
200 31.5 23.6 35.4





#42

2,4,5-Trichlorophenol

Concen: 20.735 ng/ul

RT: 13.281 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion:196 Resp: 429614

Ion Ratio Lower Upper

196 100

198 95.9 77.0 115.4

200 31.7 24.3 36.5



#43

1,1'-Biphenyl

Concen: 22.301 ng/ul

RT: 13.610 min Scan# 1738

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04




Tgt Ion:154 Resp: 1565664


Ion Ratio Lower Upper

154 100

153 40.1 32.7 49.1

76 14.2 10.7 16.1





#44

2-Chloronaphthalene

Concen: 22.251 ng/ul

RT: 13.651 min Scan# 1745

Delta R.T. 0.000 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion:162 Resp: 1219925

Ion Ratio Lower Upper

162 100

164 32.6 25.9 38.9

127 38.9 30.8 46.2



#45

2-Nitroaniline

Concen: 22.666 ng/ul

RT: 13.857 min Scan# 1780

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04




Tgt Ion: 65 Resp: 355568


Ion Ratio Lower Upper

65 100

92 70.7 59.8 89.8

138 102.0 92.3 138.5









#47
Dimethylphthalate
Concen: 21.507 ng/ul
RT: 14.222 min Scan# 1842
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:163 Resp: 1531485
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.2 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 20.780 ng/ul
RT: 14.345 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:165 Resp: 293158
Ion Ratio Lower Upper
165 100
89 62.4 48.0 72.0
121 23.1 17.5 26.3









#50
Acenaphthylene
Concen: 22.266 ng/ul
RT: 14.498 min Scan# 1889
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:152 Resp: 1942804
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1
153 13.2 10.3 15.5



#51
3-Nitroaniline
Concen: 17.690 ng/ul
RT: 14.675 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:138 Resp: 224721
Ion Ratio Lower Upper
138 100
108 11.0 8.8 13.2
92 128.4 93.4 140.0









#52
Acenaphthene
Concen: 22.343 ng/ul
RT: 14.840 min Scan# 1947
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:153 Resp: 1331511
Ion Ratio Lower Upper
153 100
152 47.9 38.0 57.0
154 88.7 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 15.555 ng/ul
RT: 14.887 min Scan# 1955
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:184 Resp: 145704
Ion Ratio Lower Upper
184 100
63 70.8 49.7 74.5
154 61.9 47.5 71.3









#55
4-Nitrophenol
Concen: 22.153 ng/ul
RT: 14.975 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:109 Resp: 216331
Ion Ratio Lower Upper
109 100
139 137.8 115.3 172.9
65 142.1 107.9 161.9



#56
Dibenzofuran
Concen: 22.216 ng/ul
RT: 15.169 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:168 Resp: 1841487
Ion Ratio Lower Upper
168 100
139 36.5 29.4 44.0









#57
2,4-Dinitrotoluene
Concen: 21.161 ng/ul
RT: 15.134 min Scan# 1997
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:165 Resp: 431987
Ion Ratio Lower Upper
165 100
63 45.8 33.1 49.7
182 2.5 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.436 ng/ul
RT: 15.392 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:232 Resp: 357357
Ion Ratio Lower Upper
232 100
131 50.2 36.8 55.2
130 2.6 1.9 2.9
166 34.0 24.6 37.0









#59
Diethylphthalate
Concen: 22.461 ng/ul
RT: 15.587 min Scan# 2074
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:149 Resp: 1595212
Ion Ratio Lower Upper
149 100
177 23.1 17.5 26.3
150 11.9 9.8 14.8



#61
Fluorene
Concen: 22.503 ng/ul
RT: 15.822 min Scan# 2114
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:166 Resp: 1499019
Ion Ratio Lower Upper
166 100
165 97.2 77.8 116.6
167 13.1 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 21.913 ng/ul
RT: 15.810 min Scan# 2112
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:204 Resp: 713218
Ion Ratio Lower Upper
204 100
206 32.5 26.1 39.1
141 58.3 44.5 66.7



#63
4-Nitroaniline
Concen: 17.407 ng/ul
RT: 15.839 min Scan# 2117
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:138 Resp: 190952
Ion Ratio Lower Upper
138 100
92 53.9 41.2 61.8
108 79.3 60.1 90.1









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.467 ng/ul
 RT: 15.898 min Scan# 2127
 Delta R.T. 0.006 min
 Lab File: BP015136.D
 Acq: 18 May 2023 09:04




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020717


Tgt Ion:198 Resp: 256531
 Ion Ratio Lower Upper
 198 100
 182 4.0 3.7 5.5
 77 28.1 21.2 31.8



#67
 N-Nitrosodiphenylamine
 Concen: 22.063 ng/ul
 RT: 16.022 min Scan# 2148
 Delta R.T. 0.000 min
 Lab File: BP015136.D
 Acq: 18 May 2023 09:04

Tgt Ion:169 Resp: 1250188
 Ion Ratio Lower Upper
 169 100
 168 68.3 53.5 80.3
 167 36.7 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 21.863 ng/ul
RT: 16.710 min Scan# 2265
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:248 Resp: 433969
Ion Ratio Lower Upper
248 100
250 98.9 79.0 118.4
141 70.9 58.6 87.8



#69
Hexachlorobenzene
Concen: 21.521 ng/ul
RT: 16.828 min Scan# 2285
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:284 Resp: 504806
Ion Ratio Lower Upper
284 100
142 35.1 27.3 40.9
249 31.1 24.6 36.8





#70

Atrazine

Concen: 21.255 ng/ul

RT: 16.975 min Scan# 2310

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion:200 Resp: 445168

Ion Ratio Lower Upper

200 100

173 27.3 21.4 32.2

215 48.8 38.1 57.1



#71

Pentachlorophenol

Concen: 20.359 ng/ul

RT: 17.175 min Scan# 2344

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04




Tgt Ion:266 Resp: 298728


Ion Ratio Lower Upper

266 100

264 62.1 51.0 76.4

268 64.3 51.8 77.6





#72

Phenanthrene

Concen: 22.266 ng/ul

RT: 17.569 min Scan# 2411

Delta R.T. 0.000 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion:178 Resp: 2336576

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 19.5 15.6 23.4



#74

Anthracene

Concen: 22.508 ng/ul

RT: 17.663 min Scan# 2427

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04




Tgt Ion:178 Resp: 2388449


Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.2

176 18.7 15.2 22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.985 ng/uL
RT: 13.575 min Scan# 1732
Delta R.T. 0.000 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:216 Resp: 600693
Ion Ratio Lower Upper
216 100
214 79.3 62.2 93.4
218 48.3 38.2 57.2



#76
Pentachlorobenzene
Concen: 21.485 ng/uL
RT: 15.092 min Scan# 1990
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:250 Resp: 589442
Ion Ratio Lower Upper
250 100
248 63.5 49.8 74.6
252 64.5 51.6 77.4









#77
 Carbazole
 Concen: 22.559 ng/ul
 RT: 17.928 min Scan# 2472
 Delta R.T. 0.006 min
 Lab File: BP015136.D
 Acq: 18 May 2023 09:04

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020717

Tgt Ion:167 Resp: 2099072




Ion	Ratio	Lower	Upper
167	100		
166	22.0	17.4	26.2
139	13.3	10.2	15.4




#78
 Di-n-butylphthalate
 Concen: 20.910 ng/ul
 RT: 18.457 min Scan# 2562
 Delta R.T. 0.000 min
 Lab File: BP015136.D
 Acq: 18 May 2023 09:04

Tgt Ion:149 Resp: 2505601

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.3	10.9
104	5.1	3.9	5.9









#80
Fluoranthene
Concen: 21.903 ng/ul
RT: 19.522 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04




Instrument :
BNA_P
ClientSampleId :
SSTD020717


Tgt Ion:202 Resp: 2604087
Ion Ratio Lower Upper
202 100
101 12.0 9.5 14.3
100 9.4 7.3 10.9



#82
Pyrene
Concen: 22.444 ng/ul
RT: 19.875 min Scan# 2803
Delta R.T. 0.006 min
Lab File: BP015136.D
Acq: 18 May 2023 09:04

Tgt Ion:202 Resp: 2755583
Ion Ratio Lower Upper
202 100
101 14.0 11.2 16.8
100 12.0 9.0 13.6





#83

Butylbenzylphthalate

Concen: 19.866 ng/ul

RT: 20.727 min Scan# 2948

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion:149 Resp: 1081289

Ion Ratio Lower Upper

149 100

91 74.5 57.4 86.2

206 20.1 16.6 25.0



#84

3,3'-Dichlorobenzidine

Concen: 19.953 ng/ul

RT: 21.510 min Scan# 3081

Delta R.T. 0.000 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04




Tgt Ion:252 Resp: 737013


Ion Ratio Lower Upper

252 100

254 64.0 50.8 76.2

126 12.0 9.8 14.8





#85

Benzo(a)anthracene

Concen: 21.651 ng/ul

RT: 21.586 min Scan# 3094

Delta R.T. 0.000 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion:228 Resp: 2517815

Ion Ratio Lower Upper

228 100

229 20.2 16.1 24.1

226 28.1 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.661 ng/ul

RT: 21.486 min Scan# 3077

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04




Tgt Ion:149 Resp: 1540645


Ion Ratio Lower Upper

149 100

167 26.9 21.8 32.6

279 3.7 3.2 4.8





#87

Chrysene

Concen: 21.575 ng/ul

RT: 21.645 min Scan# 3104

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion:228 Resp: 2398387

Ion Ratio Lower Upper

228 100

226 31.3 25.0 37.6

229 20.1 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 17.424 ng/ul




RT: 22.463 min Scan# 3243


Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Tgt Ion:149 Resp: 2589954





#90

Benzo(b)fluoranthene

Concen: 17.652 ng/ul

RT: 23.380 min Scan# 3399

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion:252 Resp: 2428711

Ion Ratio Lower Upper

252 100

253 22.2 17.4 26.2

125 10.9 8.3 12.5



#91

Benzo(k)fluoranthene

Concen: 18.324 ng/ul

RT: 23.433 min Scan# 3408

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04




Tgt Ion:252 Resp: 2420907


Ion Ratio Lower Upper

252 100

253 22.0 17.5 26.3

125 10.9 8.5 12.7





#93

Benzo(a)pyrene

Concen: 20.911 ng/ul

RT: 24.057 min Scan# 3514

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion:252 Resp: 2543148

Ion Ratio Lower Upper

252 100

253 21.9 17.2 25.8

125 11.0 9.1 13.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 14.566 ng/ul

RT: 26.892 min Scan# 3996

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04




Tgt Ion:276 Resp: 2293701


Ion Ratio Lower Upper

276 100

138 25.2 20.1 30.1

277 24.5 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 14.774 ng/ul

RT: 26.910 min Scan# 3999

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Instrument :

BNA_P

ClientSampleId :

SSTD020717





Tgt Ion:278 Resp: 1909817

Ion Ratio Lower Upper

278 100

139 16.5 13.3 19.9

279 23.9 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 13.979 ng/ul

RT: 27.727 min Scan# 4138

Delta R.T. 0.006 min

Lab File: BP015136.D

Acq: 18 May 2023 09:04

Tgt Ion:276 Resp: 1795029

Ion Ratio Lower Upper

276 100

138 22.5 17.7 26.5

277 23.9 18.8 28.2

