

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051922\
 Data File : BP010408.D
 Acq On : 19 May 2022 17:43
 Operator : CG/JU
 Sample : PB144950BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK950

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/20/2022
 Supervised By :Yogesh Patel 05/24/2022

Quant Time: May 20 00:45:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 15:04:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	107573	20.000	ng/u1	0.00
20) Naphthalene-d8	10.693	136	489041	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.528	164	339412	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.281	188	782768	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.363	240	818924	20.000	ng/u1	# 0.00
88) Perylene-d12	23.792	264	773352	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	15867m	6.287	ng/uL	0.00
4) Pyridine-d5	3.758	84	205594	28.691	ng/u1	0.00
7) Phenol-d5	7.058	99	262739	29.100	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	188346m	31.048	ng/u1	0.00
11) 2-Chlorophenol-d4	7.428	132	208282	30.147	ng/u1	0.00
15) 4-Methylphenol-d8	8.599	113	225872	29.437	ng/u1	-0.01
21) Nitrobenzene-d5	9.052	128	114153m	30.160	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	120672	30.085	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	213728	28.180	ng/u1	0.00
31) 4-Chloroaniline-d4	10.834	131	320527	28.423	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	792324	31.684	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	886407	30.730	ng/u1	0.00
54) 4-Nitrophenol-d4	14.722	143	101610	22.829	ng/u1	0.00
60) Fluorene-d10	15.522	176	686645	31.562	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.639	200	103497	22.608	ng/u1	0.00
73) Anthracene-d10	17.381	188	1124016	31.712	ng/u1	0.00
81) Pyrene-d10	19.610	212	1375812	31.635	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.633	264	1230136	31.678	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051922\
Data File : BP010408.D
Acq On : 19 May 2022 17:43
Operator : CG/JU
Sample : PB144950BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK950

Quant Time: May 20 00:45:45 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 13 15:04:31 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022

