

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051922\
 Data File : BP010422.D
 Acq On : 20 May 2022 02:16
 Operator : CG/JU
 Sample : PB144950BS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS950

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/20/2022
 Supervised By :Yogesh Patel 05/24/2022

Quant Time: May 20 05:51:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 15:04:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	179930	20.000	ng/u1	0.00
20) Naphthalene-d8	10.693	136	789350	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.528	164	526887	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.281	188	1142043	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.363	240	1108220	20.000	ng/u1	# 0.00
88) Perylene-d12	23.786	264	1013694	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	24222m	5.738	ng/uL	0.00
4) Pyridine-d5	3.758	84	356451	29.739	ng/u1	0.00
7) Phenol-d5	7.058	99	486688	32.227	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	321564m	31.692	ng/u1	0.00
11) 2-Chlorophenol-d4	7.422	132	382212	33.075	ng/u1	-0.01
15) 4-Methylphenol-d8	8.605	113	408115	31.799	ng/u1	0.00
21) Nitrobenzene-d5	9.052	128	202719	33.183	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	224581	34.689	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	402556	32.884	ng/u1	0.00
31) 4-Chloroaniline-d4	10.828	131	520199m	28.580	ng/u1	-0.01
46) Dimethylphthalate-d6	13.940	166	1269856	32.712	ng/u1	0.00
49) Acenaphthylene-d8	14.216	160	1487069	33.211	ng/u1	0.00
54) 4-Nitrophenol-d4	14.728	143	223519	32.349	ng/u1	0.00
60) Fluorene-d10	15.522	176	1092253	32.342	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.645	200	223904	33.523	ng/u1	0.00
73) Anthracene-d10	17.381	188	1684925	32.582	ng/u1	0.00
81) Pyrene-d10	19.610	212	2030694	34.504	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.633	264	1735315	34.092	ng/u1	0.00
Target Compounds						
2) 1,4-Dioxane	3.375	88	49281	11.270	ng/uL#	1
5) Pyridine	3.775	79	365523	29.344	ng/u1#	34
6) Benzaldehyde	7.034	77	328008	41.086	ng/u1	93
8) Phenol	7.087	94	523270	32.117	ng/u1#	93
10) Bis(2-Chloroethyl)ether	7.316	93	406307	31.718	ng/u1#	74
12) 2-Chlorophenol	7.458	128	391533	32.455	ng/u1	95
13) 2-Methylphenol	8.340	108	398742	32.785	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.422	45	669588	32.073	ng/u1#	82
16) Acetophenone	8.716	105	641713	31.823	ng/u1#	76
17) N-Nitroso-di-n-propyla...	8.705	70	350532	32.481	ng/u1#	68
18) 4-Methylphenol	8.669	108	434236	32.304	ng/u1	94
19) Hexachloroethane	8.975	117	166047	31.648	ng/u1#	81
22) Nitrobenzene	9.093	77	508808	31.694	ng/u1#	82
23) Isophorone	9.622	82	976444	32.906	ng/u1#	97
25) 2-Nitrophenol	9.805	139	239541	33.697	ng/u1#	86
26) 2,4-Dimethylphenol	9.869	107	494270	32.044	ng/u1#	80
27) Bis(2-Chloroethoxy)met...	10.105	93	574159	32.497	ng/u1#	98
29) 2,4-Dichlorophenol	10.346	162	404730	32.337	ng/u1#	84
30) Naphthalene	10.746	128	1308564	31.739	ng/u1	97
32) 4-Chloroaniline	10.857	127	516931	28.617	ng/u1	98
33) Hexachlorobutadiene	11.034	225	273780	30.870	ng/u1	94
34) Caprolactam	11.634	113	145617m	33.338	ng/u1	
35) 4-Chloro-3-methylphenol	11.981	107	478896	32.708	ng/u1#	69
36) 2-Methylnaphthalene	12.351	142	921005	31.555	ng/u1	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051922\
 Data File : BP010422.D
 Acq On : 20 May 2022 02:16
 Operator : CG/JU
 Sample : PB144950BS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS950


Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/20/2022
 Supervised By :Yogesh Patel 05/24/2022

Quant Time: May 20 05:51:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 15:04:31 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.569	142	939491	31.879	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.722	216	518709	32.261	ng/ul	94
40) Hexachlorocyclopentadiene	12.704	237	282935	32.367	ng/ul	95
41) 2,4,6-Trichlorophenol	12.963	196	343208	34.243	ng/ul#	85
42) 2,4,5-Trichlorophenol	13.034	196	376155	33.613	ng/ul#	87
43) 1,1'-Biphenyl	13.363	154	1261292	32.614	ng/ul#	94
44) 2-Chloronaphthalene	13.404	162	984745	32.967	ng/ul	94
45) 2-Nitroaniline	13.604	65	349307	35.532	ng/ul#	77
47) Dimethylphthalate	13.987	163	1259034	32.697	ng/ul#	92
48) 2,6-Dinitrotoluene	14.104	165	272140	34.670	ng/ul	93
50) Acenaphthylene	14.245	152	1594789	32.923	ng/ul	96
51) 3-Nitroaniline	14.428	138	254916	34.602	ng/ul#	83
52) Acenaphthene	14.587	153	1020835	32.305	ng/ul	97
53) 2,4-Dinitrophenol	14.640	184	136759	31.206	ng/ul#	76
55) 4-Nitrophenol	14.740	109	198384	32.290	ng/ul#	41
56) Dibenzofuran	14.928	168	1490131	32.178	ng/ul	99
57) 2,4-Dinitrotoluene	14.893	165	392134	34.703	ng/ul#	83
58) 2,3,4,6-Tetrachlorophenol	15.151	232	323265	33.593	ng/ul#	87
59) Diethylphthalate	15.351	149	1298365	33.045	ng/ul	93
61) Fluorene	15.575	166	1231495	32.518	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.569	204	631566	32.059	ng/ul	99
63) 4-Nitroaniline	15.598	138	273050	37.536	ng/ul#	76
66) 4,6-Dinitro-2-methylph...	15.657	198	221476	33.250	ng/ul#	87
67) N-Nitrosodiphenylamine	15.781	169	1066877	33.311	ng/ul	94
68) 4-Bromophenyl-phenylether	16.463	248	386091	32.827	ng/ul	97
69) Hexachlorobenzene	16.587	284	439879	32.514	ng/ul#	87
70) Atrazine	16.739	200	419534	32.698	ng/ul#	89
71) Pentachlorophenol	16.934	266	275127	32.411	ng/ul#	83
72) Phenanthrene	17.322	178	1978768	32.466	ng/ul	98
74) Anthracene	17.416	178	1998580	32.687	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.328	216	539404	31.079	ng/uL#	85
76) Pentachlorobenzene	14.851	250	512095	32.519	ng/uL	89
77) Carbazole	17.681	167	1760909	32.770	ng/ul#	95
78) Di-n-butylphthalate	18.233	149	2241057	34.243	ng/ul#	93
80) Fluoranthene	19.286	202	2396585	35.041	ng/ul	97
82) Pyrene	19.639	202	2449733	34.322	ng/ul	96
83) Butylbenzylphthalate	20.510	149	1041742	36.501	ng/ul#	81
84) 3,3'-Dichlorobenzidine	21.275	252	765066	34.348	ng/ul#	96
85) Benzo(a)anthracene	21.345	228	2314557	33.399	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.269	149	1534682	36.644	ng/ul#	81
87) Chrysene	21.404	228	2238106	32.833	ng/ul	99
89) Di-n-octyl phthalate	22.198	149	2614333	34.378	ng/ul	100
90) Benzo(b)fluoranthene	23.051	252	2302869	34.083	ng/ul	99
91) Benzo(k)fluoranthene	23.098	252	2113696	32.915	ng/ul#	97
93) Benzo(a)pyrene	23.686	252	2093196	33.344	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.321	276	2019795	34.201	ng/ul#	91
95) Dibenzo(a,h)anthracene	26.333	278	1723462	33.633	ng/ul#	95
96) Benzo(g,h,i)perylene	27.098	276	1681144	34.092	ng/ul#	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.270 ng/uL
RT: 3.375 min Scan# 49
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950




Tgt Ion: 88 Resp: 4928
Ion Ratio Lower Upper
88 100
43 43.1 6.8 10.2
58 93.8 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022




#5
Pyridine
Concen: 29.344 ng/uL
RT: 3.775 min Scan# 117
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Tgt Ion: 79 Resp: 365523
Ion Ratio Lower Upper
79 100
52 96.7 35.5 53.3#
51 45.4 20.2 30.4#





#6
Benzaldehyde
Concen: 41.086 ng/ul
RT: 7.034 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950




Tgt Ion: 77 Resp: 328008
Ion Ratio Lower Upper
77 100
105 92.5 77.0 115.4
106 85.3 77.0 115.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022




#8
Phenol
Concen: 32.117 ng/ul
RT: 7.087 min Scan# 680
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Tgt Ion: 94 Resp: 523270
Ion Ratio Lower Upper
94 100
65 31.9 19.0 28.6#
66 44.1 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 31.718 ng/ul
RT: 7.316 min Scan# 719
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion: 93 Resp: 406307
Ion Ratio Lower Upper
93 100
63 86.4 45.0 67.4
95 32.8 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#12
2-Chlorophenol
Concen: 32.455 ng/ul
RT: 7.458 min Scan# 743
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Tgt Ion:128 Resp: 391533
Ion Ratio Lower Upper
128 100
64 54.4 45.3 67.9
130 31.7 29.0 43.6









#13
2-Methylphenol
Concen: 32.785 ng/ul
RT: 8.340 min Scan# 893
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Instrument :
BNA_P
ClientSampleId :
SLCS950

Tgt Ion: 108 Resp: 39874
Ion Ratio Lower Upper
108 100
107 91.6 70.8 106.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 32.073 ng/ul
RT: 8.422 min Scan# 907
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Tgt Ion: 45 Resp: 669588
Ion Ratio Lower Upper
45 100
77 12.0 17.1 25.7#
79 9.5 13.2 19.8#









#16
Acetophenone
Concen: 31.823 ng/ul
RT: 8.716 min Scan# 911
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Instrument :
BNA_P
ClientSampleId :
SLCS950

Tgt Ion:105 Resp: 641711
Ion Ratio Lower Upper
105 100
77 84.2 53.4 80.2
51 32.3 13.8 20.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#17
N-Nitroso-di-n-propylamine
Concen: 32.481 ng/ul
RT: 8.705 min Scan# 955
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Tgt Ion: 70 Resp: 350532
Ion Ratio Lower Upper
70 100
42 68.7 31.8 47.8#
101 10.2 15.6 23.4#
130 21.3 13.7 20.5#









#18
4-Methylphenol
Concen: 32.304 ng/ul
RT: 8.669 min Scan# 948
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Instrument :
BNA_P
ClientSampleId :
SLCS950

Tgt Ion:108 Resp: 434230
Ion Ratio Lower Upper
108 100
107 110.2 83.6 125.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#19
Hexachloroethane
Concen: 31.648 ng/ul
RT: 8.975 min Scan# 1001
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16


Tgt Ion:117 Resp: 166047
Ion Ratio Lower Upper
117 100
201 103.9 106.5 159.7#
199 65.5 59.0 88.6





#22
Nitrobenzene
Concen: 31.694 ng/ul
RT: 9.093 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion: 77 Resp: 508800
Ion Ratio Lower Upper
77 100
123 40.3 42.3 63.5
65 14.9 5.8 8.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#23
Isophorone
Concen: 32.906 ng/ul
RT: 9.622 min Scan# 1111
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Tgt Ion: 82 Resp: 976444
Ion Ratio Lower Upper
82 100
95 7.3 7.8 11.8
138 16.4 12.9 19.3





#25
2-Nitrophenol
Concen: 33.697 ng/ul
RT: 9.805 min Scan# 1143
Delta R.T. -0.012 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Instrument :

BNA_P

ClientSampleId :





SLCS950

Tgt Ion:139 Resp: 23954
Ion Ratio Lower Upper
139 100
65 62.0 49.4 74.2
109 34.7 44.7 67.1

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#26
2,4-Dimethylphenol
Concen: 32.044 ng/ul
RT: 9.869 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16


Tgt Ion:107 Resp: 494270
Ion Ratio Lower Upper
107 100
121 47.8 34.1 51.1
122 79.1 46.3 69.5





#27
 Bis(2-Chloroethoxy)methane
 Concen: 32.497 ng/ul
 RT: 10.105 min Scan# 1193
 Delta R.T. -0.006 min
 Lab File: BP010422.D
 Acq: 20 May 2022 02:16




Instrument :
 BNA_P
 ClientSampleId :
 SLCS950



Tgt Ion: 93 Resp: 574159
 Ion Ratio Lower Upper
 93 100
 95 32.7 26.2 39.4
 123 11.2 6.2 9.4




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
 Supervised By :Yogesh Patel 05/24/2022



#29
 2,4-Dichlorophenol
 Concen: 32.337 ng/ul
 RT: 10.346 min Scan# 1234
 Delta R.T. -0.006 min
 Lab File: BP010422.D
 Acq: 20 May 2022 02:16


Tgt Ion:162 Resp: 404730
 Ion Ratio Lower Upper
 162 100
 164 64.0 37.0 55.6#
 98 37.3 31.0 46.6





#30
Naphthalene
Concen: 31.739 ng/ul
RT: 10.746 min Scan# 1302
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion:128 Resp: 1308564
Ion Ratio Lower Upper
128 100
129 10.6 8.6 13.0
127 13.2 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#32
4-Chloroaniline
Concen: 28.617 ng/ul
RT: 10.857 min Scan# 1321
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16


Tgt Ion:127 Resp: 516931
Ion Ratio Lower Upper
127 100
129 31.9 24.6 37.0





#33
Hexachlorobutadiene
Concen: 30.870 ng/ul
RT: 11.034 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950




Tgt Ion:225 Resp: 273780
Ion Ratio Lower Upper
225 100
223 61.5 53.4 80.0
227 63.2 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022




#34
Caprolactam
Concen: 33.338 ng/ul m
RT: 11.634 min Scan# 1453
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Tgt Ion:113 Resp: 145617
Ion Ratio Lower Upper
113 100
55 191.0 58.2 87.2#
56 151.2 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 32.708 ng/ul
RT: 11.981 min Scan# 1512
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion:107 Resp: 478890
Ion Ratio Lower Upper
107 100
144 24.0 15.8 23.8
142 75.3 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#36
2-Methylnaphthalene
Concen: 31.555 ng/ul
RT: 12.351 min Scan# 1575
Delta R.T. -0.012 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16


Tgt Ion:142 Resp: 921005
Ion Ratio Lower Upper
142 100
141 88.5 64.3 96.5





#37
1-Methylnaphthalene
Concen: 31.879 ng/ul
RT: 12.569 min Scan# 1612
Delta R.T. -0.012 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion:142 Resp: 93949
Ion Ratio Lower Upper
142 100
141 91.3 69.8 104.8
116 4.0 5.6 8.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.261 ng/ul
RT: 12.722 min Scan# 1638
Delta R.T. -0.012 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16


Tgt Ion:216 Resp: 518709
Ion Ratio Lower Upper
216 100
214 77.8 58.2 87.2
179 21.1 14.2 21.2
108 17.3 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 32.367 ng/ul
RT: 12.704 min Scan# 1635
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950




Tgt Ion:237 Resp: 282935
Ion Ratio Lower Upper
237 100
235 62.0 53.2 79.8
272 13.1 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022




#41
2,4,6-Trichlorophenol
Concen: 34.243 ng/ul
RT: 12.963 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Tgt Ion:196 Resp: 343208
Ion Ratio Lower Upper
196 100
198 95.0 64.1 96.1
200 29.9 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 33.613 ng/ul
RT: 13.034 min Scan# 1691
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion:196 Resp: 37615
Ion Ratio Lower Upper
196 100
198 97.0 68.6 103.0
200 30.9 32.8 49.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#43
1,1'-Biphenyl
Concen: 32.614 ng/ul
RT: 13.363 min Scan# 1747
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16


Tgt Ion:154 Resp: 1261292
Ion Ratio Lower Upper
154 100
153 39.5 28.6 43.0
76 12.8 8.5 12.7





#44
2-Chloronaphthalene
Concen: 32.967 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950




Tgt Ion:162 Resp: 984749
Ion Ratio Lower Upper
162 100
164 33.1 27.5 41.3
127 40.0 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022




#45
2-Nitroaniline
Concen: 35.532 ng/ul
RT: 13.604 min Scan# 1788
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Tgt Ion: 65 Resp: 349307
Ion Ratio Lower Upper
65 100
92 64.8 77.9 116.9#
138 89.1 81.5 122.3





#47
Dimethylphthalate
Concen: 32.697 ng/ul
RT: 13.987 min Scan# 1853
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950




Tgt Ion:163 Resp: 1259034
Ion Ratio Lower Upper
163 100
194 4.3 4.1 6.1
164 9.8 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022




#48
2,6-Dinitrotoluene
Concen: 34.670 ng/ul
RT: 14.104 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Tgt Ion:165 Resp: 272140
Ion Ratio Lower Upper
165 100
89 58.0 51.7 77.5
121 20.9 18.3 27.5





#50
Acenaphthylene
Concen: 32.923 ng/ul
RT: 14.245 min Scan# 1897
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion:152 Resp: 1594789
Ion Ratio Lower Upper
152 100
151 19.9 18.3 27.5
153 13.3 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#51
3-Nitroaniline
Concen: 34.602 ng/ul
RT: 14.428 min Scan# 1928
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16


Tgt Ion:138 Resp: 254916
Ion Ratio Lower Upper
138 100
108 12.1 15.8 23.8#
92 125.4 117.1 175.7





#52
Acenaphthene
Concen: 32.305 ng/ul
RT: 14.587 min Scan# 1955
Delta R.T. -0.012 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950




Tgt Ion:153 Resp: 1020835
Ion Ratio Lower Upper
153 100
152 46.5 34.3 51.5
154 89.2 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022




#53
2,4-Dinitrophenol
Concen: 31.206 ng/ul
RT: 14.640 min Scan# 1964
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Tgt Ion:184 Resp: 136759
Ion Ratio Lower Upper
184 100
63 76.6 34.6 52.0#
154 65.8 57.7 86.5





#55
4-Nitrophenol
Concen: 32.290 ng/ul
RT: 14.740 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion:109 Resp: 198384
Ion Ratio Lower Upper
109 100
139 117.0 61.4 92.2
65 144.6 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#56
Dibenzofuran
Concen: 32.178 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16


Tgt Ion:168 Resp: 1490131
Ion Ratio Lower Upper
168 100
139 37.1 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 34.703 ng/ul
RT: 14.893 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion:165 Resp: 392134
Ion Ratio Lower Upper
165 100
63 47.7 29.5 44.3#
182 2.9 2.8 4.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 33.593 ng/ul
RT: 15.151 min Scan# 2051
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16


Tgt Ion:232 Resp: 323265
Ion Ratio Lower Upper
232 100
131 44.8 25.4 38.2#
130 2.4 4.2 6.2#
166 32.2 28.1 42.1





#59
Diethylphthalate
Concen: 33.045 ng/ul
RT: 15.351 min Scan# 2084
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion:149 Resp: 129836
Ion Ratio Lower Upper
149 100
177 22.3 22.0 33.0
150 12.5 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#61
Fluorene
Concen: 32.518 ng/ul
RT: 15.575 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16


Tgt Ion:166 Resp: 1231495
Ion Ratio Lower Upper
166 100
165 97.0 70.3 105.5
167 13.6 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 32.059 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950




Tgt Ion:204 Resp: 631560
Ion Ratio Lower Upper
204 100
206 33.3 27.6 41.4
141 58.6 46.6 69.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022




#63
4-Nitroaniline
Concen: 37.536 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Tgt Ion:138 Resp: 273050
Ion Ratio Lower Upper
138 100
92 56.5 41.6 62.4
108 86.7 97.6 146.4





#66
4,6-Dinitro-2-methylphenol
Concen: 33.250 ng/ul
RT: 15.657 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion:198 Resp: 221470
Ion Ratio Lower Upper
198 100
182 3.9 6.0 9.0
77 28.9 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#67
N-Nitrosodiphenylamine
Concen: 33.311 ng/ul
RT: 15.781 min Scan# 2158
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Tgt Ion:169 Resp: 1066877
Ion Ratio Lower Upper
169 100
168 66.0 48.3 72.5
167 35.0 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 32.827 ng/ul
RT: 16.463 min Scan# 2275
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Instrument :
BNA_P
ClientSampleId :
SLCS950

Tgt Ion:248 Resp: 386091
Ion Ratio Lower Upper
248 100
250 99.2 79.4 119.0
141 78.0 67.2 100.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#69
Hexachlorobenzene
Concen: 32.514 ng/ul
RT: 16.587 min Scan# 2295
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Tgt Ion:284 Resp: 439879
Ion Ratio Lower Upper
284 100
142 35.3 19.2 28.8#
249 31.2 27.4 41.2





#70

Atrazine

Concen: 32.698 ng/ul

RT: 16.739 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP010422.D

Acq: 20 May 2022 02:16

Instrument :

BNA_P

ClientSampleId :





SLCS950

Tgt Ion:	200	Resp:	419534
Ion Ratio	Lower	Upper	
200	100		
173	26.9	19.2	28.8
215	49.5	32.2	48.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#71

Pentachlorophenol

Concen: 32.411 ng/ul




RT: 16.934 min Scan# 2354


Delta R.T. 0.000 min

Lab File: BP010422.D

Acq: 20 May 2022 02:16

Tgt Ion:	266	Resp:	275127
Ion Ratio	Lower	Upper	
266	100		
264	62.1	57.3	85.9
268	62.8	65.7	98.5





#72

Phenanthrene

Concen: 32.466 ng/u1

RT: 17.322 min Scan# 2420

Delta R.T. -0.006 min

Lab File: BP010422.D

Acq: 20 May 2022 02:16

Instrument :

BNA_P

ClientSampleId :

SLCS950

Tgt Ion:178 Resp: 1978768

Ion Ratio Lower Upper

178 100

179 15.3 11.7 17.5





176 18.8 14.5 21.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/20/2022

Supervised By :Yogesh Patel 05/24/2022



#74

Anthracene

Concen: 32.687 ng/u1

RT: 17.416 min Scan# 2436

Delta R.T. 0.000 min

Lab File: BP010422.D

Acq: 20 May 2022 02:16




Tgt Ion:178 Resp: 1998580


Ion Ratio Lower Upper

178 100

179 15.0 12.3 18.5

176 17.6 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 31.079 ng/uL
RT: 13.328 min Scan# 1741
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Instrument :
BNA_P
ClientSampleId :
SLCS950

Tgt Ion:216 Resp: 539404
Ion Ratio Lower Upper
216 100
214 80.0 50.4 75.6
218 47.2 35.2 52.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#76
Pentachlorobenzene
Concen: 32.519 ng/uL
RT: 14.851 min Scan# 2000
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16


Tgt Ion:250 Resp: 512095
Ion Ratio Lower Upper
250 100
248 63.5 43.5 65.3
252 64.6 46.5 69.7





#77
 Carbazole
 Concen: 32.770 ng/ul
 RT: 17.681 min Scan# 2481
 Delta R.T. -0.006 min
 Lab File: BP010422.D
 Acq: 20 May 2022 02:16




Instrument :
 BNA_P
 ClientSampleId :
 SLCS950




Tgt Ion:167 Resp: 1760909
 Ion Ratio Lower Upper
 167 100
 166 21.4 15.8 23.8
 139 12.9 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 05/20/2022
 Supervised By :Yogesh Patel 05/24/2022




#78
 Di-n-butylphthalate
 Concen: 34.243 ng/ul
 RT: 18.233 min Scan# 2575
 Delta R.T. -0.006 min
 Lab File: BP010422.D
 Acq: 20 May 2022 02:16




Tgt Ion:149 Resp: 2241057
 Ion Ratio Lower Upper
 149 100
 150 9.1 9.6 14.4#
 104 5.6 6.2 9.4#





#80
Fluoranthene
Concen: 35.041 ng/u1
RT: 19.286 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950




Tgt Ion:202 Resp: 239658
Ion Ratio Lower Upper
202 100
101 10.5 9.6 14.4
100 8.1 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022




#82
Pyrene
Concen: 34.322 ng/u1
RT: 19.639 min Scan# 2814
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Tgt Ion:202 Resp: 2449733
Ion Ratio Lower Upper
202 100
101 12.0 10.6 15.8
100 10.3 9.8 14.6





#83
Butylbenzylphthalate
Concen: 36.501 ng/ul
RT: 20.510 min Scan# 2962
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion:149 Resp: 104174
Ion Ratio Lower Upper
149 100
91 73.0 71.1 106.7
206 22.7 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#84
3,3'-Dichlorobenzidine
Concen: 34.348 ng/ul
RT: 21.275 min Scan# 3092
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16

Tgt Ion:252 Resp: 765066
Ion Ratio Lower Upper
252 100
254 64.2 49.8 74.6
126 10.5 4.3 6.5





#85

Benzo(a)anthracene

Concen: 33.399 ng/ul

RT: 21.345 min Scan# 3104

Delta R.T. 0.000 min

Lab File: BP010422.D

Acq: 20 May 2022 02:16

Instrument :

BNA_P

ClientSampleId :

SLCS950

Tgt Ion:228 Resp: 231455

Ion Ratio Lower Upper

228 100

229 20.0 18.2 27.2





226 26.3 22.8 34.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/20/2022

Supervised By :Yogesh Patel 05/24/2022



#86

Bis(2-ethylhexyl)phthalate

Concen: 36.644 ng/ul

RT: 21.269 min Scan# 3091

Delta R.T. 0.000 min

Lab File: BP010422.D

Acq: 20 May 2022 02:16




Tgt Ion:149 Resp: 1534682


Ion Ratio Lower Upper

149 100

167 27.8 32.8 49.2#


279 4.7 5.6 8.4#





#87
Chrysene
Concen: 32.833 ng/ul
RT: 21.404 min Scan# 3114
Delta R.T. 0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950



Tgt Ion:228 Resp: 2238100
Ion Ratio Lower Upper
228 100
226 29.0 22.7 34.1
229 19.5 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#89
Di-n-octyl phthalate
Concen: 34.378 ng/ul
RT: 22.198 min Scan# 3249
Delta R.T. -0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16


Tgt Ion:149 Resp: 2614333





#90
Benzo(b)fluoranthene
Concen: 34.083 ng/ul
RT: 23.051 min Scan# 3394
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950




Tgt Ion:252 Resp: 2302869
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 8.6 7.3 10.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022




#91
Benzo(k)fluoranthene
Concen: 32.915 ng/ul
RT: 23.098 min Scan# 3402
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Tgt Ion:252 Resp: 2113696
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 8.8 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 33.344 ng/ul
RT: 23.686 min Scan# 301
Delta R.T. 0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950




Tgt Ion:252 Resp: 2093190
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 9.3 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 34.201 ng/ul
RT: 26.321 min Scan# 3950
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Tgt Ion:276 Resp: 2019795
Ion Ratio Lower Upper
276 100
138 23.0 12.3 18.5#
277 25.7 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 33.633 ng/ul
RT: 26.333 min Scan# 3952
Delta R.T. 0.000 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16




Instrument :
BNA_P
ClientSampleId :
SLCS950




Tgt Ion:278 Resp: 1723462
Ion Ratio Lower Upper
278 100
139 14.1 6.6 10.0#
279 23.6 18.4 27.6#

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/20/2022
Supervised By :Yogesh Patel 05/24/2022



#96
Benzo(g,h,i)perylene
Concen: 34.092 ng/ul
RT: 27.098 min Scan# 4082
Delta R.T. 0.006 min
Lab File: BP010422.D
Acq: 20 May 2022 02:16



Tgt Ion:276 Resp: 1681144
Ion Ratio Lower Upper
276 100
138 17.8 11.6 17.4#
277 24.4 15.9 23.9#

