

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052022\
 Data File : BP010428.D
 Acq On : 20 May 2022 09:48
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020660

Quant Time: May 20 23:01:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 15:04:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	143795	20.000	ng/ul	0.00
20) Naphthalene-d8	10.693	136	645233	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.522	164	446585	20.000	ng/ul	-0.01
64) Phenanthrene-d10	17.275	188	1001643	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.363	240	998508	20.000	ng/ul	# 0.00
88) Perylene-d12	23.786	264	895868	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	203	0.060	ng/uL	0.08
4) Pyridine-d5	3.758	84	173585	18.122	ng/ul	0.00
7) Phenol-d5	7.058	99	225912	18.718	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	158862	19.591	ng/ul	0.00
11) 2-Chlorophenol-d4	7.422	132	182147	19.723	ng/ul	-0.01
15) 4-Methylphenol-d8	8.599	113	194651	18.978	ng/ul	-0.01
21) Nitrobenzene-d5	9.052	128	98346	19.694	ng/ul	0.00
24) 2-Nitrophenol-d4	9.775	143	105705	19.974	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	194726	19.460	ng/ul	0.00
31) 4-Chloroaniline-d4	10.828	131	279109	18.759	ng/ul	-0.01
46) Dimethylphthalate-d6	13.934	166	648044	19.696	ng/ul	0.00
49) Acenaphthylene-d8	14.216	160	758964	19.998	ng/ul	0.00
54) 4-Nitrophenol-d4	14.722	143	101987	17.414	ng/ul	0.00
60) Fluorene-d10	15.516	176	568373	19.856	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.640	200	104510	17.841	ng/ul	0.00
73) Anthracene-d10	17.375	188	889760	19.617	ng/ul	0.00
81) Pyrene-d10	19.610	212	1063824	20.062	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.627	264	884089	19.653	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.375	88	26183	7.492	ng/uL#	21
5) Pyridine	3.775	79	182620	18.345	ng/ul#	27
6) Benzaldehyde	7.034	77	146380	22.943	ng/ul	93
10) Bis(2-Chloroethyl)ether	7.316	93	197615	19.304	ng/ul#	77
12) 2-Chlorophenol	7.458	128	188612	19.563	ng/ul	94
13) 2-Methylphenol	8.334	108	187469	19.287	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.422	45	341092	20.444	ng/ul#	81
16) Acetophenone	8.716	105	315952	19.606	ng/ul#	73
17) N-Nitroso-di-n-propyla...	8.699	70	174634	20.248	ng/ul#	67
18) 4-Methylphenol	8.663	108	207325	19.299	ng/ul	91
19) Hexachloroethane	8.969	117	84603	20.177	ng/ul#	77
22) Nitrobenzene	9.093	77	260280	19.834	ng/ul#	81
23) Isophorone	9.622	82	478777	19.739	ng/ul	97
25) 2-Nitrophenol	9.805	139	114599	19.722	ng/ul#	86
26) 2,4-Dimethylphenol	9.869	107	249736	19.807	ng/ul#	80
27) Bis(2-Chloroethoxy)met...	10.105	93	281404	19.485	ng/ul#	96
29) 2,4-Dichlorophenol	10.340	162	199722	19.521	ng/ul#	84
30) Naphthalene	10.740	128	655427	19.448	ng/ul	97
32) 4-Chloroaniline	10.852	127	283515	19.201	ng/ul	99
33) Hexachlorobutadiene	11.034	225	142404	19.643	ng/ul	96
34) Caprolactam	11.628	113	66650	18.667	ng/ul#	1
35) 4-Chloro-3-methylphenol	11.981	107	238873	19.959	ng/ul#	72
36) 2-Methylnaphthalene	12.351	142	464808	19.482	ng/ul	90
37) 1-Methylnaphthalene	12.569	142	469253	19.479	ng/ul#	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052022\
 Data File : BP010428.D
 Acq On : 20 May 2022 09:48
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020660

Quant Time: May 20 23:01:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 15:04:31 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) 1,2,4,5-Tetrachloroben...	12.722	216	267578	19.634	ng/ul#	96
40) Hexachlorocyclopentadiene	12.704	237	138183	18.650	ng/ul	93
41) 2,4,6-Trichlorophenol	12.963	196	170170	20.032	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.034	196	186840	19.698	ng/ul#	87
43) 1,1'-Biphenyl	13.357	154	652240	19.898	ng/ul#	94
44) 2-Chloronaphthalene	13.398	162	501995	19.828	ng/ul	95
45) 2-Nitroaniline	13.604	65	172503	20.703	ng/ul#	75
47) Dimethylphthalate	13.981	163	644133	19.736	ng/ul#	93
48) 2,6-Dinitrotoluene	14.098	165	133919	20.129	ng/ul	97
50) Acenaphthylene	14.245	152	820816	19.992	ng/ul	95
51) 3-Nitroaniline	14.428	138	117777	18.861	ng/ul#	83
52) Acenaphthene	14.587	153	524373	19.578	ng/ul	96
53) 2,4-Dinitrophenol	14.640	184	53105	14.297	ng/ul#	77
55) 4-Nitrophenol	14.734	109	94431	18.134	ng/ul#	46
56) Dibenzoofuran	14.922	168	782883	19.945	ng/ul	96
57) 2,4-Dinitrotoluene	14.887	165	193653	20.219	ng/ul#	74
58) 2,3,4,6-Tetrachlorophenol	15.151	232	160877	19.724	ng/ul#	86
59) Diethylphthalate	15.345	149	657412	19.740	ng/ul	93
61) Fluorene	15.575	166	641338	19.980	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.569	204	332420	19.908	ng/ul	98
63) 4-Nitroaniline	15.592	138	116589	18.910	ng/ul#	80
66) 4,6-Dinitro-2-methylph...	15.657	198	103158	17.658	ng/ul#	92
67) N-Nitrosodiphenylamine	15.781	169	553638	19.709	ng/ul	95
68) 4-Bromophenyl-phenylether	16.463	248	200093	19.397	ng/ul	95
69) Hexachlorobenzene	16.587	284	231529	19.513	ng/ul#	89
70) Atrazine	16.739	200	209654	18.631	ng/ul#	88
71) Pentachlorophenol	16.928	266	127738	17.157	ng/ul#	83
72) Phenanthrene	17.322	178	1038374	19.425	ng/ul	98
74) Anthracene	17.410	178	1045578	19.497	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.328	216	282281	18.544	ng/ul#	84
76) Pentachlorobenzene	14.845	250	272706	19.745	ng/ul	91
77) Carbazole	17.681	167	900578	19.109	ng/ul#	96
78) Di-n-butylphthalate	18.234	149	1134790	19.770	ng/ul#	92
80) Fluoranthene	19.286	202	1248958	20.268	ng/ul	96
82) Pyrene	19.639	202	1293715	20.117	ng/ul#	94
83) Butylbenzylphthalate	20.510	149	521648	20.286	ng/ul#	80
84) 3,3'-Dichlorobenzidine	21.275	252	386556	19.262	ng/ul#	96
85) Benzo(a)anthracene	21.345	228	1231768	19.727	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.269	149	788565	20.897	ng/ul#	80
87) Chrysene	21.398	228	1200721	19.550	ng/ul	98
89) Di-n-octyl phthalate	22.198	149	1313723	19.548	ng/ul	100
90) Benzo(b)fluoranthene	23.045	252	1163449	19.484	ng/ul	100
91) Benzo(k)fluoranthene	23.098	252	1146423	20.201	ng/ul#	98
93) Benzo(a)pyrene	23.680	252	1090275	19.652	ng/ul#	99
94) Indeno(1,2,3-cd)pyrene	26.310	276	1009925	19.350	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.327	278	862491	19.045	ng/ul#	94
96) Benzo(g,h,i)perylene	27.086	276	859489	19.722	ng/ul#	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052022\
 Data File : BP010428.D
 Acq On : 20 May 2022 09:48
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020660

Quant Time: May 20 23:01:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 15:04:31 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.492 ng/uL
RT: 3.375 min Scan# 49
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion: 88 Resp: 26183
Ion Ratio Lower Upper
88 100
43 37.5 6.8 10.2#
58 81.2 28.4 42.6#



#5
Pyridine
Concen: 18.345 ng/uL
RT: 3.775 min Scan# 117
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion: 79 Resp: 182620
Ion Ratio Lower Upper
79 100
52 103.5 35.5 53.3#
51 45.5 20.2 30.4#









#6
Benzaldehyde
Concen: 22.943 ng/ul
RT: 7.034 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion: 77 Resp: 146380
Ion Ratio Lower Upper
77 100
105 90.6 77.0 115.4
106 88.0 77.0 115.4



#10
Bis(2-Chloroethyl)ether
Concen: 19.304 ng/ul
RT: 7.316 min Scan# 719
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion: 93 Resp: 197615
Ion Ratio Lower Upper
93 100
63 82.2 45.0 67.4#
95 31.6 26.3 39.5





#12

2-Chlorophenol

Concen: 19.563 ng/ul

RT: 7.458 min Scan# 743

Delta R.T. -0.006 min

Lab File: BP010428.D

Acq: 20 May 2022 09:48

Instrument :

BNA_P

ClientSampleId :

SSTD020660





Tgt Ion:128 Resp: 188612

Ion Ratio Lower Upper

128 100

64 53.6 45.3 67.9

130 31.2 29.0 43.6



#13

2-Methylphenol

Concen: 19.287 ng/ul

RT: 8.334 min Scan# 892

Delta R.T. -0.012 min

Lab File: BP010428.D




Acq: 20 May 2022 09:48


Tgt Ion:108 Resp: 187469

Ion Ratio Lower Upper

108 100

107 91.8 70.8 106.2









#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.444 ng/ul
RT: 8.422 min Scan# 907
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion: 45 Resp: 341092
Ion Ratio Lower Upper
45 100
77 12.5 17.1 25.7#
79 8.8 13.2 19.8#



#16
Acetophenone
Concen: 19.606 ng/ul
RT: 8.716 min Scan# 957
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion: 105 Resp: 315952
Ion Ratio Lower Upper
105 100
77 85.1 53.4 80.2#
51 35.6 13.8 20.8#









#17
N-Nitroso-di-n-propylamine
Concen: 20.248 ng/ul
RT: 8.699 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion: 70 Resp: 174634
Ion Ratio Lower Upper
70 100
42 71.8 31.8 47.8#
101 10.1 15.6 23.4#
130 18.2 13.7 20.5



#18
4-Methylphenol
Concen: 19.299 ng/ul
RT: 8.663 min Scan# 948
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion: 108 Resp: 207325
Ion Ratio Lower Upper
108 100
107 113.6 83.6 125.4









#19
Hexachloroethane
Concen: 20.177 ng/ul
RT: 8.969 min Scan# 1000
Delta R.T. -0.012 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:117 Resp: 84603
Ion Ratio Lower Upper
117 100
201 99.9 106.5 159.7#
199 62.5 59.0 88.6



#22
Nitrobenzene
Concen: 19.834 ng/ul
RT: 9.093 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion: 77 Resp: 260280
Ion Ratio Lower Upper
77 100
123 39.3 42.3 63.5#
65 14.7 5.8 8.8#









#23
Isophorone
Concen: 19.739 ng/ul
RT: 9.622 min Scan# 1111
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion: 82 Resp: 478777
Ion Ratio Lower Upper
82 100
95 8.0 7.8 11.8
138 16.9 12.9 19.3



#25
2-Nitrophenol
Concen: 19.722 ng/ul
RT: 9.805 min Scan# 1142
Delta R.T. -0.012 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion: 139 Resp: 114599
Ion Ratio Lower Upper
139 100
65 63.2 49.4 74.2
109 35.7 44.7 67.1#









#26
2,4-Dimethylphenol
Concen: 19.807 ng/ul
RT: 9.869 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:107 Resp: 249736
Ion Ratio Lower Upper
107 100
121 46.9 34.1 51.1
122 79.8 46.3 69.5#



#27
Bis(2-Chloroethoxy)methane
Concen: 19.485 ng/ul
RT: 10.105 min Scan# 1193
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion: 93 Resp: 281404
Ion Ratio Lower Upper
93 100
95 30.8 26.2 39.4
123 10.7 6.2 9.4#





#29

2,4-Dichlorophenol

Concen: 19.521 ng/ul

RT: 10.340 min Scan# 1233

Delta R.T. -0.012 min

Lab File: BP010428.D

Acq: 20 May 2022 09:48

Instrument :

BNA_P

ClientSampleId :

SSTD020660





Tgt Ion:162 Resp: 199722

Ion Ratio Lower Upper

162 100

164 64.7 37.0 55.6#

98 37.7 31.0 46.6



#30

Naphthalene

Concen: 19.448 ng/ul

RT: 10.740 min Scan# 1301

Delta R.T. -0.012 min

Lab File: BP010428.D

Acq: 20 May 2022 09:48




Tgt Ion:128 Resp: 655427


Ion Ratio Lower Upper

128 100

129 10.8 8.6 13.0

127 13.0 12.2 18.4









#32
4-Chloroaniline
Concen: 19.201 ng/ul
RT: 10.852 min Scan# 1320
Delta R.T. -0.012 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:127 Resp: 283515
Ion Ratio Lower Upper
127 100
129 31.5 24.6 37.0



#33
Hexachlorobutadiene
Concen: 19.643 ng/ul
RT: 11.034 min Scan# 1351
Delta R.T. -0.012 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:225 Resp: 142404
Ion Ratio Lower Upper
225 100
223 64.1 53.4 80.0
227 63.3 53.8 80.8









#34
Caprolactam
Concen: 18.667 ng/ul
RT: 11.628 min Scan# 1452
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:113 Resp: 66650
Ion Ratio Lower Upper
113 100
55 201.1 58.2 87.2#
56 155.9 68.6 103.0#



#35
4-Chloro-3-methylphenol
Concen: 19.959 ng/ul
RT: 11.981 min Scan# 1512
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:107 Resp: 238873
Ion Ratio Lower Upper
107 100
144 22.7 15.8 23.8
142 73.6 38.9 58.3#









#36
2-Methylnaphthalene
Concen: 19.482 ng/ul
RT: 12.351 min Scan# 1575
Delta R.T. -0.012 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:142 Resp: 464808
Ion Ratio Lower Upper
142 100
141 89.4 64.3 96.5



#37
1-Methylnaphthalene
Concen: 19.479 ng/ul
RT: 12.569 min Scan# 1612
Delta R.T. -0.012 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:142 Resp: 469253
Ion Ratio Lower Upper
142 100
141 89.8 69.8 104.8
116 4.1 5.6 8.4#









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 19.634 ng/ul
 RT: 12.722 min Scan# 1638
 Delta R.T. -0.012 min
 Lab File: BP010428.D
 Acq: 20 May 2022 09:48




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020660


Tgt Ion:216 Resp: 267578
 Ion Ratio Lower Upper
 216 100
 214 75.2 58.2 87.2
 179 21.7 14.2 21.2#
 108 17.3 15.3 22.9



#40
 Hexachlorocyclopentadiene
 Concen: 18.650 ng/ul
 RT: 12.704 min Scan# 1635
 Delta R.T. -0.006 min
 Lab File: BP010428.D
 Acq: 20 May 2022 09:48

Tgt Ion:237 Resp: 138183
 Ion Ratio Lower Upper
 237 100
 235 60.7 53.2 79.8
 272 11.6 10.8 16.2









#41
2,4,6-Trichlorophenol
Concen: 20.032 ng/ul
RT: 12.963 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:196 Resp: 170170
Ion Ratio Lower Upper
196 100
198 96.7 64.1 96.1#
200 29.7 19.9 29.9



#42
2,4,5-Trichlorophenol
Concen: 19.698 ng/ul
RT: 13.034 min Scan# 1691
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:196 Resp: 186840
Ion Ratio Lower Upper
196 100
198 96.8 68.6 103.0
200 30.8 32.8 49.2#









#43
1,1'-Biphenyl
Concen: 19.898 ng/ul
RT: 13.357 min Scan# 1746
Delta R.T. -0.012 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:154 Resp: 652240
Ion Ratio Lower Upper
154 100
153 39.4 28.6 43.0
76 12.8 8.5 12.7#



#44
2-Chloronaphthalene
Concen: 19.828 ng/ul
RT: 13.398 min Scan# 1753
Delta R.T. -0.012 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:162 Resp: 501995
Ion Ratio Lower Upper
162 100
164 34.1 27.5 41.3
127 40.6 37.1 55.7









#45
2-Nitroaniline
Concen: 20.703 ng/ul
RT: 13.604 min Scan# 1788
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion: 65 Resp: 172503
Ion Ratio Lower Upper
65 100
92 60.5 77.9 116.9#
138 89.4 81.5 122.3



#47
Dimethylphthalate
Concen: 19.736 ng/ul
RT: 13.981 min Scan# 1852
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion: 163 Resp: 644133
Ion Ratio Lower Upper
163 100
194 4.2 4.1 6.1
164 10.4 11.3 16.9#





#48

2,6-Dinitrotoluene

Concen: 20.129 ng/ul

RT: 14.098 min Scan# 1872

Delta R.T. -0.006 min

Lab File: BP010428.D

Acq: 20 May 2022 09:48

Instrument :

BNA_P

ClientSampleId :

SSTD020660





Tgt Ion:165 Resp: 133919

Ion Ratio Lower Upper

165 100

89 66.8 51.7 77.5

121 21.8 18.3 27.5



#50

Acenaphthylene

Concen: 19.992 ng/ul

RT: 14.245 min Scan# 1897

Delta R.T. -0.006 min

Lab File: BP010428.D

Acq: 20 May 2022 09:48




Tgt Ion:152 Resp: 820816


Ion Ratio Lower Upper

152 100

151 19.3 18.3 27.5

153 13.2 11.0 16.6









#51
3-Nitroaniline
Concen: 18.861 ng/ul
RT: 14.428 min Scan# 1928
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:138 Resp: 117777
Ion Ratio Lower Upper
138 100
108 11.4 15.8 23.8#
92 125.3 117.1 175.7



#52
Acenaphthene
Concen: 19.578 ng/ul
RT: 14.587 min Scan# 1955
Delta R.T. -0.012 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:153 Resp: 524373
Ion Ratio Lower Upper
153 100
152 47.0 34.3 51.5
154 90.0 70.0 105.0









#53
2,4-Dinitrophenol
Concen: 14.297 ng/ul
RT: 14.640 min Scan# 1964
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:184 Resp: 53105
Ion Ratio Lower Upper
184 100
63 78.8 34.6 52.0#
154 69.0 57.7 86.5



#55
4-Nitrophenol
Concen: 18.134 ng/ul
RT: 14.734 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:109 Resp: 94431
Ion Ratio Lower Upper
109 100
139 112.8 61.4 92.2#
65 140.6 65.3 97.9#









#56
Dibenzofuran
Concen: 19.945 ng/ul
RT: 14.922 min Scan# 2012
Delta R.T. -0.012 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:168 Resp: 782883
Ion Ratio Lower Upper
168 100
139 39.8 30.1 45.1



#57
2,4-Dinitrotoluene
Concen: 20.219 ng/ul
RT: 14.887 min Scan# 2006
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:165 Resp: 193653
Ion Ratio Lower Upper
165 100
63 53.8 29.5 44.3#
182 3.3 2.8 4.2









#58
2,3,4,6-Tetrachlorophenol
Concen: 19.724 ng/ul
RT: 15.151 min Scan# 2051
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Instrument :
BNA_P
ClientSampleId :
SSTD020660

Tgt Ion:232 Resp: 160877




Ion	Ratio	Lower	Upper
232	100		
131	44.9	25.4	38.2#
130	2.6	4.2	6.2#
166	31.4	28.1	42.1




#59
Diethylphthalate
Concen: 19.740 ng/ul
RT: 15.345 min Scan# 2084
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:149 Resp: 657412

Ion	Ratio	Lower	Upper
149	100		
177	22.2	22.0	33.0
150	12.4	10.2	15.2





#61

Fluorene

Concen: 19.980 ng/ul

RT: 15.575 min Scan# 2123

Delta R.T. -0.006 min

Lab File: BP010428.D

Acq: 20 May 2022 09:48

Instrument :

BNA_P

ClientSampleId :

SSTD020660





Tgt Ion:166 Resp: 641338

Ion Ratio Lower Upper

166 100

165 97.3 70.3 105.5

167 13.1 9.5 14.3



#62

4-Chlorophenyl-phenylether

Concen: 19.908 ng/ul

RT: 15.569 min Scan# 2122

Delta R.T. -0.006 min

Lab File: BP010428.D

Acq: 20 May 2022 09:48




Tgt Ion:204 Resp: 332420


Ion Ratio Lower Upper

204 100

206 32.0 27.6 41.4

141 59.0 46.6 69.8









#63
4-Nitroaniline
Concen: 18.910 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:138 Resp: 116589
Ion Ratio Lower Upper
138 100
92 56.1 41.6 62.4
108 92.9 97.6 146.4#



#66
4,6-Dinitro-2-methylphenol
Concen: 17.658 ng/ul
RT: 15.657 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:198 Resp: 103158
Ion Ratio Lower Upper
198 100
182 4.3 6.0 9.0#
77 25.7 17.7 26.5









#67
N-Nitrosodiphenylamine
Concen: 19.709 ng/ul
RT: 15.781 min Scan# 2158
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:169 Resp: 553638
Ion Ratio Lower Upper
169 100
168 65.5 48.3 72.5
167 35.1 29.6 44.4



#68
4-Bromophenyl-phenylether
Concen: 19.397 ng/ul
RT: 16.463 min Scan# 2274
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:248 Resp: 200093
Ion Ratio Lower Upper
248 100
250 96.6 79.4 119.0
141 76.5 67.2 100.8









#69
Hexachlorobenzene
Concen: 19.513 ng/ul
RT: 16.587 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:284 Resp: 231529
Ion Ratio Lower Upper
284 100
142 34.7 19.2 28.8#
249 32.4 27.4 41.2



#70
Atrazine
Concen: 18.631 ng/ul
RT: 16.739 min Scan# 2321
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:200 Resp: 209654
Ion Ratio Lower Upper
200 100
173 25.6 19.2 28.8
215 50.8 32.2 48.4#









#71
 Pentachlorophenol
 Concen: 17.157 ng/ul
 RT: 16.928 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BP010428.D
 Acq: 20 May 2022 09:48




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020660


Tgt Ion:266 Resp: 127738
 Ion Ratio Lower Upper
 266 100
 264 62.7 57.3 85.9
 268 61.6 65.7 98.5#



#72
 Phenanthrene
 Concen: 19.425 ng/ul
 RT: 17.322 min Scan# 2420
 Delta R.T. -0.006 min
 Lab File: BP010428.D
 Acq: 20 May 2022 09:48

Tgt Ion:178 Resp: 1038374
 Ion Ratio Lower Upper
 178 100
 179 15.4 11.7 17.5
 176 18.7 14.5 21.7





#74

Anthracene

Concen: 19.497 ng/uI

RT: 17.410 min Scan# 2435

Delta R.T. -0.006 min

Lab File: BP010428.D

Acq: 20 May 2022 09:48

Instrument :

BNA_P

ClientSampleId :

SSTD020660





Tgt Ion:178 Resp: 1045578

Ion Ratio Lower Upper

178 100

179 15.7 12.3 18.5

176 18.4 13.4 20.2



#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.544 ng/uI

RT: 13.328 min Scan# 1741

Delta R.T. -0.006 min

Lab File: BP010428.D

Acq: 20 May 2022 09:48




Tgt Ion:216 Resp: 282281


Ion Ratio Lower Upper

216 100

214 79.7 50.4 75.6#

218 49.0 35.2 52.8









#76
 Pentachlorobenzene
 Concen: 19.745 ng/uL
 RT: 14.845 min Scan# 1999
 Delta R.T. -0.012 min
 Lab File: BP010428.D
 Acq: 20 May 2022 09:48




Instrument :
 BNA_P
 ClientSampleId :
 SST020660


Tgt Ion:250 Resp: 272706
 Ion Ratio Lower Upper
 250 100
 248 62.8 43.5 65.3
 252 62.9 46.5 69.7



#77
 Carbazole
 Concen: 19.109 ng/uL
 RT: 17.681 min Scan# 2481
 Delta R.T. -0.006 min
 Lab File: BP010428.D
 Acq: 20 May 2022 09:48

Tgt Ion:167 Resp: 900578
 Ion Ratio Lower Upper
 167 100
 166 21.3 15.8 23.8
 139 12.8 8.3 12.5#









#78
Di-n-butylphthalate
Concen: 19.770 ng/ul
RT: 18.234 min Scan# 2575
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:149 Resp: 1134790
Ion Ratio Lower Upper
149 100
150 8.6 9.6 14.4#
104 5.5 6.2 9.4#



#80
Fluoranthene
Concen: 20.268 ng/ul
RT: 19.286 min Scan# 2754
Delta R.T. 0.000 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:202 Resp: 1248958
Ion Ratio Lower Upper
202 100
101 9.7 9.6 14.4
100 7.9 6.3 9.5









#82
 Pyrene
 Concen: 20.117 ng/ul
 RT: 19.639 min Scan# 2814
 Delta R.T. 0.000 min
 Lab File: BP010428.D
 Acq: 20 May 2022 09:48




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020660


Tgt Ion:202 Resp: 1293715
 Ion Ratio Lower Upper
 202 100
 101 11.2 10.6 15.8
 100 9.6 9.8 14.6#



#83
 Butylbenzylphthalate
 Concen: 20.286 ng/ul
 RT: 20.510 min Scan# 2962
 Delta R.T. 0.000 min
 Lab File: BP010428.D
 Acq: 20 May 2022 09:48

Tgt Ion:149 Resp: 521648
 Ion Ratio Lower Upper
 149 100
 91 71.8 71.1 106.7
 206 22.6 28.8 43.2#









#84
3,3'-Dichlorobenzidine
Concen: 19.262 ng/ul
RT: 21.275 min Scan# 3092
Delta R.T. -0.006 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:252 Resp: 386556
Ion Ratio Lower Upper
252 100
254 64.4 49.8 74.6
126 9.8 4.3 6.5#



#85
Benzo(a)anthracene
Concen: 19.727 ng/ul
RT: 21.345 min Scan# 3104
Delta R.T. 0.000 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:228 Resp: 1231768
Ion Ratio Lower Upper
228 100
229 19.1 18.2 27.2
226 26.2 22.8 34.2









#86
Bis(2-ethylhexyl)phthalate
Concen: 20.897 ng/ul
RT: 21.269 min Scan# 3091
Delta R.T. 0.000 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:149 Resp: 788565
Ion Ratio Lower Upper
149 100
167 27.4 32.8 49.2#
279 4.1 5.6 8.4#



#87
Chrysene
Concen: 19.550 ng/ul
RT: 21.398 min Scan# 3113
Delta R.T. 0.000 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:228 Resp: 1200721
Ion Ratio Lower Upper
228 100
226 29.7 22.7 34.1
229 19.4 15.2 22.8





#89

Di-n-octyl phthalate

Concen: 19.548 ng/ul

RT: 22.198 min Scan# 3249

Delta R.T. -0.006 min

Lab File: BP010428.D

Acq: 20 May 2022 09:48





Instrument :

BNA_P

ClientSampleId :

SSTD020660

Tgt Ion:149 Resp: 1313723



#90

Benzo(b)fluoranthene

Concen: 19.484 ng/ul

RT: 23.045 min Scan# 3393

Delta R.T. -0.006 min

Lab File: BP010428.D

Acq: 20 May 2022 09:48




Tgt Ion:252 Resp: 1163449


Ion Ratio Lower Upper

252 100

253 22.0 17.4 26.0

125 9.1 7.3 10.9









#91
Benzo(k)fluoranthene
Concen: 20.201 ng/ul
RT: 23.098 min Scan# 3402
Delta R.T. 0.000 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48




Instrument :
BNA_P
ClientSampleId :
SSTD020660


Tgt Ion:252 Resp: 1146423
Ion Ratio Lower Upper
252 100
253 22.1 17.9 26.9
125 8.4 5.1 7.7#



#93
Benzo(a)pyrene
Concen: 19.652 ng/ul
RT: 23.680 min Scan# 3501
Delta R.T. 0.000 min
Lab File: BP010428.D
Acq: 20 May 2022 09:48

Tgt Ion:252 Resp: 1090275
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 9.7 6.4 9.6#









#94
 Indeno(1,2,3-cd)pyrene
 Concen: 19.350 ng/ul
 RT: 26.310 min Scan# 3948
 Delta R.T. -0.012 min
 Lab File: BP010428.D
 Acq: 20 May 2022 09:48




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020660


Tgt Ion:276 Resp: 1009925
 Ion Ratio Lower Upper
 276 100
 138 21.2 12.3 18.5#
 277 25.6 19.3 28.9



#95
 Dibenzo(a,h)anthracene
 Concen: 19.045 ng/ul
 RT: 26.327 min Scan# 3951
 Delta R.T. -0.006 min
 Lab File: BP010428.D
 Acq: 20 May 2022 09:48

Tgt Ion:278 Resp: 862491
 Ion Ratio Lower Upper
 278 100
 139 14.4 6.6 10.0#
 279 24.2 18.4 27.6





#96
 Benzo(g,h,i)perylene
 Concen: 19.722 ng/ul
 RT: 27.086 min Scan# 4080
 Delta R.T. -0.006 min
 Lab File: BP010428.D
 Acq: 20 May 2022 09:48

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020660

Tgt Ion:276 Resp: 859489

Ion	Ratio	Lower	Upper
276	100		
138	18.7	11.6	17.4#
277	24.6	15.9	23.9#

