

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052022\
 Data File : BP010440.D
 Acq On : 20 May 2022 18:02
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020661

Quant Time: May 20 23:03:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 15:04:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.892	152	146532	20.000	ng/ul	0.00
20) Naphthalene-d8	10.692	136	656926	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.522	164	454006	20.000	ng/ul	-0.01
64) Phenanthrene-d10	17.274	188	1000329	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.356	240	1005857	20.000	ng/ul	# 0.00
88) Perylene-d12	23.780	264	915472	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	26395	7.678	ng/uL	0.00
4) Pyridine-d5	3.757	84	175507	17.980	ng/ul	0.00
7) Phenol-d5	7.057	99	231387	18.814	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	164479	19.905	ng/ul	0.00
11) 2-Chlorophenol-d4	7.422	132	185168	19.676	ng/ul	-0.01
15) 4-Methylphenol-d8	8.598	113	201418	19.271	ng/ul	-0.01
21) Nitrobenzene-d5	9.051	128	101476	19.959	ng/ul	0.00
24) 2-Nitrophenol-d4	9.775	143	108103	20.064	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	201207	19.749	ng/ul	0.00
31) 4-Chloroaniline-d4	10.828	131	284905	18.808	ng/ul	-0.01
46) Dimethylphthalate-d6	13.933	166	650791	19.456	ng/ul	0.00
49) Acenaphthylene-d8	14.216	160	763850	19.797	ng/ul	0.00
54) 4-Nitrophenol-d4	14.721	143	102367	17.194	ng/ul	0.00
60) Fluorene-d10	15.516	176	564108	19.385	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.639	200	102288	17.484	ng/ul	0.00
73) Anthracene-d10	17.374	188	884589	19.529	ng/ul	0.00
81) Pyrene-d10	19.604	212	1056602	19.780	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.627	264	916982	19.948	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.375	88	28289	7.944	ng/uL#	8
5) Pyridine	3.775	79	185636	18.300	ng/ul#	32
6) Benzaldehyde	7.034	77	143068	22.005	ng/ul	94
8) Phenol	7.081	94	254723	19.198	ng/ul#	92
10) Bis(2-Chloroethyl)ether	7.310	93	202598	19.421	ng/ul#	74
12) 2-Chlorophenol	7.457	128	194098	19.756	ng/ul	96
13) 2-Methylphenol	8.334	108	193891	19.575	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.422	45	342379	20.138	ng/ul#	83
16) Acetophenone	8.716	105	322750	19.654	ng/ul#	76
17) N-Nitroso-di-n-propyla...	8.698	70	180612	20.550	ng/ul#	71
18) 4-Methylphenol	8.663	108	211462	19.317	ng/ul	93
19) Hexachloroethane	8.969	117	83497	19.541	ng/ul#	79
22) Nitrobenzene	9.092	77	260332	19.485	ng/ul#	85
23) Isophorone	9.622	82	488739	19.791	ng/ul	97
25) 2-Nitrophenol	9.804	139	114428	19.342	ng/ul#	84
26) 2,4-Dimethylphenol	9.869	107	252970	19.706	ng/ul#	81
27) Bis(2-Chloroethoxy)met...	10.104	93	285166	19.394	ng/ul#	97
29) 2,4-Dichlorophenol	10.339	162	202042	19.397	ng/ul#	83
30) Naphthalene	10.739	128	664522	19.367	ng/ul	97
32) 4-Chloroaniline	10.851	127	287854	19.148	ng/ul	98
33) Hexachlorobutadiene	11.033	225	141379	19.154	ng/ul	97
34) Caprolactam	11.622	113	66141	18.195	ng/ul#	1
35) 4-Chloro-3-methylphenol	11.980	107	238387	19.564	ng/ul#	71
36) 2-Methylnaphthalene	12.351	142	471370	19.405	ng/ul	90

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052022\
 Data File : BP010440.D
 Acq On : 20 May 2022 18:02
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020661

Quant Time: May 20 23:03:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 13 15:04:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.569	142	478025	19.490	ng/ul#	97
39) 1,2,4,5-Tetrachloroben...	12.722	216	269215	19.431	ng/ul	94
40) Hexachlorocyclopentadiene	12.704	237	138668	18.410	ng/ul	91
41) 2,4,6-Trichlorophenol	12.957	196	170930	19.792	ng/ul#	85
42) 2,4,5-Trichlorophenol	13.033	196	189412	19.643	ng/ul#	86
43) 1,1'-Biphenyl	13.357	154	657122	19.719	ng/ul#	94
44) 2-Chloronaphthalene	13.398	162	506013	19.660	ng/ul	93
45) 2-Nitroaniline	13.604	65	173473	20.479	ng/ul#	76
47) Dimethylphthalate	13.980	163	643750	19.402	ng/ul#	92
48) 2,6-Dinitrotoluene	14.098	165	135560	20.042	ng/ul	97
50) Acenaphthylene	14.245	152	815843	19.546	ng/ul	95
51) 3-Nitroaniline	14.427	138	118960	18.739	ng/ul#	82
52) Acenaphthene	14.586	153	524652	19.268	ng/ul	95
53) 2,4-Dinitrophenol	14.639	184	53446	14.153	ng/ul#	78
55) 4-Nitrophenol	14.733	109	95798	18.096	ng/ul#	47
56) Dibenzofuran	14.921	168	781127	19.575	ng/ul	99
57) 2,4-Dinitrotoluene	14.886	165	193565	19.880	ng/ul#	78
58) 2,3,4,6-Tetrachlorophenol	15.151	232	159476	19.233	ng/ul#	89
59) Diethylphthalate	15.345	149	660241	19.501	ng/ul	93
61) Fluorene	15.574	166	640329	19.622	ng/ul	92
62) 4-Chlorophenyl-phenyle...	15.569	204	328156	19.332	ng/ul	98
63) 4-Nitroaniline	15.592	138	122878	19.604	ng/ul#	78
66) 4,6-Dinitro-2-methylph...	15.651	198	103409	17.724	ng/ul#	87
67) N-Nitrosodiphenylamine	15.780	169	551221	19.649	ng/ul	96
68) 4-Bromophenyl-phenylether	16.463	248	201156	19.526	ng/ul	96
69) Hexachlorobenzene	16.586	284	227734	19.218	ng/ul#	88
70) Atrazine	16.733	200	210467	18.728	ng/ul	90
71) Pentachlorophenol	16.927	266	129707	17.445	ng/ul#	84
72) Phenanthrene	17.315	178	1033776	19.364	ng/ul	99
74) Anthracene	17.410	178	1039721	19.414	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.327	216	284828	18.736	ng/ul#	87
76) Pentachlorobenzene	14.845	250	268852	19.491	ng/ul	90
77) Carbazole	17.680	167	899947	19.121	ng/ul#	96
78) Di-n-butylphthalate	18.227	149	1130791	19.726	ng/ul#	93
80) Fluoranthene	19.280	202	1248770	20.117	ng/ul	98
82) Pyrene	19.633	202	1289775	19.910	ng/ul	96
83) Butylbenzylphthalate	20.504	149	530682	20.486	ng/ul#	83
84) 3,3'-Dichlorobenzidine	21.274	252	385974	19.092	ng/ul#	97
85) Benzo(a)anthracene	21.339	228	1224579	19.469	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.268	149	780215	20.525	ng/ul#	81
87) Chrysene	21.398	228	1222919	19.766	ng/ul	99
89) Di-n-octyl phthalate	22.198	149	1322407	19.255	ng/ul	100
90) Benzo(b)fluoranthene	23.045	252	1232423	20.197	ng/ul	99
91) Benzo(k)fluoranthene	23.092	252	1122774	19.360	ng/ul#	96
93) Benzo(a)pyrene	23.674	252	1119970	19.755	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.309	276	1060240	19.879	ng/ul#	91
95) Dibenzo(a,h)anthracene	26.327	278	893274	19.302	ng/ul#	95
96) Benzo(g,h,i)perylene	27.080	276	884941	19.871	ng/ul#	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

BNA P

BNA P

ClientSampleId :

SSTD020661


Quant Time: May 20 23:03:14 2022


Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051322.M

Quant Title : SVOA CALIBRATION

QLast Update : Fri May 13 15:04:31 2022

Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.944 ng/uL
RT: 3.375 min Scan# 49
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion: 88 Resp: 28289
Ion Ratio Lower Upper
88 100
43 38.6 6.8 10.2#
58 90.3 28.4 42.6#



#5
Pyridine
Concen: 18.300 ng/uL
RT: 3.775 min Scan# 117
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion: 79 Resp: 185636
Ion Ratio Lower Upper
79 100
52 99.4 35.5 53.3#
51 44.9 20.2 30.4#









#6
Benzaldehyde
Concen: 22.005 ng/ul
RT: 7.034 min Scan# 671
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion: 77 Resp: 143068
Ion Ratio Lower Upper
77 100
105 90.5 77.0 115.4
106 89.2 77.0 115.4



#8
Phenol
Concen: 19.198 ng/ul
RT: 7.081 min Scan# 679
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion: 94 Resp: 254723
Ion Ratio Lower Upper
94 100
65 32.6 19.0 28.6#
66 45.3 34.6 51.8









#10
Bis(2-Chloroethyl)ether
Concen: 19.421 ng/ul
RT: 7.310 min Scan# 719
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion: 93 Resp: 202598
Ion Ratio Lower Upper
93 100
63 86.2 45.0 67.4#
95 32.0 26.3 39.5



#12
2-Chlorophenol
Concen: 19.756 ng/ul
RT: 7.457 min Scan# 743
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:128 Resp: 194098
Ion Ratio Lower Upper
128 100
64 55.5 45.3 67.9
130 31.5 29.0 43.6









#13
2-Methylphenol
Concen: 19.575 ng/ul
RT: 8.334 min Scan# 892
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion: 108 Resp: 193891
Ion Ratio Lower Upper
108 100
107 89.9 70.8 106.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.138 ng/ul
RT: 8.422 min Scan# 907
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion: 45 Resp: 342379
Ion Ratio Lower Upper
45 100
77 12.7 17.1 25.7#
79 10.0 13.2 19.8#









#16
Acetophenone
Concen: 19.654 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:105 Resp: 322750
Ion Ratio Lower Upper
105 100
77 82.9 53.4 80.2#
51 33.5 13.8 20.8#



#17
N-Nitroso-di-n-propylamine
Concen: 20.550 ng/ul
RT: 8.698 min Scan# 954
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion: 70 Resp: 180612
Ion Ratio Lower Upper
70 100
42 66.4 31.8 47.8#
101 10.1 15.6 23.4#
130 20.0 13.7 20.5









#18
4-Methylphenol
Concen: 19.317 ng/ul
RT: 8.663 min Scan# 948
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:108 Resp: 211462
Ion Ratio Lower Upper
108 100
107 111.3 83.6 125.4



#19
Hexachloroethane
Concen: 19.541 ng/ul
RT: 8.969 min Scan# 1000
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:117 Resp: 83497
Ion Ratio Lower Upper
117 100
201 101.2 106.5 159.7#
199 64.3 59.0 88.6









#22
Nitrobenzene
Concen: 19.485 ng/ul
RT: 9.092 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion: 77 Resp: 260332
Ion Ratio Lower Upper
77 100
123 43.0 42.3 63.5
65 15.0 5.8 8.8#



#23
Isophorone
Concen: 19.791 ng/ul
RT: 9.622 min Scan# 1111
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion: 82 Resp: 488739
Ion Ratio Lower Upper
82 100
95 7.9 7.8 11.8
138 16.6 12.9 19.3









#25
2-Nitrophenol
Concen: 19.342 ng/ul
RT: 9.804 min Scan# 1142
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:139 Resp: 114428
Ion Ratio Lower Upper
139 100
65 58.8 49.4 74.2
109 35.2 44.7 67.1#



#26
2,4-Dimethylphenol
Concen: 19.706 ng/ul
RT: 9.869 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:107 Resp: 252970
Ion Ratio Lower Upper
107 100
121 46.7 34.1 51.1
122 79.5 46.3 69.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.394 ng/ul
 RT: 10.104 min Scan# 1193
 Delta R.T. -0.006 min
 Lab File: BP010440.D
 Acq: 20 May 2022 18:02




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020661


Tgt Ion: 93 Resp: 285166
 Ion Ratio Lower Upper
 93 100
 95 31.6 26.2 39.4
 123 11.0 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 19.397 ng/ul
 RT: 10.339 min Scan# 1233
 Delta R.T. -0.012 min
 Lab File: BP010440.D
 Acq: 20 May 2022 18:02

Tgt Ion:162 Resp: 202042
 Ion Ratio Lower Upper
 162 100
 164 66.1 37.0 55.6#
 98 38.1 31.0 46.6









#30
Naphthalene
Concen: 19.367 ng/ul
RT: 10.739 min Scan# 1301
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:128 Resp: 664522
Ion Ratio Lower Upper
128 100
129 11.1 8.6 13.0
127 13.6 12.2 18.4



#32
4-Chloroaniline
Concen: 19.148 ng/ul
RT: 10.851 min Scan# 1320
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:127 Resp: 287854
Ion Ratio Lower Upper
127 100
129 31.9 24.6 37.0





#33

Hexachlorobutadiene

Concen: 19.154 ng/ul

RT: 11.033 min Scan# 11

Delta R.T. -0.012 min

Lab File: BP010440.D

Acq: 20 May 2022 18:02

Instrument :

BNA_P

ClientSampleId :

SSTD020661





Tgt Ion:225 Resp: 141379

Ion Ratio Lower Upper

225 100

223 63.6 53.4 80.0

227 65.1 53.8 80.8



#34

Caprolactam

Concen: 18.195 ng/ul

RT: 11.622 min Scan# 1451

Delta R.T. -0.012 min

Lab File: BP010440.D

Acq: 20 May 2022 18:02




Tgt Ion:113 Resp: 66141


Ion Ratio Lower Upper

113 100

55 205.1 58.2 87.2#

56 164.3 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 19.564 ng/ul
RT: 11.980 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:107 Resp: 238387
Ion Ratio Lower Upper
107 100
144 23.9 15.8 23.8#
142 73.6 38.9 58.3#



#36
2-Methylnaphthalene
Concen: 19.405 ng/ul
RT: 12.351 min Scan# 1575
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:142 Resp: 471370
Ion Ratio Lower Upper
142 100
141 89.6 64.3 96.5









#37
1-Methylnaphthalene
Concen: 19.490 ng/ul
RT: 12.569 min Scan# 1612
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:142 Resp: 478025
Ion Ratio Lower Upper
142 100
141 89.8 69.8 104.8
116 4.2 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.431 ng/ul
RT: 12.722 min Scan# 1638
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:216 Resp: 269215
Ion Ratio Lower Upper
216 100
214 78.4 58.2 87.2
179 19.4 14.2 21.2
108 16.2 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 18.410 ng/ul
RT: 12.704 min Scan# 1635
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:237	Resp:	138668
Ion Ratio	Lower	Upper
237	100	
235	58.7	53.2 79.8
272	12.2	10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 19.792 ng/ul
RT: 12.957 min Scan# 1678
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:196	Resp:	170930
Ion Ratio	Lower	Upper
196	100	
198	94.4	64.1 96.1
200	30.6	19.9 29.9





#42

2,4,5-Trichlorophenol

Concen: 19.643 ng/ul

RT: 13.033 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP010440.D

Acq: 20 May 2022 18:02

Instrument :

BNA_P

ClientSampleId :

SSTD020661





Tgt Ion:196 Resp: 189412

Ion Ratio Lower Upper

196 100

198 99.0 68.6 103.0

200 32.1 32.8 49.2#



#43

1,1'-Biphenyl

Concen: 19.719 ng/ul

RT: 13.357 min Scan# 1746

Delta R.T. -0.012 min

Lab File: BP010440.D

Acq: 20 May 2022 18:02




Tgt Ion:154 Resp: 657122


Ion Ratio Lower Upper

154 100

153 39.0 28.6 43.0

76 12.9 8.5 12.7#









#44
2-Chloronaphthalene
Concen: 19.660 ng/ul
RT: 13.398 min Scan# 1753
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:162 Resp: 506013
Ion Ratio Lower Upper
162 100
164 31.9 27.5 41.3
127 40.1 37.1 55.7



#45
2-Nitroaniline
Concen: 20.479 ng/ul
RT: 13.604 min Scan# 1788
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion: 65 Resp: 173473
Ion Ratio Lower Upper
65 100
92 61.0 77.9 116.9#
138 89.4 81.5 122.3









#47
Dimethylphthalate
Concen: 19.402 ng/ul
RT: 13.980 min Scan# 1852
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:163 Resp: 643750
Ion Ratio Lower Upper
163 100
194 4.2 4.1 6.1
164 10.0 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 20.042 ng/ul
RT: 14.098 min Scan# 1872
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:165 Resp: 135560
Ion Ratio Lower Upper
165 100
89 62.1 51.7 77.5
121 21.0 18.3 27.5









#50
Acenaphthylene
Concen: 19.546 ng/ul
RT: 14.245 min Scan# 1897
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:152 Resp: 815843
Ion Ratio Lower Upper
152 100
151 19.2 18.3 27.5
153 13.4 11.0 16.6



#51
3-Nitroaniline
Concen: 18.739 ng/ul
RT: 14.427 min Scan# 1928
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:138 Resp: 118960
Ion Ratio Lower Upper
138 100
108 11.0 15.8 23.8#
92 124.4 117.1 175.7









#52
Acenaphthene
Concen: 19.268 ng/ul
RT: 14.586 min Scan# 1955
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:153 Resp: 524652
Ion Ratio Lower Upper
153 100
152 48.2 34.3 51.5
154 90.6 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 14.153 ng/ul
RT: 14.639 min Scan# 1964
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:184 Resp: 53446
Ion Ratio Lower Upper
184 100
63 73.7 34.6 52.0#
154 67.0 57.7 86.5









#55
4-Nitrophenol
Concen: 18.096 ng/ul
RT: 14.733 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:109 Resp: 95798
Ion Ratio Lower Upper
109 100
139 114.0 61.4 92.2#
65 136.5 65.3 97.9#



#56
Dibenzofuran
Concen: 19.575 ng/ul
RT: 14.921 min Scan# 2012
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:168 Resp: 781127
Ion Ratio Lower Upper
168 100
139 38.4 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 19.880 ng/ul
RT: 14.886 min Scan# 2006
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:165 Resp: 193565
Ion Ratio Lower Upper
165 100
63 50.9 29.5 44.3#
182 2.8 2.8 4.2#



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.233 ng/ul
RT: 15.151 min Scan# 2051
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:232 Resp: 159476
Ion Ratio Lower Upper
232 100
131 42.9 25.4 38.2#
130 2.3 4.2 6.2#
166 32.8 28.1 42.1









#59
Diethylphthalate
Concen: 19.501 ng/ul
RT: 15.345 min Scan# 2084
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:149 Resp: 660241
Ion Ratio Lower Upper
149 100
177 22.2 22.0 33.0
150 12.3 10.2 15.2



#61
Fluorene
Concen: 19.622 ng/ul
RT: 15.574 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:166 Resp: 640329
Ion Ratio Lower Upper
166 100
165 95.8 70.3 105.5
167 13.7 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 19.332 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:204 Resp: 328156
Ion Ratio Lower Upper
204 100
206 32.7 27.6 41.4
141 59.1 46.6 69.8



#63
4-Nitroaniline
Concen: 19.604 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:138 Resp: 122878
Ion Ratio Lower Upper
138 100
92 53.3 41.6 62.4
108 87.2 97.6 146.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 17.724 ng/ul
RT: 15.651 min Scan# 2137
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:198 Resp: 103409
Ion Ratio Lower Upper
198 100
182 4.6 6.0 9.0#
77 29.4 17.7 26.5#



#67
N-Nitrosodiphenylamine
Concen: 19.649 ng/ul
RT: 15.780 min Scan# 2158
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:169 Resp: 551221
Ion Ratio Lower Upper
169 100
168 64.0 48.3 72.5
167 34.8 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 19.526 ng/ul
RT: 16.463 min Scan# 21156
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:248 Resp: 201156
Ion Ratio Lower Upper
248 100
250 97.4 79.4 119.0
141 77.5 67.2 100.8



#69
Hexachlorobenzene
Concen: 19.218 ng/ul
RT: 16.586 min Scan# 2295
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:284 Resp: 227734
Ion Ratio Lower Upper
284 100
142 34.0 19.2 28.8#
249 31.0 27.4 41.2









#70
Atrazine
Concen: 18.728 ng/ul
RT: 16.733 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:200 Resp: 210467
Ion Ratio Lower Upper
200 100
173 27.0 19.2 28.8
215 48.1 32.2 48.4



#71
Pentachlorophenol
Concen: 17.445 ng/ul
RT: 16.927 min Scan# 2353
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:266 Resp: 129707
Ion Ratio Lower Upper
266 100
264 62.1 57.3 85.9
268 64.8 65.7 98.5#









#72
Phenanthrene
Concen: 19.364 ng/ul
RT: 17.315 min Scan# 2419
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:178 Resp: 1033776
Ion Ratio Lower Upper
178 100
179 15.5 11.7 17.5
176 18.5 14.5 21.7



#74
Anthracene
Concen: 19.414 ng/ul
RT: 17.410 min Scan# 2435
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:178 Resp: 1039721
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 18.0 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.736 ng/uL
RT: 13.327 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:216 Resp: 284828
Ion Ratio Lower Upper
216 100
214 76.7 50.4 75.6#
218 48.2 35.2 52.8



#76
Pentachlorobenzene
Concen: 19.491 ng/uL
RT: 14.845 min Scan# 1999
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:250 Resp: 268852
Ion Ratio Lower Upper
250 100
248 63.6 43.5 65.3
252 63.4 46.5 69.7





#77

Carbazole

Concen: 19.121 ng/ul

RT: 17.680 min Scan# 2481

Delta R.T. -0.006 min

Lab File: BP010440.D

Acq: 20 May 2022 18:02

Instrument :

BNA_P

ClientSampleId :

SSTD020661





Tgt Ion:167 Resp: 899947

Ion Ratio Lower Upper

167 100

166 20.6 15.8 23.8

139 13.0 8.3 12.5#



#78

Di-n-butylphthalate

Concen: 19.726 ng/ul

RT: 18.227 min Scan# 2574

Delta R.T. -0.012 min

Lab File: BP010440.D

Acq: 20 May 2022 18:02




Tgt Ion:149 Resp: 1130791


Ion Ratio Lower Upper

149 100

150 9.0 9.6 14.4#

104 5.5 6.2 9.4#









#80
Fluoranthene
Concen: 20.117 ng/ul
RT: 19.280 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:202 Resp: 1248770
Ion Ratio Lower Upper
202 100
101 11.1 9.6 14.4
100 8.7 6.3 9.5



#82
Pyrene
Concen: 19.910 ng/ul
RT: 19.633 min Scan# 2813
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:202 Resp: 1289775
Ion Ratio Lower Upper
202 100
101 12.0 10.6 15.8
100 10.1 9.8 14.6





#83

Butylbenzylphthalate

Concen: 20.486 ng/ul

RT: 20.504 min Scan# 2961

Delta R.T. -0.006 min

Lab File: BP010440.D

Acq: 20 May 2022 18:02

Instrument :

BNA_P

ClientSampleId :

SSTD020661





Tgt Ion:149 Resp: 530682

Ion Ratio Lower Upper

149 100

91 75.2 71.1 106.7

206 21.8 28.8 43.2#



#84

3,3'-Dichlorobenzidine

Concen: 19.092 ng/ul

RT: 21.274 min Scan# 3092

Delta R.T. -0.006 min

Lab File: BP010440.D

Acq: 20 May 2022 18:02




Tgt Ion:252 Resp: 385974


Ion Ratio Lower Upper

252 100

254 63.4 49.8 74.6

126 10.4 4.3 6.5#









#85
Benzo(a)anthracene
Concen: 19.469 ng/ul
RT: 21.339 min Scan# 3104
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:228 Resp: 1224579
Ion Ratio Lower Upper
228 100
229 19.8 18.2 27.2
226 25.6 22.8 34.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.525 ng/ul
RT: 21.268 min Scan# 3091
Delta R.T. -0.000 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:149 Resp: 780215
Ion Ratio Lower Upper
149 100
167 27.9 32.8 49.2#
279 4.7 5.6 8.4#









#87
Chrysene
Concen: 19.766 ng/ul
RT: 21.398 min Scan# 3113
Delta R.T. -0.000 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:228 Resp: 1222919
Ion Ratio Lower Upper
228 100
226 28.9 22.7 34.1
229 19.3 15.2 22.8



#89
Di-n-octyl phthalate
Concen: 19.255 ng/ul
RT: 22.198 min Scan# 3249
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:149 Resp: 1322407









#90
Benzo(b)fluoranthene
Concen: 20.197 ng/ul
RT: 23.045 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:252 Resp: 1232423
Ion Ratio Lower Upper
252 100
253 21.4 17.4 26.0
125 8.5 7.3 10.9



#91
Benzo(k)fluoranthene
Concen: 19.360 ng/ul
RT: 23.092 min Scan# 3401
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:252 Resp: 1122774
Ion Ratio Lower Upper
252 100
253 21.4 17.9 26.9
125 9.2 5.1 7.7#









#93
Benzo(a)pyrene
Concen: 19.755 ng/ul
RT: 23.674 min Scan# 3500
Delta R.T. -0.006 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02




Instrument :
BNA_P
ClientSampleId :
SSTD020661


Tgt Ion:252 Resp: 1119970
Ion Ratio Lower Upper
252 100
253 21.6 17.5 26.3
125 9.6 6.4 9.6



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.879 ng/ul
RT: 26.309 min Scan# 3948
Delta R.T. -0.012 min
Lab File: BP010440.D
Acq: 20 May 2022 18:02

Tgt Ion:276 Resp: 1060240
Ion Ratio Lower Upper
276 100
138 22.5 12.3 18.5#
277 25.9 19.3 28.9









#95
 Dibenzo(a,h)anthracene
 Concen: 19.302 ng/ul
 RT: 26.327 min Scan# 3951
 Delta R.T. -0.006 min
 Lab File: BP010440.D
 Acq: 20 May 2022 18:02

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020661

Tgt Ion:278 Resp: 893274
 Ion Ratio Lower Upper
 278 100
 139 13.9 6.6 10.0#
 279 23.6 18.4 27.6



#96
 Benzo(g,h,i)perylene
 Concen: 19.871 ng/ul
 RT: 27.080 min Scan# 4079
 Delta R.T. -0.012 min
 Lab File: BP010440.D
 Acq: 20 May 2022 18:02

Tgt Ion:276 Resp: 884941
 Ion Ratio Lower Upper
 276 100
 138 17.9 11.6 17.4#
 277 22.8 15.9 23.9

