

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052025\
 Data File : BP024731.D
 Acq On : 21 May 2025 08:14
 Operator : RC/JU
 Sample : Q2062-20
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 L1-WC-5

Quant Time: May 21 08:39:06 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 13 16:21:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.651	152	112725	20.000	ng	-0.01
21) Naphthalene-d8	10.422	136	426082	20.000	ng	-0.01
39) Acenaphthene-d10	14.286	164	273143	20.000	ng	-0.02
64) Phenanthrene-d10	17.074	188	564536	20.000	ng	-0.02
76) Chrysene-d12	21.504	240	638528	20.000	ng	-0.02
86) Perylene-d12	24.803	264	740653	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.293	112	734866	111.501	ng	0.00
7) Phenol-d6	6.857	99	796901	94.739	ng	0.00
23) Nitrobenzene-d5	8.804	82	591938	70.194	ng	0.00
42) 2,4,6-Tribromophenol	15.804	330	704461	152.133	ng	0.00
45) 2-Fluorobiphenyl	12.898	172	1557549	74.707	ng	0.00
79) Terphenyl-d14	19.798	244	2980739	80.044	ng	-0.01
Target Compounds						
50) Dimethylphthalate	13.751	163	59207	2.844	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052025\
Data File : BP024731.D
Acq On : 21 May 2025 08:14
Operator : RC/JU
Sample : Q2062-20
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
L1-WC-5

Quant Time: May 21 08:39:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

