

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052225\
 Data File : BP024774.D
 Acq On : 23 May 2025 01:52
 Operator : RC/JU
 Sample : Q2078-11
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: May 23 02:58:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\BFDDTBREAKDOWN-8270.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 01 16:14:27 2010
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	10.563	152	46	40.000	ng	0.30
2) Naphthalene-d8	12.028	136	56	40.000	ng	# 0.31
3) Acenaphthene-d10	15.298	164	549	40.000	ng	# 0.39
4) Phenanthrene-d10	15.481	188	28	40.000	ng	# 0.39
8) Chrysene-d12	31.727	240	32	40.000	ng	# 0.80
9) Perylene-d12	0.000	264	0	0.000	ng	-34.56
Target Compounds						Qvalue
5) DDE	10.687	246	6	N.D.		
6) DDD	11.116	235	9	N.D.		
7) DDT	11.469	235	23	N.D.		


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052225\
Data File : BP024774.D
Acq On : 23 May 2025 01:52
Operator : RC/JU
Sample : Q2078-11
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :


Quant Time: May 23 02:58:42 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\BFDDTBREAKDOWN-8270.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 01 16:14:27 2010
Response via : Initial Calibration








#5
DDE
Concen: Below Cal
RT: 10.687 min Scan# 11
Delta R.T. 0.248 min
Lab File: BP024774.D
Acq: 23 May 2025 01:52

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:246 Resp: 6
Ion Ratio Lower Upper
246 100
248 66.7 1434.9 2152.3#
318 100.0 161.5 242.3#
176 416.7 3844.2 5766.4#



#6
DDD
Concen: Below Cal
RT: 11.116 min Scan# 1365
Delta R.T. 0.283 min
Lab File: BP024774.D
Acq: 23 May 2025 01:52


Tgt Ion:235 Resp: 9
Ion Ratio Lower Upper
235 100
237 188.9 4800.0 7200.0#
165 0.0 4800.0 7200.0#





#7
DDT
Concen: Below Cal
RT: 11.469 min Scan# 14
Delta R.T. 0.277 min
Lab File: BP024774.D
Acq: 23 May 2025 01:52

Instrument :
BNA_P
ClientSampleId :



Tgt Ion: 235 Resp: 23
Ion Ratio Lower Upper
235 100
237 30.4 0.0 0.0#
165 108.7 0.2 0.2#

