

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052323\
 Data File : BP015231.D
 Acq On : 23 May 2023 09:49
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020727

Quant Time: May 23 11:29:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 23 01:25:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.122	152	251670	20.000	ng/u1	0.00
20) Naphthalene-d8	10.951	136	1105308	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.763	164	683042	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.516	188	1447580	20.000	ng/u1	0.00
79) Chrysene-d12	21.592	240	1012372	20.000	ng/u1	0.00
88) Perylene-d12	24.151	264	1066712	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	47285	7.300	ng/uL	0.00
4) Pyridine-d5	3.869	84	334747	18.232	ng/u1	0.00
7) Phenol-d5	7.269	99	434196	18.582	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.440	67	259210	19.207	ng/u1	0.00
11) 2-Chlorophenol-d4	7.646	132	327692	19.073	ng/u1	0.00
15) 4-Methylphenol-d8	8.834	113	349814	18.787	ng/u1	0.00
21) Nitrobenzene-d5	9.298	128	169683	19.536	ng/u1	0.00
24) 2-Nitrophenol-d4	10.028	143	191889	19.856	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.563	165	353835	20.290	ng/u1	0.00
31) 4-Chloroaniline-d4	11.086	131	476390	18.463	ng/u1	0.00
46) Dimethylphthalate-d6	14.163	166	1007472	19.519	ng/u1	0.00
49) Acenaphthylene-d8	14.457	160	1157934	19.615	ng/u1	0.00
54) 4-Nitrophenol-d4	14.957	143	166876	16.650	ng/u1	0.00
60) Fluorene-d10	15.751	176	858034	19.701	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.874	200	146902	16.103	ng/u1	0.00
73) Anthracene-d10	17.616	188	1284003	19.371	ng/u1	0.00
81) Pyrene-d10	19.833	212	1313083	22.363	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.980	264	1013510	19.140	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.463	88	50340	7.233	ng/uL	97
5) Pyridine	3.887	79	346077	18.124	ng/u1	96
6) Benzaldehyde	7.251	77	162488	25.299	ng/u1	98
8) Phenol	7.298	94	451886	18.859	ng/u1	93
10) Bis(2-Chloroethyl)ether	7.534	93	361323	18.848	ng/u1	97
12) 2-Chlorophenol	7.675	128	350964	19.313	ng/u1	97
13) 2-Methylphenol	8.563	108	344690	18.613	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.657	45	465541	19.383	ng/u1	99
16) Acetophenone	8.951	105	554891	19.094	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.940	70	286353	18.788	ng/u1	96
18) 4-Methylphenol	8.898	108	378141	18.552	ng/u1	98
19) Hexachloroethane	9.210	117	139647	18.820	ng/u1	98
22) Nitrobenzene	9.340	77	431526	20.117	ng/u1	97
23) Isophorone	9.869	82	819466	18.585	ng/u1	99
25) 2-Nitrophenol	10.057	139	210765	19.888	ng/u1	94
26) 2,4-Dimethylphenol	10.110	107	417150	19.726	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.351	93	504352	18.800	ng/u1	99
29) 2,4-Dichlorophenol	10.592	162	357126	20.274	ng/u1	98
30) Naphthalene	11.004	128	1143399	19.028	ng/u1	100
32) 4-Chloroaniline	11.110	127	494598	18.651	ng/u1	98
33) Hexachlorobutadiene	11.281	225	224237	21.121	ng/u1	98
34) Caprolactam	11.886	113	112252	16.903	ng/u1	92
35) 4-Chloro-3-methylphenol	12.222	107	391574	19.687	ng/u1	100
36) 2-Methylnaphthalene	12.598	142	775544	19.154	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052323\
 Data File : BP015231.D
 Acq On : 23 May 2023 09:49
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020727

Quant Time: May 23 11:29:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 23 01:25:41 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.816	142	778980	19.158	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.963	216	431866	21.141	ng/ul	99
40) Hexachlorocyclopentadiene	12.939	237	220987	16.696	ng/ul	99
41) 2,4,6-Trichlorophenol	13.198	196	290695	20.152	ng/ul	98
42) 2,4,5-Trichlorophenol	13.275	196	320772	20.509	ng/ul	98
43) 1,1'-Biphenyl	13.598	154	1061831	20.036	ng/ul	100
44) 2-Chloronaphthalene	13.645	162	836122	20.203	ng/ul	99
45) 2-Nitroaniline	13.845	65	247735	20.920	ng/ul	91
47) Dimethylphthalate	14.210	163	1040474	19.356	ng/ul	99
48) 2,6-Dinitrotoluene	14.339	165	219878	20.647	ng/ul	96
50) Acenaphthylene	14.486	152	1269767	19.278	ng/ul	99
51) 3-Nitroaniline	14.669	138	173625	18.106	ng/ul	98
52) Acenaphthene	14.827	153	869317	19.324	ng/ul	99
53) 2,4-Dinitrophenol	14.875	184	116091	16.418	ng/ul	95
55) 4-Nitrophenol	14.969	109	140308	19.034	ng/ul	86
56) Dibenzofuran	15.157	168	1235247	19.741	ng/ul	98
57) 2,4-Dinitrotoluene	15.122	165	308840	20.042	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.380	232	262322	19.873	ng/ul	99
59) Diethylphthalate	15.569	149	1029506	19.203	ng/ul	99
61) Fluorene	15.810	166	975154	19.393	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.798	204	494513	20.127	ng/ul	97
63) 4-Nitroaniline	15.827	138	129664	15.658	ng/ul	89
66) 4,6-Dinitro-2-methylph...	15.886	198	154143	16.372	ng/ul	99
67) N-Nitrosodiphenylamine	16.010	169	836826	19.660	ng/ul	99
68) 4-Bromophenyl-phenylether	16.698	248	310295	20.811	ng/ul	99
69) Hexachlorobenzene	16.816	284	359892	20.425	ng/ul	99
70) Atrazine	16.963	200	315448	20.051	ng/ul	99
71) Pentachlorophenol	17.163	266	192925	17.504	ng/ul	96
72) Phenanthrene	17.557	178	1524117	19.335	ng/ul	99
74) Anthracene	17.651	178	1530001	19.195	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.569	216	443025	21.585	ng/ul	99
76) Pentachlorobenzene	15.080	250	435953	21.154	ng/ul	99
77) Carbazole	17.916	167	1315251	18.817	ng/ul	99
78) Di-n-butylphthalate	18.445	149	1746046	19.398	ng/ul	99
80) Fluoranthene	19.510	202	1650497	23.063	ng/ul	97
82) Pyrene	19.863	202	1649496	22.320	ng/ul	97
83) Butylbenzylphthalate	20.715	149	712167	21.737	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.498	252	420557	18.915	ng/ul	97
85) Benzo(a)anthracene	21.574	228	1399580	19.994	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.474	149	1041897	22.089	ng/ul	99
87) Chrysene	21.633	228	1308750	19.558	ng/ul	97
89) Di-n-octyl phthalate	22.445	149	1592026	21.792	ng/ul	100
90) Benzo(b)fluoranthene	23.362	252	1332568	19.706	ng/ul	99
91) Benzo(k)fluoranthene	23.409	252	1276649	19.661	ng/ul	99
93) Benzo(a)pyrene	24.033	252	1137794	19.035	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.850	276	1484671	19.183	ng/ul	97
95) Dibenzo(a,h)anthracene	26.874	278	1220549	19.212	ng/ul	99
96) Benzo(g,h,i)perylene	27.692	276	1210829	19.186	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052323\
Data File : BP015231.D
Acq On : 23 May 2023 09:49
Operator : CG/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD020727

Quant Time: May 23 11:29:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 23 01:25:41 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.233 ng/uL
RT: 3.463 min Scan# 41
Delta R.T. -0.006 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion: 88 Resp: 50340
Ion Ratio Lower Upper
88 100
43 30.3 20.4 30.6
58 69.4 55.4 83.2



#5
Pyridine
Concen: 18.124 ng/uL
RT: 3.887 min Scan# 119
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion: 79 Resp: 346077
Ion Ratio Lower Upper
79 100
52 71.1 55.4 83.2
51 36.0 25.4 38.2









#6
Benzaldehyde
Concen: 25.299 ng/ul
RT: 7.251 min Scan# 691
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion: 77 Resp: 162488
Ion Ratio Lower Upper
77 100
105 95.3 76.9 115.3
106 90.7 75.0 112.6



#8
Phenol
Concen: 18.859 ng/ul
RT: 7.298 min Scan# 699
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion: 94 Resp: 451886
Ion Ratio Lower Upper
94 100
65 28.0 20.6 31.0
66 37.2 26.1 39.1









#10
Bis(2-Chloroethyl)ether
Concen: 18.848 ng/ul
RT: 7.534 min Scan# 710
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion: 93 Resp: 361323
Ion Ratio Lower Upper
93 100
63 75.9 58.0 87.0
95 32.7 26.4 39.6



#12
2-Chlorophenol
Concen: 19.313 ng/ul
RT: 7.675 min Scan# 763
Delta R.T. -0.006 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:128 Resp: 350964
Ion Ratio Lower Upper
128 100
64 47.0 36.2 54.2
130 31.0 26.3 39.5





#13

2-Methylphenol

Concen: 18.613 ng/ul

RT: 8.563 min Scan# 91

Delta R.T. -0.000 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :





SSTD020727

Tgt Ion:108 Resp: 344690

Ion Ratio Lower Upper

108 100

107 87.9 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.383 ng/ul

RT: 8.657 min Scan# 930

Delta R.T. -0.000 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion: 45 Resp: 465541


Ion Ratio Lower Upper

45 100

77 15.1 12.8 19.2

79 11.6 9.5 14.3









#16
Acetophenone
Concen: 19.094 ng/ul
RT: 8.951 min Scan# 981
Delta R.T. -0.006 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:105 Resp: 554891
Ion Ratio Lower Upper
105 100
77 78.4 59.4 89.2
51 26.3 19.2 28.8



#17
N-Nitroso-di-n-propylamine
Concen: 18.788 ng/ul
RT: 8.940 min Scan# 978
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion: 70 Resp: 286353
Ion Ratio Lower Upper
70 100
42 55.6 41.0 61.6
101 9.6 7.9 11.9
130 20.7 16.1 24.1









#18
4-Methylphenol
Concen: 18.552 ng/ul
RT: 8.898 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:108 Resp: 378141
Ion Ratio Lower Upper
108 100
107 112.8 88.5 132.7



#19
Hexachloroethane
Concen: 18.820 ng/ul
RT: 9.210 min Scan# 1024
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:117 Resp: 139647
Ion Ratio Lower Upper
117 100
201 94.4 75.8 113.8
199 61.8 46.5 69.7









#22
Nitrobenzene
Concen: 20.117 ng/ul
RT: 9.340 min Scan# 1046
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion: 77 Resp: 431526
Ion Ratio Lower Upper
77 100
123 42.6 36.2 54.2
65 13.1 10.1 15.1



#23
Isophorone
Concen: 18.585 ng/ul
RT: 9.869 min Scan# 1136
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion: 82 Resp: 819466
Ion Ratio Lower Upper
82 100
95 7.3 5.4 8.0
138 17.2 13.6 20.4









#25
2-Nitrophenol
Concen: 19.888 ng/ul
RT: 10.057 min Scan# 1168
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:139 Resp: 210765
Ion Ratio Lower Upper
139 100
65 50.7 38.4 57.6
109 32.0 22.2 33.2



#26
2,4-Dimethylphenol
Concen: 19.726 ng/ul
RT: 10.110 min Scan# 1177
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:107 Resp: 417150
Ion Ratio Lower Upper
107 100
121 49.2 41.0 61.6
122 82.7 69.5 104.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.800 ng/ul
 RT: 10.351 min Scan# 1218
 Delta R.T. -0.000 min
 Lab File: BP015231.D
 Acq: 23 May 2023 09:49




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020727


Tgt Ion: 93 Resp: 504352
 Ion Ratio Lower Upper
 93 100
 95 32.9 26.7 40.1
 123 11.0 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 20.274 ng/ul
 RT: 10.592 min Scan# 1259
 Delta R.T. -0.000 min
 Lab File: BP015231.D
 Acq: 23 May 2023 09:49

Tgt Ion: 162 Resp: 357126
 Ion Ratio Lower Upper
 162 100
 164 63.3 52.0 78.0
 98 36.7 28.9 43.3









#30
Naphthalene
Concen: 19.028 ng/ul
RT: 11.004 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:128 Resp: 1143399
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 13.2 10.6 15.8



#32
4-Chloroaniline
Concen: 18.651 ng/ul
RT: 11.110 min Scan# 1347
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:127 Resp: 494598
Ion Ratio Lower Upper
127 100
129 33.6 25.8 38.8









#33
Hexachlorobutadiene
Concen: 21.121 ng/ul
RT: 11.281 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:225 Resp: 224237
Ion Ratio Lower Upper
225 100
223 62.7 51.0 76.6
227 63.4 52.2 78.4



#34
Caprolactam
Concen: 16.903 ng/ul
RT: 11.886 min Scan# 1479
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:113 Resp: 112252
Ion Ratio Lower Upper
113 100
55 160.9 121.2 181.8
56 121.2 90.2 135.4









#35
4-Chloro-3-methylphenol
Concen: 19.687 ng/ul
RT: 12.222 min Scan# 1536
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:107 Resp: 391574
Ion Ratio Lower Upper
107 100
144 25.3 20.3 30.5
142 78.5 63.2 94.8



#36
2-Methylnaphthalene
Concen: 19.154 ng/ul
RT: 12.598 min Scan# 1600
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:142 Resp: 775544
Ion Ratio Lower Upper
142 100
141 88.8 71.0 106.6





#37

1-Methylnaphthalene

Concen: 19.158 ng/ul

RT: 12.816 min Scan# 1637

Delta R.T. -0.000 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:142 Resp: 778980

Ion Ratio Lower Upper

142 100

141 92.3 74.6 112.0

116 4.0 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.141 ng/ul

RT: 12.963 min Scan# 1662

Delta R.T. -0.000 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Tgt Ion:216 Resp: 431866




Ion Ratio Lower Upper


216 100

214 77.6 61.4 92.0

179 20.9 15.9 23.9

108 16.3 13.7 20.5





#40

Hexachlorocyclopentadiene

Concen: 16.696 ng/ul

RT: 12.939 min Scan# 1658

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:237 Resp: 220987

Ion Ratio Lower Upper

237 100

235 62.2 49.4 74.0

272 13.7 10.4 15.6



#41

2,4,6-Trichlorophenol

Concen: 20.152 ng/ul

RT: 13.198 min Scan# 1702

Delta R.T. -0.000 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion:196 Resp: 290695


Ion Ratio Lower Upper

196 100

198 95.9 75.5 113.3

200 30.3 23.6 35.4





#42

2,4,5-Trichlorophenol

Concen: 20.509 ng/ul

RT: 13.275 min Scan# 1715

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:196 Resp: 320772

Ion Ratio Lower Upper

196 100

198 93.9 77.0 115.4

200 31.1 24.3 36.5



#43

1,1'-Biphenyl

Concen: 20.036 ng/ul

RT: 13.598 min Scan# 1770

Delta R.T. -0.000 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion:154 Resp: 1061831


Ion Ratio Lower Upper

154 100

153 40.6 32.7 49.1

76 13.3 10.7 16.1





#44

2-Chloronaphthalene

Concen: 20.203 ng/ul

RT: 13.645 min Scan# 11

Delta R.T. 0.005 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:162 Resp: 836122

Ion Ratio Lower Upper

162 100

164 33.8 25.9 38.9

127 38.8 30.8 46.2



#45

2-Nitroaniline

Concen: 20.920 ng/ul

RT: 13.845 min Scan# 1812

Delta R.T. -0.000 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion: 65 Resp: 247735


Ion Ratio Lower Upper

65 100

92 69.8 59.8 89.8

138 103.3 92.3 138.5





#47

Dimethylphthalate

Concen: 19.356 ng/ul

RT: 14.210 min Scan# 1874

Delta R.T. -0.000 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:163 Resp: 1040474

Ion Ratio Lower Upper

163 100

194 3.9 3.1 4.7

164 10.0 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 20.647 ng/ul

RT: 14.339 min Scan# 1896

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion:165 Resp: 219878


Ion Ratio Lower Upper

165 100

89 56.7 48.0 72.0

121 20.6 17.5 26.3





#50

Acenaphthylene

Concen: 19.278 ng/ul

RT: 14.486 min Scan# 1921

Delta R.T. -0.000 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:152 Resp: 1269767

Ion Ratio Lower Upper

152 100

151 19.7 16.1 24.1

153 13.1 10.3 15.5



#51

3-Nitroaniline

Concen: 18.106 ng/ul

RT: 14.669 min Scan# 1952

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion:138 Resp: 173625


Ion Ratio Lower Upper

138 100

108 11.1 8.8 13.2

92 119.4 93.4 140.0









#52
Acenaphthene
Concen: 19.324 ng/ul
RT: 14.827 min Scan# 1979
Delta R.T. 0.006 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:153 Resp: 869317
Ion Ratio Lower Upper
153 100
152 47.4 38.0 57.0
154 87.4 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 16.418 ng/ul
RT: 14.875 min Scan# 1987
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:184 Resp: 116091
Ion Ratio Lower Upper
184 100
63 63.9 49.7 74.5
154 65.9 47.5 71.3









#55
4-Nitrophenol
Concen: 19.034 ng/ul
RT: 14.969 min Scan# 2003
Delta R.T. 0.006 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:109 Resp: 140308
Ion Ratio Lower Upper
109 100
139 124.6 115.3 172.9
65 120.8 107.9 161.9



#56
Dibenzofuran
Concen: 19.741 ng/ul
RT: 15.157 min Scan# 2035
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:168 Resp: 1235247
Ion Ratio Lower Upper
168 100
139 37.7 29.4 44.0





#57

2,4-Dinitrotoluene

Concen: 20.042 ng/ul

RT: 15.122 min Scan# 2029

Delta R.T. -0.000 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:165 Resp: 308840

Ion Ratio Lower Upper

165 100

63 44.7 33.1 49.7

182 2.9 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.873 ng/ul

RT: 15.380 min Scan# 2073

Delta R.T. -0.000 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Tgt Ion:232 Resp: 262322




Ion Ratio Lower Upper


232 100

131 45.7 36.8 55.2

130 2.5 1.9 2.9

166 31.4 24.6 37.0









#59
Diethylphthalate
Concen: 19.203 ng/ul
RT: 15.569 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:149 Resp: 1029506
Ion Ratio Lower Upper
149 100
177 21.5 17.5 26.3
150 12.1 9.8 14.8



#61
Fluorene
Concen: 19.393 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. 0.006 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:166 Resp: 975154
Ion Ratio Lower Upper
166 100
165 95.6 77.8 116.6
167 13.3 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 20.127 ng/ul
RT: 15.798 min Scan# 2144
Delta R.T. 0.006 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:204 Resp: 494513
Ion Ratio Lower Upper
204 100
206 32.4 26.1 39.1
141 59.1 44.5 66.7



#63
4-Nitroaniline
Concen: 15.658 ng/ul
RT: 15.827 min Scan# 2149
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:138 Resp: 129664
Ion Ratio Lower Upper
138 100
92 54.3 41.2 61.8
108 88.0 60.1 90.1









#66
4,6-Dinitro-2-methylphenol
Concen: 16.372 ng/ul
RT: 15.886 min Scan# 2159
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:198 Resp: 154143
Ion Ratio Lower Upper
198 100
182 4.1 3.7 5.5
77 27.3 21.2 31.8



#67
N-Nitrosodiphenylamine
Concen: 19.660 ng/ul
RT: 16.010 min Scan# 2180
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:169 Resp: 836826
Ion Ratio Lower Upper
169 100
168 67.2 53.5 80.3
167 36.5 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 20.811 ng/ul
RT: 16.698 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:248 Resp: 310295
Ion Ratio Lower Upper
248 100
250 98.5 79.0 118.4
141 71.5 58.6 87.8



#69
Hexachlorobenzene
Concen: 20.425 ng/ul
RT: 16.816 min Scan# 2317
Delta R.T. -0.000 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:284 Resp: 359892
Ion Ratio Lower Upper
284 100
142 33.3 27.3 40.9
249 31.4 24.6 36.8





#70

Atrazine

Concen: 20.051 ng/ul

RT: 16.963 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:200 Resp: 315448

Ion Ratio Lower Upper

200 100

173 25.7 21.4 32.2

215 46.9 38.1 57.1



#71

Pentachlorophenol

Concen: 17.504 ng/ul

RT: 17.163 min Scan# 2376

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion:266 Resp: 192925


Ion Ratio Lower Upper

266 100

264 59.1 51.0 76.4

268 63.5 51.8 77.6





#72

Phenanthrene

Concen: 19.335 ng/u1

RT: 17.557 min Scan# 2443

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:178 Resp: 1524117

Ion Ratio Lower Upper

178 100

179 15.1 12.2 18.2

176 19.0 15.6 23.4



#74

Anthracene

Concen: 19.195 ng/u1

RT: 17.651 min Scan# 2459

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion:178 Resp: 1530001


Ion Ratio Lower Upper

178 100

179 14.9 12.2 18.2

176 18.6 15.2 22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.585 ng/uL
RT: 13.569 min Scan# 1765
Delta R.T. 0.005 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:216 Resp: 443025
Ion Ratio Lower Upper
216 100
214 77.4 62.2 93.4
218 48.8 38.2 57.2



#76
Pentachlorobenzene
Concen: 21.154 ng/uL
RT: 15.080 min Scan# 2022
Delta R.T. 0.006 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:250 Resp: 435953
Ion Ratio Lower Upper
250 100
248 63.0 49.8 74.6
252 64.6 51.6 77.4





#77

Carbazole

Concen: 18.817 ng/ul

RT: 17.916 min Scan# 2504

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:167 Resp: 1315251

Ion Ratio Lower Upper

167 100

166 21.5 17.4 26.2

139 13.2 10.2 15.4



#78

Di-n-butylphthalate

Concen: 19.398 ng/ul

RT: 18.445 min Scan# 2594

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion:149 Resp: 1746046


Ion Ratio Lower Upper

149 100

150 9.0 7.3 10.9

104 5.2 3.9 5.9





#80

Fluoranthene

Concen: 23.063 ng/u1

RT: 19.510 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:202 Resp: 1650497

Ion Ratio Lower Upper

202 100

101 11.0 9.5 14.3

100 8.1 7.3 10.9



#82

Pyrene

Concen: 22.320 ng/u1

RT: 19.863 min Scan# 2835

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion:202 Resp: 1649496


Ion Ratio Lower Upper

202 100

101 12.5 11.2 16.8

100 10.5 9.0 13.6





#83

Butylbenzylphthalate

Concen: 21.737 ng/ul

RT: 20.715 min Scan# 2980

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:149 Resp: 712167

Ion Ratio Lower Upper

149 100

91 70.6 57.4 86.2

206 21.0 16.6 25.0



#84

3,3'-Dichlorobenzidine

Concen: 18.915 ng/ul

RT: 21.498 min Scan# 3113

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion:252 Resp: 420557


Ion Ratio Lower Upper

252 100

254 65.6 50.8 76.2

126 10.8 9.8 14.8





#85

Benzo(a)anthracene

Concen: 19.994 ng/ul

RT: 21.574 min Scan# 3126

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:228 Resp: 1399580

Ion Ratio Lower Upper

228 100

229 19.6 16.1 24.1

226 27.1 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.089 ng/ul

RT: 21.474 min Scan# 3109

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion:149 Resp: 1041897


Ion Ratio Lower Upper

149 100

167 27.4 21.8 32.6

279 4.6 3.2 4.8





#87

Chrysene

Concen: 19.558 ng/ul

RT: 21.633 min Scan# 3135

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:228 Resp: 1308750

Ion Ratio Lower Upper

228 100

226 29.3 25.0 37.6

229 19.3 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 21.792 ng/ul




RT: 22.445 min Scan# 3274


Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Tgt Ion:149 Resp: 1592026









#90
Benzo(b)fluoranthene
Concen: 19.706 ng/ul
RT: 23.362 min Scan# 3429
Delta R.T. 0.011 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:252 Resp: 1332568
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.2
125 9.0 8.3 12.5



#91
Benzo(k)fluoranthene
Concen: 19.661 ng/ul
RT: 23.409 min Scan# 3438
Delta R.T. 0.006 min
Lab File: BP015231.D
Acq: 23 May 2023 09:49

Tgt Ion:252 Resp: 1276649
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 9.8 8.5 12.7





#93

Benzo(a)pyrene

Concen: 19.035 ng/ul

RT: 24.033 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:252 Resp: 1137794

Ion Ratio Lower Upper

252 100

253 21.9 17.2 25.8

125 11.1 9.1 13.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.183 ng/ul

RT: 26.850 min Scan# 4023

Delta R.T. 0.006 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49




Tgt Ion:276 Resp: 1484671


Ion Ratio Lower Upper

276 100

138 22.1 20.1 30.1

277 25.0 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 19.212 ng/ul

RT: 26.874 min Scan# 4027

Delta R.T. 0.011 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:278 Resp: 1220549

Ion Ratio Lower Upper

278 100

139 15.4 13.3 19.9

279 24.2 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 19.186 ng/ul

RT: 27.692 min Scan# 4166

Delta R.T. 0.017 min

Lab File: BP015231.D

Acq: 23 May 2023 09:49

Tgt Ion:276 Resp: 1210829

Ion Ratio Lower Upper

276 100

138 20.4 17.7 26.5

277 24.1 18.8 28.2

