

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052323\
 Data File : BP015269.D
 Acq On : 24 May 2023 10:27
 Operator : CG/JU
 Sample : PB152977BS
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS977

Quant Time: May 24 17:44:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 23 01:25:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.116	152	238448	20.000	ng/u1	0.00
20) Naphthalene-d8	10.946	136	1053979	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.757	164	652099	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.510	188	1422563	20.000	ng/u1	0.00
79) Chrysene-d12	21.586	240	975102	20.000	ng/u1	0.00
88) Perylene-d12	24.145	264	1004516	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	41945	6.835	ng/uL	0.00
4) Pyridine-d5	3.864	84	374079	21.504	ng/u1	0.00
7) Phenol-d5	7.269	99	783882	35.407	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.440	67	465835	36.432	ng/u1	0.00
11) 2-Chlorophenol-d4	7.646	132	624348	38.355	ng/u1	0.00
15) 4-Methylphenol-d8	8.834	113	654547	37.103	ng/u1	0.00
21) Nitrobenzene-d5	9.299	128	316310	38.192	ng/u1	0.00
24) 2-Nitrophenol-d4	10.022	143	363869	39.485	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.563	165	654832	39.379	ng/u1	0.00
31) 4-Chloroaniline-d4	11.081	131	335170	13.623	ng/u1	0.00
46) Dimethylphthalate-d6	14.163	166	1940765	39.385	ng/u1	0.00
49) Acenaphthylene-d8	14.451	160	2193192	38.916	ng/u1	0.00
54) 4-Nitrophenol-d4	14.951	143	333045	34.807	ng/u1	0.00
60) Fluorene-d10	15.751	176	1586485	38.155	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.875	200	323016	36.030	ng/u1	0.00
73) Anthracene-d10	17.610	188	2389125	36.677	ng/u1	0.00
81) Pyrene-d10	19.828	212	2519850	44.556	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.980	264	1997831	40.065	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.464	88	87737	13.305	ng/uL	95
5) Pyridine	3.881	79	526360	29.093	ng/u1	94
6) Benzaldehyde	7.246	77	457632	75.204	ng/u1	98
8) Phenol	7.299	94	805248	35.470	ng/u1	92
10) Bis(2-Chloroethyl)ether	7.534	93	629214	34.642	ng/u1	97
12) 2-Chlorophenol	7.675	128	619046	35.954	ng/u1	97
13) 2-Methylphenol	8.563	108	613069	34.941	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.652	45	814160	35.778	ng/u1	99
16) Acetophenone	8.952	105	967868	35.152	ng/u1	94
17) N-Nitroso-di-n-propyla...	8.940	70	497014	34.419	ng/u1	94
18) 4-Methylphenol	8.899	108	673772	34.889	ng/u1	99
19) Hexachloroethane	9.210	117	247338	35.182	ng/u1	88
22) Nitrobenzene	9.340	77	763709	37.336	ng/u1	98
23) Isophorone	9.869	82	1452712	34.552	ng/u1	99
25) 2-Nitrophenol	10.057	139	374575	37.067	ng/u1	94
26) 2,4-Dimethylphenol	10.110	107	738015	36.598	ng/u1	95
27) Bis(2-Chloroethoxy)met...	10.346	93	893194	34.915	ng/u1	99
29) 2,4-Dichlorophenol	10.593	162	633293	37.703	ng/u1	100
30) Naphthalene	10.999	128	1985026	34.642	ng/u1	100
32) 4-Chloroaniline	11.104	127	653746	25.853	ng/u1	99
33) Hexachlorobutadiene	11.275	225	394151	38.934	ng/u1	98
34) Caprolactam	11.899	113	134350	21.215	ng/u1	89
35) 4-Chloro-3-methylphenol	12.222	107	715373	37.718	ng/u1	99
36) 2-Methylnaphthalene	12.593	142	1372212	35.542	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052323\
 Data File : BP015269.D
 Acq On : 24 May 2023 10:27
 Operator : CG/JU
 Sample : PB152977BS
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS977

Quant Time: May 24 17:44:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 23 01:25:41 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.810	142	1381350	35.628	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.957	216	769847	39.474	ng/ul	98
40) Hexachlorocyclopentadiene	12.934	237	365802	28.949	ng/ul	97
41) 2,4,6-Trichlorophenol	13.198	196	525723	38.174	ng/ul	96
42) 2,4,5-Trichlorophenol	13.269	196	585707	39.225	ng/ul	99
43) 1,1'-Biphenyl	13.593	154	1864726	36.856	ng/ul	99
44) 2-Chloronaphthalene	13.640	162	1470468	37.216	ng/ul	99
45) 2-Nitroaniline	13.845	65	455917	40.328	ng/ul	93
47) Dimethylphthalate	14.210	163	1889580	36.821	ng/ul	100
48) 2,6-Dinitrotoluene	14.334	165	412239	40.547	ng/ul	99
50) Acenaphthylene	14.481	152	2237900	35.589	ng/ul	99
51) 3-Nitroaniline	14.669	138	391996	42.818	ng/ul	100
52) Acenaphthene	14.822	153	1526154	35.535	ng/ul	99
53) 2,4-Dinitrophenol	14.875	184	211297	31.301	ng/ul	94
55) 4-Nitrophenol	14.969	109	276798	39.331	ng/ul	83
56) Dibenzofuran	15.157	168	2168274	36.297	ng/ul	97
57) 2,4-Dinitrotoluene	15.122	165	570514	38.779	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.381	232	473630	37.584	ng/ul	98
59) Diethylphthalate	15.569	149	1896401	37.051	ng/ul	100
61) Fluorene	15.804	166	1692114	35.248	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.792	204	868725	37.036	ng/ul	95
63) 4-Nitroaniline	15.834	138	378195	47.838	ng/ul	90
66) 4,6-Dinitro-2-methylph...	15.887	198	313769	33.912	ng/ul	97
67) N-Nitrosodiphenylamine	16.010	169	1472431	35.201	ng/ul	99
68) 4-Bromophenyl-phenylether	16.692	248	568389	38.791	ng/ul	99
69) Hexachlorobenzene	16.810	284	658439	38.026	ng/ul	99
70) Atrazine	16.957	200	361926	23.409	ng/ul	99
71) Pentachlorophenol	17.157	266	377987	34.898	ng/ul	99
72) Phenanthrene	17.557	178	2733732	35.290	ng/ul	99
74) Anthracene	17.645	178	2693323	34.383	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.563	216	789786	39.157	ng/ul	99
76) Pentachlorobenzene	15.075	250	751099	37.088	ng/ul	100
77) Carbazole	17.910	167	2347801	34.181	ng/ul	100
78) Di-n-butylphthalate	18.439	149	3210035	36.290	ng/ul	99
80) Fluoranthene	19.504	202	2996827	43.476	ng/ul	97
82) Pyrene	19.857	202	2941351	41.321	ng/ul	98
83) Butylbenzylphthalate	20.710	149	1343480	42.573	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.498	252	785334	36.671	ng/ul	98
85) Benzo(a)anthracene	21.569	228	2512265	37.261	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.469	149	1975267	43.478	ng/ul	99
87) Chrysene	21.627	228	2347330	36.420	ng/ul	99
89) Di-n-octyl phthalate	22.439	149	3072912	44.667	ng/ul	100
90) Benzo(b)fluoranthene	23.357	252	2382598	37.414	ng/ul	99
91) Benzo(k)fluoranthene	23.410	252	2383970	38.987	ng/ul	98
93) Benzo(a)pyrene	24.033	252	2082509	36.998	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.857	276	2710868	37.196	ng/ul	98
95) Dibenzo(a,h)anthracene	26.874	278	2252436	37.649	ng/ul	97
96) Benzo(g,h,i)perylene	27.692	276	2235191	37.610	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052323\
 Data File : BP015269.D
 Acq On : 24 May 2023 10:27
 Operator : CG/JU
 Sample : PB152977BS
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS977

Quant Time: May 24 17:44:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 23 01:25:41 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 13.305 ng/uL
RT: 3.464 min Scan# 41
Delta R.T. -0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion: 88 Resp: 87737
Ion Ratio Lower Upper
88 100
43 29.2 20.4 30.6
58 73.1 55.4 83.2



#5
Pyridine
Concen: 29.093 ng/uL
RT: 3.881 min Scan# 118
Delta R.T. -0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion: 79 Resp: 526360
Ion Ratio Lower Upper
79 100
52 74.0 55.4 83.2
51 35.7 25.4 38.2









#6
Benzaldehyde
Concen: 75.204 ng/ul
RT: 7.246 min Scan# 690
Delta R.T. -0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion: 77 Resp: 457632
Ion Ratio Lower Upper
77 100
105 97.5 76.9 115.3
106 91.9 75.0 112.6



#8
Phenol
Concen: 35.470 ng/ul
RT: 7.299 min Scan# 699
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion: 94 Resp: 805248
Ion Ratio Lower Upper
94 100
65 28.0 20.6 31.0
66 38.8 26.1 39.1









#10
Bis(2-Chloroethyl)ether
Concen: 34.642 ng/ul
RT: 7.534 min Scan# 738
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion: 93 Resp: 629214
Ion Ratio Lower Upper
93 100
63 75.6 58.0 87.0
95 33.3 26.4 39.6



#12
2-Chlorophenol
Concen: 35.954 ng/ul
RT: 7.675 min Scan# 763
Delta R.T. -0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:128 Resp: 619046
Ion Ratio Lower Upper
128 100
64 47.9 36.2 54.2
130 32.1 26.3 39.5





#13

2-Methylphenol

Concen: 34.941 ng/ul

RT: 8.563 min Scan# 914

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :





SLCS977

Tgt Ion:108 Resp: 613069

Ion Ratio Lower Upper

108 100

107 87.6 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 35.778 ng/ul

RT: 8.652 min Scan# 929

Delta R.T. -0.006 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27




Tgt Ion: 45 Resp: 814160


Ion Ratio Lower Upper

45 100

77 15.3 12.8 19.2

79 12.1 9.5 14.3









#16
Acetophenone
Concen: 35.152 ng/ul
RT: 8.952 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:105 Resp: 967868
Ion Ratio Lower Upper
105 100
77 79.0 59.4 89.2
51 26.8 19.2 28.8



#17
N-Nitroso-di-n-propylamine
Concen: 34.419 ng/ul
RT: 8.940 min Scan# 978
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion: 70 Resp: 497014
Ion Ratio Lower Upper
70 100
42 57.6 41.0 61.6
101 9.9 7.9 11.9
130 20.5 16.1 24.1









#18
4-Methylphenol
Concen: 34.889 ng/ul
RT: 8.899 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:108 Resp: 673772
Ion Ratio Lower Upper
108 100
107 109.6 88.5 132.7



#19
Hexachloroethane
Concen: 35.182 ng/ul
RT: 9.210 min Scan# 1024
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:117 Resp: 247338
Ion Ratio Lower Upper
117 100
201 105.6 75.8 113.8
199 67.2 46.5 69.7









#22
Nitrobenzene
Concen: 37.336 ng/ul
RT: 9.340 min Scan# 1045
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion: 77 Resp: 763709
Ion Ratio Lower Upper
77 100
123 44.1 36.2 54.2
65 13.3 10.1 15.1



#23
Isophorone
Concen: 34.552 ng/ul
RT: 9.869 min Scan# 1136
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion: 82 Resp: 1452712
Ion Ratio Lower Upper
82 100
95 7.5 5.4 8.0
138 16.7 13.6 20.4









#25
2-Nitrophenol
Concen: 37.067 ng/ul
RT: 10.057 min Scan# 1167
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:139 Resp: 374575
Ion Ratio Lower Upper
139 100
65 50.7 38.4 57.6
109 32.7 22.2 33.2



#26
2,4-Dimethylphenol
Concen: 36.598 ng/ul
RT: 10.110 min Scan# 1177
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:107 Resp: 738015
Ion Ratio Lower Upper
107 100
121 49.4 41.0 61.6
122 81.1 69.5 104.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 34.915 ng/ul
 RT: 10.346 min Scan# 1217
 Delta R.T. -0.006 min
 Lab File: BP015269.D
 Acq: 24 May 2023 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SLCS977


Tgt Ion: 93 Resp: 893194
 Ion Ratio Lower Upper
 93 100
 95 32.8 26.7 40.1
 123 10.7 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 37.703 ng/ul
 RT: 10.593 min Scan# 1259
 Delta R.T. -0.000 min
 Lab File: BP015269.D
 Acq: 24 May 2023 10:27

Tgt Ion:162 Resp: 633293
 Ion Ratio Lower Upper
 162 100
 164 64.8 52.0 78.0
 98 35.9 28.9 43.3









#30
Naphthalene
Concen: 34.642 ng/ul
RT: 10.999 min Scan# 1328
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:128 Resp: 1985026
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 13.0 10.6 15.8



#32
4-Chloroaniline
Concen: 25.853 ng/ul
RT: 11.104 min Scan# 1346
Delta R.T. -0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:127 Resp: 653746
Ion Ratio Lower Upper
127 100
129 31.9 25.8 38.8





#33

Hexachlorobutadiene

Concen: 38.934 ng/ul

RT: 11.275 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:225 Resp: 394151

Ion Ratio Lower Upper

225 100

223 62.9 51.0 76.6

227 63.1 52.2 78.4



#34

Caprolactam

Concen: 21.215 ng/ul

RT: 11.899 min Scan# 1481

Delta R.T. 0.012 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27




Tgt Ion:113 Resp: 134350


Ion Ratio Lower Upper

113 100

55 163.8 121.2 181.8

56 125.7 90.2 135.4









#35
4-Chloro-3-methylphenol
Concen: 37.718 ng/ul
RT: 12.222 min Scan# 1536
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:107 Resp: 715373
Ion Ratio Lower Upper
107 100
144 25.1 20.3 30.5
142 77.6 63.2 94.8



#36
2-Methylnaphthalene
Concen: 35.542 ng/ul
RT: 12.593 min Scan# 1599
Delta R.T. -0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:142 Resp: 1372212
Ion Ratio Lower Upper
142 100
141 89.1 71.0 106.6





#37

1-Methylnaphthalene

Concen: 35.628 ng/ul

RT: 12.810 min Scan# 1636

Delta R.T. -0.006 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:142 Resp: 1381350

Ion Ratio Lower Upper

142 100

141 93.6 74.6 112.0

116 4.3 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 39.474 ng/ul

RT: 12.957 min Scan# 1661

Delta R.T. -0.006 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Tgt Ion:216 Resp: 769847




Ion Ratio Lower Upper


216 100

214 78.9 61.4 92.0

179 20.9 15.9 23.9

108 18.0 13.7 20.5









#40
Hexachlorocyclopentadiene
Concen: 28.949 ng/ul
RT: 12.934 min Scan# 1657
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:237 Resp: 365802
Ion Ratio Lower Upper
237 100
235 64.2 49.4 74.0
272 12.3 10.4 15.6



#41
2,4,6-Trichlorophenol
Concen: 38.174 ng/ul
RT: 13.198 min Scan# 1702
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:196 Resp: 525723
Ion Ratio Lower Upper
196 100
198 97.4 75.5 113.3
200 32.1 23.6 35.4





#42

2,4,5-Trichlorophenol

Concen: 39.225 ng/ul

RT: 13.269 min Scan# 1713

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:196 Resp: 585707

Ion Ratio Lower Upper

196 100

198 94.6 77.0 115.4

200 30.8 24.3 36.5



#43

1,1'-Biphenyl

Concen: 36.856 ng/ul

RT: 13.593 min Scan# 1769

Delta R.T. -0.006 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27




Tgt Ion:154 Resp: 1864726


Ion Ratio Lower Upper

154 100

153 40.5 32.7 49.1

76 13.1 10.7 16.1





#44

2-Chloronaphthalene

Concen: 37.216 ng/ul

RT: 13.640 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:162 Resp: 1470468

Ion Ratio Lower Upper

162 100

164 33.1 25.9 38.9

127 39.3 30.8 46.2



#45

2-Nitroaniline

Concen: 40.328 ng/ul

RT: 13.845 min Scan# 1812

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27




Tgt Ion: 65 Resp: 455917


Ion Ratio Lower Upper

65 100

92 70.6 59.8 89.8

138 106.5 92.3 138.5









#47
Dimethylphthalate
Concen: 36.821 ng/ul
RT: 14.210 min Scan# 1874
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:163 Resp: 1889580
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.1 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 40.547 ng/ul
RT: 14.334 min Scan# 1895
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:165 Resp: 412239
Ion Ratio Lower Upper
165 100
89 60.5 48.0 72.0
121 21.6 17.5 26.3









#50
Acenaphthylene
Concen: 35.589 ng/ul
RT: 14.481 min Scan# 1920
Delta R.T. -0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:152 Resp: 2237900
Ion Ratio Lower Upper
152 100
151 20.3 16.1 24.1
153 13.2 10.3 15.5



#51
3-Nitroaniline
Concen: 42.818 ng/ul
RT: 14.669 min Scan# 1952
Delta R.T. 0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:138 Resp: 391996
Ion Ratio Lower Upper
138 100
108 10.5 8.8 13.2
92 116.7 93.4 140.0









#52
Acenaphthene
Concen: 35.535 ng/ul
RT: 14.822 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:153 Resp: 1526154
Ion Ratio Lower Upper
153 100
152 47.9 38.0 57.0
154 89.6 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 31.301 ng/ul
RT: 14.875 min Scan# 1987
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:184 Resp: 211297
Ion Ratio Lower Upper
184 100
63 64.6 49.7 74.5
154 66.4 47.5 71.3









#55
4-Nitrophenol
Concen: 39.331 ng/ul
RT: 14.969 min Scan# 2002
Delta R.T. 0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:109 Resp: 276798
Ion Ratio Lower Upper
109 100
139 116.8 115.3 172.9
65 120.7 107.9 161.9



#56
Dibenzofuran
Concen: 36.297 ng/ul
RT: 15.157 min Scan# 2035
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:168 Resp: 2168274
Ion Ratio Lower Upper
168 100
139 38.2 29.4 44.0





#57

2,4-Dinitrotoluene

Concen: 38.779 ng/ul

RT: 15.122 min Scan# 2028

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:165 Resp: 570514

Ion Ratio Lower Upper

165 100

63 45.0 33.1 49.7

182 2.8 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 37.584 ng/ul

RT: 15.381 min Scan# 2073

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Tgt Ion:232 Resp: 473630




Ion Ratio Lower Upper


232 100

131 43.7 36.8 55.2

130 2.6 1.9 2.9

166 30.3 24.6 37.0









#59
Diethylphthalate
Concen: 37.051 ng/ul
RT: 15.569 min Scan# 2104
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:149 Resp: 1896401
Ion Ratio Lower Upper
149 100
177 21.9 17.5 26.3
150 12.2 9.8 14.8



#61
Fluorene
Concen: 35.248 ng/ul
RT: 15.804 min Scan# 2145
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:166 Resp: 1692114
Ion Ratio Lower Upper
166 100
165 97.1 77.8 116.6
167 13.0 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 37.036 ng/ul
RT: 15.792 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:204 Resp: 868725
Ion Ratio Lower Upper
204 100
206 33.0 26.1 39.1
141 60.7 44.5 66.7



#63
4-Nitroaniline
Concen: 47.838 ng/ul
RT: 15.834 min Scan# 2150
Delta R.T. 0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:138 Resp: 378195
Ion Ratio Lower Upper
138 100
92 51.6 41.2 61.8
108 88.9 60.1 90.1









#66
4,6-Dinitro-2-methylphenol
Concen: 33.912 ng/ul
RT: 15.887 min Scan# 2158
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:198 Resp: 313769
Ion Ratio Lower Upper
198 100
182 3.8 3.7 5.5
77 27.9 21.2 31.8



#67
N-Nitrosodiphenylamine
Concen: 35.201 ng/ul
RT: 16.010 min Scan# 2180
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:169 Resp: 1472431
Ion Ratio Lower Upper
169 100
168 67.7 53.5 80.3
167 36.6 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 38.791 ng/ul
RT: 16.692 min Scan# 2296
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:248 Resp: 568389
Ion Ratio Lower Upper
248 100
250 98.1 79.0 118.4
141 73.8 58.6 87.8



#69
Hexachlorobenzene
Concen: 38.026 ng/ul
RT: 16.810 min Scan# 2316
Delta R.T. -0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:284 Resp: 658439
Ion Ratio Lower Upper
284 100
142 34.7 27.3 40.9
249 31.6 24.6 36.8





#70

Atrazine

Concen: 23.409 ng/ul

RT: 16.957 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:200 Resp: 361926

Ion Ratio Lower Upper

200 100

173 25.9 21.4 32.2

215 47.8 38.1 57.1



#71

Pentachlorophenol

Concen: 34.898 ng/ul

RT: 17.157 min Scan# 2375

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27




Tgt Ion:266 Resp: 377987


Ion Ratio Lower Upper

266 100

264 62.0 51.0 76.4

268 64.6 51.8 77.6









#72
Phenanthrene
Concen: 35.290 ng/ul
RT: 17.557 min Scan# 2442
Delta R.T. 0.006 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:178 Resp: 2733732
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.2
176 19.2 15.6 23.4



#74
Anthracene
Concen: 34.383 ng/ul
RT: 17.645 min Scan# 2458
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:178 Resp: 2693323
Ion Ratio Lower Upper
178 100
179 15.8 12.2 18.2
176 18.7 15.2 22.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 39.157 ng/uL

RT: 13.563 min Scan# 1764

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:216 Resp: 789786

Ion Ratio Lower Upper

216 100

214 78.5 62.2 93.4

218 48.9 38.2 57.2



#76

Pentachlorobenzene

Concen: 37.088 ng/uL

RT: 15.075 min Scan# 2021

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27




Tgt Ion:250 Resp: 751099


Ion Ratio Lower Upper

250 100

248 62.1 49.8 74.6

252 64.0 51.6 77.4





#77

Carbazole

Concen: 34.181 ng/ul

RT: 17.910 min Scan# 2503

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:167 Resp: 2347801

Ion Ratio Lower Upper

167 100

166 21.8 17.4 26.2

139 12.9 10.2 15.4



#78

Di-n-butylphthalate

Concen: 36.290 ng/ul

RT: 18.439 min Scan# 2593

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27




Tgt Ion:149 Resp: 3210035


Ion Ratio Lower Upper

149 100

150 9.3 7.3 10.9

104 5.4 3.9 5.9





#80

Fluoranthene

Concen: 43.476 ng/ul

RT: 19.504 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:202 Resp: 2996827

Ion Ratio Lower Upper

202 100

101 10.7 9.5 14.3

100 8.3 7.3 10.9



#82

Pyrene

Concen: 41.321 ng/ul

RT: 19.857 min Scan# 2834

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27




Tgt Ion:202 Resp: 2941351


Ion Ratio Lower Upper

202 100

101 13.0 11.2 16.8

100 10.5 9.0 13.6





#83

Butylbenzylphthalate

Concen: 42.573 ng/ul

RT: 20.710 min Scan# 2979

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:149 Resp: 1343480

Ion Ratio Lower Upper

149 100

91 73.6 57.4 86.2

206 20.6 16.6 25.0



#84

3,3'-Dichlorobenzidine

Concen: 36.671 ng/ul

RT: 21.498 min Scan# 3113

Delta R.T. 0.006 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27




Tgt Ion:252 Resp: 785334


Ion Ratio Lower Upper

252 100

254 64.3 50.8 76.2

126 10.3 9.8 14.8





#85

Benzo(a)anthracene

Concen: 37.261 ng/ul

RT: 21.569 min Scan# 3125

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNL_P

ClientSampleId :

SLCS977





Tgt Ion:228 Resp: 2512265

Ion Ratio Lower Upper

228 100

229 20.0 16.1 24.1

226 27.0 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 43.478 ng/ul

RT: 21.469 min Scan# 3108

Delta R.T. -0.000 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27




Tgt Ion:149 Resp: 1975267


Ion Ratio Lower Upper

149 100

167 27.8 21.8 32.6

279 4.4 3.2 4.8









#87
Chrysene
Concen: 36.420 ng/ul
RT: 21.627 min Scan# 3134
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27




Instrument :
BNA_P
ClientSampleId :
SLCS977


Tgt Ion:228 Resp: 2347330
Ion Ratio Lower Upper
228 100
226 30.2 25.0 37.6
229 20.0 16.1 24.1



#89
Di-n-octyl phthalate
Concen: 44.667 ng/ul
RT: 22.439 min Scan# 3273
Delta R.T. -0.000 min
Lab File: BP015269.D
Acq: 24 May 2023 10:27

Tgt Ion:149 Resp: 3072912





#90

Benzo(b)fluoranthene

Concen: 37.414 ng/ul

RT: 23.357 min Scan# 3428

Delta R.T. 0.006 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:252 Resp: 2382598

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.2

125 9.4 8.3 12.5



#91

Benzo(k)fluoranthene

Concen: 38.987 ng/ul

RT: 23.410 min Scan# 3438

Delta R.T. 0.006 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27




Tgt Ion:252 Resp: 2383970


Ion Ratio Lower Upper

252 100

253 21.7 17.5 26.3

125 8.9 8.5 12.7





#93

Benzo(a)pyrene

Concen: 36.998 ng/ul

RT: 24.033 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:252 Resp: 2082509

Ion Ratio Lower Upper

252 100

253 21.6 17.2 25.8

125 10.4 9.1 13.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 37.196 ng/ul

RT: 26.857 min Scan# 4024

Delta R.T. 0.012 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27




Tgt Ion:276 Resp: 2710868


Ion Ratio Lower Upper

276 100

138 23.9 20.1 30.1

277 25.4 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 37.649 ng/ul

RT: 26.874 min Scan# 4024

Delta R.T. 0.012 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Instrument :

BNA_P

ClientSampleId :

SLCS977





Tgt Ion:278 Resp: 2252436

Ion Ratio Lower Upper

278 100

139 14.3 13.3 19.9

279 23.3 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 37.610 ng/ul

RT: 27.692 min Scan# 4166

Delta R.T. 0.018 min

Lab File: BP015269.D

Acq: 24 May 2023 10:27

Tgt Ion:276 Resp: 2235191

Ion Ratio Lower Upper

276 100

138 19.4 17.7 26.5

277 23.0 18.8 28.2

