

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052323\
 Data File : BP015256.D
 Acq On : 24 May 2023 00:22
 Operator : CG/JU
 Sample : 02861-11
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 H0FL4

Quant Time: May 24 00:58:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 23 01:25:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.116	152	232495	20.000	ng/u1	0.00
20) Naphthalene-d8	10.946	136	1028194	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.757	164	655922	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.510	188	1373906	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	981588	20.000	ng/u1	0.00
88) Perylene-d12	24.139	264	1036086	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	24676	4.124	ng/uL	0.00
4) Pyridine-d5	3.875	84	69067	4.072	ng/u1	0.00
7) Phenol-d5	7.263	99	122986	5.697	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.434	67	375179	30.093	ng/u1	0.00
11) 2-Chlorophenol-d4	7.640	132	374021	23.565	ng/u1	0.00
15) 4-Methylphenol-d8	8.828	113	248087	14.423	ng/u1	0.00
21) Nitrobenzene-d5	9.293	128	249941	30.935	ng/u1	0.00
24) 2-Nitrophenol-d4	10.022	143	279618	31.104	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.563	165	473844	29.210	ng/u1	0.00
31) 4-Chloroaniline-d4	11.081	131	303059	12.627	ng/u1	0.00
46) Dimethylphthalate-d6	14.157	166	1765445	35.619	ng/u1	0.00
49) Acenaphthylene-d8	14.451	160	1828106	32.249	ng/u1	0.00
54) 4-Nitrophenol-d4	14.951	143	40155	4.172	ng/u1	0.00
60) Fluorene-d10	15.745	176	1437386	34.368	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.863	200	248250	28.671	ng/u1	0.00
73) Anthracene-d10	17.610	188	2242543	35.646	ng/u1	0.00
81) Pyrene-d10	19.827	212	2475599	43.485	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.974	264	2043470	39.731	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052323\
Data File : BP015256.D
Acq On : 24 May 2023 00:22
Operator : CG/JU
Sample : 02861-11
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
H0FL4

Quant Time: May 24 00:58:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 23 01:25:41 2023
Response via : Initial Calibration

