

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052323\
 Data File : BP015242.D
 Acq On : 23 May 2023 15:59
 Operator : CG/JU
 Sample : 02861-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 H0FK6

Quant Time: May 23 16:56:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 23 01:25:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.122	152	204340	20.000	ng/u1	0.00
20) Naphthalene-d8	10.951	136	909122	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.757	164	575647	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.510	188	1201323	20.000	ng/u1	0.00
79) Chrysene-d12	21.586	240	840297	20.000	ng/u1	0.00
88) Perylene-d12	24.139	264	871617	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.428	96	23707	4.508	ng/uL	0.00
4) Pyridine-d5	3.875	84	103212	6.923	ng/u1	0.00
7) Phenol-d5	7.269	99	118378	6.239	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.440	67	373587	34.094	ng/u1	0.00
11) 2-Chlorophenol-d4	7.646	132	359543	25.774	ng/u1	0.00
15) 4-Methylphenol-d8	8.828	113	241148	15.951	ng/u1	0.00
21) Nitrobenzene-d5	9.299	128	251384	35.189	ng/u1	0.00
24) 2-Nitrophenol-d4	10.022	143	272073	34.228	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.563	165	448193	31.247	ng/u1	0.00
31) 4-Chloroaniline-d4	11.087	131	496898	23.414	ng/u1	0.00
46) Dimethylphthalate-d6	14.163	166	1665882	38.297	ng/u1	0.00
49) Acenaphthylene-d8	14.451	160	1791338	36.007	ng/u1	0.00
54) 4-Nitrophenol-d4	14.957	143	34685	4.106	ng/u1	0.00
60) Fluorene-d10	15.751	176	1358611	37.014	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.869	200	227199	30.010	ng/u1	0.00
73) Anthracene-d10	17.610	188	2223527	40.421	ng/u1	0.00
81) Pyrene-d10	19.827	212	2342152	48.058	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.974	264	1884170	43.547	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052323\
Data File : BP015242.D
Acq On : 23 May 2023 15:59
Operator : CG/JU
Sample : 02861-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
H0FK6

Quant Time: May 23 16:56:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 23 01:25:41 2023
Response via : Initial Calibration

