

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052423\
 Data File : BP015281.D
 Acq On : 24 May 2023 20:01
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP052423

Quant Time: May 25 01:44:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 25 01:41:43 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	132	0.00
2	1,4-Dioxane	40.000	38.744	3.1	121	0.00
3	Pyridine	40.000	40.892	-2.2	128	0.00
4	n-Nitrosodimethylamine	40.000	40.165	-0.4	129	0.00
5 S	2-Fluorophenol	80.000	78.905	1.4	125	0.00
6	Aniline	40.000	41.538	-3.8	135	0.00
7 S	Phenol-d6	80.000	81.422	-1.8	132	0.00
8	2-Chlorophenol	40.000	40.493	-1.2	132	0.00
9	Benzaldehyde	40.000	41.399	-3.5	135	0.00
10 C	Phenol	40.000	40.825	-2.1	134	0.00
11	bis(2-Chloroethyl)ether	40.000	40.068	-0.2	132	0.00
12	1,3-Dichlorobenzene	40.000	40.033	-0.1	132	0.00
13 C	1,4-Dichlorobenzene	40.000	40.117	-0.3	133	0.00
14	1,2-Dichlorobenzene	40.000	39.730	0.7	132	0.00
15	Benzyl Alcohol	40.000	43.067	-7.7	137	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.924	0.2	131	0.00
17	2-Methylphenol	40.000	41.555	-3.9	134	0.00
18	Hexachloroethane	40.000	40.207	-0.5	133	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.016	-2.5	132	0.00
20	3+4-Methylphenols	40.000	41.531	-3.8	132	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	134	0.00
22	Acetophenone	40.000	40.774	-1.9	134	0.00
23 S	Nitrobenzene-d5	80.000	82.400	-3.0	134	0.00
24	Nitrobenzene	40.000	41.005	-2.5	134	0.00
25	Isophorone	40.000	41.071	-2.7	131	0.00
26 C	2-Nitrophenol	40.000	43.442	-8.6	138	0.00
27	2,4-Dimethylphenol	40.000	41.160	-2.9	131	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.089	-2.7	132	0.00
29 C	2,4-Dichlorophenol	40.000	41.680	-4.2	134	0.00
30	1,2,4-Trichlorobenzene	40.000	40.435	-1.1	134	0.00
31	Naphthalene	40.000	40.101	-0.3	134	0.00
32	Benzoic acid	40.000	41.348	-3.4	137	0.01
33	4-Chloroaniline	40.000	41.577	-3.9	130	0.00
34 C	Hexachlorobutadiene	40.000	40.720	-1.8	136	0.00
35	Caprolactam	40.000	38.848	2.9	116	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.058	-2.6	128	0.00
37	2-Methylnaphthalene	40.000	39.900	0.3	130	0.00
38	1-Methylnaphthalene	40.000	39.769	0.6	129	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	125	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.199	-5.5	132	0.00
41 P	Hexachlorocyclopentadiene	40.000	46.087	-15.2	134	0.00
42 S	2,4,6-Tribromophenol	80.000	80.558	-0.7	118	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.460	-8.7	129	0.00
44	2,4,5-Trichlorophenol	40.000	42.830	-7.1	127	0.00
45 S	2-Fluorobiphenyl	80.000	81.789	-2.2	127	0.00
46	1,1'-Biphenyl	40.000	41.711	-4.3	129	0.00
47	2-Chloronaphthalene	40.000	41.214	-3.0	128	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052423\
 Data File : BP015281.D
 Acq On : 24 May 2023 20:01
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP052423

Quant Time: May 25 01:44:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 25 01:41:43 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.701	-6.8	121	0.00
49	Acenaphthylene	40.000	41.200	-3.0	123	0.00
50	Dimethylphthalate	40.000	39.284	1.8	118	0.00
51	2,6-Dinitrotoluene	40.000	41.438	-3.6	120	0.00
52 C	Acenaphthene	40.000	40.509	-1.3	123	0.00
53	3-Nitroaniline	40.000	41.543	-3.9	115	0.00
54 P	2,4-Dinitrophenol	40.000	38.162	4.6	116	0.00
55	Dibenzofuran	40.000	39.983	0.0	121	0.00
56 P	4-Nitrophenol	40.000	38.094	4.8	111	0.00
57	2,4-Dinitrotoluene	40.000	41.003	-2.5	114	0.00
58	Fluorene	40.000	39.162	2.1	118	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.144	-0.4	116	0.00
60	Diethylphthalate	40.000	38.977	2.6	115	0.00
61	4-Chlorophenyl-phenylether	40.000	39.624	0.9	121	0.00
62	4-Nitroaniline	40.000	37.328	6.7	109	0.00
63	Azobenzene	40.000	39.295	1.8	117	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	111	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.382	-11.0	111	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.556	-3.9	113	0.00
67	4-Bromophenyl-phenylether	40.000	42.786	-7.0	117	0.00
68	Hexachlorobenzene	40.000	42.181	-5.5	116	0.00
69	Atrazine	40.000	41.466	-3.7	107	0.00
70 C	Pentachlorophenol	40.000	44.336	-10.8	114	0.00
71	Phenanthrene	40.000	40.193	-0.5	110	0.00
72	Anthracene	40.000	40.830	-2.1	110	0.00
73	Carbazole	40.000	40.242	-0.6	105	0.00
74	Di-n-butylphthalate	40.000	40.932	-2.3	107	0.00
75 C	Fluoranthene	40.000	39.040	2.4	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	46.337	-15.8	97	0.00
78	Pyrene	40.000	41.500	-3.8	103	0.00
79 S	Terphenyl-d14	80.000	81.818	-2.3	103	0.00
80	Butylbenzylphthalate	40.000	42.699	-6.7	101	0.00
81	Benzo(a)anthracene	40.000	41.225	-3.1	100	0.00
82	3,3'-Dichlorobenzidine	40.000	43.115	-7.8	101	0.00
83	Chrysene	40.000	40.795	-2.0	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.232	-8.1	101	0.00
85 c	Di-n-octyl phthalate	40.000	44.095	-10.2	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.587	-4.0	102	0.00
88	Benzo(b)fluoranthene	40.000	39.154	2.1	100	0.00
89	Benzo(k)fluoranthene	40.000	39.482	1.3	101	0.00
90 C	Benzo(a)pyrene	40.000	40.369	-0.9	102	0.00
91	Dibenzo(a,h)anthracene	40.000	41.687	-4.2	102	0.00
92	Benzo(g,h,i)perylene	40.000	41.918	-4.8	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052423\
Data File : BP015281.D
Acq On : 24 May 2023 20:01
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP052423

Quant Time: May 25 01:44:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 25 01:41:43 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0