

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP052820\
 Data File : BP002304.D
 Acq On : 28 May 2020 11:59
 Operator : CG/JU
 Sample : L2744-13
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 NM42-COMP-2020-0-6

Quant Time: May 28 12:42:11 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP052720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 28 11:01:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	312310	20.00	ng	0.00
21) Naphthalene-d8	10.38	136	1268931	20.00	ng	0.00
39) Acenaphthene-d10	14.25	164	729262	20.00	ng	0.00
64) Phenanthrene-d10	17.00	188	1486856	20.00	ng	0.00
76) Chrysene-d12	21.10	240	1356827	20.00	ng	0.00
86) Perylene-d12	23.29	264	1428184	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.22	112	1520585	91.44	ng	0.00
7) Phenol-d6	6.79	99	2138656	94.54	ng	0.00
23) Nitrobenzene-d5	8.75	82	1476383	69.95	ng	0.00
42) 2,4,6-Tribromophenol	15.74	330	657811	95.17	ng	0.00
45) 2-Fluorobiphenyl	12.87	172	3024797	73.19	ng	0.00
79) Terphenyl-d14	19.59	244	3872096	69.11	ng	0.00
Target Compounds						Qvalue
50) Dimethylphthalate	13.71	163	401262	8.129	ng	99
71) Phenanthrene	17.04	178	337433	4.791	ng	98
75) Fluoranthene	19.02	202	605513	7.465	ng	99
78) Pyrene	19.37	202	534847	6.330	ng	100
81) Benzo(a)anthracene	21.08	228	284650	3.665	ng	95
83) Chrysene	21.13	228	257931	3.460	ng	98
88) Benzo(b)fluoranthene	22.64	252	343087	4.406	ng	# 96
90) Benzo(a)pyrene	23.19	252	235983	3.267	ng	# 97
92) Benzo(g,h,i)perylene	26.17	276	161327	2.194	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP052820\
Data File : BP002304.D
Acq On : 28 May 2020 11:59
Operator : CG/JU
Sample : L2744-13
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
NM42-COMP-2020-0-6

Quant Time: May 28 12:42:11 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP052720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 28 11:01:43 2020
Response via : Initial Calibration

