


Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP052820\
 Data File : BP002300.D
 Acq On : 28 May 2020 09:13
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP052720.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu May 28 11:01:43 2020



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2474

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.9	61220	PASS
68	69	0.00	2	1.6	936	PASS
69	198	0.00	100	36.5	58950	PASS
70	69	0.00	2	0.6	326	PASS
127	198	10	80	45.8	73850	PASS
197	198	0.00	2	0.2	379	PASS
198	198	100	100	100.0	161322	PASS
199	198	5	9	7.4	11886	PASS
275	198	10	60	23.6	38008	PASS
365	198	1	100	2.7	4435	PASS
441	198	0.01	100	11.7	18855	PASS
442	198	50	100	62.3	100470	PASS
443	442	15	24	21.5	21600	PASS