



2,4-Dinitrophenol



Response = 2.060e-001 * Amt - 6.847e-002
 Coef of Det (r^2) = 0.997823 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP052822.M
 Calibration Table Last Updated: Sat May 28 02:18:04 2022

Benzoic acid


Response Ratio



$R = 1.062e-002 A^2 + 2.063e-001 A - 7.698e-002$
 Coef of Det (r^2) = 0.991916 Curve Fit: Quadratic
 Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP052822.M
 Calibration Table Last Updated: Sat May 28 02:18:04 2022

4-Nitrophenol

Response Ratio



$$\text{Response} = 2.641\text{e-}001 * \text{Amt} - 4.751\text{e-}002$$


Coef of Det (r^2) = 0.996684 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP052822.M

Calibration Table Last Updated: Sat May 28 02:18:04 2022

4-Nitroaniline

Response Ratio




$$\text{Response} = 3.340\text{e-}001 * \text{Amt} - 3.562\text{e-}002$$

Coef of Det (r^2) = 0.998292 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP052822.M

Calibration Table Last Updated: Sat May 28 02:18:04 2022

4,6-Dinitro-2-methylphenol



Response = $1.405e-001 \cdot \text{Amt} - 3.233e-002$
 Coef of Det (r^2) = 0.995056 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP052822.M
 Calibration Table Last Updated: Sat May 28 02:18:04 2022