

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052822\
 Data File : BP010588.D
 Acq On : 28 May 2022 01:13
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP052822

Quant Time: May 28 02:37:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 28 02:18:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.466	0.469	-0.6	94	0.00
3	Pyridine	1.304	1.353	-3.8	95	0.00
4	n-Nitrosodimethylamine	0.795	0.762	4.2	91	0.00
5 S	2-Fluorophenol	1.096	1.128	-2.9	95	0.00
6	Aniline	1.831	1.862	-1.7	96	0.00
7 S	Phenol-d6	1.556	1.588	-2.1	95	0.00
8	2-Chlorophenol	1.237	1.233	0.3	94	0.00
9	Benzaldehyde	1.024	0.949	7.3	97	0.00
10 C	Phenol	1.609	1.629	-1.2	96	0.00
11	bis(2-Chloroethyl)ether	1.276	1.253	1.8	92	0.00
12	1,3-Dichlorobenzene	1.434	1.414	1.4	93	0.00
13 C	1,4-Dichlorobenzene	1.469	1.448	1.4	93	0.00
14	1,2-Dichlorobenzene	1.407	1.392	1.1	94	0.00
15	Benzyl Alcohol	1.248	1.277	-2.3	95	0.00
16	2,2'-oxybis(1-Chloropropane	2.385	2.234	6.3	90	-0.01
17	2-Methylphenol	1.080	1.090	-0.9	94	0.00
18	Hexachloroethane	0.568	0.554	2.5	92	0.00
19 P	n-Nitroso-di-n-propylamine	1.135	1.133	0.2	97	0.00
20	3+4-Methylphenols	1.485	1.521	-2.4	95	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
22	Acetophenone	0.522	0.520	0.4	94	0.00
23 S	Nitrobenzene-d5	0.423	0.429	-1.4	95	0.00
24	Nitrobenzene	0.427	0.431	-0.9	94	0.00
25	Isophorone	0.707	0.729	-3.1	97	0.00
26 C	2-Nitrophenol	0.167	0.176	-5.4	97	0.00
27	2,4-Dimethylphenol	0.248	0.257	-3.6	96	0.00
28	bis(2-Chloroethoxy)methane	0.390	0.392	-0.5	94	0.00
29 C	2,4-Dichlorophenol	0.298	0.311	-4.4	95	0.00
30	1,2,4-Trichlorobenzene	0.350	0.349	0.3	94	0.00
31	Naphthalene	1.054	1.054	0.0	94	0.00
32	Benzoic acid	0.174	0.203	-16.7	111	0.00
33	4-Chloroaniline	0.407	0.427	-4.9	98	0.00
34 C	Hexachlorobutadiene	0.236	0.226	4.2	92	0.00
35	Caprolactam	0.086	0.107	-24.4	111	0.00
36 C	4-Chloro-3-methylphenol	0.339	0.367	-8.3	101	0.00
37	2-Methylnaphthalene	0.735	0.743	-1.1	96	0.00
38	1-Methylnaphthalene	0.718	0.725	-1.0	96	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	0.622	0.601	3.4	95	0.00
41 P	Hexachlorocyclopentadiene	0.331	0.312	5.7	90	0.00
42 S	2,4,6-Tribromophenol	0.226	0.253	-11.9	108	0.00
43 C	2,4,6-Trichlorophenol	0.388	0.399	-2.8	97	0.00
44	2,4,5-Trichlorophenol	0.439	0.453	-3.2	100	0.00
45 S	2-Fluorobiphenyl	1.409	1.371	2.7	97	0.00
46	1,1'-Biphenyl	1.492	1.453	2.6	97	0.00
47	2-Chloronaphthalene	1.173	1.150	2.0	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052822\
 Data File : BP010588.D
 Acq On : 28 May 2022 01:13
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP052822

Quant Time: May 28 02:37:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 28 02:18:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.406	0.439	-8.1	103	0.00
49	Acenaphthylene	1.801	1.852	-2.8	102	0.00
50	Dimethylphthalate	1.440	1.520	-5.6	104	0.00
51	2,6-Dinitrotoluene	0.306	0.333	-8.8	105	0.00
52 C	Acenaphthene	1.106	1.114	-0.7	101	0.00
53	3-Nitroaniline	0.300	0.343	-14.3	108	0.00
54 P	2,4-Dinitrophenol	0.158	0.185	-17.1	111	0.00
55	Dibenzofuran	1.749	1.787	-2.2	101	0.00
56 P	4-Nitrophenol	0.226	0.278	-23.0	114	0.00
57	2,4-Dinitrotoluene	0.399	0.463	-16.0	109	0.00
58	Fluorene	1.430	1.483	-3.7	102	0.00
59	2,3,4,6-Tetrachlorophenol	0.377	0.412	-9.3	106	0.00
60	Diethylphthalate	1.433	1.545	-7.8	105	0.00
61	4-Chlorophenyl-phenylether	0.718	0.738	-2.8	102	0.00
62	4-Nitroaniline	0.291	0.368	-26.5#	113	0.00
63	Azobenzene	1.514	1.584	-4.6	102	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	0.117	0.125	-6.8	114	0.00
66 c	n-Nitrosodiphenylamine	0.602	0.581	3.5	105	0.00
67	4-Bromophenyl-phenylether	0.218	0.209	4.1	105	0.00
68	Hexachlorobenzene	0.240	0.230	4.2	106	0.00
69	Atrazine	0.195	0.207	-6.2	111	0.00
70 C	Pentachlorophenol	0.138	0.146	-5.8	110	0.00
71	Phenanthrene	1.097	1.095	0.2	107	0.00
72	Anthracene	1.050	1.066	-1.5	108	0.00
73	Carbazole	0.893	0.967	-8.3	111	0.00
74	Di-n-butylphthalate	1.135	1.225	-7.9	108	0.00
75 C	Fluoranthene	1.191	1.295	-8.7	111	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
77	Benzydine	0.384	0.414	-7.8	121	0.00
78	Pyrene	1.394	1.349	3.2	112	0.00
79 S	Terphenyl-d14	1.065	1.027	3.6	112	0.00
80	Butylbenzylphthalate	0.549	0.582	-6.0	114	0.00
81	Benzo(a)anthracene	1.312	1.328	-1.2	114	0.00
82	3,3'-Dichlorobenzidine	0.360	0.390	-8.3	115	0.00
83	Chrysene	1.287	1.285	0.2	113	0.00
84	Bis(2-ethylhexyl)phthalate	0.774	0.849	-9.7	115	0.00
85 c	Di-n-octyl phthalate	1.289	1.391	-7.9	113	0.00
86 I	Perylene-d12	1.000	1.000	0.0	112	0.00
87	Indeno(1,2,3-cd)pyrene	1.458	1.453	0.3	111	0.00
88	Benzo(b)fluoranthene	1.259	1.256	0.2	109	0.00
89	Benzo(k)fluoranthene	1.276	1.322	-3.6	118	0.00
90 C	Benzo(a)pyrene	1.042	1.069	-2.6	113	0.00
91	Dibenzo(a,h)anthracene	1.219	1.217	0.2	112	0.00
92	Benzo(g,h,i)perylene	1.212	1.200	1.0	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052822\
Data File : BP010588.D
Acq On : 28 May 2022 01:13
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP052822

Quant Time: May 28 02:37:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 28 02:18:04 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range SPCC's out = 0 CCC's out = 0