

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP052920\
 Data File : BP002320.D
 Acq On : 29 May 2020 18:02
 Operator : CG/JU
 Sample : L2773-17 2X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 NM33-COMP-2020-0-6

Manual Integrations
 APPROVED

mohammad
 6/1/2020 12:24:24 PM

Quant Time: May 30 01:05:20 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP052720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 29 15:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	320391	20.00	ng	0.00
21) Naphthalene-d8	10.38	136	1319472	20.00	ng	0.00
39) Acenaphthene-d10	14.25	164	776980	20.00	ng	0.00
64) Phenanthrene-d10	17.00	188	1614580	20.00	ng	0.00
76) Chrysene-d12	21.10	240	1496231	20.00	ng	0.00
86) Perylene-d12	23.31	264	1514540	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.22	112	1352067	79.26	ng	0.00
7) Phenol-d6	6.79	99	1878729	80.96	ng	0.00
23) Nitrobenzene-d5	8.75	82	1238813	56.45	ng	0.00
42) 2,4,6-Tribromophenol	15.75	330	597406	81.13	ng	0.00
45) 2-Fluorobiphenyl	12.87	172	2604934	59.16	ng	0.00
79) Terphenyl-d14	19.59	244	3521001	56.99	ng	0.00
Target Compounds						
50) Dimethylphthalate	13.71	163	249035	4.735	ng	99
71) Phenanthrene	17.05	178	247594	3.237	ng	99
75) Fluoranthene	19.03	202	489792	5.560	ng	99
78) Pyrene	19.38	202	469110	5.035	ng	99
80) Butylbenzylphthalate	20.28	149	111040	2.993	ng	98
81) Benzo(a)anthracene	21.09	228	248700	2.904	ng	99
83) Chrysene	21.14	228	213752	2.600	ng	97
88) Benzo(b)fluoranthene	22.65	252	299768m	3.630	ng	
90) Benzo(a)pyrene	23.21	252	223216	2.914	ng	# 96
92) Benzo(g,h,i)perylene	26.20	276	161648	2.073	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP052920\
Data File : BP002320.D
Acq On : 29 May 2020 18:02
Operator : CG/JU
Sample : L2773-17 2X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
NM33-COMP-2020-0-6

Manual Integrations
APPROVED

mohammad
6/1/2020 12:24:24 PM

Quant Time: May 30 01:05:20 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP052720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 29 15:09:03 2020
Response via : Initial Calibration

