

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052922\
 Data File : BP010632.D
 Acq On : 29 May 2022 22:33
 Operator : CG/JU
 Sample : N2934-05DL 20X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 P025-SS001-1824-01DL

Quant Time: May 30 02:08:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 28 02:18:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	158180	20.000	ng	0.00
21) Naphthalene-d8	10.669	136	635418	20.000	ng	0.00
39) Acenaphthene-d10	14.504	164	393871	20.000	ng	0.00
64) Phenanthrene-d10	17.257	188	822593	20.000	ng	0.00
76) Chrysene-d12	21.345	240	719159	20.000	ng	0.00
86) Perylene-d12	23.756	264	730019	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.452	112	53633	6.186	ng	0.00
7) Phenol-d6	7.051	99	65997	5.362	ng	0.01
23) Nitrobenzene-d5	9.034	82	31042	2.310	ng	0.00
42) 2,4,6-Tribromophenol	16.004	330	24088	5.403	ng	0.01
45) 2-Fluorobiphenyl	13.128	172	99059	3.569	ng	0.00
79) Terphenyl-d14	19.810	244	153375	4.006	ng	0.00
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	21.251	149	2669759	95.953	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP052922\
Data File : BP010632.D
Acq On : 29 May 2022 22:33
Operator : CG/JU
Sample : N2934-05DL 20X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
P025-SS001-1824-01DL

Quant Time: May 30 02:08:33 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 28 02:18:04 2022
Response via : Initial Calibration

