

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060122\
 Data File : BP010661.D
 Acq On : 01 Jun 2022 20:39
 Operator : CG/JU
 Sample : N3080-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 WC-4

Quant Time: Jun 02 10:06:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 28 02:18:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	97931	20.000	ng	-0.01
21) Naphthalene-d8	10.657	136	418441	20.000	ng	-0.01
39) Acenaphthene-d10	14.492	164	294545	20.000	ng	-0.01
64) Phenanthrene-d10	17.251	188	668909	20.000	ng	0.00
76) Chrysene-d12	21.339	240	774115	20.000	ng	0.00
86) Perylene-d12	23.745	264	817500	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.446	112	886026	165.053	ng	0.00
7) Phenol-d6	7.040	99	1119836	146.943	ng	0.00
23) Nitrobenzene-d5	9.022	82	894291	101.043	ng	0.00
42) 2,4,6-Tribromophenol	15.992	330	621682	186.461	ng	0.00
45) 2-Fluorobiphenyl	13.116	172	1943975	93.669	ng	-0.01
79) Terphenyl-d14	19.809	244	4054057	98.367	ng	0.00

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060122\
Data File : BP010661.D
Acq On : 01 Jun 2022 20:39
Operator : CG/JU
Sample : N3080-02
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
WC-4

Quant Time: Jun 02 10:06:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 28 02:18:04 2022
Response via : Initial Calibration

