

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060221\
 Data File : BP005668.D
 Acq On : 02 Jun 2021 16:40
 Operator : CG/JU
 Sample : M2580-16
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 18-B6(0-6)

Quant Time: Jun 02 17:13:35 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 16:11:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.987	152	300226	20.00	ng	0.00
21) Naphthalene-d8	10.793	136	1124209	20.00	ng	0.00
39) Acenaphthene-d10	14.610	164	631101	20.00	ng	0.00
64) Phenanthrene-d10	17.351	188	1204459	20.00	ng	# 0.00
76) Chrysene-d12	21.422	240	1350023	20.00	ng	# 0.00
86) Perylene-d12	23.869	264	1627456	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.546	112	1646678	85.27	ng	0.00
7) Phenol-d6	7.152	99	2068803	78.59	ng	0.00
23) Nitrobenzene-d5	9.152	82	1321069	65.35	ng	0.00
42) 2,4,6-Tribromophenol	16.093	330	1045001	129.13	ng	0.00
45) 2-Fluorobiphenyl	13.240	172	3121777	65.46	ng	0.00
79) Terphenyl-d14	19.898	244	4176082	58.04	ng	0.00
Target Compounds						Qvalue
26) 2-Nitrophenol	9.916	139	23	4.16	ng	# 1
32) Benzoic acid	10.087	122	192	5.03	ng	# 56
48) 2-Nitroaniline	13.710	65	175	4.50	ng	# 69
50) Dimethylphthalate	14.063	163	323632	6.45	ng	99
51) 2,6-Dinitrotoluene	14.446	165	419	5.09	ng	90
52) Acenaphthene	14.675	154	111588	2.64	ng	98
53) 3-Nitroaniline	14.522	138	159	3.67	ng	# 73
54) 2,4-Dinitrophenol	14.722	184	81	6.54	ng	# 1
56) 4-Nitrophenol	14.587	139	1115	3.86	ng	94
57) 2,4-Dinitrotoluene	14.963	165	215	4.27	ng	# 45
58) Fluorene	15.651	166	134059	2.59	ng	96
62) 4-Nitroaniline	15.657	138	2156	3.80	ng	# 15
71) Phenanthrene	17.398	178	2059714	27.45	ng	100
72) Anthracene	17.487	178	536091	7.24	ng	99
73) Carbazole	17.751	167	220456	3.02	ng	99
75) Fluoranthene	19.351	202	3945576	43.80	ng	96
78) Pyrene	19.704	202	3654289	38.36	ng	99
80) Butylbenzylphthalate	20.569	149	17397	4.54	ng	99
81) Benzo(a)anthracene	21.404	228	2201132	22.22	ng	97
83) Chrysene	21.463	228	1987427	20.72	ng	97
87) Indeno(1,2,3-cd)pyrene	26.433	276	1696438	12.25	ng	# 95
88) Benzo(b)fluoranthene	23.122	252	4035655	34.18	ng	# 95
89) Benzo(k)fluoranthene	23.122	252	4035655	36.07	ng	# 93
90) Benzo(a)pyrene	23.763	252	2270854	21.14	ng	# 95
91) Dibenzo(a,h)anthracene	26.439	278	449510	3.74	ng	# 81
92) Benzo(g,h,i)perylene	27.227	276	1658891	14.31	ng	# 94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060221\
Data File : BP005668.D
Acq On : 02 Jun 2021 16:40
Operator : CG/JU
Sample : M2580-16
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)

Quant Time: Jun 02 17:13:35 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 27 16:11:11 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.987 min Scan# 833
Delta R.T. -0.006 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion:152 Resp: 300226
Ion Ratio Lower Upper
152 100
150 154.0 119.8 179.8
115 58.0 49.0 73.4



#5
2-Fluorophenol
Concen: 85.27 ng
RT: 5.546 min Scan# 418
Delta R.T. -0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:112 Resp: 1646678
Ion Ratio Lower Upper
112 100
64 52.8 33.6 50.4#
63 28.8 18.6 27.8#









#7
Phenol-d6
Concen: 78.59 ng
RT: 7.152 min Scan# 691
Delta R.T. 0.006 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)

Tgt Ion: 99 Resp: 2068803




Ion	Ratio	Lower	Upper
99	100		
42	16.6	12.9	19.3
71	34.1	26.7	40.1




#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.793 min Scan# 1310
Delta R.T. -0.006 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:136 Resp: 1124209

Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.9	13.3
54	6.2	5.0	7.4
68	5.4	3.7	5.5









#23
Nitrobenzene-d5
Concen: 65.35 ng
RT: 9.152 min Scan# 1031
Delta R.T. -0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)

Tgt Ion: 82 Resp: 1321069




Ion	Ratio	Lower	Upper
82	100		
128	45.3	30.5	45.7
54	43.0	33.9	50.9




#26
2-Nitrophenol
Concen: 4.16 ng
RT: 9.916 min Scan# 1161
Delta R.T. 0.005 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:139 Resp: 23

Ion	Ratio	Lower	Upper
139	100		
109	281.8	40.7	61.1#
65	579.5	47.8	71.6#









#32
Benzoic acid
Concen: 5.03 ng
RT: 10.087 min Scan# 1190
Delta R.T. -0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion:122 Resp: 192
Ion Ratio Lower Upper
122 100
105 96.5 105.1 145.1#
77 156.6 69.6 109.6#



#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.610 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:164 Resp: 631101
Ion Ratio Lower Upper
164 100
162 101.4 73.5 110.3
160 44.9 40.6 61.0









#42
2,4,6-Tribromophenol
Concen: 129.13 ng
RT: 16.093 min Scan# 2211
Delta R.T. -0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion:330 Resp: 1045001
Ion Ratio Lower Upper
330 100
332 96.9 76.3 114.5
141 24.6 22.8 34.2



#45
2-Fluorobiphenyl
Concen: 65.46 ng
RT: 13.240 min Scan# 1726
Delta R.T. -0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:172 Resp: 3121777
Ion Ratio Lower Upper
172 100
171 35.3 28.1 42.1
170 23.7 19.6 29.4









#48
2-Nitroaniline
Concen: 4.50 ng
RT: 13.710 min Scan# 1806
Delta R.T. 0.017 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion: 65 Resp: 175
Ion Ratio Lower Upper
65 100
92 54.2 48.8 73.2
138 44.7 69.5 104.3#



#50
Dimethylphthalate
Concen: 6.45 ng
RT: 14.063 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:163 Resp: 323632
Ion Ratio Lower Upper
163 100
194 4.4 3.9 5.9
164 10.5 8.1 12.1









#51
2,6-Dinitrotoluene
Concen: 5.09 ng
RT: 14.446 min Scan# 1931
Delta R.T. 0.259 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion:165 Resp: 419
Ion Ratio Lower Upper
165 100
63 84.2 67.4 101.0
89 109.5 73.8 110.6



#52
Acenaphthene
Concen: 2.64 ng
RT: 14.675 min Scan# 1970
Delta R.T. -0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:154 Resp: 111588
Ion Ratio Lower Upper
154 100
153 112.3 90.9 136.3
152 57.0 48.5 72.7









#53
3-Nitroaniline
Concen: 3.67 ng
RT: 14.522 min Scan# 1944
Delta R.T. 0.012 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion:138 Resp: 159
Ion Ratio Lower Upper
138 100
108 85.1 10.2 15.4#
92 120.9 108.1 162.1



#54
2,4-Dinitrophenol
Concen: 6.54 ng
RT: 14.722 min Scan# 1978
Delta R.T. 0.012 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:184 Resp: 81
Ion Ratio Lower Upper
184 100
63 499.0 51.2 76.8#
154 1154.4 55.8 83.8#









#56
4-Nitrophenol
Concen: 3.86 ng
RT: 14.587 min Scan# 1955
Delta R.T. -0.217 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion:139 Resp: 1115
Ion Ratio Lower Upper
139 100
109 151.9 122.8 162.8
65 134.5 110.0 150.0



#57
2,4-Dinitrotoluene
Concen: 4.27 ng
RT: 14.963 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:165 Resp: 215
Ion Ratio Lower Upper
165 100
63 41.1 56.7 85.1#
89 45.2 85.8 128.6#
182 95.4 3.0 4.6#





#58

Fluorene

Concen: 2.59 ng

RT: 15.651 min Scan# 2136

Delta R.T. -0.006 min

Lab File: BP005668.D

Acq: 02 Jun 2021 16:40

Tgt Ion:166 Resp: 134059

Ion Ratio Lower Upper

166 100

165 102.9 78.8 118.2





167 14.6 10.9 16.3

Instrument :

BNA_P

ClientSampleId :

18-B6(0-6)



#62

4-Nitroaniline

Concen: 3.80 ng

RT: 15.657 min Scan# 2137

Delta R.T. -0.012 min

Lab File: BP005668.D

Acq: 02 Jun 2021 16:40




Tgt Ion:138 Resp: 2156


Ion Ratio Lower Upper

138 100

92 33.4 50.9 90.9#

108 33.7 154.7 194.7#









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.351 min Scan# 2425
Delta R.T. 0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion:188 Resp: 1204459
Ion Ratio Lower Upper
188 100
94 7.7 5.0 7.6#
80 8.4 5.9 8.9



#71
Phenanthrene
Concen: 27.45 ng
RT: 17.398 min Scan# 2433
Delta R.T. 0.006 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:178 Resp: 2059714
Ion Ratio Lower Upper
178 100
176 19.4 15.4 23.2
179 15.2 12.3 18.5









#72
 Anthracene
 Concen: 7.24 ng
 RT: 17.487 min Scan# 2448
 Delta R.T. -0.000 min
 Lab File: BP005668.D
 Acq: 02 Jun 2021 16:40




Instrument :
 BNA_P
 ClientSampleId :
 18-B6(0-6)


Tgt Ion:178	Resp: 536091
Ion Ratio	Lower Upper
178 100	
176 18.9	15.2 22.8
179 15.1	12.6 19.0



#73
 Carbazole
 Concen: 3.02 ng
 RT: 17.751 min Scan# 2493
 Delta R.T. -0.000 min
 Lab File: BP005668.D
 Acq: 02 Jun 2021 16:40

Tgt Ion:167	Resp: 220456
Ion Ratio	Lower Upper
167 100	
166 22.0	17.4 26.0
139 12.4	10.4 15.6









#75
Fluoranthene
Concen: 43.80 ng
RT: 19.351 min Scan# 2765
Delta R.T. -0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion:202 Resp: 3945576
Ion Ratio Lower Upper
202 100
101 9.5 0.0 26.1
203 18.2 0.0 37.2



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.422 min Scan# 3117
Delta R.T. -0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:240 Resp: 1350023
Ion Ratio Lower Upper
240 100
120 8.9 4.2 6.2#
236 25.7 20.2 30.2





#78

Pyrene

Concen: 38.36 ng

RT: 19.704 min Scan# 2825





Delta R.T. 0.006 min

Lab File: BP005668.D

Acq: 02 Jun 2021 16:40

Tgt Ion:202 Resp: 3654289

Ion	Ratio	Lower	Upper
202	100		
200	21.3	16.9	25.3
203	18.2	13.8	20.6



#79

Terphenyl-d14

Concen: 58.04 ng

RT: 19.898 min Scan# 2858




Delta R.T. -0.000 min


Lab File: BP005668.D

Acq: 02 Jun 2021 16:40

Tgt Ion:244 Resp: 4176082

Ion	Ratio	Lower	Upper
244	100		
212	6.9	5.8	8.6
122	8.4	5.2	7.8









#80
Butylbenzylphthalate
Concen: 4.54 ng
RT: 20.569 min Scan# 2972
Delta R.T. -0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion:149 Resp: 17397
Ion Ratio Lower Upper
149 100
91 76.2 60.2 90.4
206 30.4 23.9 35.9



#81
Benzo(a)anthracene
Concen: 22.22 ng
RT: 21.404 min Scan# 3114
Delta R.T. -0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:228 Resp: 2201132
Ion Ratio Lower Upper
228 100
226 27.9 21.2 31.8
229 21.0 15.6 23.4









#83
Chrysene
Concen: 20.72 ng
RT: 21.463 min Scan# 3124
Delta R.T. 0.006 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion:228 Resp: 1987427
Ion Ratio Lower Upper
228 100
226 30.7 24.8 37.2
229 22.6 15.4 23.0



#86
Perylene-d12
Concen: 20.00 ng
RT: 23.869 min Scan# 3533
Delta R.T. 0.006 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:264 Resp: 1627456
Ion Ratio Lower Upper
264 100
260 24.2 20.6 30.8
265 23.3 19.0 28.6









#87
Indeno(1,2,3-cd)pyrene
Concen: 12.25 ng
RT: 26.433 min Scan# 3969
Delta R.T. -0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion:276 Resp: 1696438
Ion Ratio Lower Upper
276 100
138 15.8 8.4 12.6#
277 24.8 20.6 31.0



#88
Benzo(b)fluoranthene
Concen: 34.18 ng
RT: 23.122 min Scan# 3406
Delta R.T. 0.006 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:252 Resp: 4035655
Ion Ratio Lower Upper
252 100
253 23.9 18.2 27.2
125 9.7 3.4 5.0#









#89
Benzo(k)fluoranthene
Concen: 36.07 ng
RT: 23.122 min Scan# 3406
Delta R.T. -0.047 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40




Instrument :
BNA_P
ClientSampleId :
18-B6(0-6)


Tgt Ion:252 Resp: 4035655
Ion Ratio Lower Upper
252 100
253 23.9 17.2 25.8
125 9.7 3.8 5.6#



#90
Benzo(a)pyrene
Concen: 21.14 ng
RT: 23.763 min Scan# 3515
Delta R.T. 0.006 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:252 Resp: 2270854
Ion Ratio Lower Upper
252 100
253 23.0 17.1 25.7
125 9.2 3.8 5.6#









#91
Dibenzo(a,h)anthracene
Concen: 3.74 ng
RT: 26.439 min Scan# 3970
Delta R.T. -0.012 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Instrument :
BNA_P
ClientSampled :
18-B6(0-6)

Tgt Ion:278 Resp: 449510
Ion Ratio Lower Upper
278 100
139 19.8 5.0 7.6#
279 29.5 18.6 28.0#



#92
Benzo(g,h,i)perylene
Concen: 14.31 ng
RT: 27.227 min Scan# 4104
Delta R.T. -0.000 min
Lab File: BP005668.D
Acq: 02 Jun 2021 16:40

Tgt Ion:276 Resp: 1658891
Ion Ratio Lower Upper
276 100
277 24.7 19.8 29.8
138 16.8 7.7 11.5#

