

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP060221\  
 Data File : BP005672.D  
 Acq On : 02 Jun 2021 18:56  
 Operator : CG/JU  
 Sample : M2580-05 2X  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 18-B2(2-4)

Quant Time: Jun 03 02:42:32 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP052721.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu May 27 16:11:11 2021  
 Response via : Initial Calibration


| Compound                    | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|-------|----------|
| -----                       |        |      |          |       |       |          |
| Internal Standards          |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.993  | 152  | 383955   | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.798 | 136  | 1341744  | 20.00 | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.610 | 164  | 734198   | 20.00 | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.357 | 188  | 1443026  | 20.00 | ng    | # 0.00   |
| 76) Chrysene-d12            | 21.421 | 240  | 1708599  | 20.00 | ng    | # 0.00   |
| 86) Perylene-d12            | 23.874 | 264  | 2034122  | 20.00 | ng    | 0.01     |
|                             |        |      |          |       |       |          |
| System Monitoring Compounds |        |      |          |       |       |          |
| 5) 2-Fluorophenol           | 5.552  | 112  | 889178   | 36.00 | ng    | 0.00     |
| 7) Phenol-d6                | 7.151  | 99   | 1026504  | 30.49 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.151  | 82   | 645790   | 26.77 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 16.098 | 330  | 443386   | 47.10 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.239 | 172  | 1407484  | 25.37 | ng    | 0.00     |
| 79) Terphenyl-d14           | 19.898 | 244  | 2062967  | 22.66 | ng    | 0.00     |
|                             |        |      |          |       |       |          |
| Target Compounds            |        |      |          |       |       | Qvalue   |
| 26) 2-Nitrophenol           | 9.910  | 139  | 50       | 4.16  | ng    | # 32     |
| 32) Benzoic acid            | 10.040 | 122  | 598      | 5.06  | ng    | # 53     |
| 48) 2-Nitroaniline          | 13.745 | 65   | 129      | 4.50  | ng    | # 52     |
| 50) Dimethylphthalate       | 14.063 | 163  | 172088   | 2.95  | ng    | 98       |
| 51) 2,6-Dinitrotoluene      | 14.451 | 165  | 206      | 5.07  | ng    | 90       |
| 53) 3-Nitroaniline          | 14.516 | 138  | 115      | 3.66  | ng    | # 34     |
| 54) 2,4-Dinitrophenol       | 14.710 | 184  | 113      | 6.55  | ng    | # 50     |
| 56) 4-Nitrophenol           | 14.804 | 139  | 245      | 3.77  | ng    | 91       |
| 57) 2,4-Dinitrotoluene      | 14.957 | 165  | 243      | 4.27  | ng    | # 41     |
| 62) 4-Nitroaniline          | 15.657 | 138  | 848      | 3.69  | ng    | # 17     |
| 71) Phenanthrene            | 17.398 | 178  | 490443   | 5.46  | ng    | 100      |
| 75) Fluoranthene            | 19.357 | 202  | 1249749  | 11.58 | ng    | 98       |
| 78) Pyrene                  | 19.704 | 202  | 1202596  | 9.97  | ng    | 99       |
| 80) Butylbenzylphthalate    | 20.568 | 149  | 16454    | 4.44  | ng    | 94       |
| 81) Benzo(a)anthracene      | 21.410 | 228  | 792066   | 6.32  | ng    | 96       |
| 83) Chrysene                | 21.462 | 228  | 686649   | 5.66  | ng    | 96       |
| 87) Indeno(1,2,3-cd)pyrene  | 26.439 | 276  | 645485   | 3.73  | ng    | # 95     |
| 88) Benzo(b)fluoranthene    | 23.121 | 252  | 1143740  | 7.75  | ng    | # 92     |
| 89) Benzo(k)fluoranthene    | 23.121 | 252  | 1143740  | 8.18  | ng    | # 90     |
| 90) Benzo(a)pyrene          | 23.762 | 252  | 834451   | 6.21  | ng    | # 90     |
| 92) Benzo(g,h,i)perylene    | 27.239 | 276  | 644067   | 4.45  | ng    | # 93     |
| -----                       |        |      |          |       |       |          |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP060221\  
Data File : BP005672.D  
Acq On : 02 Jun 2021 18:56  
Operator : CG/JU  
Sample : M2580-05 2X  
Misc :  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)

Quant Time: Jun 03 02:42:32 2021  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP052721.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu May 27 16:11:11 2021  
Response via : Initial Calibration









#1  
1,4-Dichlorobenzene-d4  
Concen: 20.00 ng  
RT: 7.993 min Scan# 834  
Delta R.T. -0.000 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion:152 Resp: 383955  
Ion Ratio Lower Upper  
152 100  
150 152.8 119.8 179.8  
115 55.1 49.0 73.4



#5  
2-Fluorophenol  
Concen: 36.00 ng  
RT: 5.552 min Scan# 419  
Delta R.T. 0.006 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion:112 Resp: 889178  
Ion Ratio Lower Upper  
112 100  
64 52.1 33.6 50.4#  
63 28.7 18.6 27.8#









#7  
Phenol-d6  
Concen: 30.49 ng  
RT: 7.151 min Scan# 691  
Delta R.T. 0.006 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)

Tgt Ion: 99 Resp: 1026504




| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 99  | 100   |       |       |
| 42  | 17.0  | 12.9  | 19.3  |
| 71  | 34.3  | 26.7  | 40.1  |




#21  
Naphthalene-d8  
Concen: 20.00 ng  
RT: 10.798 min Scan# 1311  
Delta R.T. -0.001 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion: 136 Resp: 1341744

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 136 | 100   |       |       |
| 137 | 10.9  | 8.9   | 13.3  |
| 54  | 6.2   | 5.0   | 7.4   |
| 68  | 5.4   | 3.7   | 5.5   |









#23  
Nitrobenzene-d5  
Concen: 26.77 ng  
RT: 9.151 min Scan# 1031  
Delta R.T. -0.000 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion: 82 Resp: 645790  
Ion Ratio Lower Upper  
82 100  
128 44.7 30.5 45.7  
54 42.4 33.9 50.9



#26  
2-Nitrophenol  
Concen: 4.16 ng  
RT: 9.910 min Scan# 1160  
Delta R.T. -0.001 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion: 139 Resp: 50  
Ion Ratio Lower Upper  
139 100  
109 89.8 40.7 61.1#  
65 119.3 47.8 71.6#









#32  
Benzoic acid  
Concen: 5.06 ng  
RT: 10.040 min Scan# 1182  
Delta R.T. -0.047 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion:122 Resp: 598  
Ion Ratio Lower Upper  
122 100  
105 63.0 105.1 145.1#  
77 123.7 69.6 109.6#



#39  
Acenaphthene-d10  
Concen: 20.00 ng  
RT: 14.610 min Scan# 1959  
Delta R.T. -0.000 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion:164 Resp: 734198  
Ion Ratio Lower Upper  
164 100  
162 99.6 73.5 110.3  
160 44.8 40.6 61.0









#42  
2,4,6-Tribromophenol  
Concen: 47.10 ng  
RT: 16.098 min Scan# 2212  
Delta R.T. 0.006 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion:330 Resp: 443386  
Ion Ratio Lower Upper  
330 100  
332 98.7 76.3 114.5  
141 25.6 22.8 34.2



#45  
2-Fluorobiphenyl  
Concen: 25.37 ng  
RT: 13.239 min Scan# 1726  
Delta R.T. -0.000 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion:172 Resp: 1407484  
Ion Ratio Lower Upper  
172 100  
171 34.8 28.1 42.1  
170 23.6 19.6 29.4









#48  
2-Nitroaniline  
Concen: 4.50 ng  
RT: 13.745 min Scan# 1812  
Delta R.T. 0.052 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion: 65 Resp: 129  
Ion Ratio Lower Upper  
65 100  
92 76.1 48.8 73.2#  
138 24.8 69.5 104.3#



#50  
Dimethylphthalate  
Concen: 2.95 ng  
RT: 14.063 min Scan# 1866  
Delta R.T. -0.006 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion: 163 Resp: 172088  
Ion Ratio Lower Upper  
163 100  
194 4.0 3.9 5.9  
164 10.6 8.1 12.1









#51  
2,6-Dinitrotoluene  
Concen: 5.07 ng  
RT: 14.451 min Scan# 1932  
Delta R.T. 0.264 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion:165 Resp: 206  
Ion Ratio Lower Upper  
165 100  
63 74.5 67.4 101.0  
89 83.5 73.8 110.6



#53  
3-Nitroaniline  
Concen: 3.66 ng  
RT: 14.516 min Scan# 1943  
Delta R.T. 0.006 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion:138 Resp: 115  
Ion Ratio Lower Upper  
138 100  
108 68.5 10.2 15.4#  
92 64.9 108.1 162.1#









#54  
2,4-Dinitrophenol  
Concen: 6.55 ng  
RT: 14.710 min Scan# 1976  
Delta R.T. -0.000 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion:184 Resp: 113  
Ion Ratio Lower Upper  
184 100  
63 74.2 51.2 76.8  
154 139.2 55.8 83.8#



#56  
4-Nitrophenol  
Concen: 3.77 ng  
RT: 14.804 min Scan# 1992  
Delta R.T. 0.000 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion:139 Resp: 245  
Ion Ratio Lower Upper  
139 100  
109 149.2 122.8 162.8  
65 145.2 110.0 150.0









#57  
2,4-Dinitrotoluene  
Concen: 4.27 ng  
RT: 14.957 min Scan# 2018  
Delta R.T. -0.006 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion:165 Resp: 243  
Ion Ratio Lower Upper  
165 100  
63 23.8 56.7 85.1#  
89 50.9 85.8 128.6#  
182 88.3 3.0 4.6#



#62  
4-Nitroaniline  
Concen: 3.69 ng  
RT: 15.657 min Scan# 2137  
Delta R.T. -0.012 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion:138 Resp: 848  
Ion Ratio Lower Upper  
138 100  
92 38.0 50.9 90.9#  
108 35.5 154.7 194.7#









#64  
Phenanthrene-d10  
Concen: 20.00 ng  
RT: 17.357 min Scan# 2426  
Delta R.T. 0.006 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion:188 Resp: 1443026  
Ion Ratio Lower Upper  
188 100  
94 7.6 5.0 7.6#  
80 8.2 5.9 8.9



#71  
Phenanthrene  
Concen: 5.46 ng  
RT: 17.398 min Scan# 2433  
Delta R.T. 0.006 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion:178 Resp: 490443  
Ion Ratio Lower Upper  
178 100  
176 19.5 15.4 23.2  
179 15.3 12.3 18.5









#75  
Fluoranthene  
Concen: 11.58 ng  
RT: 19.357 min Scan# 2766  
Delta R.T. 0.006 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion:202 Resp: 1249749  
Ion Ratio Lower Upper  
202 100  
101 8.7 0.0 26.1  
203 17.5 0.0 37.2



#76  
Chrysene-d12  
Concen: 20.00 ng  
RT: 21.421 min Scan# 3117  
Delta R.T. -0.001 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion:240 Resp: 1708599  
Ion Ratio Lower Upper  
240 100  
120 9.1 4.2 6.2#  
236 26.1 20.2 30.2





#78

Pyrene

Concen: 9.97 ng

RT: 19.704 min Scan# 2825

Delta R.T. 0.006 min

Lab File: BP005672.D

Acq: 02 Jun 2021 18:56

Tgt Ion:202 Resp: 1202596





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 202 | 100   |       |       |
| 200 | 21.1  | 16.9  | 25.3  |
| 203 | 17.9  | 13.8  | 20.6  |

Instrument :

BNA\_P

ClientSampleId :

18-B2(2-4)



#79

Terphenyl-d14

Concen: 22.66 ng

RT: 19.898 min Scan# 2858




Delta R.T. -0.000 min


Lab File: BP005672.D

Acq: 02 Jun 2021 18:56

Tgt Ion:244 Resp: 2062967

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 244 | 100   |       |       |
| 212 | 6.7   | 5.8   | 8.6   |
| 122 | 8.3   | 5.2   | 7.8#  |









#80  
Butylbenzylphthalate  
Concen: 4.44 ng  
RT: 20.568 min Scan# 2972  
Delta R.T. -0.001 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion:149 Resp: 16454  
Ion Ratio Lower Upper  
149 100  
91 80.4 60.2 90.4  
206 27.2 23.9 35.9



#81  
Benzo(a)anthracene  
Concen: 6.32 ng  
RT: 21.410 min Scan# 3115  
Delta R.T. 0.006 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion:228 Resp: 792066  
Ion Ratio Lower Upper  
228 100  
226 28.7 21.2 31.8  
229 21.6 15.6 23.4









#83  
Chrysene  
Concen: 5.66 ng  
RT: 21.462 min Scan# 3124  
Delta R.T. 0.006 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion:228 Resp: 686649  
Ion Ratio Lower Upper  
228 100  
226 29.7 24.8 37.2  
229 22.8 15.4 23.0



#86  
Perylene-d12  
Concen: 20.00 ng  
RT: 23.874 min Scan# 3534  
Delta R.T. 0.011 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Tgt Ion:264 Resp: 2034122  
Ion Ratio Lower Upper  
264 100  
260 24.4 20.6 30.8  
265 22.7 19.0 28.6





#87

Indeno(1,2,3-cd)pyrene

Concen: 3.73 ng

RT: 26.439 min Scan# 3970

Delta R.T. 0.006 min

Lab File: BP005672.D

Acq: 02 Jun 2021 18:56





|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 276   | Resp: | 645485 |
| Ion Ratio | Lower | Upper |        |
| 276       | 100   |       |        |
| 138       | 16.4  | 8.4   | 12.6#  |
| 277       | 25.1  | 20.6  | 31.0   |

Instrument :

BNA\_P

ClientSampleId :

18-B2(2-4)



#88

Benzo(b)fluoranthene

Concen: 7.75 ng




RT: 23.121 min Scan# 3406


Delta R.T. 0.006 min

Lab File: BP005672.D

Acq: 02 Jun 2021 18:56

|           |       |       |         |
|-----------|-------|-------|---------|
| Tgt Ion:  | 252   | Resp: | 1143740 |
| Ion Ratio | Lower | Upper |         |
| 252       | 100   |       |         |
| 253       | 25.3  | 18.2  | 27.2    |
| 125       | 10.8  | 3.4   | 5.0#    |









#89  
Benzo(k)fluoranthene  
Concen: 8.18 ng  
RT: 23.121 min Scan# 3406  
Delta R.T. -0.047 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56




Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)


Tgt Ion:252 Resp: 1143740  
Ion Ratio Lower Upper  
252 100  
253 25.3 17.2 25.8  
125 10.8 3.8 5.6#



#90  
Benzo(a)pyrene  
Concen: 6.21 ng  
RT: 23.762 min Scan# 3515  
Delta R.T. 0.006 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56


Tgt Ion:252 Resp: 834451  
Ion Ratio Lower Upper  
252 100  
253 24.9 17.1 25.7  
125 11.1 3.8 5.6#





#92  
Benzo(g,h,i)perylene  
Concen: 4.45 ng  
RT: 27.239 min Scan# 4106  
Delta R.T. 0.011 min  
Lab File: BP005672.D  
Acq: 02 Jun 2021 18:56

Instrument :  
BNA\_P  
ClientSampleId :  
18-B2(2-4)



Tgt Ion: 276 Resp: 644067  
Ion Ratio Lower Upper  
276 100  
277 25.7 19.8 29.8  
138 16.6 7.7 11.5#

