

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060221\
 Data File : BP005664.D
 Acq On : 02 Jun 2021 13:44
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 02 14:34:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 16:11:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.589	0.538	8.7	107	0.00
3	Pyridine	1.457	1.460	-0.2	108	0.00
4	n-Nitrosodimethylamine	0.649	0.648	0.2	114	0.00
5 S	2-Fluorophenol	1.286	1.252	2.6	112	0.00
6	Aniline	2.050	1.909	6.9	106	0.00
7 S	Phenol-d6	1.754	1.680	4.2	109	0.00
8	2-Chlorophenol	1.440	1.392	3.3	111	0.00
9	Benzaldehyde	0.759	0.686	9.6	103	0.00
10 C	Phenol	1.790	1.686	5.8	108	0.00
11	bis(2-Chloroethyl)ether	1.382	1.292	6.5	108	0.00
12	1,3-Dichlorobenzene	1.635	1.514	7.4	108	0.00
13 C	1,4-Dichlorobenzene	1.657	1.532	7.5	108	0.00
14	1,2-Dichlorobenzene	1.585	1.471	7.2	107	0.00
15	Benzyl Alcohol	1.258	1.196	4.9	107	0.00
16	2,2'-oxybis(1-Chloropropane	1.582	1.482	6.3	107	0.00
17	2-Methylphenol	1.166	1.093	6.3	107	0.00
18	Hexachloroethane	0.588	0.550	6.5	108	0.00
19 P	n-Nitroso-di-n-propylamine	1.115	1.060	4.9	107	0.00
20	3+4-Methylphenols	1.553	1.472	5.2	107	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
22	Acetophenone	0.533	0.506	5.1	106	0.00
23 S	Nitrobenzene-d5	0.360	0.394	-9.4	118	0.00
24	Nitrobenzene	0.399	0.419	-5.0	113	0.00
25	Isophorone	0.798	0.756	5.3	105	0.00
26 C	2-Nitrophenol	0.128	0.174	-35.9#	139	0.00
27	2,4-Dimethylphenol	0.306	0.290	5.2	105	0.00
28	bis(2-Chloroethoxy)methane	0.430	0.401	6.7	104	0.00
29 C	2,4-Dichlorophenol	0.330	0.332	-0.6	110	0.00
30	1,2,4-Trichlorobenzene	0.390	0.367	5.9	105	0.00
31	Naphthalene	1.191	1.107	7.1	104	0.00
32	Benzoic acid	0.133	0.200	-50.4#	142	0.01
33	4-Chloroaniline	0.496	0.450	9.3	101	0.00
34 C	Hexachlorobutadiene	0.244	0.237	2.9	109	0.00
35	Caprolactam	0.090	0.098	-8.9	110	0.00
36 C	4-Chloro-3-methylphenol	0.347	0.339	2.3	107	0.00
37	2-Methylnaphthalene	0.844	0.781	7.5	103	0.00
38	1-Methylnaphthalene	0.801	0.735	8.2	103	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	0.667	0.657	1.5	106	0.00
41 P	Hexachlorocyclopentadiene	0.371	0.322	13.2	91	0.00
42 S	2,4,6-Tribromophenol	0.256	0.305	-19.1	124	0.00
43 C	2,4,6-Trichlorophenol	0.405	0.443	-9.4	114	0.00
44	2,4,5-Trichlorophenol	0.438	0.477	-8.9	113	0.00
45 S	2-Fluorobiphenyl	1.511	1.463	3.2	104	0.00
46	1,1'-Biphenyl	1.624	1.530	5.8	102	0.00
47	2-Chloronaphthalene	1.320	1.246	5.6	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060221\
 Data File : BP005664.D
 Acq On : 02 Jun 2021 13:44
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 02 14:34:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 16:11:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.257	0.343	-33.5#	129	0.00
49	Acenaphthylene	2.058	1.943	5.6	102	0.00
50	Dimethylphthalate	1.591	1.506	5.3	102	0.00
51	2,6-Dinitrotoluene	0.297	0.334	-12.5	112	0.00
52 C	Acenaphthene	1.338	1.192	10.9	96	0.00
53	3-Nitroaniline	0.293	0.349	-19.1	114	0.00
54 P	2,4-Dinitrophenol	0.118	0.139	-17.8	122	0.00
55	Dibenzofuran	2.050	1.919	6.4	101	0.00
56 P	4-Nitrophenol	0.241	0.282	-17.0	112	0.00
57	2,4-Dinitrotoluene	0.355	0.443	-24.8	116	0.00
58	Fluorene	1.639	1.552	5.3	102	0.00
59	2,3,4,6-Tetrachlorophenol	0.379	0.426	-12.4	119	0.00
60	Diethylphthalate	1.596	1.485	7.0	101	0.00
61	4-Chlorophenyl-phenylether	0.819	0.782	4.5	104	0.00
62	4-Nitroaniline	0.307	0.365	-18.9	113	0.00
63	Azobenzene	1.671	1.571	6.0	101	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	0.095	0.110	-15.8	123	0.00
66 c	n-Nitrosodiphenylamine	0.681	0.659	3.2	100	0.00
67	4-Bromophenyl-phenylether	0.241	0.243	-0.8	105	0.00
68	Hexachlorobenzene	0.287	0.283	1.4	103	0.00
69	Atrazine	0.195	0.205	-5.1	102	0.00
70 C	Pentachlorophenol	0.144	0.169	-17.4	117	0.00
71	Phenanthrene	1.246	1.170	6.1	99	0.00
72	Anthracene	1.230	1.171	4.8	99	0.00
73	Carbazole	1.211	1.134	6.4	98	0.00
74	Di-n-butylphthalate	1.297	1.239	4.5	98	0.00
75 C	Fluoranthene	1.496	1.643	-9.8	115	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	125	0.00
77	Benzydine	0.380	0.358	5.8	93	0.00
78	Pyrene	1.411	1.263	10.5	116	0.00
79 S	Terphenyl-d14	1.066	1.030	3.4	124	0.00
80	Butylbenzylphthalate	0.505	0.471	6.7	114	0.00
81	Benzo(a)anthracene	1.468	1.344	8.4	117	0.00
82	3,3'-Dichlorobenzidine	0.474	0.470	0.8	121	0.00
83	Chrysene	1.421	1.295	8.9	117	0.00
84	Bis(2-ethylhexyl)phthalate	0.797	0.723	9.3	112	0.00
85 c	Di-n-octyl phthalate	1.340	1.252	6.6	115	0.00
86 I	Perylene-d12	1.000	1.000	0.0	128	0.00
87	Indeno(1,2,3-cd)pyrene	1.702	1.632	4.1	124	0.00
88	Benzo(b)fluoranthene	1.451	1.319	9.1	120	0.00
89	Benzo(k)fluoranthene	1.375	1.281	6.8	119	0.00
90 C	Benzo(a)pyrene	1.320	1.240	6.1	122	0.00
91	Dibenzo(a,h)anthracene	1.477	1.408	4.7	124	0.00
92	Benzo(g,h,i)perylene	1.424	1.360	4.5	125	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060221\
Data File : BP005664.D
Acq On : 02 Jun 2021 13:44
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 02 14:34:11 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 27 16:11:11 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1