

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060221\
 Data File : BP005664.D
 Acq On : 02 Jun 2021 13:44
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 02 14:34:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 16:11:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	112	0.00
2	1,4-Dioxane	40.000	36.537	8.7	107	0.00
3	Pyridine	40.000	40.068	-0.2	108	0.00
4	n-Nitrosodimethylamine	40.000	39.918	0.2	114	0.00
5 S	2-Fluorophenol	80.000	77.852	2.7	112	0.00
6	Aniline	40.000	37.251	6.9	106	0.00
7 S	Phenol-d6	80.000	76.639	4.2	109	0.00
8	2-Chlorophenol	40.000	38.657	3.4	111	0.00
9	Benzaldehyde	40.000	36.131	9.7	103	0.00
10 C	Phenol	40.000	37.657	5.9	108	0.00
11	bis(2-Chloroethyl)ether	40.000	37.383	6.5	108	0.00
12	1,3-Dichlorobenzene	40.000	37.032	7.4	108	0.00
13 C	1,4-Dichlorobenzene	40.000	36.993	7.5	108	0.00
14	1,2-Dichlorobenzene	40.000	37.117	7.2	107	0.00
15	Benzyl Alcohol	40.000	38.043	4.9	107	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.458	6.4	107	0.00
17	2-Methylphenol	40.000	37.515	6.2	107	0.00
18	Hexachloroethane	40.000	37.429	6.4	108	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.011	5.0	107	0.00
20	3+4-Methylphenols	40.000	37.908	5.2	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	109	0.00
22	Acetophenone	40.000	37.957	5.1	106	0.00
23 S	Nitrobenzene-d5	80.000	87.607	-9.5	118	0.00
24	Nitrobenzene	40.000	41.972	-4.9	113	0.00
25	Isophorone	40.000	37.897	5.3	105	0.00
26 C	2-Nitrophenol	40.000	45.740	-14.4	139	0.00
27	2,4-Dimethylphenol	40.000	37.900	5.3	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.363	6.6	104	0.00
29 C	2,4-Dichlorophenol	40.000	40.286	-0.7	110	0.00
30	1,2,4-Trichlorobenzene	40.000	37.613	6.0	105	0.00
31	Naphthalene	40.000	37.180	7.1	104	0.00
32	Benzoic acid	40.000	47.339	-18.3	142	0.01
33	4-Chloroaniline	40.000	36.331	9.2	101	0.00
34 C	Hexachlorobutadiene	40.000	38.804	3.0	109	0.00
35	Caprolactam	40.000	43.426	-8.6	110	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.049	2.4	107	0.00
37	2-Methylnaphthalene	40.000	37.007	7.5	103	0.00
38	1-Methylnaphthalene	40.000	36.723	8.2	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.383	1.5	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	34.628	13.4	91	0.00
42 S	2,4,6-Tribromophenol	80.000	95.097	-18.9	124	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.756	-9.4	114	0.00
44	2,4,5-Trichlorophenol	40.000	43.561	-8.9	113	0.00
45 S	2-Fluorobiphenyl	80.000	77.450	3.2	104	0.00
46	1,1'-Biphenyl	40.000	37.678	5.8	102	0.00
47	2-Chloronaphthalene	40.000	37.746	5.6	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060221\
 Data File : BP005664.D
 Acq On : 02 Jun 2021 13:44
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 02 14:34:11 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 16:11:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.146	-10.4	129	0.00
49	Acenaphthylene	40.000	37.781	5.5	102	0.00
50	Dimethylphthalate	40.000	37.868	5.3	102	0.00
51	2,6-Dinitrotoluene	40.000	39.101	2.2	112	0.00
52 C	Acenaphthene	40.000	35.634	10.9	96	0.00
53	3-Nitroaniline	40.000	40.809	-2.0	114	0.00
54 P	2,4-Dinitrophenol	40.000	41.570	-3.9	122	0.00
55	Dibenzofuran	40.000	37.433	6.4	101	0.00
56 P	4-Nitrophenol	40.000	39.930	0.2	112	0.00
57	2,4-Dinitrotoluene	40.000	41.431	-3.6	116	0.00
58	Fluorene	40.000	37.885	5.3	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	45.029	-12.6	119	0.00
60	Diethylphthalate	40.000	37.223	6.9	101	0.00
61	4-Chlorophenyl-phenylether	40.000	38.187	4.5	104	0.00
62	4-Nitroaniline	40.000	40.712	-1.8	113	0.00
63	Azobenzene	40.000	37.607	6.0	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.868	-12.2	123	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.668	3.3	100	0.00
67	4-Bromophenyl-phenylether	40.000	40.353	-0.9	105	0.00
68	Hexachlorobenzene	40.000	39.416	1.5	103	0.00
69	Atrazine	40.000	42.140	-5.4	102	0.00
70 C	Pentachlorophenol	40.000	47.020	-17.6	117	0.00
71	Phenanthrene	40.000	37.563	6.1	99	0.00
72	Anthracene	40.000	38.065	4.8	99	0.00
73	Carbazole	40.000	37.449	6.4	98	0.00
74	Di-n-butylphthalate	40.000	38.212	4.5	98	0.00
75 C	Fluoranthene	40.000	43.948	-9.9	115	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	125	0.00
77	Benzydine	40.000	37.663	5.8	93	0.00
78	Pyrene	40.000	35.787	10.5	116	0.00
79 S	Terphenyl-d14	80.000	77.320	3.4	124	0.00
80	Butylbenzylphthalate	40.000	34.323	14.2	114	0.00
81	Benzo(a)anthracene	40.000	36.636	8.4	117	0.00
82	3,3'-Dichlorobenzidine	40.000	39.662	0.8	121	0.00
83	Chrysene	40.000	36.442	8.9	117	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.280	9.3	112	0.00
85 c	Di-n-octyl phthalate	40.000	37.384	6.5	115	0.00
86 I	Perylene-d12	20.000	20.000	0.0	128	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.368	4.1	124	0.00
88	Benzo(b)fluoranthene	40.000	36.357	9.1	120	0.00
89	Benzo(k)fluoranthene	40.000	37.270	6.8	119	0.00
90 C	Benzo(a)pyrene	40.000	37.563	6.1	122	0.00
91	Dibenzo(a,h)anthracene	40.000	38.141	4.6	124	0.00
92	Benzo(g,h,i)perylene	40.000	38.206	4.5	125	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060221\
Data File : BP005664.D
Acq On : 02 Jun 2021 13:44
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 02 14:34:11 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 27 16:11:11 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0