

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060221\
 Data File : BP005678.D
 Acq On : 02 Jun 2021 22:20
 Operator : CG/JU
 Sample : M2580-03 5X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 18-B1(4-6)

Manual Integrations
 APPROVED

mohammad
 6/3/2021 11:08:38 AM

Quant Time: Jun 03 02:44:34 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 16:11:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.993	152	230509	20.00	ng	0.00
21) Naphthalene-d8	10.799	136	883749	20.00	ng	0.00
39) Acenaphthene-d10	14.616	164	562561	20.00	ng	0.00
64) Phenanthrene-d10	17.363	188	1225651	20.00	ng	# 0.01
76) Chrysene-d12	21.439	240	1288048	20.00	ng	# 0.02
86) Perylene-d12	23.892	264	1618216	20.00	ng	0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.552	112	196458	13.25	ng	0.00
7) Phenol-d6	7.152	99	233362	11.55	ng	0.00
23) Nitrobenzene-d5	9.152	82	146562	9.22	ng	0.00
42) 2,4,6-Tribromophenol	16.098	330	144481	20.03	ng	0.00
45) 2-Fluorobiphenyl	13.246	172	381757	8.98	ng	0.00
79) Terphenyl-d14	19.904	244	670756	9.77	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	10.852	128	3444758	65.46	ng	98
37) 2-Methylnaphthalene	12.457	142	1311843	35.18	ng	# 86
38) 1-Methylnaphthalene	12.675	142	810585	22.91	ng	# 96
46) 1,1'-Biphenyl	13.451	154	333403	7.30	ng	96
49) Acenaphthylene	14.340	152	416510	7.20	ng	# 95
52) Acenaphthene	14.681	154	1530188	40.67	ng	96
55) Dibenzofuran	15.016	168	1932076	33.50	ng	97
58) Fluorene	15.663	166	2355146	51.08	ng	100
71) Phenanthrene	17.416	178	12895424	168.88	ng	95
72) Anthracene	17.498	178	4298754	57.02	ng	100
73) Carbazole	17.763	167	1973034	26.59	ng	99
75) Fluoranthene	19.380	202	13773571	150.27	ng	92
78) Pyrene	19.722	202	12665187	139.35	ng	95
80) Butylbenzylphthalate	20.575	149	810166	24.27	ng	90
81) Benzo(a)anthracene	21.422	228	7437781	78.68	ng	95
83) Chrysene	21.474	228	6627727	72.41	ng	96
84) Bis(2-ethylhexyl)phtha...	21.333	149	1288862	25.10	ng	98
87) Indeno(1,2,3-cd)pyrene	26.480	276	4793117	34.81	ng	# 95
88) Benzo(b)fluoranthene	23.145	252	9021649	76.84	ng	# 97
89) Benzo(k)fluoranthene	23.186	252	3312071m	29.77	ng	
90) Benzo(a)pyrene	23.786	252	6935484	64.93	ng	# 93
91) Dibenzo(a,h)anthracene	26.486	278	1315028	11.01	ng	# 89
92) Benzo(g,h,i)perylene	27.280	276	4607583	39.98	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060221\
Data File : BP005678.D
Acq On : 02 Jun 2021 22:20
Operator : CG/JU
Sample : M2580-03 5X
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)

Manual Integrations
APPROVED

mohammad
6/3/2021 11:08:38 AM

Quant Time: Jun 03 02:44:34 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 27 16:11:11 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.993 min Scan# 834
Delta R.T. -0.000 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20





Tgt Ion:152 Resp: 230509

Ion	Ratio	Lower	Upper
152	100		
150	157.3	119.8	179.8
115	55.9	49.0	73.4

Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)

Manual Integrations
APPROVED




mohammad
6/3/2021 11:08:38 AM




#5
2-Fluorophenol
Concen: 13.25 ng
RT: 5.552 min Scan# 419
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20

Tgt Ion:112 Resp: 196458


Ion	Ratio	Lower	Upper
112	100		
64	53.8	33.6	50.4#
63	28.9	18.6	27.8#





#7
Phenol-d6
Concen: 11.55 ng
RT: 7.152 min Scan# 691
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20




Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)




Tgt Ion: 99 Resp: 233362
Ion Ratio Lower Upper
99 100
42 17.8 12.9 19.3
71 40.0 26.7 40.1

Manual Integrations
APPROVED



mohammad
6/3/2021 11:08:38 AM




#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.799 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20



Tgt Ion: 136 Resp: 883749
Ion Ratio Lower Upper
136 100
137 11.4 8.9 13.3
54 6.2 5.0 7.4
68 5.3 3.7 5.5









#23
Nitrobenzene-d5
Concen: 9.22 ng
RT: 9.152 min Scan# 1031
Delta R.T. -0.000 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20

Tgt Ion: 82 Resp: 146562
Ion Ratio Lower Upper
82 100
128 43.1 30.5 45.7
54 42.7 33.9 50.9

Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)




Manual Integrations
APPROVED


mohammad
6/3/2021 11:08:38 AM



#31
Naphthalene
Concen: 65.46 ng
RT: 10.852 min Scan# 1320
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20


Tgt Ion: 128 Resp: 3444758
Ion Ratio Lower Upper
128 100
129 11.0 8.0 12.0
127 13.0 11.1 16.7





#37
2-Methylnaphthalene
Concen: 35.18 ng
RT: 12.457 min Scan# 1593
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20




Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)




Tgt Ion:142 Resp: 1311843
Ion Ratio Lower Upper
142 100
141 87.5 74.7 112.1
115 35.4 43.8 65.8#

Manual Integrations
APPROVED



mohammad
6/3/2021 11:08:38 AM




#38
1-Methylnaphthalene
Concen: 22.91 ng
RT: 12.675 min Scan# 1630
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20



Tgt Ion:142 Resp: 810585
Ion Ratio Lower Upper
142 100
141 93.1 77.0 115.6
116 3.8 5.0 7.6#









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.616 min Scan# 1960
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20

Tgt Ion:164 Resp: 562561
Ion Ratio Lower Upper
164 100
162 99.5 73.5 110.3
160 44.8 40.6 61.0

Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)




Manual Integrations
APPROVED


mohammad
6/3/2021 11:08:38 AM



#42
2,4,6-Tribromophenol
Concen: 20.03 ng
RT: 16.098 min Scan# 2212
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20

Tgt Ion:330 Resp: 144481
Ion Ratio Lower Upper
330 100
332 97.1 76.3 114.5
141 26.6 22.8 34.2









#45
2-Fluorobiphenyl
Concen: 8.98 ng
RT: 13.246 min Scan# 1727
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20

Tgt Ion:	172	Resp:	381757
Ion Ratio	Lower	Upper	
172	100		
171	34.6	28.1	42.1
170	22.3	19.6	29.4

Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)




Manual Integrations
APPROVED


mohammad
6/3/2021 11:08:38 AM



#46
1,1'-Biphenyl
Concen: 7.30 ng
RT: 13.451 min Scan# 1762
Delta R.T. -0.000 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20

Tgt Ion:	154	Resp:	333403
Ion Ratio	Lower	Upper	
154	100		
153	41.6	18.9	58.9
76	11.6	0.0	32.1









#49
Acenaphthylene
Concen: 7.20 ng
RT: 14.340 min Scan# 1913
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20

Tgt Ion:152 Resp: 416510
Ion Ratio Lower Upper
152 100
151 20.7 15.8 23.6
153 16.5 10.5 15.7#

Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)




Manual Integrations
APPROVED


mohammad
6/3/2021 11:08:38 AM



#52
Acenaphthene
Concen: 40.67 ng
RT: 14.681 min Scan# 1971
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20


Tgt Ion:154 Resp: 1530188
Ion Ratio Lower Upper
154 100
153 112.4 90.9 136.3
152 53.7 48.5 72.7





#55
Dibenzofuran
Concen: 33.50 ng
RT: 15.016 min Scan# 2028
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20




Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)




Tgt Ion:168 Resp: 1932076
Ion Ratio Lower Upper
168 100
139 36.5 31.4 47.0
169 13.9 11.4 17.0

Manual Integrations
APPROVED



mohammad
6/3/2021 11:08:38 AM




#58
Fluorene
Concen: 51.08 ng
RT: 15.663 min Scan# 2138
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20



Tgt Ion:166 Resp: 2355146
Ion Ratio Lower Upper
166 100
165 98.5 78.8 118.2
167 13.5 10.9 16.3









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.363 min Scan# 2427
Delta R.T. 0.012 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20

Tgt Ion:188 Resp: 1225651
Ion Ratio Lower Upper
188 100
94 8.0 5.0 7.6#
80 8.6 5.9 8.9

Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)




Manual Integrations
APPROVED


mohammad
6/3/2021 11:08:38 AM



#71
Phenanthrene
Concen: 168.88 ng
RT: 17.416 min Scan# 2436
Delta R.T. 0.023 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20


Tgt Ion:178 Resp:12895424
Ion Ratio Lower Upper
178 100
176 21.7 15.4 23.2
179 17.1 12.3 18.5





#72
 Anthracene
 Concen: 57.02 ng
 RT: 17.498 min Scan# 2450
 Delta R.T. 0.012 min
 Lab File: BP005678.D
 Acq: 02 Jun 2021 22:20




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(4-6)



Tgt Ion:178 Resp: 4298754
 Ion Ratio Lower Upper
 178 100
 176 19.2 15.2 22.8
 179 15.9 12.6 19.0




Manual Integrations
 APPROVED


mohammad
 6/3/2021 11:08:38 AM




#73
 Carbazole
 Concen: 26.59 ng
 RT: 17.763 min Scan# 2495
 Delta R.T. 0.012 min
 Lab File: BP005678.D
 Acq: 02 Jun 2021 22:20

Tgt Ion:167 Resp: 1973034
 Ion Ratio Lower Upper
 167 100
 166 22.1 17.4 26.0
 139 12.3 10.4 15.6









#75
Fluoranthene
Concen: 150.27 ng
RT: 19.380 min Scan# 2770
Delta R.T. 0.029 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20





Tgt Ion:202 Resp:13773571
Ion Ratio Lower Upper
202 100
101 9.3 0.0 26.1
203 20.5 0.0 37.2



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.439 min Scan# 3120
Delta R.T. 0.017 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20




Tgt Ion:240 Resp: 1288048
Ion Ratio Lower Upper
240 100
120 8.7 4.2 6.2#
236 26.3 20.2 30.2



Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)


Manual Integrations
APPROVED

mohammad
6/3/2021 11:08:38 AM



#78
Pyrene
Concen: 139.35 ng
RT: 19.722 min Scan# 2828
Delta R.T. 0.023 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20




Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)




Tgt Ion:202 Resp:12665187
Ion Ratio Lower Upper
202 100
200 23.5 16.9 25.3
203 19.8 13.8 20.6

Manual Integrations
APPROVED



mohammad
6/3/2021 11:08:38 AM




#79
Terphenyl-d14
Concen: 9.77 ng
RT: 19.904 min Scan# 2859
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20



Tgt Ion:244 Resp: 670756
Ion Ratio Lower Upper
244 100
212 7.1 5.8 8.6
122 8.2 5.2 7.8#









#80
Butylbenzylphthalate
Concen: 24.27 ng
RT: 20.575 min Scan# 2973
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20

Tgt Ion:	149	Resp:	810166
Ion	Ratio	Lower	Upper
149	100		
91	66.8	60.2	90.4
206	24.5	23.9	35.9

Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)




Manual Integrations
APPROVED


mohammad
6/3/2021 11:08:38 AM



#81
Benzo(a)anthracene
Concen: 78.68 ng
RT: 21.422 min Scan# 3117
Delta R.T. 0.018 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20

Tgt Ion:	228	Resp:	7437781
Ion	Ratio	Lower	Upper
228	100		
226	29.3	21.2	31.8
229	21.8	15.6	23.4









#83
Chrysene
Concen: 72.41 ng
RT: 21.474 min Scan# 3126
Delta R.T. 0.018 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20

Tgt Ion:228 Resp: 6627727
Ion Ratio Lower Upper
228 100
226 31.8 24.8 37.2
229 23.0 15.4 23.0

Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)




Manual Integrations
APPROVED


mohammad
6/3/2021 11:08:38 AM



#84
Bis(2-ethylhexyl)phthalate
Concen: 25.10 ng
RT: 21.333 min Scan# 3102
Delta R.T. 0.006 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20


Tgt Ion:149 Resp: 1288862
Ion Ratio Lower Upper
149 100
167 28.7 23.8 35.8
279 6.4 5.7 8.5





#86
Perylene-d12
Concen: 20.00 ng
RT: 23.892 min Scan# 3537
Delta R.T. 0.029 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20




Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)



Tgt Ion:264 Resp: 1618216
Ion Ratio Lower Upper
264 100
260 24.3 20.6 30.8
265 24.7 19.0 28.6




Manual Integrations
APPROVED


mohammad
6/3/2021 11:08:38 AM



#87
Indeno(1,2,3-cd)pyrene
Concen: 34.81 ng
RT: 26.480 min Scan# 3977
Delta R.T. 0.047 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20


Tgt Ion:276 Resp: 4793117
Ion Ratio Lower Upper
276 100
138 15.7 8.4 12.6#
277 25.0 20.6 31.0





#88
Benzo(b)fluoranthene
Concen: 76.84 ng
RT: 23.145 min Scan# 3410
Delta R.T. 0.029 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20




Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)



Tgt Ion:252 Resp: 9021649
Ion Ratio Lower Upper
252 100
253 23.5 18.2 27.2
125 8.7 3.4 5.0#




Manual Integrations
APPROVED


mohammad
6/3/2021 11:08:38 AM



#89
Benzo(k)fluoranthene
Concen: 29.77 ng m
RT: 23.186 min Scan# 3417
Delta R.T. 0.018 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20


Tgt Ion:252 Resp: 3312071
Ion Ratio Lower Upper
252 100
253 24.1 17.2 25.8
125 8.6 3.8 5.6#





#90
Benzo(a)pyrene
Concen: 64.93 ng
RT: 23.786 min Scan# 3519
Delta R.T. 0.029 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20




Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)



Tgt Ion:252 Resp: 6935484
Ion Ratio Lower Upper
252 100
253 23.8 17.1 25.7
125 9.4 3.8 5.6#




Manual Integrations
APPROVED


mohammad
6/3/2021 11:08:38 AM



#91
Dibenzo(a,h)anthracene
Concen: 11.01 ng
RT: 26.486 min Scan# 3978
Delta R.T. 0.035 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20


Tgt Ion:278 Resp: 1315028
Ion Ratio Lower Upper
278 100
139 0.0 5.0 7.6#
279 27.7 18.6 28.0





#92
Benzo(g,h,i)perylene
Concen: 39.98 ng
RT: 27.280 min Scan# 4113
Delta R.T. 0.053 min
Lab File: BP005678.D
Acq: 02 Jun 2021 22:20

Instrument :
BNA_P
ClientSampleId :
18-B1(4-6)



Tgt Ion:276 Resp: 4607583
Ion Ratio Lower Upper
276 100
277 24.3 19.8 29.8
138 15.7 7.7 11.5#

Manual Integrations
APPROVED

mohammad
6/3/2021 11:08:38 AM

