

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060222\
 Data File : BP010663.D
 Acq On : 02 Jun 2022 10:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 02 15:19:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 28 02:18:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	-0.01
2	1,4-Dioxane	40.000	35.431	11.4	93	0.00
3	Pyridine	40.000	39.826	0.4	102	0.00
4	n-Nitrosodimethylamine	40.000	36.321	9.2	97	0.00
5 S	2-Fluorophenol	80.000	82.988	-3.7	107	0.00
6	Aniline	40.000	43.351	-8.4	115	-0.01
7 S	Phenol-d6	80.000	87.150	-8.9	114	0.00
8	2-Chlorophenol	40.000	41.861	-4.7	111	0.00
9	Benzaldehyde	40.000	36.798	8.0	108	-0.01
10 C	Phenol	40.000	43.410	-8.5	115	0.00
11	bis(2-Chloroethyl)ether	40.000	41.524	-3.8	109	0.00
12	1,3-Dichlorobenzene	40.000	39.825	0.4	106	0.00
13 C	1,4-Dichlorobenzene	40.000	40.300	-0.7	107	-0.01
14	1,2-Dichlorobenzene	40.000	40.132	-0.3	107	-0.01
15	Benzyl Alcohol	40.000	44.638	-11.6	116	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.060	4.8	102	-0.01
17	2-Methylphenol	40.000	44.155	-10.4	116	0.00
18	Hexachloroethane	40.000	39.675	0.8	105	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.863	-9.7	119	-0.01
20	3+4-Methylphenols	40.000	44.892	-12.2	116	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	115	-0.01
22	Acetophenone	40.000	40.844	-2.1	118	0.00
23 S	Nitrobenzene-d5	80.000	79.342	0.8	114	0.00
24	Nitrobenzene	40.000	39.811	0.5	113	0.00
25	Isophorone	40.000	43.252	-8.1	124	0.00
26 C	2-Nitrophenol	40.000	43.861	-9.7	123	-0.01
27	2,4-Dimethylphenol	40.000	42.729	-6.8	121	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.012	-2.5	118	-0.01
29 C	2,4-Dichlorophenol	40.000	42.138	-5.3	117	0.00
30	1,2,4-Trichlorobenzene	40.000	38.243	4.4	110	-0.01
31	Naphthalene	40.000	40.190	-0.5	116	0.00
32	Benzoic acid	40.000	40.331	-0.8	128	0.01
33	4-Chloroaniline	40.000	43.466	-8.7	124	-0.01
34 C	Hexachlorobutadiene	40.000	38.201	4.5	112	-0.01
35	Caprolactam	40.000	52.964	-32.4#	145	0.01
36 C	4-Chloro-3-methylphenol	40.000	45.394	-13.5	130	0.00
37	2-Methylnaphthalene	40.000	41.203	-3.0	120	0.00
38	1-Methylnaphthalene	40.000	41.479	-3.7	120	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	125	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.970	5.1	119	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.167	-2.9	124	-0.01
42 S	2,4,6-Tribromophenol	80.000	97.394	-21.7	148	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.449	-3.6	124	0.00
44	2,4,5-Trichlorophenol	40.000	40.822	-2.1	125	0.00
45 S	2-Fluorobiphenyl	80.000	76.634	4.2	121	0.00
46	1,1'-Biphenyl	40.000	39.019	2.5	123	-0.01
47	2-Chloronaphthalene	40.000	38.672	3.3	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060222\
 Data File : BP010663.D
 Acq On : 02 Jun 2022 10:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 02 15:19:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 28 02:18:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.540	-8.8	132	0.00
49	Acenaphthylene	40.000	40.486	-1.2	127	-0.01
50	Dimethylphthalate	40.000	41.558	-3.9	130	0.00
51	2,6-Dinitrotoluene	40.000	43.485	-8.7	133	0.00
52 C	Acenaphthene	40.000	39.904	0.2	126	0.00
53	3-Nitroaniline	40.000	46.859	-17.1	140	0.00
54 P	2,4-Dinitrophenol	40.000	44.283	-10.7	148	0.00
55	Dibenzofuran	40.000	40.531	-1.3	127	-0.01
56 P	4-Nitrophenol	40.000	44.378	-10.9	139	0.00
57	2,4-Dinitrotoluene	40.000	46.368	-15.9	137	0.00
58	Fluorene	40.000	41.687	-4.2	130	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.067	-10.2	135	0.00
60	Diethylphthalate	40.000	43.175	-7.9	133	0.00
61	4-Chlorophenyl-phenylether	40.000	41.462	-3.7	131	0.00
62	4-Nitroaniline	40.000	47.356	-18.4	147	0.00
63	Azobenzene	40.000	41.906	-4.8	130	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	135	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.025	-5.1	150	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.955	2.6	133	0.00
67	4-Bromophenyl-phenylether	40.000	39.535	1.2	136	0.00
68	Hexachlorobenzene	40.000	41.114	-2.8	143	0.00
69	Atrazine	40.000	43.557	-8.9	142	0.00
70 C	Pentachlorophenol	40.000	43.389	-8.5	142	0.00
71	Phenanthrene	40.000	39.888	0.3	135	0.00
72	Anthracene	40.000	40.683	-1.7	136	0.00
73	Carbazole	40.000	43.642	-9.1	140	0.00
74	Di-n-butylphthalate	40.000	44.029	-10.1	139	0.00
75 C	Fluoranthene	40.000	43.677	-9.2	140	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	137	0.00
77	Benzydine	40.000	47.300	-18.2	162	0.00
78	Pyrene	40.000	39.608	1.0	140	0.00
79 S	Terphenyl-d14	80.000	80.953	-1.2	143	0.00
80	Butylbenzylphthalate	40.000	44.259	-10.6	146	0.00
81	Benzo(a)anthracene	40.000	40.242	-0.6	138	0.00
82	3,3'-Dichlorobenzidine	40.000	43.377	-8.4	140	0.00
83	Chrysene	40.000	39.542	1.1	136	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.907	-9.8	140	0.00
85 c	Di-n-octyl phthalate	40.000	43.371	-8.4	138	0.00
86 I	Perylene-d12	20.000	20.000	0.0	127	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	34.632	13.4	109	0.00
88	Benzo(b)fluoranthene	40.000	42.334	-5.8	130	0.00
89	Benzo(k)fluoranthene	40.000	40.393	-1.0	130	0.00
90 C	Benzo(a)pyrene	40.000	40.542	-1.4	125	0.00
91	Dibenzo(a,h)anthracene	40.000	34.993	12.5	110	0.00
92	Benzo(g,h,i)perylene	40.000	33.199	17.0	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060222\
Data File : BP010663.D
Acq On : 02 Jun 2022 10:31
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 02 15:19:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 28 02:18:04 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0