

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060324\
 Data File : BP020487.D
 Acq On : 03 Jun 2024 11:11
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 03 13:08:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 03:41:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.451	0.508	-12.6	107	0.00
3	Pyridine	1.269	1.298	-2.3	94	0.00
4	n-Nitrosodimethylamine	0.616	0.619	-0.5	92	0.00
5 S	2-Fluorophenol	1.117	1.213	-8.6	100	0.00
6	Aniline	1.596	1.461	8.5	84	0.00
7 S	Phenol-d6	1.575	1.511	4.1	88	0.00
8	2-Chlorophenol	1.252	1.252	0.0	92	0.00
9	Benzaldehyde	1.010	1.091	-8.0	114	0.00
10 C	Phenol	1.614	1.551	3.9	88	0.00
11	bis(2-Chloroethyl)ether	1.331	1.257	5.6	87	-0.01
12	1,3-Dichlorobenzene	1.475	1.520	-3.1	96	0.00
13 C	1,4-Dichlorobenzene	1.499	1.537	-2.5	95	0.00
14	1,2-Dichlorobenzene	1.447	1.449	-0.1	93	0.00
15	Benzyl Alcohol	1.164	1.025	11.9	80	-0.01
16	2,2'-oxybis(1-Chloropropane	1.987	1.726	13.1	81	0.00
17	2-Methylphenol	1.095	1.001	8.6	83	0.00
18	Hexachloroethane	0.548	0.558	-1.8	94	0.00
19 P	n-Nitroso-di-n-propylamine	1.075	0.876	18.5	74	-0.01
20	3+4-Methylphenols	1.489	1.324	11.1	81	-0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	78	0.00
22	Acetophenone	0.496	0.504	-1.6	79	-0.01
23 S	Nitrobenzene-d5	0.365	0.386	-5.8	81	-0.01
24	Nitrobenzene	0.371	0.388	-4.6	81	-0.01
25	Isophorone	0.710	0.648	8.7	70	-0.01
26 C	2-Nitrophenol	0.128	0.136	-6.3	76	-0.01
27	2,4-Dimethylphenol	0.214	0.214	0.0	76	-0.01
28	bis(2-Chloroethoxy)methane	0.436	0.424	2.8	76	0.00
29 C	2,4-Dichlorophenol	0.288	0.297	-3.1	78	-0.01
30	1,2,4-Trichlorobenzene	0.342	0.366	-7.0	84	0.00
31	Naphthalene	1.022	1.054	-3.1	81	0.00
32	Benzoic acid	0.181	0.151	16.6	64	-0.02
33	4-Chloroaniline	0.396	0.372	6.1	73	-0.01
34 C	Hexachlorobutadiene	0.215	0.238	-10.7	87	0.00
35	Caprolactam	0.105	0.098	6.7	70	-0.02
36 C	4-Chloro-3-methylphenol	0.337	0.297	11.9	68	-0.01
37	2-Methylnaphthalene	0.733	0.686	6.4	74	-0.01
38	1-Methylnaphthalene	0.726	0.676	6.9	74	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	65	0.00
40	1,2,4,5-Tetrachlorobenzene	0.579	0.666	-15.0	75	0.00
41 P	Hexachlorocyclopentadiene	0.203	0.245	-20.7	76	-0.01
42 S	2,4,6-Tribromophenol	0.208	0.199	4.3	59	0.00
43 C	2,4,6-Trichlorophenol	0.350	0.385	-10.0	69	0.00
44	2,4,5-Trichlorophenol	0.404	0.432	-6.9	67	0.00
45 S	2-Fluorobiphenyl	1.342	1.474	-9.8	73	0.00
46	1,1'-Biphenyl	1.411	1.509	-6.9	69	0.00
47	2-Chloronaphthalene	1.122	1.198	-6.8	69	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060324\
 Data File : BP020487.D
 Acq On : 03 Jun 2024 11:11
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 03 13:08:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 03:41:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.304	0.306	-0.7	62	0.00
49	Acenaphthylene	1.665	1.716	-3.1	68	0.00
50	Dimethylphthalate	1.410	1.347	4.5	63	0.00
51	2,6-Dinitrotoluene	0.301	0.299	0.7	62	0.00
52 C	Acenaphthene	1.050	1.084	-3.2	69	0.00
53	3-Nitroaniline	0.302	0.297	1.7	62	0.00
54 P	2,4-Dinitrophenol	0.104	0.092	11.5	59	-0.01
55	Dibenzofuran	1.664	1.668	-0.2	66	0.00
56 P	4-Nitrophenol	0.234	0.228	2.6	61	0.00
57	2,4-Dinitrotoluene	0.396	0.382	3.5	60	0.00
58	Fluorene	1.336	1.295	3.1	65	0.00
59	2,3,4,6-Tetrachlorophenol	0.330	0.324	1.8	61	0.00
60	Diethylphthalate	1.433	1.298	9.4	59	-0.01
61	4-Chlorophenyl-phenylether	0.689	0.668	3.0	64	0.00
62	4-Nitroaniline	0.311	0.292	6.1	61	-0.01
63	Azobenzene	1.338	1.262	5.7	62	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	57	0.00
65	4,6-Dinitro-2-methylphenol	0.078	0.072	7.7	53	-0.01
66 c	n-Nitrosodiphenylamine	0.550	0.598	-8.7	63	-0.01
67	4-Bromophenyl-phenylether	0.205	0.217	-5.9	61	0.00
68	Hexachlorobenzene	0.241	0.259	-7.5	62	0.00
69	Atrazine	0.193	0.192	0.5	56	-0.02
70 C	Pentachlorophenol	0.146	0.149	-2.1	58	0.00
71	Phenanthrene	0.983	1.026	-4.4	60	0.00
72	Anthracene	0.965	1.025	-6.2	61	0.00
73	Carbazole	0.933	0.984	-5.5	61	0.00
74	Di-n-butylphthalate	1.130	1.129	0.1	55	0.00
75 C	Fluoranthene	1.163	1.234	-6.1	61	-0.01
76 I	Chrysene-d12	1.000	1.000	0.0	75	-0.02
77	Benzidine	0.610	0.509	16.6	52	0.00
78	Pyrene	1.504	1.307	13.1	63	-0.01
79 S	Terphenyl-d14	1.201	1.069	11.0	63	0.00
80	Butylbenzylphthalate	0.549	0.483	12.0	60	0.00
81	Benzo(a)anthracene	1.317	1.353	-2.7	77	-0.02
82	3,3'-Dichlorobenzidine	0.391	0.462	-18.2	83	-0.02
83	Chrysene	1.241	1.289	-3.9	77	-0.02
84	Bis(2-ethylhexyl)phthalate	0.856	0.811	5.3	66	-0.01
85 c	Di-n-octyl phthalate	1.244	1.344	-8.0	80	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	101	-0.01
87	Indeno(1,2,3-cd)pyrene	1.216	1.454	-19.6	115	-0.02
88	Benzo(b)fluoranthene	1.246	1.224	1.8	98	-0.02
89	Benzo(k)fluoranthene	1.215	1.202	1.1	100	-0.01
90 C	Benzo(a)pyrene	1.032	1.088	-5.4	104	0.00
91	Dibenzo(a,h)anthracene	1.014	1.200	-18.3	114	-0.02
92	Benzo(g,h,i)perylene	1.014	1.220	-20.3	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060324\
Data File : BP020487.D
Acq On : 03 Jun 2024 11:11
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 03 13:08:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 03:41:03 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0