

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060324\
 Data File : BP020487.D
 Acq On : 03 Jun 2024 11:11
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 03 13:08:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 03:41:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	93	0.00
2	1,4-Dioxane	40.000	44.991	-12.5	107	0.00
3	Pyridine	40.000	40.927	-2.3	94	0.00
4	n-Nitrosodimethylamine	40.000	40.169	-0.4	92	0.00
5 S	2-Fluorophenol	80.000	86.840	-8.6	100	0.00
6	Aniline	40.000	36.618	8.5	84	0.00
7 S	Phenol-d6	80.000	76.733	4.1	88	0.00
8	2-Chlorophenol	40.000	40.003	-0.0	92	0.00
9	Benzaldehyde	40.000	43.170	-7.9	114	0.00
10 C	Phenol	40.000	38.436	3.9	88	0.00
11	bis(2-Chloroethyl)ether	40.000	37.781	5.5	87	-0.01
12	1,3-Dichlorobenzene	40.000	41.217	-3.0	96	0.00
13 C	1,4-Dichlorobenzene	40.000	41.011	-2.5	95	0.00
14	1,2-Dichlorobenzene	40.000	40.065	-0.2	93	0.00
15	Benzyl Alcohol	40.000	35.252	11.9	80	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	34.745	13.1	81	0.00
17	2-Methylphenol	40.000	36.568	8.6	83	0.00
18	Hexachloroethane	40.000	40.773	-1.9	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	32.575	18.6	74	-0.01
20	3+4-Methylphenols	40.000	35.554	11.1	81	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	78	0.00
22	Acetophenone	40.000	40.623	-1.6	79	-0.01
23 S	Nitrobenzene-d5	80.000	84.483	-5.6	81	-0.01
24	Nitrobenzene	40.000	41.892	-4.7	81	-0.01
25	Isophorone	40.000	36.527	8.7	70	-0.01
26 C	2-Nitrophenol	40.000	38.036	4.9	76	-0.01
27	2,4-Dimethylphenol	40.000	39.902	0.2	76	-0.01
28	bis(2-Chloroethoxy)methane	40.000	38.828	2.9	76	0.00
29 C	2,4-Dichlorophenol	40.000	41.309	-3.3	78	-0.01
30	1,2,4-Trichlorobenzene	40.000	42.811	-7.0	84	0.00
31	Naphthalene	40.000	41.234	-3.1	81	0.00
32	Benzoic acid	40.000	32.286	19.3	64	-0.02
33	4-Chloroaniline	40.000	37.608	6.0	73	-0.01
34 C	Hexachlorobutadiene	40.000	44.181	-10.5	87	0.00
35	Caprolactam	40.000	37.056	7.4	70	-0.02
36 C	4-Chloro-3-methylphenol	40.000	35.267	11.8	68	-0.01
37	2-Methylnaphthalene	40.000	37.445	6.4	74	-0.01
38	1-Methylnaphthalene	40.000	37.217	7.0	74	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	65	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	46.061	-15.2	75	0.00
41 P	Hexachlorocyclopentadiene	40.000	48.225	-20.6	76	-0.01
42 S	2,4,6-Tribromophenol	80.000	76.606	4.2	59	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.005	-10.0	69	0.00
44	2,4,5-Trichlorophenol	40.000	42.843	-7.1	67	0.00
45 S	2-Fluorobiphenyl	80.000	87.880	-9.8	73	0.00
46	1,1'-Biphenyl	40.000	42.772	-6.9	69	0.00
47	2-Chloronaphthalene	40.000	42.693	-6.7	69	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060324\
 Data File : BP020487.D
 Acq On : 03 Jun 2024 11:11
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 03 13:08:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 03:41:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.294	-0.7	62	0.00
49	Acenaphthylene	40.000	41.231	-3.1	68	0.00
50	Dimethylphthalate	40.000	38.205	4.5	63	0.00
51	2,6-Dinitrotoluene	40.000	39.737	0.7	62	0.00
52 C	Acenaphthene	40.000	41.288	-3.2	69	0.00
53	3-Nitroaniline	40.000	39.306	1.7	62	0.00
54 P	2,4-Dinitrophenol	40.000	35.416	11.5	59	-0.01
55	Dibenzofuran	40.000	40.101	-0.3	66	0.00
56 P	4-Nitrophenol	40.000	39.053	2.4	61	0.00
57	2,4-Dinitrotoluene	40.000	38.540	3.7	60	0.00
58	Fluorene	40.000	38.796	3.0	65	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.204	2.0	61	0.00
60	Diethylphthalate	40.000	36.224	9.4	59	-0.01
61	4-Chlorophenyl-phenylether	40.000	38.760	3.1	64	0.00
62	4-Nitroaniline	40.000	37.511	6.2	61	-0.01
63	Azobenzene	40.000	37.729	5.7	62	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	57	0.00
65	4,6-Dinitro-2-methylphenol	40.000	35.141	12.1	53	-0.01
66 c	n-Nitrosodiphenylamine	40.000	43.492	-8.7	63	-0.01
67	4-Bromophenyl-phenylether	40.000	42.448	-6.1	61	0.00
68	Hexachlorobenzene	40.000	43.002	-7.5	62	0.00
69	Atrazine	40.000	39.675	0.8	56	-0.02
70 C	Pentachlorophenol	40.000	40.856	-2.1	58	0.00
71	Phenanthrene	40.000	41.756	-4.4	60	0.00
72	Anthracene	40.000	42.507	-6.3	61	0.00
73	Carbazole	40.000	42.167	-5.4	61	0.00
74	Di-n-butylphthalate	40.000	39.976	0.1	55	0.00
75 C	Fluoranthene	40.000	42.451	-6.1	61	-0.01
76 I	Chrysene-d12	20.000	20.000	0.0	75	-0.02
77	Benzydine	40.000	31.023	22.4	52	0.00
78	Pyrene	40.000	34.754	13.1	63	-0.01
79 S	Terphenyl-d14	80.000	71.152	11.1	63	0.00
80	Butylbenzylphthalate	40.000	32.707	18.2	60	0.00
81	Benzo(a)anthracene	40.000	41.089	-2.7	77	-0.02
82	3,3'-Dichlorobenzidine	40.000	47.233	-18.1	83	-0.02
83	Chrysene	40.000	41.567	-3.9	77	-0.02
84	Bis(2-ethylhexyl)phthalate	40.000	37.904	5.2	66	-0.01
85 c	Di-n-octyl phthalate	40.000	43.216	-8.0	80	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	101	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	47.832	-19.6	115	-0.02
88	Benzo(b)fluoranthene	40.000	39.297	1.8	98	-0.02
89	Benzo(k)fluoranthene	40.000	39.551	1.1	100	-0.01
90 C	Benzo(a)pyrene	40.000	42.187	-5.5	104	0.00
91	Dibenzo(a,h)anthracene	40.000	47.331	-18.3	114	-0.02
92	Benzo(g,h,i)perylene	40.000	48.125	-20.3	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060324\
Data File : BP020487.D
Acq On : 03 Jun 2024 11:11
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 03 13:08:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 03:41:03 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0