

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060421\
 Data File : BP005711.D
 Acq On : 04 Jun 2021 14:11
 Operator : CG/JU
 Sample : M2611-01MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 T-1MS

Quant Time: Jun 04 14:49:18 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 16:11:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.981	152	126115	20.000	ng	-0.01
21) Naphthalene-d8	10.787	136	496745	20.000	ng	-0.01
39) Acenaphthene-d10	14.604	164	314597	20.000	ng	0.00
64) Phenanthrene-d10	17.351	188	678940	20.000	ng	# 0.00
76) Chrysene-d12	21.421	240	740509	20.000	ng	# 0.00
86) Perylene-d12	23.857	264	859144	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.540	112	643373	79.314	ng	0.00
7) Phenol-d6	7.146	99	804602	72.760	ng	0.00
23) Nitrobenzene-d5	9.146	82	509738	57.068	ng	0.00
42) 2,4,6-Tribromophenol	16.092	330	403859	100.112	ng	0.00
45) 2-Fluorobiphenyl	13.234	172	1159502	48.777	ng	0.00
79) Terphenyl-d14	19.892	244	1899976	48.145	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.405	88	111942	30.152	ng	89
3) Pyridine	3.816	79	292477	31.824	ng	# 91
4) n-Nitrosodimethylamine	3.722	42	164449	40.177	ng	# 27
6) Aniline	7.305	93	532409	41.186	ng	96
8) 2-Chlorophenol	7.546	128	402772	44.344	ng	95
9) Benzaldehyde	7.116	77	248498	51.903	ng	92
10) Phenol	7.169	94	489219	43.331	ng	98
11) bis(2-Chloroethyl)ether	7.399	93	351287	40.298	ng	99
12) 1,3-Dichlorobenzene	7.869	146	419719	40.710	ng	99
13) 1,4-Dichlorobenzene	8.016	146	426238	40.793	ng	98
14) 1,2-Dichlorobenzene	8.334	146	411907	41.220	ng	99
15) Benzyl Alcohol	8.222	79	350653	44.206	ng	# 84
16) 2,2'-oxybis(1-Chloropr...	8.516	45	375390	37.619	ng	85
17) 2-Methylphenol	8.428	107	320887	43.659	ng	97
18) Hexachloroethane	9.063	117	154983	41.798	ng	92
19) n-Nitroso-di-n-propyla...	8.793	70	279172	39.689	ng	98
20) 3+4-Methylphenols	8.757	107	442127	45.153	ng	95
22) Acetophenone	8.804	105	561703	42.392	ng	# 97
24) Nitrobenzene	9.187	77	425780	42.915	ng	91
25) Isophorone	9.716	82	737926	37.212	ng	# 98
26) 2-Nitrophenol	9.899	139	208640	54.395	ng	# 83
27) 2,4-Dimethylphenol	9.963	122	359952	47.290	ng	98
28) bis(2-Chloroethoxy)met...	10.193	93	449842	42.149	ng	98
29) 2,4-Dichlorophenol	10.440	162	379510	46.358	ng	99
30) 1,2,4-Trichlorobenzene	10.651	180	396907	40.934	ng	98
31) Naphthalene	10.840	128	1163338	39.327	ng	98
32) Benzoic acid	10.122	122	264441	61.462	ng	94
33) 4-Chloroaniline	10.945	127	364391	29.583	ng	# 91
34) Hexachlorobutadiene	11.122	225	251661	41.554	ng	98
35) Caprolactam	11.740	113	123662	55.160	ng	# 66
36) 4-Chloro-3-methylphenol	12.069	107	393741	45.693	ng	95
37) 2-Methylnaphthalene	12.445	142	844430	40.287	ng	# 88
38) 1-Methylnaphthalene	12.663	142	778429	39.143	ng	# 96
40) 1,2,4,5-Tetrachloroben...	12.810	216	463803	44.195	ng	97
41) Hexachlorocyclopentadiene	12.787	237	507843	86.921	ng	98
43) 2,4,6-Trichlorophenol	13.045	196	316292	49.691	ng	95

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060421\
 Data File : BP005711.D
 Acq On : 04 Jun 2021 14:11
 Operator : CG/JU
 Sample : M2611-01MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 T-1MS


Quant Time: Jun 04 14:49:18 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 16:11:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.122	196	349475	50.771	ng	# 94
46) 1,1'-Biphenyl	13.445	154	1081828	42.350	ng	97
47) 2-Chloronaphthalene	13.487	162	838858	40.399	ng	99
48) 2-Nitroaniline	13.687	65	248948	50.272	ng	# 77
49) Acenaphthylene	14.328	152	1359419	42.002	ng	99
50) Dimethylphthalate	14.063	163	1064278	42.532	ng	99
51) 2,6-Dinitrotoluene	14.181	165	233350	42.831	ng	# 65
52) Acenaphthene	14.669	154	790714	37.583	ng	97
53) 3-Nitroaniline	14.510	138	198327	37.171	ng	81
54) 2,4-Dinitrophenol	14.722	184	241402	103.320	ng	93
55) Dibenzofuran	15.004	168	1285193	39.853	ng	97
56) 4-Nitrophenol	14.822	139	451575	95.963	ng	# 59
57) 2,4-Dinitrotoluene	14.969	165	325465	47.638	ng	# 66
58) Fluorene	15.651	166	1039537	40.318	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.145	232	301247	50.584	ng	95
60) Diethylphthalate	15.422	149	1065985	42.458	ng	98
61) 4-Chlorophenyl-phenyle...	15.645	204	529932	41.110	ng	99
62) 4-Nitroaniline	15.669	138	258589	45.347	ng	# 48
63) Azobenzene	15.933	77	1003024	38.150	ng	96
65) 4,6-Dinitro-2-methylph...	15.733	198	161200	47.224	ng	91
66) n-Nitrosodiphenylamine	15.857	169	916747	39.626	ng	# 93
67) 4-Bromophenyl-phenylether	16.533	248	353259	43.191	ng	96
68) Hexachlorobenzene	16.657	284	425094	43.650	ng	# 95
69) Atrazine	16.810	200	347981	52.633	ng	96
70) Pentachlorophenol	17.004	266	521888	106.849	ng	96
71) Phenanthrene	17.392	178	1707717	40.373	ng	100
72) Anthracene	17.480	178	1756310	42.053	ng	100
73) Carbazole	17.751	167	1629813	39.651	ng	100
74) Di-n-butylphthalate	18.298	149	1928972	43.823	ng	# 91
75) Fluoranthene	19.351	202	2120834	41.769	ng	98
77) Benzidine	19.527	184	1334101	94.864	ng	99
78) Pyrene	19.698	202	2182165	41.763	ng	99
80) Butylbenzylphthalate	20.563	149	910998	43.525	ng	# 92
81) Benzo(a)anthracene	21.404	228	2187255	40.248	ng	99
82) 3,3'-Dichlorobenzidine	21.333	252	808245	46.044	ng	# 96
83) Chrysene	21.457	228	2133705	40.546	ng	99
84) Bis(2-ethylhexyl)phtha...	21.321	149	1296296	43.917	ng	# 96
85) Di-n-octyl phthalate	22.251	149	2309703	46.558	ng	# 96
87) Indeno(1,2,3-cd)pyrene	26.439	276	3088280	42.250	ng	# 92
88) Benzo(b)fluoranthene	23.110	252	2412126	38.696	ng	# 98
89) Benzo(k)fluoranthene	23.163	252	2379043	40.276	ng	# 98
90) Benzo(a)pyrene	23.751	252	2379810	41.965	ng	# 97
91) Dibenzo(a,h)anthracene	26.451	278	2625514	41.389	ng	# 95
92) Benzo(g,h,i)perylene	27.221	276	2598792	42.474	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 04 14:49:18 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 27 16:11:11 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.981 min Scan# 832
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampled :
T-1MS


Tgt Ion:152 Resp: 126115
Ion Ratio Lower Upper
152 100
150 154.0 119.8 179.8
115 57.6 49.0 73.4



#2
1,4-Dioxane
Concen: 30.152 ng
RT: 3.405 min Scan# 54
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion: 88 Resp: 111942
Ion Ratio Lower Upper
88 100
58 62.7 42.5 63.7
43 27.5 19.8 29.8









#3
Pyridine
Concen: 31.824 ng
RT: 3.816 min Scan# 124
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion: 79 Resp: 292477
Ion Ratio Lower Upper
79 100
52 55.8 40.3 60.5
51 30.8 20.0 30.0#



#4
n-Nitrosodimethylamine
Concen: 40.177 ng
RT: 3.722 min Scan# 108
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion: 42 Resp: 164449
Ion Ratio Lower Upper
42 100
74 136.4 122.6 183.8
44 6.0 146.9 220.3#









#5
2-Fluorophenol
Concen: 79.314 ng
RT: 5.540 min Scan# 417
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:112 Resp: 643373
Ion Ratio Lower Upper
112 100
64 53.1 33.6 50.4#
63 29.8 18.6 27.8#



#6
Aniline
Concen: 41.186 ng
RT: 7.305 min Scan# 717
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion: 93 Resp: 532409
Ion Ratio Lower Upper
93 100
66 37.1 27.0 40.4
65 18.6 14.4 21.6









#7
Phenol-d6
Concen: 72.760 ng
RT: 7.146 min Scan# 690
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion: 99 Resp: 804602
Ion Ratio Lower Upper
99 100
42 17.9 12.9 19.3
71 35.8 26.7 40.1



#8
2-Chlorophenol
Concen: 44.344 ng
RT: 7.546 min Scan# 758
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion: 128 Resp: 402772
Ion Ratio Lower Upper
128 100
130 31.6 14.1 54.1
64 40.0 17.2 57.2









#9
Benzaldehyde
Concen: 51.903 ng
RT: 7.116 min Scan# 685
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion: 77 Resp: 248498
Ion Ratio Lower Upper
77 100
105 98.1 89.4 129.4
106 94.3 79.9 119.9



#10
Phenol
Concen: 43.331 ng
RT: 7.169 min Scan# 694
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11


Tgt Ion: 94 Resp: 489219
Ion Ratio Lower Upper
94 100
65 28.4 8.0 48.0
66 38.7 17.3 57.3








#11
bis(2-Chloroethyl)ether
Concen: 40.298 ng
RT: 7.399 min Scan# 733
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Instrument :
BNA_P
ClientSampled :
T-1MS







Tgt Ion: 93 Resp: 351287
Ion Ratio Lower Upper
93 100
63 70.3 49.6 89.6
95 33.0 14.4 54.4



#12
1,3-Dichlorobenzene
Concen: 40.710 ng
RT: 7.869 min Scan# 813
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:146 Resp: 419719
Ion Ratio Lower Upper
146 100
148 64.0 51.4 77.0
75 32.2 23.9 35.9









#13
1,4-Dichlorobenzene
Concen: 40.793 ng
RT: 8.016 min Scan# 838
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:146 Resp: 426238
Ion Ratio Lower Upper
146 100
148 64.8 52.2 78.4
111 39.7 34.2 51.4



#14
1,2-Dichlorobenzene
Concen: 41.220 ng
RT: 8.334 min Scan# 892
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:146 Resp: 411907
Ion Ratio Lower Upper
146 100
148 64.1 51.5 77.3
111 41.5 34.0 51.0









#15
Benzyl Alcohol
Concen: 44.206 ng
RT: 8.222 min Scan# 873
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampled :
T-1MS


Tgt Ion: 79 Resp: 350653
Ion Ratio Lower Upper
79 100
108 77.1 64.0 96.0
77 65.1 72.7 109.1#



#16
2,2'-oxybis(1-Chloropropane)
Concen: 37.619 ng
RT: 8.516 min Scan# 923
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion: 45 Resp: 375390
Ion Ratio Lower Upper
45 100
77 18.7 7.0 47.0
79 14.7 1.0 41.0









#17
2-Methylphenol
Concen: 43.659 ng
RT: 8.428 min Scan# 908
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:	107	Resp:	320887
Ion Ratio	Lower	Upper	
107	100		
108	112.3	85.8	128.8
77	50.1	41.4	62.0
79	47.5	38.6	57.8



#18
Hexachloroethane
Concen: 41.798 ng
RT: 9.063 min Scan# 1016
Delta R.T. -0.018 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:	117	Resp:	154983
Ion Ratio	Lower	Upper	
117	100		
119	97.8	80.1	120.1
201	108.2	98.8	148.2









#19
n-Nitroso-di-n-propylamine
Concen: 39.689 ng
RT: 8.793 min Scan# 970
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:	70	Resp:	279172
Ion	Ratio	Lower	Upper
70	100		
42	51.3	39.7	59.5
101	9.7	7.4	11.0
130	22.5	17.2	25.8



#20
3+4-Methylphenols
Concen: 45.153 ng
RT: 8.757 min Scan# 964
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:	107	Resp:	442127
Ion	Ratio	Lower	Upper
107	100		
108	90.3	69.5	109.5
77	34.7	21.5	61.5
79	31.2	16.8	56.8









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.787 min Scan# 1309
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:	136	Resp:	496745
Ion Ratio	Lower	Upper	
136	100		
137	10.9	8.9	13.3
54	6.7	5.0	7.4
68	5.5	3.7	5.5



#22
Acetophenone
Concen: 42.392 ng
RT: 8.804 min Scan# 972
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:	105	Resp:	561703
Ion Ratio	Lower	Upper	
105	100		
71	3.5	0.9	1.3
51	24.2	18.2	27.2
120	22.3	17.0	25.6









#23
Nitrobenzene-d5
Concen: 57.068 ng
RT: 9.146 min Scan# 1030
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion: 82 Resp: 509738
Ion Ratio Lower Upper
82 100
128 43.8 30.5 45.7
54 43.4 33.9 50.9



#24
Nitrobenzene
Concen: 42.915 ng
RT: 9.187 min Scan# 1037
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion: 77 Resp: 425780
Ion Ratio Lower Upper
77 100
123 46.3 31.7 47.5
65 13.1 11.5 17.3









#25
Isophorone
Concen: 37.212 ng
RT: 9.716 min Scan# 1127
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion: 82 Resp: 737926
Ion Ratio Lower Upper
82 100
95 7.4 7.5 11.3#
138 18.2 14.6 22.0



#26
2-Nitrophenol
Concen: 54.395 ng
RT: 9.899 min Scan# 1158
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion: 139 Resp: 208640
Ion Ratio Lower Upper
139 100
109 32.7 40.7 61.1#
65 53.3 47.8 71.6









#27
2,4-Dimethylphenol
Concen: 47.290 ng
RT: 9.963 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:122 Resp: 359952
Ion Ratio Lower Upper
122 100
107 115.1 93.8 140.8
121 58.6 45.6 68.4



#28
bis(2-Chloroethoxy)methane
Concen: 42.149 ng
RT: 10.193 min Scan# 1208
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion: 93 Resp: 449842
Ion Ratio Lower Upper
93 100
95 32.9 24.9 37.3
123 11.9 9.5 14.3









#29
2,4-Dichlorophenol
Concen: 46.358 ng
RT: 10.440 min Scan# 1250
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:162 Resp: 379510
Ion Ratio Lower Upper
162 100
164 64.4 44.1 84.1
98 33.3 12.6 52.6



#30
1,2,4-Trichlorobenzene
Concen: 40.934 ng
RT: 10.651 min Scan# 1286
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:180 Resp: 396907
Ion Ratio Lower Upper
180 100
182 95.8 77.2 115.8
145 29.9 21.7 32.5









#31
Naphthalene
Concen: 39.327 ng
RT: 10.840 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:128 Resp: 1163338
Ion Ratio Lower Upper
128 100
129 11.0 8.0 12.0
127 13.2 11.1 16.7



#32
Benzoic acid
Concen: 61.462 ng
RT: 10.122 min Scan# 1196
Delta R.T. 0.035 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:122 Resp: 264441
Ion Ratio Lower Upper
122 100
105 117.8 105.1 145.1
77 94.4 69.6 109.6









#33
4-Chloroaniline
Concen: 29.583 ng
RT: 10.945 min Scan# 1336
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:	127	Resp:	364391
Ion Ratio	Lower	Upper	
127	100		
129	31.8	25.5	38.3
65	28.6	30.0	45.0#
92	18.9	20.2	30.2#



#34
Hexachlorobutadiene
Concen: 41.554 ng
RT: 11.122 min Scan# 1366
Delta R.T. -0.018 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:	225	Resp:	251661
Ion Ratio	Lower	Upper	
225	100		
223	62.4	51.1	76.7
227	64.7	50.2	75.2









#35
Caprolactam
Concen: 55.160 ng
RT: 11.740 min Scan# 1471
Delta R.T. 0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:113 Resp: 123662
Ion Ratio Lower Upper
113 100
55 134.3 72.0 112.0#
56 105.6 65.1 105.1#



#36
4-Chloro-3-methylphenol
Concen: 45.693 ng
RT: 12.069 min Scan# 1527
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:107 Resp: 393741
Ion Ratio Lower Upper
107 100
144 26.6 20.1 30.1
142 82.2 62.2 93.4









#37
2-Methylnaphthalene
Concen: 40.287 ng
RT: 12.445 min Scan# 1591
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:142 Resp: 844430
Ion Ratio Lower Upper
142 100
141 89.1 74.7 112.1
115 37.1 43.8 65.8#



#38
1-Methylnaphthalene
Concen: 39.143 ng
RT: 12.663 min Scan# 1628
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:142 Resp: 778429
Ion Ratio Lower Upper
142 100
141 92.9 77.0 115.6
116 3.9 5.0 7.6#









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.604 min Scan# 1958
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:164 Resp: 314597
Ion Ratio Lower Upper
164 100
162 99.8 73.5 110.3
160 43.1 40.6 61.0



#40
1,2,4,5-Tetrachlorobenzene
Concen: 44.195 ng
RT: 12.810 min Scan# 1653
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:216 Resp: 463803
Ion Ratio Lower Upper
216 100
214 78.1 60.6 91.0
179 20.2 15.3 22.9
108 20.4 13.8 20.6









#41
Hexachlorocyclopentadiene
Concen: 86.921 ng
RT: 12.787 min Scan# 1649
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:237 Resp: 507843
Ion Ratio Lower Upper
237 100
235 61.8 40.9 80.9
272 12.8 0.0 34.6



#42
2,4,6-Tribromophenol
Concen: 100.112 ng
RT: 16.092 min Scan# 2211
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:330 Resp: 403859
Ion Ratio Lower Upper
330 100
332 96.0 76.3 114.5
141 30.5 22.8 34.2









#43
2,4,6-Trichlorophenol
Concen: 49.691 ng
RT: 13.045 min Scan# 1693
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:196 Resp: 316292
Ion Ratio Lower Upper
196 100
198 96.8 81.4 122.2
200 31.1 28.2 42.2



#44
2,4,5-Trichlorophenol
Concen: 50.771 ng
RT: 13.122 min Scan# 1706
Delta R.T. 0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:196 Resp: 349475
Ion Ratio Lower Upper
196 100
198 95.4 71.0 106.4
97 46.9 38.5 57.7
132 25.2 15.6 23.4#









#45
2-Fluorobiphenyl
Concen: 48.777 ng
RT: 13.234 min Scan# 1725
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Instrument :
BNA_P
ClientSampleId :
T-1MS

Tgt Ion:172 Resp: 1159502




Ion	Ratio	Lower	Upper
172	100		
171	35.2	28.1	42.1
170	23.7	19.6	29.4




#46
1,1'-Biphenyl
Concen: 42.350 ng
RT: 13.445 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:154 Resp: 1081828

Ion	Ratio	Lower	Upper
154	100		
153	41.2	18.9	58.9
76	11.7	0.0	32.1









#47
2-Chloronaphthalene
Concen: 40.399 ng
RT: 13.487 min Scan# 1768
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:162 Resp: 838858
Ion Ratio Lower Upper
162 100
127 38.6 31.4 47.0
164 33.1 26.6 39.8



#48
2-Nitroaniline
Concen: 50.272 ng
RT: 13.687 min Scan# 1802
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion: 65 Resp: 248948
Ion Ratio Lower Upper
65 100
92 71.3 48.8 73.2
138 114.5 69.5 104.3#









#49
Acenaphthylene
Concen: 42.002 ng
RT: 14.328 min Scan# 1911
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:152 Resp: 1359419
Ion Ratio Lower Upper
152 100
151 20.4 15.8 23.6
153 13.5 10.5 15.7



#50
Dimethylphthalate
Concen: 42.532 ng
RT: 14.063 min Scan# 1866
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:163 Resp: 1064278
Ion Ratio Lower Upper
163 100
194 4.2 3.9 5.9
164 10.4 8.1 12.1









#51
2,6-Dinitrotoluene
Concen: 42.831 ng
RT: 14.181 min Scan# 1886
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:165 Resp: 233350
Ion Ratio Lower Upper
165 100
63 54.4 67.4 101.0#
89 56.8 73.8 110.6#



#52
Acenaphthene
Concen: 37.583 ng
RT: 14.669 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:154 Resp: 790714
Ion Ratio Lower Upper
154 100
153 114.6 90.9 136.3
152 54.4 48.5 72.7









#53
3-Nitroaniline
Concen: 37.171 ng
RT: 14.510 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:138 Resp: 198327
Ion Ratio Lower Upper
138 100
108 10.4 10.2 15.4
92 110.7 108.1 162.1



#54
2,4-Dinitrophenol
Concen: 103.320 ng
RT: 14.722 min Scan# 1978
Delta R.T. 0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:184 Resp: 241402
Ion Ratio Lower Upper
184 100
63 59.5 51.2 76.8
154 63.5 55.8 83.8









#55
Dibenzenofuran
Concen: 39.853 ng
RT: 15.004 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:168 Resp: 1285193
Ion Ratio Lower Upper
168 100
139 37.7 31.4 47.0
169 13.1 11.4 17.0



#56
4-Nitrophenol
Concen: 95.963 ng
RT: 14.822 min Scan# 1995
Delta R.T. 0.018 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:139 Resp: 451575
Ion Ratio Lower Upper
139 100
109 82.0 122.8 162.8#
65 92.7 110.0 150.0#









#57
2,4-Dinitrotoluene
Concen: 47.638 ng
RT: 14.969 min Scan# 2020
Delta R.T. 0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:165 Resp: 325465
Ion Ratio Lower Upper
165 100
63 41.0 56.7 85.1#
89 71.6 85.8 128.6#
182 2.9 3.0 4.6#



#58
Fluorene
Concen: 40.318 ng
RT: 15.651 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:166 Resp: 1039537
Ion Ratio Lower Upper
166 100
165 95.9 78.8 118.2
167 13.6 10.9 16.3









#59
2,3,4,6-Tetrachlorophenol
Concen: 50.584 ng
RT: 15.145 min Scan# 2050
Delta R.T. -0.083 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:232 Resp: 301247
Ion Ratio Lower Upper
232 100
131 37.8 26.5 39.7
130 2.4 2.1 3.1
166 23.9 19.4 29.0



#60
Diethylphthalate
Concen: 42.458 ng
RT: 15.422 min Scan# 2097
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:149 Resp: 1065985
Ion Ratio Lower Upper
149 100
177 22.7 19.2 28.8
150 12.6 9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 41.110 ng
RT: 15.645 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:204 Resp: 529932
Ion Ratio Lower Upper
204 100
206 33.4 26.0 39.0
141 55.4 44.2 66.2



#62
4-Nitroaniline
Concen: 45.347 ng
RT: 15.669 min Scan# 2139
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:138 Resp: 258589
Ion Ratio Lower Upper
138 100
92 50.2 50.9 90.9#
108 87.1 154.7 194.7#









#63
Azobenzene
Concen: 38.150 ng
RT: 15.933 min Scan# 2184
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion: 77 Resp: 1003024
Ion Ratio Lower Upper
77 100
182 25.4 5.5 45.5
105 19.5 0.0 36.5
51 27.4 10.3 50.3



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.351 min Scan# 2425
Delta R.T. 0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:188 Resp: 678940
Ion Ratio Lower Upper
188 100
94 7.9 5.0 7.6#
80 8.8 5.9 8.9









#65
4,6-Dinitro-2-methylphenol
Concen: 47.224 ng
RT: 15.733 min Scan# 2150
Delta R.T. 0.005 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:198 Resp: 161200
Ion Ratio Lower Upper
198 100
51 35.1 10.3 50.3
105 38.8 13.6 53.6



#66
n-Nitrosodiphenylamine
Concen: 39.626 ng
RT: 15.857 min Scan# 2171
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:169 Resp: 916747
Ion Ratio Lower Upper
169 100
168 66.8 55.2 82.8
167 36.0 36.6 54.8#









#67
4-Bromophenyl-phenylether
Concen: 43.191 ng
RT: 16.533 min Scan# 2286
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:248 Resp: 353259
Ion Ratio Lower Upper
248 100
250 98.6 77.5 116.3
141 69.0 50.1 75.1



#68
Hexachlorobenzene
Concen: 43.650 ng
RT: 16.657 min Scan# 2307
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:284 Resp: 425094
Ion Ratio Lower Upper
284 100
142 31.5 20.8 31.2#
249 29.6 23.7 35.5









#69
Atrazine
Concen: 52.633 ng
RT: 16.810 min Scan# 2333
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:200 Resp: 347981
Ion Ratio Lower Upper
200 100
173 25.5 0.7 40.7
215 50.3 29.0 69.0



#70
Pentachlorophenol
Concen: 106.849 ng
RT: 17.004 min Scan# 2366
Delta R.T. 0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:266 Resp: 521888
Ion Ratio Lower Upper
266 100
268 64.5 49.4 74.0
264 63.3 53.7 80.5









#71
Phenanthrene
Concen: 40.373 ng
RT: 17.392 min Scan# 2432
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:178 Resp: 1707717
Ion Ratio Lower Upper
178 100
176 19.2 15.4 23.2
179 15.4 12.3 18.5



#72
Anthracene
Concen: 42.053 ng
RT: 17.480 min Scan# 2447
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:178 Resp: 1756310
Ion Ratio Lower Upper
178 100
176 18.8 15.2 22.8
179 15.6 12.6 19.0









#73
 Carbazole
 Concen: 39.651 ng
 RT: 17.751 min Scan# 2493
 Delta R.T. -0.000 min
 Lab File: BP005711.D
 Acq: 04 Jun 2021 14:11

Instrument :
 BNA_P
 ClientSampleId :
 T-1MS

Tgt Ion:167 Resp: 1629813




Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.4	26.0
139	12.9	10.4	15.6





#74
 Di-n-butylphthalate
 Concen: 43.823 ng
 RT: 18.298 min Scan# 2586
 Delta R.T. -0.006 min
 Lab File: BP005711.D
 Acq: 04 Jun 2021 14:11

Tgt Ion:149 Resp: 1928972

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.6	11.4
104	5.2	8.9	13.3











#77
Benzidine
Concen: 94.864 ng
RT: 19.527 min Scan# 2795
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:184 Resp: 1334101
Ion Ratio Lower Upper
184 100
185 14.2 11.1 16.7
183 11.7 9.8 14.8



#78
Pyrene
Concen: 41.763 ng
RT: 19.698 min Scan# 2824
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:202 Resp: 2182165
Ion Ratio Lower Upper
202 100
200 21.3 16.9 25.3
203 17.8 13.8 20.6





#79

Terphenyl-d14

Concen: 48.145 ng

RT: 19.892 min Scan# 2857

Delta R.T. -0.006 min

Lab File: BP005711.D

Acq: 04 Jun 2021 14:11

Instrument :

BNA_P

ClientSampleId :

T-1MS





Tgt Ion:244 Resp: 1899976

Ion Ratio Lower Upper

244 100

212 7.2 5.8 8.6

122 8.8 5.2 7.8#



#80

Butylbenzylphthalate

Concen: 43.525 ng

RT: 20.563 min Scan# 2971

Delta R.T. -0.006 min

Lab File: BP005711.D

Acq: 04 Jun 2021 14:11




Tgt Ion:149 Resp: 910998


Ion Ratio Lower Upper

149 100

91 69.9 60.2 90.4

206 23.0 23.9 35.9#









#81
Benzo(a)anthracene
Concen: 40.248 ng
RT: 21.404 min Scan# 3114
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:228 Resp: 2187255
Ion Ratio Lower Upper
228 100
226 27.2 21.2 31.8
229 19.8 15.6 23.4



#82
3,3'-Dichlorobenzidine
Concen: 46.044 ng
RT: 21.333 min Scan# 3102
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:252 Resp: 808245
Ion Ratio Lower Upper
252 100
254 64.9 49.9 74.9
126 9.3 5.5 8.3#





#83

Chrysene

Concen: 40.546 ng

RT: 21.457 min Scan# 3123

Delta R.T. -0.000 min

Lab File: BP005711.D

Acq: 04 Jun 2021 14:11





Tgt Ion:228 Resp: 2133705

Ion Ratio Lower Upper

228 100

226 30.4 24.8 37.2

229 19.9 15.4 23.0



#84

Bis(2-ethylhexyl)phthalate

Concen: 43.917 ng

RT: 21.321 min Scan# 3100

Delta R.T. -0.006 min

Lab File: BP005711.D

Acq: 04 Jun 2021 14:11




Tgt Ion:149 Resp: 1296296


Ion Ratio Lower Upper

149 100

167 28.0 23.8 35.8

279 5.6 5.7 8.5#









#85
Di-n-octyl phthalate
Concen: 46.558 ng
RT: 22.251 min Scan# 3258
Delta R.T. -0.012 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:149 Resp: 2309703
Ion Ratio Lower Upper
149 100
167 1.6 1.5 2.3
43 8.0 5.1 7.7#



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.857 min Scan# 3531
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:264 Resp: 859144
Ion Ratio Lower Upper
264 100
260 24.2 20.6 30.8
265 22.4 19.0 28.6









#87
Indeno(1,2,3-cd)pyrene
Concen: 42.250 ng
RT: 26.439 min Scan# 3970
Delta R.T. 0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:276 Resp: 3088280
Ion Ratio Lower Upper
276 100
138 19.8 8.4 12.6#
277 25.5 20.6 31.0



#88
Benzo(b)fluoranthene
Concen: 38.696 ng
RT: 23.110 min Scan# 3404
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:252 Resp: 2412126
Ion Ratio Lower Upper
252 100
253 22.9 18.2 27.2
125 8.6 3.4 5.0#









#89
Benzo(k)fluoranthene
Concen: 40.276 ng
RT: 23.163 min Scan# 3413
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11




Instrument :
BNA_P
ClientSampleId :
T-1MS


Tgt Ion:252 Resp: 2379043
Ion Ratio Lower Upper
252 100
253 21.9 17.2 25.8
125 8.0 3.8 5.6#



#90
Benzo(a)pyrene
Concen: 41.965 ng
RT: 23.751 min Scan# 3513
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:252 Resp: 2379810
Ion Ratio Lower Upper
252 100
253 22.0 17.1 25.7
125 8.9 3.8 5.6#









#91
Dibenzo(a,h)anthracene
Concen: 41.389 ng
RT: 26.451 min Scan# 3972
Delta R.T. -0.000 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Instrument :
BNA_P
ClientSampleId :
T-1MS

Tgt Ion:278 Resp: 2625514
Ion Ratio Lower Upper
278 100
139 12.9 5.0 7.6#
279 23.7 18.6 28.0



#92
Benzo(g,h,i)perylene
Concen: 42.474 ng
RT: 27.221 min Scan# 4103
Delta R.T. -0.006 min
Lab File: BP005711.D
Acq: 04 Jun 2021 14:11

Tgt Ion:276 Resp: 2598792
Ion Ratio Lower Upper
276 100
277 24.0 19.8 29.8
138 16.6 7.7 11.5#

