

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060424\
 Data File : BP020499.D
 Acq On : 04 Jun 2024 16:44
 Operator : MA/JU
 Sample : P2687-09
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 72-11998

Quant Time: Jun 04 17:24:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 03:41:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	267600	20.000	ng	0.00
21) Naphthalene-d8	10.534	136	1037480	20.000	ng	0.00
39) Acenaphthene-d10	14.381	164	656828	20.000	ng	-0.01
64) Phenanthrene-d10	17.192	188	1443895	20.000	ng	-0.01
76) Chrysene-d12	21.651	240	1352829	20.000	ng	-0.02
86) Perylene-d12	25.004	264	1475279	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	1758574	117.662	ng	0.00
7) Phenol-d6	6.928	99	2173896	103.138	ng	0.00
23) Nitrobenzene-d5	8.916	82	1535600	81.022	ng	0.00
42) 2,4,6-Tribromophenol	15.893	330	929412	136.345	ng	-0.02
45) 2-Fluorobiphenyl	12.999	172	3258263	73.949	ng	0.00
79) Terphenyl-d14	19.934	244	6315652	77.717	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060424\
Data File : BP020499.D
Acq On : 04 Jun 2024 16:44
Operator : MA/JU
Sample : P2687-09
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
72-11998

Quant Time: Jun 04 17:24:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 03:41:03 2024
Response via : Initial Calibration

