

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060424\
 Data File : BP020494.D
 Acq On : 04 Jun 2024 12:27
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 04 13:06:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 03:41:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	120	0.00
2	1,4-Dioxane	40.000	41.305	-3.3	126	0.00
3	Pyridine	40.000	42.060	-5.2	124	0.00
4	n-Nitrosodimethylamine	40.000	41.298	-3.2	122	0.00
5 S	2-Fluorophenol	80.000	85.132	-6.4	126	0.00
6	Aniline	40.000	40.474	-1.2	119	0.00
7 S	Phenol-d6	80.000	84.123	-5.2	124	0.00
8	2-Chlorophenol	40.000	42.274	-5.7	125	0.00
9	Benzaldehyde	40.000	44.997	-12.5	153	0.00
10 C	Phenol	40.000	42.037	-5.1	124	0.00
11	bis(2-Chloroethyl)ether	40.000	40.856	-2.1	121	0.00
12	1,3-Dichlorobenzene	40.000	41.626	-4.1	125	0.00
13 C	1,4-Dichlorobenzene	40.000	41.354	-3.4	124	0.00
14	1,2-Dichlorobenzene	40.000	40.881	-2.2	122	0.00
15	Benzyl Alcohol	40.000	41.544	-3.9	121	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.825	2.9	116	0.00
17	2-Methylphenol	40.000	41.460	-3.7	121	0.00
18	Hexachloroethane	40.000	41.588	-4.0	124	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.478	1.3	115	0.00
20	3+4-Methylphenols	40.000	42.062	-5.2	122	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	117	0.01
22	Acetophenone	40.000	40.038	-0.1	117	0.00
23 S	Nitrobenzene-d5	80.000	84.850	-6.1	122	0.00
24	Nitrobenzene	40.000	41.486	-3.7	120	0.00
25	Isophorone	40.000	40.175	-0.4	116	0.01
26 C	2-Nitrophenol	40.000	42.859	-7.1	130	0.00
27	2,4-Dimethylphenol	40.000	41.294	-3.2	118	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.055	-0.1	117	0.00
29 C	2,4-Dichlorophenol	40.000	43.142	-7.9	123	0.00
30	1,2,4-Trichlorobenzene	40.000	42.055	-5.1	123	0.00
31	Naphthalene	40.000	40.499	-1.2	119	0.01
32	Benzoic acid	40.000	43.669	-9.2	139	0.02
33	4-Chloroaniline	40.000	40.415	-1.0	117	0.00
34 C	Hexachlorobutadiene	40.000	42.419	-6.0	125	0.01
35	Caprolactam	40.000	48.496	-21.2	137	0.01
36 C	4-Chloro-3-methylphenol	40.000	41.556	-3.9	120	0.00
37	2-Methylnaphthalene	40.000	40.007	-0.0	118	-0.01
38	1-Methylnaphthalene	40.000	39.860	0.4	119	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	114	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.928	-4.8	121	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.982	-10.0	122	-0.02
42 S	2,4,6-Tribromophenol	80.000	91.637	-14.5	124	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.535	-11.3	122	0.00
44	2,4,5-Trichlorophenol	40.000	44.059	-10.1	121	0.00
45 S	2-Fluorobiphenyl	80.000	80.199	-0.2	117	0.01
46	1,1'-Biphenyl	40.000	40.263	-0.7	115	-0.01
47	2-Chloronaphthalene	40.000	40.787	-2.0	117	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060424\
 Data File : BP020494.D
 Acq On : 04 Jun 2024 12:27
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 04 13:06:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 03:41:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.699	-11.7	121	0.00
49	Acenaphthylene	40.000	41.125	-2.8	119	0.00
50	Dimethylphthalate	40.000	40.810	-2.0	118	0.00
51	2,6-Dinitrotoluene	40.000	43.415	-8.5	120	0.00
52 C	Acenaphthene	40.000	40.310	-0.8	118	0.00
53	3-Nitroaniline	40.000	43.260	-8.1	119	0.00
54 P	2,4-Dinitrophenol	40.000	44.476	-11.2	140	0.01
55	Dibenzofuran	40.000	40.801	-2.0	118	0.00
56 P	4-Nitrophenol	40.000	43.973	-9.9	121	0.01
57	2,4-Dinitrotoluene	40.000	45.932	-14.8	125	0.00
58	Fluorene	40.000	40.074	-0.2	118	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.573	-11.4	121	0.00
60	Diethylphthalate	40.000	41.281	-3.2	119	0.00
61	4-Chlorophenyl-phenylether	40.000	40.421	-1.1	118	0.00
62	4-Nitroaniline	40.000	42.960	-7.4	123	0.00
63	Azobenzene	40.000	39.759	0.6	115	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	113	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.932	-7.3	135	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.062	-5.2	120	0.00
67	4-Bromophenyl-phenylether	40.000	41.505	-3.8	119	0.00
68	Hexachlorobenzene	40.000	41.661	-4.2	120	0.00
69	Atrazine	40.000	42.020	-5.1	118	0.00
70 C	Pentachlorophenol	40.000	43.731	-9.3	123	0.00
71	Phenanthrene	40.000	40.618	-1.5	116	0.00
72	Anthracene	40.000	41.533	-3.8	117	0.00
73	Carbazole	40.000	40.828	-2.1	117	0.00
74	Di-n-butylphthalate	40.000	42.764	-6.9	117	0.00
75 C	Fluoranthene	40.000	40.583	-1.5	115	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	115	-0.01
77	Benzydine	40.000	30.166	24.6	78	0.00
78	Pyrene	40.000	42.079	-5.2	117	0.00
79 S	Terphenyl-d14	80.000	86.017	-7.5	117	-0.01
80	Butylbenzylphthalate	40.000	41.336	-3.3	119	0.00
81	Benzo(a)anthracene	40.000	41.237	-3.1	119	-0.02
82	3,3'-Dichlorobenzidine	40.000	44.071	-10.2	119	-0.02
83	Chrysene	40.000	41.731	-4.3	120	-0.01
84	Bis(2-ethylhexyl)phthalate	40.000	42.789	-7.0	114	-0.02
85 c	Di-n-octyl phthalate	40.000	41.843	-4.6	119	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	117	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	45.411	-13.5	126	-0.05
88	Benzo(b)fluoranthene	40.000	42.064	-5.2	122	-0.02
89	Benzo(k)fluoranthene	40.000	42.068	-5.2	123	0.00
90 C	Benzo(a)pyrene	40.000	43.277	-8.2	124	0.00
91	Dibenzo(a,h)anthracene	40.000	44.852	-12.1	125	-0.02
92	Benzo(g,h,i)perylene	40.000	46.029	-15.1	128	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060424\
Data File : BP020494.D
Acq On : 04 Jun 2024 12:27
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 04 13:06:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 03:41:03 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0