




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060524\
Data File : BP020510.D
Acq On : 05 Jun 2024 09:17
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 05 19:21:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 03:41:03 2024
Response via : Initial Calibration



TIC: BP020510.D\data.ms

(70) Pentachlorophenol (C)




16.827min (-0.001) 388630.29 ng

response 508710

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.60	64.56
264.00	63.30	62.01
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060524\
Data File : BP020510.D
Acq On : 05 Jun 2024 09:17
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 05 19:21:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 03:41:03 2024
Response via : Initial Calibration



TIC: BP020510.D\data.ms

(77) Benzidine

19.527min (-0.006) 1334174.52 ng

response 3466910

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	14.71
183.00	11.70	12.08
0.00	0.00	0.00