

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060622\
 Data File : BP010701.D
 Acq On : 06 Jun 2022 12:41
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 06 15:40:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 06 15:39:44 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	62	0.00
2	1,4-Dioxane	40.000	37.287	6.8	58	0.00
3	Pyridine	40.000	43.511	-8.8	66	0.00
4	n-Nitrosodimethylamine	40.000	39.251	1.9	62	0.00
5 S	2-Fluorophenol	80.000	86.392	-8.0	66	0.00
6	Aniline	40.000	46.719	-16.8	73	0.00
7 S	Phenol-d6	80.000	92.810	-16.0	72	0.00
8	2-Chlorophenol	40.000	43.875	-9.7	69	0.00
9	Benzaldehyde	40.000	38.872	2.8	68	0.00
10 C	Phenol	40.000	46.199	-15.5	73	0.00
11	bis(2-Chloroethyl)ether	40.000	43.437	-8.6	67	0.00
12	1,3-Dichlorobenzene	40.000	40.653	-1.6	64	0.00
13 C	1,4-Dichlorobenzene	40.000	41.604	-4.0	65	0.00
14	1,2-Dichlorobenzene	40.000	41.853	-4.6	66	0.00
15	Benzyl Alcohol	40.000	47.453	-18.6	73	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.595	3.5	61	0.00
17	2-Methylphenol	40.000	47.254	-18.1	73	0.00
18	Hexachloroethane	40.000	40.705	-1.8	64	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	47.675	-19.2	77	0.00
20	3+4-Methylphenols	40.000	49.319	-23.3	76	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	71	0.00
22	Acetophenone	40.000	42.189	-5.5	75	0.00
23 S	Nitrobenzene-d5	80.000	82.104	-2.6	73	0.00
24	Nitrobenzene	40.000	40.574	-1.4	71	0.00
25	Isophorone	40.000	44.594	-11.5	79	0.00
26 C	2-Nitrophenol	40.000	45.069	-12.7	78	0.00
27	2,4-Dimethylphenol	40.000	44.251	-10.6	78	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.482	-3.7	74	0.00
29 C	2,4-Dichlorophenol	40.000	43.239	-8.1	75	0.00
30	1,2,4-Trichlorobenzene	40.000	38.863	2.8	69	0.00
31	Naphthalene	40.000	40.902	-2.3	73	0.00
32	Benzoic acid	40.000	43.610	-9.0	88	0.00
33	4-Chloroaniline	40.000	46.162	-15.4	82	0.00
34 C	Hexachlorobutadiene	40.000	38.297	4.3	69	0.00
35	Caprolactam	40.000	61.165	-52.9#	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	48.611	-21.5#	86	0.00
37	2-Methylnaphthalene	40.000	43.085	-7.7	78	0.00
38	1-Methylnaphthalene	40.000	43.520	-8.8	78	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	81	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.046	4.9	77	0.00
41 P	Hexachlorocyclopentadiene	40.000	26.920	32.7#	52	0.00
42 S	2,4,6-Tribromophenol	80.000	104.985	-31.2#	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.904	-7.3	83	0.00
44	2,4,5-Trichlorophenol	40.000	41.927	-4.8	83	0.00
45 S	2-Fluorobiphenyl	80.000	77.443	3.2	79	0.00
46	1,1'-Biphenyl	40.000	39.260	1.9	80	0.00
47	2-Chloronaphthalene	40.000	39.611	1.0	81	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060622\
 Data File : BP010701.D
 Acq On : 06 Jun 2022 12:41
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 06 15:40:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 06 15:39:44 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	47.176	-17.9	92	0.00
49	Acenaphthylene	40.000	42.466	-6.2	86	0.00
50	Dimethylphthalate	40.000	44.349	-10.9	90	0.00
51	2,6-Dinitrotoluene	40.000	45.830	-14.6	90	0.00
52 C	Acenaphthene	40.000	41.486	-3.7	85	0.00
53	3-Nitroaniline	40.000	51.284	-28.2#	99	0.00
54 P	2,4-Dinitrophenol	40.000	46.244	-15.6	101	0.00
55	Dibenzofuran	40.000	42.323	-5.8	86	0.00
56 P	4-Nitrophenol	40.000	49.112	-22.8	100	0.00
57	2,4-Dinitrotoluene	40.000	51.342	-28.4#	98	0.00
58	Fluorene	40.000	44.328	-10.8	89	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	47.657	-19.1	94	0.00
60	Diethylphthalate	40.000	46.269	-15.7	92	0.00
61	4-Chlorophenyl-phenylether	40.000	43.155	-7.9	88	0.00
62	4-Nitroaniline	40.000	53.449	-33.6#	108	0.00
63	Azobenzene	40.000	44.214	-10.5	88	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.688	-6.7	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.998	2.5	93	0.00
67	4-Bromophenyl-phenylether	40.000	39.566	1.1	94	0.00
68	Hexachlorobenzene	40.000	40.971	-2.4	99	0.00
69	Atrazine	40.000	44.616	-11.5	102	0.00
70 C	Pentachlorophenol	40.000	42.506	-6.3	97	0.00
71	Phenanthrene	40.000	40.873	-2.2	96	0.00
72	Anthracene	40.000	42.056	-5.1	98	0.00
73	Carbazole	40.000	46.249	-15.6	103	0.00
74	Di-n-butylphthalate	40.000	44.998	-12.5	99	0.00
75 C	Fluoranthene	40.000	46.587	-16.5	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	108	0.00
77	Benzydine	40.000	53.525	-33.8#	144	0.00
78	Pyrene	40.000	37.620	6.0	105	0.00
79 S	Terphenyl-d14	80.000	77.699	2.9	108	0.00
80	Butylbenzylphthalate	40.000	42.379	-5.9	109	0.00
81	Benzo(a)anthracene	40.000	41.339	-3.3	111	0.00
82	3,3'-Dichlorobenzidine	40.000	47.879	-19.7	122	0.00
83	Chrysene	40.000	40.518	-1.3	109	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.479	-6.2	106	0.00
85 c	Di-n-octyl phthalate	40.000	43.547	-8.9	109	0.00
86 I	Perylene-d12	20.000	20.000	0.0	116	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.921	-4.8	121	0.00
88	Benzo(b)fluoranthene	40.000	40.639	-1.6	114	0.00
89	Benzo(k)fluoranthene	40.000	39.730	0.7	117	0.00
90 C	Benzo(a)pyrene	40.000	41.286	-3.2	117	0.00
91	Dibenzo(a,h)anthracene	40.000	42.050	-5.1	121	0.00
92	Benzo(g,h,i)perylene	40.000	41.116	-2.8	120	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060622\
Data File : BP010701.D
Acq On : 06 Jun 2022 12:41
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 06 15:40:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 06 15:39:44 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 1