

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060624\
 Data File : BP020537.D
 Acq On : 06 Jun 2024 11:30
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 06 13:04:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 03:41:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	38.888	2.8	101	0.00
3	Pyridine	40.000	37.821	5.4	95	0.00
4	n-Nitrosodimethylamine	40.000	35.134	12.2	88	0.00
5 S	2-Fluorophenol	80.000	79.375	0.8	100	0.00
6	Aniline	40.000	35.045	12.4	87	0.00
7 S	Phenol-d6	80.000	73.509	8.1	92	0.00
8	2-Chlorophenol	40.000	39.401	1.5	99	0.00
9	Benzaldehyde	40.000	32.971	17.6	96	0.00
10 C	Phenol	40.000	36.314	9.2	91	0.00
11	bis(2-Chloroethyl)ether	40.000	36.616	8.5	92	-0.02
12	1,3-Dichlorobenzene	40.000	39.667	0.8	101	0.00
13 C	1,4-Dichlorobenzene	40.000	39.172	2.1	100	0.00
14	1,2-Dichlorobenzene	40.000	39.062	2.3	99	0.00
15	Benzyl Alcohol	40.000	35.391	11.5	88	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	32.893	17.8	84	-0.01
17	2-Methylphenol	40.000	36.741	8.1	92	0.00
18	Hexachloroethane	40.000	39.798	0.5	101	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.643	13.4	86	-0.01
20	3+4-Methylphenols	40.000	37.422	6.4	93	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	92	0.00
22	Acetophenone	40.000	38.830	2.9	90	-0.01
23 S	Nitrobenzene-d5	80.000	79.659	0.4	90	-0.01
24	Nitrobenzene	40.000	38.862	2.8	89	-0.01
25	Isophorone	40.000	37.210	7.0	85	0.00
26 C	2-Nitrophenol	40.000	45.166	-12.9	109	-0.01
27	2,4-Dimethylphenol	40.000	38.565	3.6	87	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.777	8.1	85	0.00
29 C	2,4-Dichlorophenol	40.000	41.376	-3.4	93	0.00
30	1,2,4-Trichlorobenzene	40.000	41.020	-2.6	95	0.00
31	Naphthalene	40.000	39.544	1.1	92	0.00
32	Benzoic acid	40.000	42.892	-7.2	108	0.02
33	4-Chloroaniline	40.000	35.227	11.9	81	0.00
34 C	Hexachlorobutadiene	40.000	43.408	-8.5	101	0.00
35	Caprolactam	40.000	32.955	17.6	74	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.496	8.8	83	0.00
37	2-Methylnaphthalene	40.000	36.947	7.6	86	-0.02
38	1-Methylnaphthalene	40.000	36.517	8.7	86	0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	81	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	44.180	-10.4	91	0.00
41 P	Hexachlorocyclopentadiene	40.000	52.363	-30.9#	103	-0.02
42 S	2,4,6-Tribromophenol	80.000	86.895	-8.6	84	0.00
43 C	2,4,6-Trichlorophenol	40.000	46.283	-15.7	91	0.00
44	2,4,5-Trichlorophenol	40.000	43.327	-8.3	85	0.00
45 S	2-Fluorobiphenyl	80.000	84.312	-5.4	88	0.00
46	1,1'-Biphenyl	40.000	41.457	-3.6	84	-0.01
47	2-Chloronaphthalene	40.000	41.681	-4.2	85	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060624\
 Data File : BP020537.D
 Acq On : 06 Jun 2024 11:30
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 06 13:04:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 03:41:03 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.772	-1.9	78	0.00
49	Acenaphthylene	40.000	39.992	0.0	82	0.00
50	Dimethylphthalate	40.000	37.224	6.9	77	0.00
51	2,6-Dinitrotoluene	40.000	39.493	1.3	78	0.00
52 C	Acenaphthene	40.000	39.800	0.5	83	0.00
53	3-Nitroaniline	40.000	35.808	10.5	70	0.01
54 P	2,4-Dinitrophenol	40.000	44.528	-11.3	100	0.00
55	Dibenzofuran	40.000	38.164	4.6	79	0.00
56 P	4-Nitrophenol	40.000	35.278	11.8	69	0.01
57	2,4-Dinitrotoluene	40.000	37.505	6.2	73	-0.01
58	Fluorene	40.000	37.630	5.9	79	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.741	-4.4	81	0.00
60	Diethylphthalate	40.000	35.038	12.4	72	-0.02
61	4-Chlorophenyl-phenylether	40.000	39.035	2.4	81	0.00
62	4-Nitroaniline	40.000	31.703	20.7	65	0.00
63	Azobenzene	40.000	34.451	13.9	71	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	67	-0.01
65	4,6-Dinitro-2-methylphenol	40.000	46.573	-16.4	88	0.00
66 c	n-Nitrosodiphenylamine	40.000	43.949	-9.9	75	-0.01
67	4-Bromophenyl-phenylether	40.000	45.246	-13.1	77	-0.02
68	Hexachlorobenzene	40.000	44.067	-10.2	75	0.00
69	Atrazine	40.000	36.882	7.8	62	-0.02
70 C	Pentachlorophenol	40.000	44.380	-11.0	74	0.00
71	Phenanthrene	40.000	40.514	-1.3	69	0.00
72	Anthracene	40.000	39.948	0.1	67	-0.02
73	Carbazole	40.000	36.464	8.8	62	-0.01
74	Di-n-butylphthalate	40.000	36.100	9.7	58	-0.02
75 C	Fluoranthene	40.000	35.788	10.5	60	-0.02
76 I	Chrysene-d12	20.000	20.000	0.0	71	-0.03
77	Benzydine	40.000	24.143	39.6#	36	0.00
78	Pyrene	40.000	35.128	12.2	60	-0.02
79 S	Terphenyl-d14	80.000	71.481	10.6	59	-0.01
80	Butylbenzylphthalate	40.000	35.695	10.8	62	-0.02
81	Benzo(a)anthracene	40.000	38.785	3.0	69	-0.03
82	3,3'-Dichlorobenzidine	40.000	48.438	-21.1	80	-0.03
83	Chrysene	40.000	39.945	0.1	70	-0.02
84	Bis(2-ethylhexyl)phthalate	40.000	38.294	4.3	62	-0.03
85 c	Di-n-octyl phthalate	40.000	48.786	-22.0#	85	-0.04
86 I	Perylene-d12	20.000	20.000	0.0	102	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	46.450	-16.1	112	-0.05
88	Benzo(b)fluoranthene	40.000	36.524	8.7	92	-0.02
89	Benzo(k)fluoranthene	40.000	36.969	7.6	94	-0.01
90 C	Benzo(a)pyrene	40.000	40.182	-0.5	100	-0.01
91	Dibenzo(a,h)anthracene	40.000	46.108	-15.3	112	-0.03
92	Benzo(g,h,i)perylene	40.000	45.988	-15.0	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060624\
Data File : BP020537.D
Acq On : 06 Jun 2024 11:30
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jun 06 13:04:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 03:41:03 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1