

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
 Data File : BP005741.D
 Acq On : 07 Jun 2021 14:22
 Operator : CG/JU
 Sample : M2580-18MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS

Quant Time: Jun 07 16:30:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 12:13:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.981	152	149637	20.000	ng	0.00
21) Naphthalene-d8	10.787	136	574072	20.000	ng	0.00
39) Acenaphthene-d10	14.604	164	347062	20.000	ng	0.00
64) Phenanthrene-d10	17.345	188	712580	20.000	ng	# 0.00
76) Chrysene-d12	21.410	240	657906	20.000	ng	# 0.00
86) Perylene-d12	23.845	264	644128	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.546	112	1180950	122.701	ng	0.01
7) Phenol-d6	7.152	99	1280883	97.623	ng	0.01
23) Nitrobenzene-d5	9.146	82	1035774	100.341	ng	0.00
42) 2,4,6-Tribromophenol	16.092	330	751418	168.844	ng	0.00
45) 2-Fluorobiphenyl	13.234	172	2264594	86.354	ng	0.00
79) Terphenyl-d14	19.886	244	3234418	92.250	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.417	88	173741	39.442	ng	# 40
3) Pyridine	3.828	79	319752	29.323	ng	94
4) n-Nitrosodimethylamine	3.728	42	215108	44.292	ng	# 26
6) Aniline	7.305	93	310624	20.252	ng	96
8) 2-Chlorophenol	7.546	128	506216	46.972	ng	96
9) Benzaldehyde	7.116	77	246684	43.425	ng	92
10) Phenol	7.175	94	489743	36.558	ng	98
11) bis(2-Chloroethyl)ether	7.399	93	452302	43.730	ng	96
12) 1,3-Dichlorobenzene	7.869	146	529723	43.303	ng	98
13) 1,4-Dichlorobenzene	8.016	146	541722	43.696	ng	98
14) 1,2-Dichlorobenzene	8.334	146	521638	43.995	ng	99
15) Benzyl Alcohol	8.222	79	445475	47.332	ng	# 84
16) 2,2'-oxybis(1-Chloropr...	8.510	45	435685	36.798	ng	88
17) 2-Methylphenol	8.428	107	389247	44.635	ng	96
18) Hexachloroethane	9.063	117	197510	44.894	ng	93
19) n-Nitroso-di-n-propyla...	8.793	70	343475	41.155	ng	98
20) 3+4-Methylphenols	8.757	107	513751	44.220	ng	94
22) Acetophenone	8.810	105	716191	46.771	ng	# 98
24) Nitrobenzene	9.187	77	535725	46.723	ng	90
25) Isophorone	9.716	82	931108	40.629	ng	98
26) 2-Nitrophenol	9.899	139	273975	61.241	ng	# 79
27) 2,4-Dimethylphenol	9.963	122	446141	50.718	ng	99
28) bis(2-Chloroethoxy)met...	10.193	93	565334	45.836	ng	98
29) 2,4-Dichlorophenol	10.434	162	477155	50.435	ng	99
30) 1,2,4-Trichlorobenzene	10.651	180	511082	45.609	ng	98
31) Naphthalene	10.840	128	1468636	42.960	ng	98
32) Benzoic acid	10.140	122	261601	53.334	ng	97
33) 4-Chloroaniline	10.946	127	119363	8.385	ng	# 90
34) Hexachlorobutadiene	11.122	225	333152	47.600	ng	98
35) Caprolactam	11.757	113	65061	25.112	ng	# 79
36) 4-Chloro-3-methylphenol	12.075	107	477412	47.940	ng	95
37) 2-Methylnaphthalene	12.440	142	1058763	43.708	ng	# 88
38) 1-Methylnaphthalene	12.657	142	977654	42.539	ng	# 94
40) 1,2,4,5-Tetrachloroben...	12.810	216	585811	50.599	ng	96
41) Hexachlorocyclopentadiene	12.787	237	657025	101.936	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
 Data File : BP005741.D
 Acq On : 07 Jun 2021 14:22
 Operator : CG/JU
 Sample : M2580-18MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS

Quant Time: Jun 07 16:30:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 12:13:53 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	13.045	196	392420	55.884	ng	93
44) 2,4,5-Trichlorophenol	13.116	196	424627	55.919	ng #	94
46) 1,1'-Biphenyl	13.440	154	1345048	47.729	ng	98
47) 2-Chloronaphthalene	13.487	162	1039486	45.379	ng	100
48) 2-Nitroaniline	13.687	65	296326	53.887	ng #	76
49) Acenaphthylene	14.328	152	1648717	46.176	ng	100
50) Dimethylphthalate	14.063	163	1425287	51.630	ng	99
51) 2,6-Dinitrotoluene	14.181	165	290343	47.660	ng #	66
52) Acenaphthene	14.669	154	959530	41.340	ng	97
53) 3-Nitroaniline	14.510	138	181804	31.504	ng	79
54) 2,4-Dinitrophenol	14.722	184	347097	132.681	ng	97
55) Dibenzofuran	14.998	168	1539045	43.261	ng	99
56) 4-Nitrophenol	14.828	139	435929	84.440	ng #	59
57) 2,4-Dinitrotoluene	14.969	165	398137	52.360	ng #	65
58) Fluorene	15.645	166	1205572	42.384	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.145	232	356304	54.232	ng	95
60) Diethylphthalate	15.416	149	1320141	47.662	ng	97
61) 4-Chlorophenyl-phenyle...	15.639	204	633535	44.550	ng	98
62) 4-Nitroaniline	15.669	138	292416	46.391	ng #	50
63) Azobenzene	15.934	77	1177242	40.588	ng	94
65) 4,6-Dinitro-2-methylph...	15.733	198	225606	57.569	ng	92
66) n-Nitrosodiphenylamine	15.857	169	1079468	44.457	ng #	92
67) 4-Bromophenyl-phenylether	16.533	248	425418	49.558	ng	98
68) Hexachlorobenzene	16.657	284	503354	49.246	ng	95
69) Atrazine	16.804	200	420503	60.600	ng	96
70) Pentachlorophenol	16.998	266	584484	114.016	ng	96
71) Phenanthrene	17.386	178	1932962	43.541	ng	100
72) Anthracene	17.480	178	1974550	45.046	ng	100
73) Carbazole	17.745	167	1776617	41.182	ng	99
74) Di-n-butylphthalate	18.286	149	2218349	48.018	ng #	92
75) Fluoranthene	19.345	202	2262855	42.462	ng	97
77) Benzidine	19.516	184	84558	6.768	ng	98
78) Pyrene	19.698	202	2297495	49.491	ng	99
80) Butylbenzylphthalate	20.557	149	963951	51.048	ng #	88
81) Benzo(a)anthracene	21.398	228	2117347	43.853	ng	99
82) 3,3'-Dichlorobenzidine	21.321	252	460386	29.520	ng #	97
83) Chrysene	21.451	228	2047579	43.795	ng	99
84) Bis(2-ethylhexyl)phtha...	21.310	149	1354832	51.663	ng #	95
85) Di-n-octyl phthalate	22.239	149	2262589	51.335	ng	98
87) Indeno(1,2,3-cd)pyrene	26.415	276	2253676	41.124	ng #	93
88) Benzo(b)fluoranthene	23.104	252	2114566	45.246	ng #	98
89) Benzo(k)fluoranthene	23.151	252	1945026	43.920	ng #	97
90) Benzo(a)pyrene	23.739	252	1927951	45.346	ng #	97
91) Dibenzo(a,h)anthracene	26.433	278	1911584	40.194	ng #	95
92) Benzo(g,h,i)perylene	27.203	276	1911219	41.664	ng #	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
 Data File : BP005741.D
 Acq On : 07 Jun 2021 14:22
 Operator : CG/JU
 Sample : M2580-18MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampled :
 18-B1(0-2)MS

Quant Time: Jun 07 16:30:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 12:13:53 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.981 min Scan# 832
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS


Tgt Ion:	152	Resp:	149637
Ion	Ratio	Lower	Upper
152	100		
150	156.7	119.8	179.8
115	60.1	49.0	73.4



1,4-Dioxane
 Concen: 39.442 ng
 RT: 3.417 min Scan# 56
 Delta R.T. 0.018 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:	88	Resp:	173741
Ion	Ratio	Lower	Upper
88	100		
58	61.2	42.5	63.7
43	107.3	19.8	29.8#









Pyridine
 Concen: 29.323 ng
 RT: 3.828 min Scan# 126
 Delta R.T. 0.012 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS


Tgt Ion:	79	Resp:	319752
Ion	Ratio	Lower	Upper
79	100		
52	53.2	40.3	60.5
51	29.5	20.0	30.0



n-Nitrosodimethylamine
 Concen: 44.292 ng
 RT: 3.728 min Scan# 109
 Delta R.T. 0.012 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:	42	Resp:	215108
Ion	Ratio	Lower	Upper
42	100		
74	134.0	122.6	183.8
44	5.3	146.9	220.3#









2-Fluorophenol
 Concen: 122.701 ng
 RT: 5.546 min Scan# 418
 Delta R.T. 0.012 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS


Tgt Ion:	112	Resp:	1180950
Ion	Ratio	Lower	Upper
112	100		
64	51.3	33.6	50.4#
63	28.6	18.6	27.8#



Aniline
 Concen: 20.252 ng
 RT: 7.305 min Scan# 717
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:	93	Resp:	310624
Ion	Ratio	Lower	Upper
93	100		
66	36.0	27.0	40.4
65	19.1	14.4	21.6









Phenol-d6
 Concen: 97.623 ng
 RT: 7.152 min Scan# 691
 Delta R.T. 0.012 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS

Tgt Ion: 99 Resp: 1280883




Ion	Ratio	Lower	Upper
99	100		
42	16.9	12.9	19.3
71	37.6	26.7	40.1




2-Chlorophenol
 Concen: 46.972 ng
 RT: 7.546 min Scan# 758
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion: 128 Resp: 506216

Ion	Ratio	Lower	Upper
128	100		
130	32.0	14.1	54.1
64	39.5	17.2	57.2









Benzaldehyde
 Concen: 43.425 ng
 RT: 7.116 min Scan# 685
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS

Tgt Ion:	77	Resp:	246684
Ion	Ratio	Lower	Upper
77	100		
105	98.3	89.4	129.4
106	94.4	79.9	119.9




Phenol
 Concen: 36.558 ng
 RT: 7.175 min Scan# 695
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:	94	Resp:	489743
Ion	Ratio	Lower	Upper
94	100		
65	28.2	8.0	48.0
66	39.7	17.3	57.3






Abundance Scan 735 (7.411 min): BP005627.D\data.ms (-729) #1




bis(2-Chloroethyl)ether
Concen: 43.730 ng
RT: 7.399 min Scan# 733
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	93	Resp:	452302
Ion	Ratio	Lower	Upper
93	100		
63	66.8	49.6	89.6
95	32.2	14.4	54.4







Abundance Scan 815 (7.881 min): BP005627.D\data.ms (-804) #1



1,3-Dichlorobenzene
Concen: 43.303 ng
RT: 7.869 min Scan# 813
Delta R.T. 0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	146	Resp:	529723
Ion	Ratio	Lower	Upper
146	100		
148	65.1	51.4	77.0
75	32.5	23.9	35.9









1,4-Dichlorobenzene
 Concen: 43.696 ng
 RT: 8.016 min Scan# 838
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS

Tgt Ion:	146	Resp:	541722
Ion	Ratio	Lower	Upper
146	100		
148	65.3	52.2	78.4
111	40.1	34.2	51.4




1,2-Dichlorobenzene
 Concen: 43.995 ng
 RT: 8.334 min Scan# 892
 Delta R.T. -0.000 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:	146	Resp:	521638
Ion	Ratio	Lower	Upper
146	100		
148	65.5	51.5	77.3
111	42.5	34.0	51.0






Abundance Scan 874 (8.228 min): BP005627.D\data.ms (-862) #15




Benzyl Alcohol
Concen: 47.332 ng
RT: 8.222 min Scan# 873
Delta R.T. 0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	79	Resp:	445475
Ion	Ratio	Lower	Upper
79	100		
108	76.1	64.0	96.0
77	66.6	72.7	109.1#







Abundance Scan 925 (8.528 min): BP005627.D\data.ms (-914) #16



2,2'-oxybis(1-Chloropropane)
Concen: 36.798 ng
RT: 8.510 min Scan# 922
Delta R.T. 0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	45	Resp:	435685
Ion	Ratio	Lower	Upper
45	100		
77	20.3	7.0	47.0
79	15.5	1.0	41.0









2-Methylphenol
 Concen: 44.635 ng
 RT: 8.428 min Scan# 908
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS


Tgt Ion:	107	Resp:	389247
Ion	Ratio	Lower	Upper
107	100		
108	112.3	85.8	128.8
77	48.4	41.4	62.0
79	47.2	38.6	57.8



Hexachloroethane
 Concen: 44.894 ng
 RT: 9.063 min Scan# 1016
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:	117	Resp:	197510
Ion	Ratio	Lower	Upper
117	100		
119	97.2	80.1	120.1
201	112.2	98.8	148.2









n-Nitroso-di-n-propylamine
 Concen: 41.155 ng
 RT: 8.793 min Scan# 970
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS


Tgt Ion:	70	Resp:	343475
Ion	Ratio	Lower	Upper
70	100		
42	50.6	39.7	59.5
101	10.1	7.4	11.0
130	23.0	17.2	25.8



3+4-Methylphenols
 Concen: 44.220 ng
 RT: 8.757 min Scan# 964
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:	107	Resp:	513751
Ion	Ratio	Lower	Upper
107	100		
108	90.9	69.5	109.5
77	35.1	21.5	61.5
79	30.5	16.8	56.8









Naphthalene-d8
Concen: 20.000 ng
RT: 10.787 min Scan# 1309
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22




Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	136	Resp:	574072
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.9	13.3
54	6.1	5.0	7.4
68	4.7	3.7	5.5




Acetophenone
Concen: 46.771 ng
RT: 8.810 min Scan# 973
Delta R.T. 0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	105	Resp:	716191
Ion	Ratio	Lower	Upper
105	100		
71	1.9	0.9	1.3#
51	22.8	18.2	27.2
120	23.3	17.0	25.6



Abundance Scan 1031 (9.152 min): BP005627.D\data.ms (-1031-)





Nitrobenzene-d5
Concen: 100.341 ng
RT: 9.146 min Scan# 1030
Delta R.T. 0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22


Tgt Ion:	82	Resp:	1035774
Ion	Ratio	Lower	Upper
82	100		
128	44.6	30.5	45.7
54	42.6	33.9	50.9

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS


Abundance Scan 1030 (9.146 min): BP005741.D\data.ms



Abundance Scan 1030 (9.146 min): BP005741.D\data.ms (-978) (-)





Abundance Scan 1038 (9.193 min): BP005627.D\data.ms (-1038-)




Nitrobenzene
Concen: 46.723 ng
RT: 9.187 min Scan# 1037
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	77	Resp:	535725
Ion	Ratio	Lower	Upper
77	100		
123	47.2	31.7	47.5
65	13.0	11.5	17.3


Abundance Scan 1037 (9.187 min): BP005741.D\data.ms



Abundance Scan 1037 (9.187 min): BP005741.D\data.ms (-986) (-)






Abundance Scan 1129 (9.728 min): BP005627.D\data.ms (-1125) (-)




Isophorone
Concen: 40.629 ng
RT: 9.716 min Scan# 1127
Delta R.T. 0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	82	Resp:	931108
Ion	Ratio	Lower	Upper
82	100		
95	7.8	7.5	11.3
138	18.8	14.6	22.0






Abundance Scan 1160 (9.911 min): BP005627.D\data.ms (-1140) (-)




2-Nitrophenol
Concen: 61.241 ng
RT: 9.899 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	139	Resp:	273975
Ion	Ratio	Lower	Upper
139	100		
109	33.2	40.7	61.1#
65	46.4	47.8	71.6#






Abundance Scan 1169 (9.963 min): BP005627.D\data.ms (-1167)




2,4-Dimethylphenol
Concen: 50.718 ng
RT: 9.963 min Scan# 1169
Delta R.T. 0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	122	Resp:	446141
Ion	Ratio	Lower	Upper
122	100		
107	117.6	93.8	140.8
121	59.0	45.6	68.4







Abundance Scan 1210 (10.205 min): BP005627.D\data.ms (-1203)



bis(2-Chloroethoxy)methane
Concen: 45.836 ng
RT: 10.193 min Scan# 1208
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	93	Resp:	565334
Ion	Ratio	Lower	Upper
93	100		
95	32.7	24.9	37.3
123	12.1	9.5	14.3









2,4-Dichlorophenol
Concen: 50.435 ng
RT: 10.434 min Scan# 1249
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22




Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS


Tgt Ion:	162	Resp:	477155
Ion	Ratio	Lower	Upper
162	100		
164	64.2	44.1	84.1
98	33.6	12.6	52.6



1,2,4-Trichlorobenzene
Concen: 45.609 ng
RT: 10.651 min Scan# 1286
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	180	Resp:	511082
Ion	Ratio	Lower	Upper
180	100		
182	98.1	77.2	115.8
145	29.8	21.7	32.5









Naphthalene
 Concen: 42.960 ng
 RT: 10.840 min Scan# 1318
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS

Tgt Ion:128 Resp: 1468636




Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.0	12.0
127	13.5	11.1	16.7




Benzoic acid
 Concen: 53.334 ng
 RT: 10.140 min Scan# 1199
 Delta R.T. 0.029 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:122 Resp: 261601

Ion	Ratio	Lower	Upper
122	100		
105	124.5	105.1	145.1
77	94.7	69.6	109.6









Scan 1337 (10.952 min): BP005627.D\data.ms (-1285) (-)
 4-Chloroaniline
 Concen: 8.385 ng
 RT: 10.946 min Scan# 1336
 Delta R.T. -0.000 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS


Tgt Ion:	127	Resp:	119363
Ion	Ratio	Lower	Upper
127	100		
129	31.4	25.5	38.3
65	28.5	30.0	45.0#
92	17.3	20.2	30.2#



Scan 1369 (11.140 min): BP005627.D\data.ms (-1315) (-)
 Hexachlorobutadiene
 Concen: 47.600 ng
 RT: 11.122 min Scan# 1366
 Delta R.T. -0.000 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:	225	Resp:	333152
Ion	Ratio	Lower	Upper
225	100		
223	62.0	51.1	76.7
227	64.2	50.2	75.2









Scan 1470 (11.734 min): BP005627.D\data.ms (-1435) (-)
 Caprolactam
 Concen: 25.112 ng
 RT: 11.757 min Scan# 1474
 Delta R.T. 0.023 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS


Tgt Ion:	113	Resp:	65061
Ion	Ratio	Lower	Upper
113	100		
55	121.8	72.0	112.0#
56	94.0	65.1	105.1



Scan 1527 (12.069 min): BP005627.D\data.ms (-1536) (-)
 4-Chloro-3-methylphenol
 Concen: 47.940 ng
 RT: 12.075 min Scan# 1528
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:	107	Resp:	477412
Ion	Ratio	Lower	Upper
107	100		
144	27.2	20.1	30.1
142	82.7	62.2	93.4









2-Methylnaphthalene
Concen: 43.708 ng
RT: 12.440 min Scan# 1590
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22




Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS


Tgt Ion:	142	Resp:	1058763
Ion	Ratio	Lower	Upper
142	100		
141	88.1	74.7	112.1
115	39.0	43.8	65.8#



1-Methylnaphthalene
Concen: 42.539 ng
RT: 12.657 min Scan# 1627
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	142	Resp:	977654
Ion	Ratio	Lower	Upper
142	100		
141	90.9	77.0	115.6
116	4.2	5.0	7.6#









Acenaphthene-d10
Concen: 20.000 ng
RT: 14.604 min Scan# 1958
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22




Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	164	Resp:	347062
Ion	Ratio	Lower	Upper
164	100		
162	100.8	73.5	110.3
160	44.7	40.6	61.0




1,2,4,5-Tetrachlorobenzene
Concen: 50.599 ng
RT: 12.810 min Scan# 1653
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	216	Resp:	585811
Ion	Ratio	Lower	Upper
216	100		
214	78.6	60.6	91.0
179	20.5	15.3	22.9
108	20.3	13.8	20.6





Abundance Scan 1651 (12.799 min): BP005627.D\data.ms (-1641) (-)




Hexachlorocyclopentadiene
Concen: 101.936 ng
RT: 12.787 min Scan# 1649
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS


Tgt Ion:	237	Resp:	657025
Ion	Ratio	Lower	Upper
237	100		
235	63.5	40.9	80.9
272	13.0	0.0	34.6



Abundance Scan 1649 (12.787 min): BP005741.D\data.ms (-1598) (-)





Abundance Scan 2211 (16.093 min): BP005627.D\data.ms (-2102) (-)





2,4,6-Tribromophenol
Concen: 168.844 ng
RT: 16.092 min Scan# 2211
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	330	Resp:	751418
Ion	Ratio	Lower	Upper
330	100		
332	97.2	76.3	114.5
141	30.3	22.8	34.2



Abundance Scan 2211 (16.092 min): BP005741.D\data.ms (-2160) (-)









2,4,6-Trichlorophenol
Concen: 55.884 ng
RT: 13.045 min Scan# 1693
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22




Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	196	Resp:	392420
Ion	Ratio	Lower	Upper
196	100		
198	95.1	81.4	122.2
200	30.2	28.2	42.2




2,4,5-Trichlorophenol
Concen: 55.919 ng
RT: 13.116 min Scan# 1705
Delta R.T. -0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	196	Resp:	424627
Ion	Ratio	Lower	Upper
196	100		
198	95.8	71.0	106.4
97	48.1	38.5	57.7
132	25.7	15.6	23.4#






Abundance Scan 1726 (13.240 min): BP005627.D\data.ms (-1745) (-)




2-Fluorobiphenyl
Concen: 86.354 ng
RT: 13.234 min Scan# 1725
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:172 Resp: 2264594
Ion Ratio Lower Upper
172 100
171 35.3 28.1 42.1
170 23.4 19.6 29.4






Abundance Scan 1762 (13.451 min): BP005627.D\data.ms (-1745) (-)




1,1'-Biphenyl
Concen: 47.729 ng
RT: 13.440 min Scan# 1760
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:154 Resp: 1345048
Ion Ratio Lower Upper
154 100
153 39.9 18.9 58.9
76 11.4 0.0 32.1






Abundance Scan 1769 (13.493 min): BP005627.D\data.ms (-1757) (-)




2-Chloronaphthalene
Concen: 45.379 ng
RT: 13.487 min Scan# 1768
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:162 Resp: 1039486
Ion Ratio Lower Upper
162 100
127 39.3 31.4 47.0
164 33.0 26.6 39.8







Abundance Scan 1803 (13.693 min): BP005627.D\data.ms (-1744) (-)



2-Nitroaniline
Concen: 53.887 ng
RT: 13.687 min Scan# 1802
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion: 65 Resp: 296326
Ion Ratio Lower Upper
65 100
92 70.9 48.8 73.2
138 116.5 69.5 104.3#









Acenaphthylene
 Concen: 46.176 ng
 RT: 14.328 min Scan# 1911
 Delta R.T. -0.000 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS

Tgt Ion:	152	Resp:	1648717
Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.8	23.6
153	13.1	10.5	15.7




Dimethylphthalate
 Concen: 51.630 ng
 RT: 14.063 min Scan# 1866
 Delta R.T. 0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:	163	Resp:	1425287
Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.9	5.9
164	10.2	8.1	12.1



Abundance Scan 1887 (14.187 min): BP005627.D\data.ms (-1835) (-)





2,6-Dinitrotoluene
Concen: 47.660 ng
RT: 14.181 min Scan# 1886
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22


Tgt Ion:	165	Resp:	290343
Ion	Ratio	Lower	Upper
165	100		
63	53.4	67.4	101.0#
89	59.1	73.8	110.6#

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS


Abundance Scan 1886 (14.181 min): BP005741.D\data.ms



Abundance Scan 1886 (14.181 min): BP005741.D\data.ms (-1835) (-)





Abundance Scan 1970 (14.675 min): BP005627.D\data.ms (-1952) (-)




Acenaphthene
Concen: 41.340 ng
RT: 14.669 min Scan# 1969
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	154	Resp:	959530
Ion	Ratio	Lower	Upper
154	100		
153	113.1	90.9	136.3
152	53.8	48.5	72.7


Abundance Scan 1969 (14.669 min): BP005741.D\data.ms



Abundance Scan 1969 (14.669 min): BP005741.D\data.ms (-1918) (-)



Abundance Scan 1942 (14.510 min): BP005627.D\data.ms (-1953) (-)





3-Nitroaniline
Concen: 31.504 ng
RT: 14.510 min Scan# 1942
Delta R.T. 0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22


Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	138	Resp:	181804
Ion	Ratio	Lower	Upper
138	100		
108	10.6	10.2	15.4
92	108.7	108.1	162.1


Abundance Scan 1942 (14.510 min): BP005741.D\data.ms



Abundance Scan 1942 (14.510 min): BP005741.D\data.ms (-1890) (-)





Abundance Scan 1976 (14.710 min): BP005627.D\data.ms (-1956) (-)




2,4-Dinitrophenol
Concen: 132.681 ng
RT: 14.722 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	184	Resp:	347097
Ion	Ratio	Lower	Upper
184	100		
63	62.1	51.2	76.8
154	67.1	55.8	83.8


Abundance Scan 1978 (14.722 min): BP005741.D\data.ms



Abundance Scan 1978 (14.722 min): BP005741.D\data.ms (-1927) (-)



Abundance Scan 2027 (15.010 min): BP005627.D\data.ms (-1974) (-)





Dibenzenofuran
Concen: 43.261 ng
RT: 14.998 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22


Tgt Ion:	168	Resp:	1539045
Ion	Ratio	Lower	Upper
168	100		
139	39.7	31.4	47.0
169	13.5	11.4	17.0

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS


Abundance Scan 2025 (14.998 min): BP005741.D\data.ms



Abundance Scan 2025 (14.998 min): BP005741.D\data.ms (-1974) (-)





Abundance Scan 1992 (14.804 min): BP005627.D\data.ms (-1956) (-)




4-Nitrophenol
Concen: 84.440 ng
RT: 14.828 min Scan# 1996
Delta R.T. 0.012 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22


Tgt Ion:	139	Resp:	435929
Ion	Ratio	Lower	Upper
139	100		
109	84.6	122.8	162.8#
65	89.8	110.0	150.0#

Abundance Scan 1996 (14.828 min): BP005741.D\data.ms



Abundance Scan 1996 (14.828 min): BP005741.D\data.ms (-1943) (-)









2,4-Dinitrotoluene
Concen: 52.360 ng
RT: 14.969 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22




Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	165	Resp:	398137
Ion	Ratio	Lower	Upper
165	100		
63	41.0	56.7	85.1#
89	71.2	85.8	128.6#
182	2.9	3.0	4.6#




Fluorene
Concen: 42.384 ng
RT: 15.645 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	166	Resp:	1205572
Ion	Ratio	Lower	Upper
166	100		
165	97.1	78.8	118.2
167	13.2	10.9	16.3






Abundance Scan 2064 (15.228 min): BP005627.D\data.ms (-2059) (-)




2,3,4,6-Tetrachlorophenol
Concen: 54.232 ng
RT: 15.145 min Scan# 2050
Delta R.T. -0.082 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	232	Resp:	356304
Ion	Ratio	Lower	Upper
232	100		
131	38.3	26.5	39.7
130	2.5	2.1	3.1
166	23.9	19.4	29.0






Abundance Scan 2098 (15.428 min): BP005627.D\data.ms (-2060) (-)




Diethylphthalate
Concen: 47.662 ng
RT: 15.416 min Scan# 2096
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	149	Resp:	1320141
Ion	Ratio	Lower	Upper
149	100		
177	22.1	19.2	28.8
150	12.5	9.9	14.9






Abundance Scan 2136 (15.651 min): BP005627.D\data.ms (-2158) (-)




4-Chlorophenyl-phenylether
Concen: 44.550 ng
RT: 15.639 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	204	Resp:	633535
Ion	Ratio	Lower	Upper
204	100		
206	32.3	26.0	39.0
141	57.6	44.2	66.2







Abundance Scan 2139 (15.669 min): BP005627.D\data.ms (-2159) (-)



4-Nitroaniline
Concen: 46.391 ng
RT: 15.669 min Scan# 2139
Delta R.T. 0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	138	Resp:	292416
Ion	Ratio	Lower	Upper
138	100		
92	49.6	50.9	90.9#
108	90.7	154.7	194.7#









Azobenzene
Concen: 40.588 ng
RT: 15.934 min Scan# 2184
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion: 77 Resp: 1177242




Ion	Ratio	Lower	Upper
77	100		
182	28.0	5.5	45.5
105	19.8	0.0	36.5
51	27.6	10.3	50.3




Phenanthrene-d10
Concen: 20.000 ng
RT: 17.345 min Scan# 2424
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion: 188 Resp: 712580

Ion	Ratio	Lower	Upper
188	100		
94	7.8	5.0	7.6#
80	8.6	5.9	8.9



Abundance Scan 2149 (15.728 min): BP005627.D\data.ms (-2149) (-)





4,6-Dinitro-2-methylphenol
Concen: 57.569 ng
RT: 15.733 min Scan# 2150
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22


Tgt Ion:	198	Resp:	225606
Ion	Ratio	Lower	Upper
198	100		
51	34.5	10.3	50.3
105	38.3	13.6	53.6

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS


Abundance Scan 2150 (15.733 min): BP005741.D\data.ms



Abundance Scan 2150 (15.733 min): BP005741.D\data.ms (-2099) (-)





Abundance Scan 2172 (15.863 min): BP005627.D\data.ms (-2168) (-)




n-Nitrosodiphenylamine
Concen: 44.457 ng
RT: 15.857 min Scan# 2171
Delta R.T. 0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	169	Resp:	1079468
Ion	Ratio	Lower	Upper
169	100		
168	66.2	55.2	82.8
167	36.3	36.6	54.8#


Abundance Scan 2171 (15.857 min): BP005741.D\data.ms



Abundance Scan 2171 (15.857 min): BP005741.D\data.ms (-2119) (-)






Abundance Scan 2288 (16.545 min): BP005627.D\data.ms (-2256) (-)




4-Bromophenyl-phenylether
Concen: 49.558 ng
RT: 16.533 min Scan# 2286
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:248 Resp: 425418
Ion Ratio Lower Upper
248 100
250 96.9 77.5 116.3
141 67.3 50.1 75.1






Abundance Scan 2307 (16.657 min): BP005627.D\data.ms (-2256) (-)




Hexachlorobenzene
Concen: 49.246 ng
RT: 16.657 min Scan# 2307
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:284 Resp: 503354
Ion Ratio Lower Upper
284 100
142 30.7 20.8 31.2
249 30.3 23.7 35.5



Abundance Scan 2333 (16.810 min): BP005627.D\data.ms (-2359) (-)





Atrazine
Concen: 60.600 ng
RT: 16.804 min Scan# 2332
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22


Tgt Ion:	200	Resp:	420503
Ion	Ratio	Lower	Upper
200	100		
173	25.1	0.7	40.7
215	47.8	29.0	69.0

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS


Abundance Scan 2332 (16.804 min): BP005741.D\data.ms



Abundance Scan 2332 (16.804 min): BP005741.D\data.ms (-2281) (-)





Abundance Scan 2365 (16.998 min): BP005627.D\data.ms (-2378) (-)




Pentachlorophenol
Concen: 114.016 ng
RT: 16.998 min Scan# 2365
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	266	Resp:	584484
Ion	Ratio	Lower	Upper
266	100		
268	62.6	49.4	74.0
264	62.3	53.7	80.5


Abundance Scan 2365 (16.998 min): BP005741.D\data.ms



Abundance Scan 2365 (16.998 min): BP005741.D\data.ms (-2314) (-)






Abundance Scan 2432 (17.392 min): BP005627.D\data.ms (-2472) (-)




Phenanthrene
Concen: 43.541 ng
RT: 17.386 min Scan# 2431
Delta R.T. -0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:178 Resp: 1932962
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.5 12.3 18.5







Abundance Scan 2448 (17.487 min): BP005627.D\data.ms (-2472) (-)



Anthracene
Concen: 45.046 ng
RT: 17.480 min Scan# 2447
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:178 Resp: 1974550
Ion Ratio Lower Upper
178 100
176 18.8 15.2 22.8
179 15.6 12.6 19.0









Carbazole
 Concen: 41.182 ng
 RT: 17.745 min Scan# 2492
 Delta R.T. -0.000 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS

Tgt Ion:	167	Resp:	1776617
Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.4	26.0
139	13.1	10.4	15.6




Di-n-butylphthalate
 Concen: 48.018 ng
 RT: 18.286 min Scan# 2584
 Delta R.T. -0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:	149	Resp:	2218349
Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.6	11.4
104	5.5	8.9	13.3#






Abundance Scan 2765 (19.351 min): BP005627.D\data.ms (-2765) (-)




Fluoranthene
Concen: 42.462 ng
RT: 19.345 min Scan# 2764
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	202	Resp:	2262855
Ion	Ratio	Lower	Upper
202	100		
101	8.6	0.0	26.1
203	17.5	0.0	37.2






Abundance Scan 3117 (21.422 min): BP005627.D\data.ms (-3105) (-)




Chrysene-d12
Concen: 20.000 ng
RT: 21.410 min Scan# 3115
Delta R.T. -0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	240	Resp:	657906
Ion	Ratio	Lower	Upper
240	100		
120	9.0	4.2	6.2#
236	25.8	20.2	30.2






Abundance Scan 2795 (19.528 min): BP005627.D\data.ms (-2795) (-)




Benzidine
Concen: 6.768 ng
RT: 19.516 min Scan# 2793
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	184	Resp:	84558
Ion	Ratio	Lower	Upper
184	100		
185	13.9	11.1	16.7
183	10.5	9.8	14.8






Abundance Scan 2824 (19.698 min): BP005627.D\data.ms (-2824) (-)




Pyrene
Concen: 49.491 ng
RT: 19.698 min Scan# 2824
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	202	Resp:	2297495
Ion	Ratio	Lower	Upper
202	100		
200	21.4	16.9	25.3
203	17.6	13.8	20.6






Abundance Scan 2858 (19.898 min): BP005627.D\data.ms (-2879) (-)




Terphenyl-d14
Concen: 92.250 ng
RT: 19.886 min Scan# 2856
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:244 Resp: 3234418
Ion Ratio Lower Upper
244 100
212 7.6 5.8 8.6
122 8.7 5.2 7.8#






Abundance Scan 2972 (20.569 min): BP005627.D\data.ms (-2980) (-)




Butylbenzylphthalate
Concen: 51.048 ng
RT: 20.557 min Scan# 2970
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:149 Resp: 963951
Ion Ratio Lower Upper
149 100
91 64.8 60.2 90.4
206 22.9 23.9 35.9#






Abundance Scan 3114 (21.404 min): BP005627.D\data.ms (-3061) (-)




Benzo(a)anthracene
Concen: 43.853 ng
RT: 21.398 min Scan# 3113
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:228 Resp: 2117347
Ion Ratio Lower Upper
228 100
226 27.4 21.2 31.8
229 19.9 15.6 23.4






Abundance Scan 3102 (21.333 min): BP005627.D\data.ms (-3062) (-)




3,3'-Dichlorobenzidine
Concen: 29.520 ng
RT: 21.321 min Scan# 3100
Delta R.T. -0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:252 Resp: 460386
Ion Ratio Lower Upper
252 100
254 64.8 49.9 74.9
126 9.3 5.5 8.3#






Abundance Scan 3123 (21.457 min): BP005627.D\data.ms (-3123) (-)




Chrysene
Concen: 43.795 ng
RT: 21.451 min Scan# 3122
Delta R.T. -0.000 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:228 Resp: 2047579
Ion Ratio Lower Upper
228 100
226 30.6 24.8 37.2
229 19.8 15.4 23.0






Abundance Scan 3101 (21.328 min): BP005627.D\data.ms (-3101) (-)




Bis(2-ethylhexyl)phthalate
Concen: 51.663 ng
RT: 21.310 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:149 Resp: 1354832
Ion Ratio Lower Upper
149 100
167 27.4 23.8 35.8
279 5.2 5.7 8.5#






Abundance Scan 3260 (22.263 min): BP005627.D\data.ms (-3260) (-)




Di-n-octyl phthalate
Concen: 51.335 ng
RT: 22.239 min Scan# 3256
Delta R.T. -0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	149	Resp:	2262589
Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.5	2.3
43	7.3	5.1	7.7






Abundance Scan 3532 (23.863 min): BP005627.D\data.ms (-3532) (-)




Perylene-d12
Concen: 20.000 ng
RT: 23.845 min Scan# 3529
Delta R.T. -0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	264	Resp:	644128
Ion	Ratio	Lower	Upper
264	100		
260	24.3	20.6	30.8
265	21.8	19.0	28.6






Abundance Scan 3969 (26.433 min): BP005627.D\data.ms (-3969) (-)




Indeno(1,2,3-cd)pyrene
Concen: 41.124 ng
RT: 26.415 min Scan# 3966
Delta R.T. -0.012 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:	276	Resp:	2253676
Ion	Ratio	Lower	Upper
276	100		
138	19.4	8.4	12.6#
277	25.7	20.6	31.0







Abundance Scan 3405 (23.116 min): BP005627.D\data.ms (-3405) (-)



Benzo(b)fluoranthene
Concen: 45.246 ng
RT: 23.104 min Scan# 3403
Delta R.T. -0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:	252	Resp:	2114566
Ion	Ratio	Lower	Upper
252	100		
253	22.5	18.2	27.2
125	8.3	3.4	5.0#









Scan 3414 (23.169 min): BP005627.D\data.ms (-3409) (-)
 Benzo(k)fluoranthene
 Concen: 43.920 ng
 RT: 23.151 min Scan# 3411
 Delta R.T. -0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22




Instrument :
 BNA_P
 ClientSampleId :
 18-B1(0-2)MS

Tgt Ion:	252	Resp:	1945026
Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.2	25.8
125	8.0	3.8	5.6#




Scan 3514 (23.757 min): BP005627.D\data.ms (-3509) (-)
 Benzo(a)pyrene
 Concen: 45.346 ng
 RT: 23.739 min Scan# 3511
 Delta R.T. -0.006 min
 Lab File: BP005741.D
 Acq: 07 Jun 2021 14:22

Tgt Ion:	252	Resp:	1927951
Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.1	25.7
125	8.7	3.8	5.6#






Abundance Scan 3972 (26.451 min): BP005627.D\data.ms (-3991) (-)




Dibenzo(a,h)anthracene
Concen: 40.194 ng
RT: 26.433 min Scan# 3969
Delta R.T. -0.006 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Instrument :
BNA_P
ClientSampleId :
18-B1(0-2)MS

Tgt Ion:278 Resp: 1911584
Ion Ratio Lower Upper
278 100
139 12.7 5.0 7.6#
279 24.0 18.6 28.0



Abundance Scan 4104 (27.227 min): BP005627.D\data.ms (-4092) (-)



Benzo(g,h,i)perylene
Concen: 41.664 ng
RT: 27.203 min Scan# 4100
Delta R.T. -0.012 min
Lab File: BP005741.D
Acq: 07 Jun 2021 14:22

Tgt Ion:276 Resp: 1911219
Ion Ratio Lower Upper
276 100
277 24.2 19.8 29.8
138 16.9 7.7 11.5#

