

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
 Data File : BP005757.D
 Acq On : 08 Jun 2021 01:14
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Jun 08 03:23:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 12:13:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	126009	20.00	ng	0.00
21) Naphthalene-d8	10.775	136	502662	20.00	ng	-0.01
39) Acenaphthene-d10	14.592	164	335328	20.00	ng	-0.01
64) Phenanthrene-d10	17.339	188	692454	20.00	ng	0.00
76) Chrysene-d12	21.404	240	786606	20.00	ng	#-0.01
86) Perylene-d12	23.827	264	929113	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	635798	78.45	ng	0.00
7) Phenol-d6	7.134	99	799855	72.39	ng	0.00
23) Nitrobenzene-d5	9.140	82	755011	83.53	ng	0.00
42) 2,4,6-Tribromophenol	16.080	330	447572	104.09	ng	-0.01
45) 2-Fluorobiphenyl	13.228	172	1737521	68.57	ng	0.00
79) Terphenyl-d14	19.880	244	2955429	70.50	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.399	88	133898	36.10	ng	92
3) Pyridine	3.810	79	357306	38.91	ng	# 93
4) n-Nitrosodimethylamine	3.716	42	160017	39.13	ng	# 24
6) Aniline	7.293	93	445552	34.50	ng	94
8) 2-Chlorophenol	7.534	128	349241	38.48	ng	96
9) Benzaldehyde	7.104	77	179032	37.43	ng	92
10) Phenol	7.163	94	399970	35.46	ng	96
11) bis(2-Chloroethyl)ether	7.393	93	296371	34.03	ng	97
12) 1,3-Dichlorobenzene	7.857	146	378941	36.79	ng	99
13) 1,4-Dichlorobenzene	8.004	146	386032	36.98	ng	97
14) 1,2-Dichlorobenzene	8.328	146	374393	37.50	ng	98
15) Benzyl Alcohol	8.210	79	311297	39.28	ng	# 82
16) 2,2'-oxybis(1-Chloropr...	8.504	45	287627	28.85	ng	89
17) 2-Methylphenol	8.416	107	281799	38.37	ng	97
18) Hexachloroethane	9.051	117	140745	37.99	ng	92
19) n-Nitroso-di-n-propyla...	8.787	70	252642	35.95	ng	98
20) 3+4-Methylphenols	8.745	107	386537	39.51	ng	95
22) Acetophenone	8.798	105	480538	35.84	ng	# 98
24) Nitrobenzene	9.175	77	391710	39.02	ng	92
25) Isophorone	9.710	82	715985	35.68	ng	97
26) 2-Nitrophenol	9.892	139	197731	51.21	ng	# 80
27) 2,4-Dimethylphenol	9.951	122	292104	37.92	ng	99
28) bis(2-Chloroethoxy)met...	10.187	93	390032	36.12	ng	98
29) 2,4-Dichlorophenol	10.428	162	364576	44.01	ng	99
30) 1,2,4-Trichlorobenzene	10.639	180	391620	39.91	ng	98
31) Naphthalene	10.828	128	1108783	37.04	ng	97
32) Benzoic acid	10.092	122	106406	27.46	ng	94
33) 4-Chloroaniline	10.934	127	475655	38.16	ng	# 91
34) Hexachlorobutadiene	11.116	225	255491	41.69	ng	98
35) Caprolactam	11.739	113	115564	50.94	ng	# 73
36) 4-Chloro-3-methylphenol	12.057	107	380544	43.64	ng	95
37) 2-Methylnaphthalene	12.433	142	743883	35.07	ng	# 88
38) 1-Methylnaphthalene	12.651	142	714454	35.50	ng	# 97
40) 1,2,4,5-Tetrachloroben...	12.798	216	400948	35.84	ng	97
41) Hexachlorocyclopentadiene	12.775	237	179460	28.82	ng	97
43) 2,4,6-Trichlorophenol	13.039	196	292336	43.09	ng	94

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
 Data File : BP005757.D
 Acq On : 08 Jun 2021 01:14
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Quant Time: Jun 08 03:23:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 12:13:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.110	196	311823	42.50	ng	# 92
46) 1,1'-Biphenyl	13.433	154	930357	34.17	ng	98
47) 2-Chloronaphthalene	13.475	162	770416	34.81	ng	99
48) 2-Nitroaniline	13.680	65	228197	43.86	ng	# 69
49) Acenaphthylene	14.316	152	1280787	37.13	ng	99
50) Dimethylphthalate	14.057	163	1008102	37.80	ng	99
51) 2,6-Dinitrotoluene	14.175	165	239867	41.49	ng	# 65
52) Acenaphthene	14.657	154	768048	34.25	ng	96
53) 3-Nitroaniline	14.498	138	253304	43.82	ng	80
54) 2,4-Dinitrophenol	14.710	184	29643	17.68	ng	98
55) Dibenzofuran	14.992	168	1267943	36.89	ng	99
56) 4-Nitrophenol	14.810	139	204388	42.90	ng	# 59
57) 2,4-Dinitrotoluene	14.957	165	339768	46.75	ng	# 67
58) Fluorene	15.639	166	978389	35.60	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.139	232	301410	47.48	ng	94
60) Diethylphthalate	15.410	149	1049599	39.22	ng	98
61) 4-Chlorophenyl-phenyle...	15.627	204	492583	35.85	ng	96
62) 4-Nitroaniline	15.657	138	274551	45.18	ng	# 53
63) Azobenzene	15.922	77	993773	35.46	ng	95
65) 4,6-Dinitro-2-methylph...	15.722	198	89297	30.46	ng	91
66) n-Nitrosodiphenylamine	15.845	169	913887	38.73	ng	# 92
67) 4-Bromophenyl-phenylether	16.522	248	320440	38.41	ng	95
68) Hexachlorobenzene	16.645	284	383296	38.59	ng	# 94
69) Atrazine	16.798	200	289076	42.87	ng	96
70) Pentachlorophenol	16.986	266	213260	42.81	ng	93
71) Phenanthrene	17.380	178	1515127	35.12	ng	100
72) Anthracene	17.469	178	1509634	35.44	ng	99
73) Carbazole	17.739	167	1490355	35.55	ng	99
74) Di-n-butylphthalate	18.280	149	1670712	37.22	ng	# 92
75) Fluoranthene	19.333	202	1922565	37.13	ng	98
77) Benzidine	19.504	184	686330	45.94	ng	99
78) Pyrene	19.686	202	1974726	35.58	ng	99
80) Butylbenzylphthalate	20.545	149	842119	38.41	ng	# 91
81) Benzo(a)anthracene	21.386	228	2104198	36.45	ng	99
82) 3,3'-Dichlorobenzidine	21.315	252	776455	41.64	ng	# 97
83) Chrysene	21.439	228	2031599	36.34	ng	99
84) Bis(2-ethylhexyl)phtha...	21.304	149	1174262	37.45	ng	# 96
85) Di-n-octyl phthalate	22.227	149	2197960	41.71	ng	98
87) Indeno(1,2,3-cd)pyrene	26.392	276	2663475	33.69	ng	# 94
88) Benzo(b)fluoranthene	23.086	252	2379856	35.30	ng	# 97
89) Benzo(k)fluoranthene	23.139	252	2281048	35.71	ng	# 98
90) Benzo(a)pyrene	23.727	252	2228013	36.33	ng	# 97
91) Dibenzo(a,h)anthracene	26.409	278	2310709	33.68	ng	# 95
92) Benzo(g,h,i)perylene	27.180	276	2196092	33.19	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 08 03:23:13 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 07 12:13:53 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.969 min Scan# 830
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 126009
Ion Ratio Lower Upper
152 100
150 153.5 119.8 179.8
115 58.4 49.0 73.4



#2
1,4-Dioxane
Concen: 36.10 ng
RT: 3.399 min Scan# 53
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 88 Resp: 133898
Ion Ratio Lower Upper
88 100
58 60.7 42.5 63.7
43 26.2 19.8 29.8









#3
Pyridine
Concen: 38.91 ng
RT: 3.810 min Scan# 123
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 357306
Ion Ratio Lower Upper
79 100
52 53.6 40.3 60.5
51 30.6 20.0 30.0#



#4
n-Nitrosodimethylamine
Concen: 39.13 ng
RT: 3.716 min Scan# 107
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 42 Resp: 160017
Ion Ratio Lower Upper
42 100
74 130.2 122.6 183.8
44 5.6 146.9 220.3#









#5
2-Fluorophenol
Concen: 78.45 ng
RT: 5.528 min Scan# 415
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 635798
Ion Ratio Lower Upper
112 100
64 51.2 33.6 50.4#
63 29.0 18.6 27.8#



#6
Aniline
Concen: 34.50 ng
RT: 7.293 min Scan# 715
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 93 Resp: 445552
Ion Ratio Lower Upper
93 100
66 38.3 27.0 40.4
65 18.7 14.4 21.6









#7
Phenol-d6
Concen: 72.39 ng
RT: 7.134 min Scan# 688
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 99 Resp: 799855
Ion Ratio Lower Upper
99 100
42 17.0 12.9 19.3
71 38.6 26.7 40.1



#8
2-Chlorophenol
Concen: 38.48 ng
RT: 7.534 min Scan# 756
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 128 Resp: 349241
Ion Ratio Lower Upper
128 100
130 32.0 14.1 54.1
64 39.4 17.2 57.2









#9
Benzaldehyde
Concen: 37.43 ng
RT: 7.104 min Scan# 683
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 179032
Ion Ratio Lower Upper
77 100
105 98.5 89.4 129.4
106 94.9 79.9 119.9



#10
Phenol
Concen: 35.46 ng
RT: 7.163 min Scan# 693
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 94 Resp: 399970
Ion Ratio Lower Upper
94 100
65 28.6 8.0 48.0
66 40.7 17.3 57.3









#11
bis(2-Chloroethyl)ether
Concen: 34.03 ng
RT: 7.393 min Scan# 732
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 296371
Ion Ratio Lower Upper
93 100
63 66.9 49.6 89.6
95 32.3 14.4 54.4



#12
1,3-Dichlorobenzene
Concen: 36.79 ng
RT: 7.857 min Scan# 811
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:146 Resp: 378941
Ion Ratio Lower Upper
146 100
148 64.2 51.4 77.0
75 31.8 23.9 35.9





#13

1,4-Dichlorobenzene

Concen: 36.98 ng

RT: 8.004 min Scan# 836

Delta R.T. -0.006 min

Lab File: BP005757.D

Acq: 08 Jun 2021 01:14





Tgt Ion:146 Resp: 386032

Ion Ratio Lower Upper

146 100

148 64.5 52.2 78.4

111 39.4 34.2 51.4



#14

1,2-Dichlorobenzene

Concen: 37.50 ng

RT: 8.328 min Scan# 891

Delta R.T. -0.006 min

Lab File: BP005757.D

Acq: 08 Jun 2021 01:14




Tgt Ion:146 Resp: 374393


Ion Ratio Lower Upper

146 100

148 65.8 51.5 77.3

111 40.9 34.0 51.0





#15

Benzyl Alcohol

Concen: 39.28 ng

RT: 8.210 min Scan# 871

Delta R.T. -0.006 min

Lab File: BP005757.D

Acq: 08 Jun 2021 01:14

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040





Tgt Ion: 79 Resp: 311297

Ion Ratio Lower Upper

79 100

108 74.3 64.0 96.0

77 64.3 72.7 109.1#



#16

2,2'-oxybis(1-Chloropropane)

Concen: 28.85 ng

RT: 8.504 min Scan# 921

Delta R.T. -0.000 min

Lab File: BP005757.D

Acq: 08 Jun 2021 01:14




Tgt Ion: 45 Resp: 287627


Ion Ratio Lower Upper

45 100

77 20.7 7.0 47.0

79 16.5 1.0 41.0





#17

2-Methylphenol

Concen: 38.37 ng

RT: 8.416 min Scan# 906

Delta R.T. -0.006 min

Lab File: BP005757.D

Acq: 08 Jun 2021 01:14

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040

Tgt Ion:107 Resp: 281799





Ion Ratio Lower Upper

107 100

108 110.8 85.8 128.8

77 49.6 41.4 62.0

79 47.9 38.6 57.8



#18

Hexachloroethane

Concen: 37.99 ng

RT: 9.051 min Scan# 1014

Delta R.T. -0.006 min

Lab File: BP005757.D

Acq: 08 Jun 2021 01:14




Tgt Ion:117 Resp: 140745


Ion Ratio Lower Upper

117 100

119 97.8 80.1 120.1

201 109.4 98.8 148.2









#19
n-Nitroso-di-n-propylamine
Concen: 35.95 ng
RT: 8.787 min Scan# 969
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 252642




Ion	Ratio	Lower	Upper
70	100		
42	49.9	39.7	59.5
101	10.6	7.4	11.0
130	24.4	17.2	25.8




#20
3+4-Methylphenols
Concen: 39.51 ng
RT: 8.745 min Scan# 962
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 107 Resp: 386537

Ion	Ratio	Lower	Upper
107	100		
108	90.0	69.5	109.5
77	34.8	21.5	61.5
79	30.7	16.8	56.8









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.775 min Scan# 1307
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	136	Resp:	502662
Ion Ratio	Lower	Upper	
136	100		
137	10.7	8.9	13.3
54	5.7	5.0	7.4
68	5.1	3.7	5.5



#22
Acetophenone
Concen: 35.84 ng
RT: 8.798 min Scan# 971
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	105	Resp:	480538
Ion Ratio	Lower	Upper	
105	100		
71	3.1	0.9	1.3
51	22.6	18.2	27.2
120	22.9	17.0	25.6









#23
Nitrobenzene-d5
Concen: 83.53 ng
RT: 9.140 min Scan# 1029
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 755011
Ion Ratio Lower Upper
82 100
128 44.7 30.5 45.7
54 41.7 33.9 50.9



#24
Nitrobenzene
Concen: 39.02 ng
RT: 9.175 min Scan# 1035
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 77 Resp: 391710
Ion Ratio Lower Upper
77 100
123 45.6 31.7 47.5
65 13.1 11.5 17.3









#25
Isophorone
Concen: 35.68 ng
RT: 9.710 min Scan# 1126
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	82	Resp:	715985
Ion	Ratio	Lower	Upper
82	100		
95	8.4	7.5	11.3
138	19.5	14.6	22.0



#26
2-Nitrophenol
Concen: 51.21 ng
RT: 9.892 min Scan# 1157
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	139	Resp:	197731
Ion	Ratio	Lower	Upper
139	100		
109	35.2	40.7	61.1#
65	45.7	47.8	71.6#









#27
2,4-Dimethylphenol
Concen: 37.92 ng
RT: 9.951 min Scan# 1167
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:122 Resp: 292104
Ion Ratio Lower Upper
122 100
107 116.5 93.8 140.8
121 56.4 45.6 68.4



#28
bis(2-Chloroethoxy)methane
Concen: 36.12 ng
RT: 10.187 min Scan# 1207
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 93 Resp: 390032
Ion Ratio Lower Upper
93 100
95 32.2 24.9 37.3
123 12.1 9.5 14.3









#29
2,4-Dichlorophenol
Concen: 44.01 ng
RT: 10.428 min Scan# 1248
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 364576
Ion Ratio Lower Upper
162 100
164 64.3 44.1 84.1
98 33.8 12.6 52.6



#30
1,2,4-Trichlorobenzene
Concen: 39.91 ng
RT: 10.639 min Scan# 1284
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:180 Resp: 391620
Ion Ratio Lower Upper
180 100
182 94.9 77.2 115.8
145 28.7 21.7 32.5









#31
Naphthalene
Concen: 37.04 ng
RT: 10.828 min Scan# 1316
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 1108783
Ion Ratio Lower Upper
128 100
129 11.0 8.0 12.0
127 12.9 11.1 16.7



#32
Benzoic acid
Concen: 27.46 ng
RT: 10.092 min Scan# 1191
Delta R.T. -0.018 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:122 Resp: 106406
Ion Ratio Lower Upper
122 100
105 118.9 105.1 145.1
77 95.8 69.6 109.6









#33
4-Chloroaniline
Concen: 38.16 ng
RT: 10.934 min Scan# 1334
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:127	Resp: 475655
Ion Ratio	Lower Upper
127 100	
129 32.4	25.5 38.3
65 29.5	30.0 45.0#
92 19.3	20.2 30.2#



#34
Hexachlorobutadiene
Concen: 41.69 ng
RT: 11.116 min Scan# 1365
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt	Ion:225	Resp:	255491
Ion	Ratio	Lower	Upper
225	100		
223	61.8	51.1	76.7
227	64.3	50.2	75.2









#35
Caprolactam
Concen: 50.94 ng
RT: 11.739 min Scan# 1471
Delta R.T. 0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:113 Resp: 115564
Ion Ratio Lower Upper
113 100
55 126.8 72.0 112.0#
56 101.0 65.1 105.1



#36
4-Chloro-3-methylphenol
Concen: 43.64 ng
RT: 12.057 min Scan# 1525
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:107 Resp: 380544
Ion Ratio Lower Upper
107 100
144 26.9 20.1 30.1
142 82.6 62.2 93.4









#37
2-Methylnaphthalene
Concen: 35.07 ng
RT: 12.433 min Scan# 1589
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 743883
Ion Ratio Lower Upper
142 100
141 87.9 74.7 112.1
115 38.6 43.8 65.8#



#38
1-Methylnaphthalene
Concen: 35.50 ng
RT: 12.651 min Scan# 1626
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:142 Resp: 714454
Ion Ratio Lower Upper
142 100
141 93.8 77.0 115.6
116 4.2 5.0 7.6#









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.592 min Scan# 1956
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 335328
Ion Ratio Lower Upper
164 100
162 101.3 73.5 110.3
160 44.4 40.6 61.0



#40
1,2,4,5-Tetrachlorobenzene
Concen: 35.84 ng
RT: 12.798 min Scan# 1651
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:216 Resp: 400948
Ion Ratio Lower Upper
216 100
214 78.0 60.6 91.0
179 20.8 15.3 22.9
108 17.8 13.8 20.6









#41
Hexachlorocyclopentadiene
Concen: 28.82 ng
RT: 12.775 min Scan# 1647
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	237	Resp:	179460
Ion Ratio	Lower	Upper	
237	100		
235	62.8	40.9	80.9
272	13.4	0.0	34.6



#42
2,4,6-Tribromophenol
Concen: 104.09 ng
RT: 16.080 min Scan# 2209
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	330	Resp:	447572
Ion Ratio	Lower	Upper	
330	100		
332	97.4	76.3	114.5
141	30.4	22.8	34.2









#43
2,4,6-Trichlorophenol
Concen: 43.09 ng
RT: 13.039 min Scan# 1692
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 292336
Ion Ratio Lower Upper
196 100
198 95.6 81.4 122.2
200 31.4 28.2 42.2



#44
2,4,5-Trichlorophenol
Concen: 42.50 ng
RT: 13.110 min Scan# 1704
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:196 Resp: 311823
Ion Ratio Lower Upper
196 100
198 97.7 71.0 106.4
97 48.7 38.5 57.7
132 26.5 15.6 23.4#









#45
2-Fluorobiphenyl
Concen: 68.57 ng
RT: 13.228 min Scan# 1724
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:172 Resp: 1737521




Ion	Ratio	Lower	Upper
172	100		
171	34.8	28.1	42.1
170	23.3	19.6	29.4




#46
1,1'-Biphenyl
Concen: 34.17 ng
RT: 13.433 min Scan# 1759
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:154 Resp: 930357

Ion	Ratio	Lower	Upper
154	100		
153	40.1	18.9	58.9
76	11.1	0.0	32.1









#47
2-Chloronaphthalene
Concen: 34.81 ng
RT: 13.475 min Scan# 1766
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162	Resp:	770416
Ion Ratio	Lower	Upper
162	100	
127	38.5	31.4 47.0
164	32.2	26.6 39.8



#48
2-Nitroaniline
Concen: 43.86 ng
RT: 13.680 min Scan# 1801
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion: 65	Resp:	228197
Ion Ratio	Lower	Upper
65	100	
92	74.8	48.8 73.2#
138	123.2	69.5 104.3#









#49
Acenaphthylene
Concen: 37.13 ng
RT: 14.316 min Scan# 1909
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 1280787
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.6
153 13.2 10.5 15.7



#50
Dimethylphthalate
Concen: 37.80 ng
RT: 14.057 min Scan# 1865
Delta R.T. -0.000 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:163 Resp: 1008102
Ion Ratio Lower Upper
163 100
194 4.4 3.9 5.9
164 10.2 8.1 12.1









#51
2,6-Dinitrotoluene
Concen: 41.49 ng
RT: 14.175 min Scan# 1885
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 239867
Ion Ratio Lower Upper
165 100
63 53.0 67.4 101.0#
89 58.6 73.8 110.6#



#52
Acenaphthene
Concen: 34.25 ng
RT: 14.657 min Scan# 1967
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:154 Resp: 768048
Ion Ratio Lower Upper
154 100
153 111.6 90.9 136.3
152 53.5 48.5 72.7









#53
3-Nitroaniline
Concen: 43.82 ng
RT: 14.498 min Scan# 1940
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 253304
Ion Ratio Lower Upper
138 100
108 10.5 10.2 15.4
92 110.1 108.1 162.1



#54
2,4-Dinitrophenol
Concen: 17.68 ng
RT: 14.710 min Scan# 1976
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:184 Resp: 29643
Ion Ratio Lower Upper
184 100
63 67.6 51.2 76.8
154 69.5 55.8 83.8









#55
Dibenzofuran
Concen: 36.89 ng
RT: 14.992 min Scan# 2024
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 1267943
Ion Ratio Lower Upper
168 100
139 39.5 31.4 47.0
169 13.6 11.4 17.0



#56
4-Nitrophenol
Concen: 42.90 ng
RT: 14.810 min Scan# 1993
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:139 Resp: 204388
Ion Ratio Lower Upper
139 100
109 84.0 122.8 162.8#
65 89.7 110.0 150.0#









#57
2,4-Dinitrotoluene
Concen: 46.75 ng
RT: 14.957 min Scan# 2018
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 339768
Ion Ratio Lower Upper
165 100
63 42.0 56.7 85.1#
89 72.3 85.8 128.6#
182 2.7 3.0 4.6#



#58
Fluorene
Concen: 35.60 ng
RT: 15.639 min Scan# 2134
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:166 Resp: 978389
Ion Ratio Lower Upper
166 100
165 98.4 78.8 118.2
167 13.3 10.9 16.3









#59
2,3,4,6-Tetrachlorophenol
Concen: 47.48 ng
RT: 15.139 min Scan# 2049
Delta R.T. -0.088 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:232	Resp:	301410
Ion Ratio	Lower	Upper
232	100	
131	38.7	26.5 39.7
130	2.6	2.1 3.1
166	23.7	19.4 29.0



#60
Diethylphthalate
Concen: 39.22 ng
RT: 15.410 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:149	Resp:	1049599
Ion Ratio	Lower	Upper
149	100	
177	22.5	19.2 28.8
150	12.8	9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 35.85 ng
RT: 15.627 min Scan# 2132
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 492583
Ion Ratio Lower Upper
204 100
206 33.0 26.0 39.0
141 59.9 44.2 66.2



#62
4-Nitroaniline
Concen: 45.18 ng
RT: 15.657 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:138 Resp: 274551
Ion Ratio Lower Upper
138 100
92 51.5 50.9 90.9
108 95.3 154.7 194.7#









#63
Azobenzene
Concen: 35.46 ng
RT: 15.922 min Scan# 2182
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 77	Resp: 993773
Ion Ratio	Lower Upper
77 100	
182 27.2	5.5 45.5
105 20.0	0.0 36.5
51 28.1	10.3 50.3



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.339 min Scan# 2423
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:188	Resp: 692454
Ion Ratio	Lower Upper
188 100	
94 7.1	5.0 7.6
80 7.6	5.9 8.9









#65
4,6-Dinitro-2-methylphenol
Concen: 30.46 ng
RT: 15.722 min Scan# 2148
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 89297
Ion Ratio Lower Upper
198 100
51 33.5 10.3 50.3
105 39.9 13.6 53.6



#66
n-Nitrosodiphenylamine
Concen: 38.73 ng
RT: 15.845 min Scan# 2169
Delta R.T. -0.006 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:169 Resp: 913887
Ion Ratio Lower Upper
169 100
168 66.1 55.2 82.8
167 35.9 36.6 54.8#









#67
4-Bromophenyl-phenylether
Concen: 38.41 ng
RT: 16.522 min Scan# 2284
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 320440
Ion Ratio Lower Upper
248 100
250 96.8 77.5 116.3
141 72.5 50.1 75.1



#68
Hexachlorobenzene
Concen: 38.59 ng
RT: 16.645 min Scan# 2305
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:284 Resp: 383296
Ion Ratio Lower Upper
284 100
142 31.9 20.8 31.2#
249 30.5 23.7 35.5





#69

Atrazine

Concen: 42.87 ng

RT: 16.798 min Scan# 2331

Delta R.T. -0.006 min

Lab File: BP005757.D

Acq: 08 Jun 2021 01:14

Tgt Ion:200 Resp: 289076

Ion Ratio Lower Upper

200 100

173 25.2 0.7 40.7





215 49.8 29.0 69.0

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040



#70

Pentachlorophenol

Concen: 42.81 ng

RT: 16.986 min Scan# 2363

Delta R.T. -0.012 min

Lab File: BP005757.D

Acq: 08 Jun 2021 01:14




Tgt Ion:266 Resp: 213260


Ion Ratio Lower Upper

266 100

268 63.7 49.4 74.0

264 59.0 53.7 80.5









#71
Phenanthrene
Concen: 35.12 ng
RT: 17.380 min Scan# 2430
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 1515127
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.4 12.3 18.5



#72
Anthracene
Concen: 35.44 ng
RT: 17.469 min Scan# 2445
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:178 Resp: 1509634
Ion Ratio Lower Upper
178 100
176 18.8 15.2 22.8
179 15.5 12.6 19.0









#73
 Carbazole
 Concen: 35.55 ng
 RT: 17.739 min Scan# 2491
 Delta R.T. -0.006 min
 Lab File: BP005757.D
 Acq: 08 Jun 2021 01:14




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:167 Resp: 1490355
 Ion Ratio Lower Upper
 167 100
 166 20.9 17.4 26.0
 139 13.1 10.4 15.6



#74
 Di-n-butylphthalate
 Concen: 37.22 ng
 RT: 18.280 min Scan# 2583
 Delta R.T. -0.012 min
 Lab File: BP005757.D
 Acq: 08 Jun 2021 01:14

Tgt Ion:149 Resp: 1670712
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.6 11.4
 104 5.5 8.9 13.3#









#75
Fluoranthene
Concen: 37.13 ng
RT: 19.333 min Scan# 2762
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	202	Resp:	1922565
Ion	Ratio	Lower	Upper
202	100		
101	8.3	0.0	26.1
203	17.7	0.0	37.2



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.404 min Scan# 3114
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:	240	Resp:	786606
Ion	Ratio	Lower	Upper
240	100		
120	8.2	4.2	6.2#
236	26.1	20.2	30.2









#77
Benzidine
Concen: 45.94 ng
RT: 19.504 min Scan# 2791
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 686330
Ion Ratio Lower Upper
184 100
185 14.0 11.1 16.7
183 11.8 9.8 14.8



#78
Pyrene
Concen: 35.58 ng
RT: 19.686 min Scan# 2822
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:202 Resp: 1974726
Ion Ratio Lower Upper
202 100
200 20.9 16.9 25.3
203 17.6 13.8 20.6









#79
 Terphenyl-d14
 Concen: 70.50 ng
 RT: 19.880 min Scan# 2855
 Delta R.T. -0.006 min
 Lab File: BP005757.D
 Acq: 08 Jun 2021 01:14




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:244 Resp: 2955429
 Ion Ratio Lower Upper
 244 100
 212 7.6 5.8 8.6
 122 8.1 5.2 7.8#



#80
 Butylbenzylphthalate
 Concen: 38.41 ng
 RT: 20.545 min Scan# 2968
 Delta R.T. -0.012 min
 Lab File: BP005757.D
 Acq: 08 Jun 2021 01:14

Tgt Ion:149 Resp: 842119
 Ion Ratio Lower Upper
 149 100
 91 69.0 60.2 90.4
 206 22.2 23.9 35.9#









#81
Benzo(a)anthracene
Concen: 36.45 ng
RT: 21.386 min Scan# 3111
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 2104198
Ion Ratio Lower Upper
228 100
226 27.4 21.2 31.8
229 19.8 15.6 23.4



#82
3,3'-Dichlorobenzidine
Concen: 41.64 ng
RT: 21.315 min Scan# 3099
Delta R.T. -0.012 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:252 Resp: 776455
Ion Ratio Lower Upper
252 100
254 64.3 49.9 74.9
126 9.3 5.5 8.3#





#83

Chrysene

Concen: 36.34 ng

RT: 21.439 min Scan# 3120





Delta R.T. -0.012 min

Lab File: BP005757.D

Acq: 08 Jun 2021 01:14

Tgt Ion:228 Resp: 2031599

Ion	Ratio	Lower	Upper
228	100		
226	30.3	24.8	37.2
229	19.8	15.4	23.0



#84

Bis(2-ethylhexyl)phthalate

Concen: 37.45 ng

RT: 21.304 min Scan# 3097




Delta R.T. -0.012 min


Lab File: BP005757.D

Acq: 08 Jun 2021 01:14

Tgt Ion:149 Resp: 1174262

Ion	Ratio	Lower	Upper
149	100		
167	27.5	23.8	35.8
279	5.4	5.7	8.5#









#85
Di-n-octyl phthalate
Concen: 41.71 ng
RT: 22.227 min Scan# 3254
Delta R.T. -0.018 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 2197960
Ion Ratio Lower Upper
149 100
167 1.6 1.5 2.3
43 7.1 5.1 7.7



#86
Perylene-d12
Concen: 20.00 ng
RT: 23.827 min Scan# 3526
Delta R.T. -0.024 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:264 Resp: 929113
Ion Ratio Lower Upper
264 100
260 24.3 20.6 30.8
265 21.4 19.0 28.6









#87
 Indeno(1,2,3-cd)pyrene
 Concen: 33.69 ng
 RT: 26.392 min Scan# 3962
 Delta R.T. -0.036 min
 Lab File: BP005757.D
 Acq: 08 Jun 2021 01:14




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:276 Resp: 2663475
 Ion Ratio Lower Upper
 276 100
 138 18.4 8.4 12.6#
 277 25.6 20.6 31.0



#88
 Benzo(b)fluoranthene
 Concen: 35.30 ng
 RT: 23.086 min Scan# 3400
 Delta R.T. -0.024 min
 Lab File: BP005757.D
 Acq: 08 Jun 2021 01:14

Tgt Ion:252 Resp: 2379856
 Ion Ratio Lower Upper
 252 100
 253 21.9 18.2 27.2
 125 7.4 3.4 5.0#









#89
Benzo(k)fluoranthene
Concen: 35.71 ng
RT: 23.139 min Scan# 3409
Delta R.T. -0.018 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 2281048
Ion Ratio Lower Upper
252 100
253 21.9 17.2 25.8
125 7.2 3.8 5.6#



#90
Benzo(a)pyrene
Concen: 36.33 ng
RT: 23.727 min Scan# 3509
Delta R.T. -0.018 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:252 Resp: 2228013
Ion Ratio Lower Upper
252 100
253 21.9 17.1 25.7
125 7.9 3.8 5.6#









#91
Dibenzo(a,h)anthracene
Concen: 33.68 ng
RT: 26.409 min Scan# 3965
Delta R.T. -0.030 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 2310709
Ion Ratio Lower Upper
278 100
139 11.9 5.0 7.6#
279 24.2 18.6 28.0



#92
Benzo(g,h,i)perylene
Concen: 33.19 ng
RT: 27.180 min Scan# 4096
Delta R.T. -0.036 min
Lab File: BP005757.D
Acq: 08 Jun 2021 01:14

Tgt Ion:276 Resp: 2196092
Ion Ratio Lower Upper
276 100
277 23.5 19.8 29.8
138 15.9 7.7 11.5#

