

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
 Data File : BP005745.D
 Acq On : 07 Jun 2021 16:39
 Operator : CG/JU
 Sample : M2580-21
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 18-B2(0-2)

Quant Time: Jun 07 17:23:14 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 12:13:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	132209	20.000	ng	0.00
21) Naphthalene-d8	10.781	136	512614	20.000	ng	0.00
39) Acenaphthene-d10	14.598	164	334145	20.000	ng	0.00
64) Phenanthrene-d10	17.339	188	731438	20.000	ng	# 0.00
76) Chrysene-d12	21.404	240	850491	20.000	ng	#-0.01
86) Perylene-d12	23.839	264	963524	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.540	112	1112379	130.811	ng	0.00
7) Phenol-d6	7.146	99	1212816	104.620	ng	0.00
23) Nitrobenzene-d5	9.140	82	987939	107.182	ng	0.00
42) 2,4,6-Tribromophenol	16.086	330	813353	189.826	ng	0.00
45) 2-Fluorobiphenyl	13.228	172	2197682	87.042	ng	0.00
79) Terphenyl-d14	19.886	244	3798476	83.806	ng	0.00
Target Compounds						
50) Dimethylphthalate	14.051	163	138887	5.226	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
Data File : BP005745.D
Acq On : 07 Jun 2021 16:39
Operator : CG/JU
Sample : M2580-21
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
18-B2(0-2)

Quant Time: Jun 07 17:23:14 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 07 12:13:53 2021
Response via : Initial Calibration

