

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
 Data File : BP005749.D
 Acq On : 07 Jun 2021 18:57
 Operator : CG/JU
 Sample : M2580-25
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 18-B3(2-4)

Quant Time: Jun 08 01:15:44 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 12:13:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	114864	20.000	ng	0.00
21) Naphthalene-d8	10.781	136	449688	20.000	ng	0.00
39) Acenaphthene-d10	14.598	164	299422	20.000	ng	0.00
64) Phenanthrene-d10	17.339	188	689357	20.000	ng	# 0.00
76) Chrysene-d12	21.404	240	843401	20.000	ng	#-0.01
86) Perylene-d12	23.833	264	988561	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.540	112	1024614	138.685	ng	0.00
7) Phenol-d6	7.140	99	1170086	116.176	ng	0.00
23) Nitrobenzene-d5	9.140	82	901349	111.471	ng	0.00
42) 2,4,6-Tribromophenol	16.086	330	800934	208.605	ng	0.00
45) 2-Fluorobiphenyl	13.228	172	2043077	90.302	ng	0.00
79) Terphenyl-d14	19.880	244	3783714	84.182	ng	0.00
Target Compounds						
50) Dimethylphthalate	14.051	163	167795	7.045	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
Data File : BP005749.D
Acq On : 07 Jun 2021 18:57
Operator : CG/JU
Sample : M2580-25
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
18-B3(2-4)

Quant Time: Jun 08 01:15:44 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 07 12:13:53 2021
Response via : Initial Calibration

