

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
 Data File : BP005751.D
 Acq On : 07 Jun 2021 20:06
 Operator : CG/JU
 Sample : M2580-27
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 18-B4(0-2)

Quant Time: Jun 08 01:16:12 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 12:13:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	145070	20.000	ng	0.00
21) Naphthalene-d8	10.781	136	556245	20.000	ng	# 0.00
39) Acenaphthene-d10	14.598	164	354300	20.000	ng	0.00
64) Phenanthrene-d10	17.340	188	787809	20.000	ng	# 0.00
76) Chrysene-d12	21.398	240	909697	20.000	ng	#-0.02
86) Perylene-d12	23.833	264	1053145	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.540	112	1168172	125.194	ng	0.00
7) Phenol-d6	7.140	99	1273497	100.115	ng	0.00
23) Nitrobenzene-d5	9.140	82	1064055	106.385	ng	0.00
42) 2,4,6-Tribromophenol	16.081	330	848362	186.734	ng	-0.01
45) 2-Fluorobiphenyl	13.228	172	2344519	87.575	ng	0.00
79) Terphenyl-d14	19.881	244	4024072	83.005	ng	0.00
Target Compounds						
50) Dimethylphthalate	14.051	163	132730	4.710	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
Data File : BP005751.D
Acq On : 07 Jun 2021 20:06
Operator : CG/JU
Sample : M2580-27
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
18-B4(0-2)

Quant Time: Jun 08 01:16:12 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 07 12:13:53 2021
Response via : Initial Calibration

