

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
 Data File : BP005736.D
 Acq On : 07 Jun 2021 11:03
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Jun 07 12:14:39 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 12:13:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	99587	20.000	ng	0.00
21) Naphthalene-d8	10.787	136	389140	20.000	ng	0.00
39) Acenaphthene-d10	14.604	164	283253	20.000	ng	0.00
64) Phenanthrene-d10	17.345	188	611793	20.000	ng	0.00
76) Chrysene-d12	21.416	240	725668	20.000	ng	# 0.00
86) Perylene-d12	23.851	264	877209	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.534	112	491013	76.656	ng	0.00
7) Phenol-d6	7.140	99	658070	75.362	ng	0.00
23) Nitrobenzene-d5	9.140	82	589527	84.252	ng	0.00
42) 2,4,6-Tribromophenol	16.092	330	441757	121.625	ng	0.00
45) 2-Fluorobiphenyl	13.234	172	1602040	74.851	ng	0.00
79) Terphenyl-d14	19.886	244	2889667	74.721	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.399	88	101425	34.597	ng	92
3) Pyridine	3.817	79	268028	36.933	ng	95
4) n-Nitrosodimethylamine	3.717	42	112694	34.866	ng	# 28
6) Aniline	7.299	93	369096	36.158	ng	95
8) 2-Chlorophenol	7.540	128	281019	39.181	ng	96
9) Benzaldehyde	7.111	77	131690	34.833	ng	91
10) Phenol	7.169	94	329773	36.989	ng	98
11) bis(2-Chloroethyl)ether	7.399	93	248594	36.114	ng	94
12) 1,3-Dichlorobenzene	7.863	146	303389	37.266	ng	99
13) 1,4-Dichlorobenzene	8.010	146	309151	37.469	ng	96
14) 1,2-Dichlorobenzene	8.334	146	298590	37.839	ng	98
15) Benzyl Alcohol	8.216	79	245636	39.216	ng	# 83
16) 2,2'-oxybis(1-Chloropr...	8.505	45	244763	31.062	ng	86
17) 2-Methylphenol	8.422	107	222048	38.259	ng	97
18) Hexachloroethane	9.057	117	111938	38.230	ng	90
19) n-Nitroso-di-n-propyla...	8.787	70	204626	36.840	ng	98
20) 3+4-Methylphenols	8.752	107	304283	39.353	ng	94
22) Acetophenone	8.805	105	381834	36.786	ng	# 98
24) Nitrobenzene	9.187	77	308927	39.747	ng	# 90
25) Isophorone	9.710	82	572139	36.830	ng	98
26) 2-Nitrophenol	9.899	139	151340	50.675	ng	# 78
27) 2,4-Dimethylphenol	9.957	122	227025	38.074	ng	98
28) bis(2-Chloroethoxy)met...	10.193	93	296998	35.523	ng	99
29) 2,4-Dichlorophenol	10.434	162	269701	42.055	ng	97
30) 1,2,4-Trichlorobenzene	10.652	180	288279	37.952	ng	99
31) Naphthalene	10.834	128	844548	36.444	ng	98
32) Benzoic acid	10.110	122	133042	41.266	ng	97
33) 4-Chloroaniline	10.946	127	357827	37.083	ng	# 90
34) Hexachlorobutadiene	11.122	225	191039	40.267	ng	98
35) Caprolactam	11.734	113	90322	51.429	ng	# 76
36) 4-Chloro-3-methylphenol	12.069	107	296125	43.868	ng	92
37) 2-Methylnaphthalene	12.440	142	680159	41.423	ng	# 88
38) 1-Methylnaphthalene	12.657	142	637484	40.920	ng	# 94
40) 1,2,4,5-Tetrachloroben...	12.810	216	380593	40.279	ng	98
41) Hexachlorocyclopentadiene	12.787	237	181792	34.558	ng	99
43) 2,4,6-Trichlorophenol	13.045	196	262050	45.725	ng	94


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP060721\
 Data File : BP005736.D
 Acq On : 07 Jun 2021 11:03
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Jun 07 12:14:39 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP052721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 12:13:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.122	196	283843	45.800	ng	# 90
46) 1,1'-Biphenyl	13.440	154	829314	36.057	ng	99
47) 2-Chloronaphthalene	13.487	162	686789	36.736	ng	99
48) 2-Nitroaniline	13.687	65	175156	40.265	ng	# 69
49) Acenaphthylene	14.328	152	1061730	36.435	ng	99
50) Dimethylphthalate	14.057	163	856136	38.000	ng	99
51) 2,6-Dinitrotoluene	14.181	165	195719	40.245	ng	# 59
52) Acenaphthene	14.669	154	639811	33.775	ng	97
53) 3-Nitroaniline	14.504	138	196035	40.450	ng	80
54) 2,4-Dinitrophenol	14.722	184	94618	48.663	ng	89
55) Dibenzofuran	14.998	168	1068141	36.788	ng	98
56) 4-Nitrophenol	14.816	139	157673	39.507	ng	# 57
57) 2,4-Dinitrotoluene	14.969	165	268885	44.063	ng	# 58
58) Fluorene	15.651	166	840583	36.210	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.228	232	255565	47.662	ng	94
60) Diethylphthalate	15.416	149	837819	37.063	ng	99
61) 4-Chlorophenyl-phenyle...	15.639	204	449079	38.693	ng	98
62) 4-Nitroaniline	15.663	138	208119	40.920	ng	# 48
63) Azobenzene	15.934	77	739161	31.225	ng	91
65) 4,6-Dinitro-2-methylph...	15.734	198	162923	51.141	ng	97
66) n-Nitrosodiphenylamine	15.851	169	754179	36.177	ng	# 91
67) 4-Bromophenyl-phenylether	16.534	248	314742	42.705	ng	97
68) Hexachlorobenzene	16.657	284	387622	44.171	ng	99
69) Atrazine	16.804	200	256067	42.982	ng	98
70) Pentachlorophenol	16.998	266	186360	42.342	ng	94
71) Phenanthrene	17.392	178	1364695	35.805	ng	100
72) Anthracene	17.481	178	1364490	36.257	ng	100
73) Carbazole	17.745	167	1298392	35.055	ng	100
74) Di-n-butylphthalate	18.292	149	1463846	36.906	ng	# 91
75) Fluoranthene	19.345	202	1733518	37.888	ng	97
77) Benzidine	19.516	184	679073	49.275	ng	99
78) Pyrene	19.698	202	1796625	35.087	ng	99
80) Butylbenzylphthalate	20.557	149	717666	35.799	ng	# 88
81) Benzo(a)anthracene	21.398	228	1941353	36.453	ng	99
82) 3,3'-Dichlorobenzidine	21.327	252	725526	42.177	ng	96
83) Chrysene	21.451	228	1858689	36.042	ng	98
84) Bis(2-ethylhexyl)phtha...	21.316	149	1024263	35.410	ng	99
85) Di-n-octyl phthalate	22.245	149	1882502	38.723	ng	99
87) Indeno(1,2,3-cd)pyrene	26.427	276	2752216	36.877	ng	# 95
88) Benzo(b)fluoranthene	23.110	252	2268533	35.643	ng	# 97
89) Benzo(k)fluoranthene	23.157	252	2049115	33.976	ng	# 98
90) Benzo(a)pyrene	23.745	252	2062563	35.622	ng	# 97
91) Dibenzo(a,h)anthracene	26.439	278	2380332	36.751	ng	# 96
92) Benzo(g,h,i)perylene	27.215	276	2316250	37.077	ng	# 95





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.975 min Scan# 831
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 99587
Ion Ratio Lower Upper
152 100
150 155.9 119.8 179.8
115 60.4 49.0 73.4



#2
1,4-Dioxane
Concen: 34.597 ng
RT: 3.399 min Scan# 53
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion: 88 Resp: 101425
Ion Ratio Lower Upper
88 100
58 60.6 42.5 63.7
43 25.5 19.8 29.8









#3
Pyridine
Concen: 36.933 ng
RT: 3.817 min Scan# 124
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 268028
Ion Ratio Lower Upper
79 100
52 52.7 40.3 60.5
51 28.8 20.0 30.0



#4
n-Nitrosodimethylamine
Concen: 34.866 ng
RT: 3.717 min Scan# 107
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion: 42 Resp: 112694
Ion Ratio Lower Upper
42 100
74 138.5 122.6 183.8
44 6.3 146.9 220.3









#5
2-Fluorophenol
Concen: 76.656 ng
RT: 5.534 min Scan# 416
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 491013
Ion Ratio Lower Upper
112 100
64 51.5 33.6 50.4#
63 28.1 18.6 27.8#



#6
Aniline
Concen: 36.158 ng
RT: 7.299 min Scan# 716
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion: 93 Resp: 369096
Ion Ratio Lower Upper
93 100
66 37.6 27.0 40.4
65 18.5 14.4 21.6









#7
Phenol-d6
Concen: 75.362 ng
RT: 7.140 min Scan# 689
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 658070




Ion	Ratio	Lower	Upper
99	100		
42	16.7	12.9	19.3
71	36.1	26.7	40.1




#8
2-Chlorophenol
Concen: 39.181 ng
RT: 7.540 min Scan# 757
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion: 128 Resp: 281019

Ion	Ratio	Lower	Upper
128	100		
130	32.6	14.1	54.1
64	40.8	17.2	57.2









#9
Benzaldehyde
Concen: 34.833 ng
RT: 7.111 min Scan# 684
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 131690
Ion Ratio Lower Upper
77 100
105 98.4 89.4 129.4
106 92.1 79.9 119.9



#10
Phenol
Concen: 36.989 ng
RT: 7.169 min Scan# 694
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion: 94 Resp: 329773
Ion Ratio Lower Upper
94 100
65 27.8 8.0 48.0
66 38.9 17.3 57.3









#11
bis(2-Chloroethyl)ether
Concen: 36.114 ng
RT: 7.399 min Scan# 733
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 248594
Ion Ratio Lower Upper
93 100
63 64.4 49.6 89.6
95 31.5 14.4 54.4



#12
1,3-Dichlorobenzene
Concen: 37.266 ng
RT: 7.863 min Scan# 812
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:146 Resp: 303389
Ion Ratio Lower Upper
146 100
148 64.0 51.4 77.0
75 31.9 23.9 35.9









#13
1,4-Dichlorobenzene
Concen: 37.469 ng
RT: 8.010 min Scan# 837
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 309151
Ion Ratio Lower Upper
146 100
148 63.0 52.2 78.4
111 39.9 34.2 51.4



#14
1,2-Dichlorobenzene
Concen: 37.839 ng
RT: 8.334 min Scan# 892
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:146 Resp: 298590
Ion Ratio Lower Upper
146 100
148 62.9 51.5 77.3
111 40.6 34.0 51.0









#15
Benzyl Alcohol
Concen: 39.216 ng
RT: 8.216 min Scan# 872
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 245636
Ion Ratio Lower Upper
79 100
108 75.4 64.0 96.0
77 65.5 72.7 109.1#



#16
2,2'-oxybis(1-Chloropropane)
Concen: 31.062 ng
RT: 8.505 min Scan# 921
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion: 45 Resp: 244763
Ion Ratio Lower Upper
45 100
77 20.2 7.0 47.0
79 14.4 1.0 41.0









#17
2-Methylphenol
Concen: 38.259 ng
RT: 8.422 min Scan# 907
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	107	Resp:	222048
Ion	Ratio	Lower	Upper
107	100		
108	112.3	85.8	128.8
77	49.9	41.4	62.0
79	48.1	38.6	57.8



#18
Hexachloroethane
Concen: 38.230 ng
RT: 9.057 min Scan# 1015
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:	117	Resp:	111938
Ion	Ratio	Lower	Upper
117	100		
119	94.8	80.1	120.1
201	107.6	98.8	148.2









#19
n-Nitroso-di-n-propylamine
Concen: 36.840 ng
RT: 8.787 min Scan# 969
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 70 Resp: 204626
Ion Ratio Lower Upper
70 100
42 50.2 39.7 59.5
101 10.4 7.4 11.0
130 24.0 17.2 25.8



#20
3+4-Methylphenols
Concen: 39.353 ng
RT: 8.752 min Scan# 963
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion: 107 Resp: 304283
Ion Ratio Lower Upper
107 100
108 90.8 69.5 109.5
77 34.4 21.5 61.5
79 30.4 16.8 56.8









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.787 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 389140




Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.9	13.3
54	6.0	5.0	7.4
68	5.0	3.7	5.5




#22
Acetophenone
Concen: 36.786 ng
RT: 8.805 min Scan# 972
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:105 Resp: 381834

Ion	Ratio	Lower	Upper
105	100		
71	2.0	0.9	1.3#
51	22.7	18.2	27.2
120	22.9	17.0	25.6









#23
Nitrobenzene-d5
Concen: 84.252 ng
RT: 9.140 min Scan# 1029
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 589527
Ion Ratio Lower Upper
82 100
128 44.4 30.5 45.7
54 41.7 33.9 50.9



#24
Nitrobenzene
Concen: 39.747 ng
RT: 9.187 min Scan# 1037
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion: 77 Resp: 308927
Ion Ratio Lower Upper
77 100
123 47.8 31.7 47.5#
65 13.6 11.5 17.3









#25
Isophorone
Concen: 36.830 ng
RT: 9.710 min Scan# 1126
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 572139
Ion Ratio Lower Upper
82 100
95 7.7 7.5 11.3
138 18.4 14.6 22.0



#26
2-Nitrophenol
Concen: 50.675 ng
RT: 9.899 min Scan# 1158
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion: 139 Resp: 151340
Ion Ratio Lower Upper
139 100
109 32.2 40.7 61.1#
65 46.4 47.8 71.6#









#27
2,4-Dimethylphenol
Concen: 38.074 ng
RT: 9.957 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:122 Resp: 227025
Ion Ratio Lower Upper
122 100
107 118.0 93.8 140.8
121 60.7 45.6 68.4



#28
bis(2-Chloroethoxy)methane
Concen: 35.523 ng
RT: 10.193 min Scan# 1208
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion: 93 Resp: 296998
Ion Ratio Lower Upper
93 100
95 31.6 24.9 37.3
123 12.0 9.5 14.3









#29
2,4-Dichlorophenol
Concen: 42.055 ng
RT: 10.434 min Scan# 1249
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 269701




Ion	Ratio	Lower	Upper
162	100		
164	66.5	44.1	84.1
98	34.2	12.6	52.6




#30
1,2,4-Trichlorobenzene
Concen: 37.952 ng
RT: 10.652 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:180 Resp: 288279

Ion	Ratio	Lower	Upper
180	100		
182	96.2	77.2	115.8
145	29.3	21.7	32.5









#31
Naphthalene
Concen: 36.444 ng
RT: 10.834 min Scan# 1317
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	128	Resp:	844548
Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.0	12.0
127	13.2	11.1	16.7



#32
Benzoic acid
Concen: 41.266 ng
RT: 10.110 min Scan# 1194
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:	122	Resp:	133042
Ion	Ratio	Lower	Upper
122	100		
105	119.3	105.1	145.1
77	90.3	69.6	109.6









#33
4-Chloroaniline
Concen: 37.083 ng
RT: 10.946 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	127	Resp:	357827
Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.5	38.3
65	28.3	30.0	45.0#
92	18.3	20.2	30.2#



#34
Hexachlorobutadiene
Concen: 40.267 ng
RT: 11.122 min Scan# 1366
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:	225	Resp:	191039
Ion	Ratio	Lower	Upper
225	100		
223	64.2	51.1	76.7
227	66.3	50.2	75.2









#35
Caprolactam
Concen: 51.429 ng
RT: 11.734 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:113 Resp: 90322
Ion Ratio Lower Upper
113 100
55 122.5 72.0 112.0#
56 98.5 65.1 105.1



#36
4-Chloro-3-methylphenol
Concen: 43.868 ng
RT: 12.069 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:107 Resp: 296125
Ion Ratio Lower Upper
107 100
144 27.7 20.1 30.1
142 86.0 62.2 93.4









#37
2-Methylnaphthalene
Concen: 41.423 ng
RT: 12.440 min Scan# 1590
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 680159
Ion Ratio Lower Upper
142 100
141 89.5 74.7 112.1
115 35.8 43.8 65.8#



#38
1-Methylnaphthalene
Concen: 40.920 ng
RT: 12.657 min Scan# 1627
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:142 Resp: 637484
Ion Ratio Lower Upper
142 100
141 90.9 77.0 115.6
116 3.8 5.0 7.6#









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.604 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 283253
Ion Ratio Lower Upper
164 100
162 101.6 73.5 110.3
160 43.8 40.6 61.0



#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.279 ng
RT: 12.810 min Scan# 1653
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:216 Resp: 380593
Ion Ratio Lower Upper
216 100
214 78.1 60.6 91.0
179 19.5 15.3 22.9
108 16.0 13.8 20.6









#41
Hexachlorocyclopentadiene
Concen: 34.558 ng
RT: 12.787 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:237 Resp: 181792
Ion Ratio Lower Upper
237 100
235 62.1 40.9 80.9
272 14.5 0.0 34.6



#42
2,4,6-Tribromophenol
Concen: 121.625 ng
RT: 16.092 min Scan# 2211
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:330 Resp: 441757
Ion Ratio Lower Upper
330 100
332 96.5 76.3 114.5
141 23.7 22.8 34.2









#43
2,4,6-Trichlorophenol
Concen: 45.725 ng
RT: 13.045 min Scan# 1693
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 262050
Ion Ratio Lower Upper
196 100
198 96.0 81.4 122.2
200 30.9 28.2 42.2



#44
2,4,5-Trichlorophenol
Concen: 45.800 ng
RT: 13.122 min Scan# 1706
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:196 Resp: 283843
Ion Ratio Lower Upper
196 100
198 98.4 71.0 106.4
97 41.7 38.5 57.7
132 24.1 15.6 23.4#









#45
2-Fluorobiphenyl
Concen: 74.851 ng
RT: 13.234 min Scan# 1725
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:172 Resp: 1602040
Ion Ratio Lower Upper
172 100
171 34.9 28.1 42.1
170 23.4 19.6 29.4



#46
1,1'-Biphenyl
Concen: 36.057 ng
RT: 13.440 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:154 Resp: 829314
Ion Ratio Lower Upper
154 100
153 39.7 18.9 58.9
76 11.3 0.0 32.1









#47
2-Chloronaphthalene
Concen: 36.736 ng
RT: 13.487 min Scan# 1768
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 686789
Ion Ratio Lower Upper
162 100
127 38.2 31.4 47.0
164 33.0 26.6 39.8



#48
2-Nitroaniline
Concen: 40.265 ng
RT: 13.687 min Scan# 1802
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion: 65 Resp: 175156
Ion Ratio Lower Upper
65 100
92 75.1 48.8 73.2#
138 123.7 69.5 104.3#









#49
Acenaphthylene
Concen: 36.435 ng
RT: 14.328 min Scan# 1911
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 1061730
Ion Ratio Lower Upper
152 100
151 20.2 15.8 23.6
153 13.1 10.5 15.7



#50
Dimethylphthalate
Concen: 38.000 ng
RT: 14.057 min Scan# 1865
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:163 Resp: 856136
Ion Ratio Lower Upper
163 100
194 4.3 3.9 5.9
164 10.2 8.1 12.1









#51
2,6-Dinitrotoluene
Concen: 40.245 ng
RT: 14.181 min Scan# 1886
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 195719
Ion Ratio Lower Upper
165 100
63 47.4 67.4 101.0#
89 53.2 73.8 110.6#



#52
Acenaphthene
Concen: 33.775 ng
RT: 14.669 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:154 Resp: 639811
Ion Ratio Lower Upper
154 100
153 112.9 90.9 136.3
152 54.4 48.5 72.7









#53
3-Nitroaniline
Concen: 40.450 ng
RT: 14.504 min Scan# 1941
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 196035
Ion Ratio Lower Upper
138 100
108 10.7 10.2 15.4
92 109.3 108.1 162.1



#54
2,4-Dinitrophenol
Concen: 48.663 ng
RT: 14.722 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:184 Resp: 94618
Ion Ratio Lower Upper
184 100
63 54.4 51.2 76.8
154 62.2 55.8 83.8









#55
Dibenzenofuran
Concen: 36.788 ng
RT: 14.998 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 1068141
Ion Ratio Lower Upper
168 100
139 37.7 31.4 47.0
169 13.5 11.4 17.0



#56
4-Nitrophenol
Concen: 39.507 ng
RT: 14.816 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:139 Resp: 157673
Ion Ratio Lower Upper
139 100
109 83.2 122.8 162.8#
65 87.0 110.0 150.0#









#57
2,4-Dinitrotoluene
Concen: 44.063 ng
RT: 14.969 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 268885
Ion Ratio Lower Upper
165 100
63 34.9 56.7 85.1#
89 63.7 85.8 128.6#
182 2.8 3.0 4.6#



#58
Fluorene
Concen: 36.210 ng
RT: 15.651 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:166 Resp: 840583
Ion Ratio Lower Upper
166 100
165 97.8 78.8 118.2
167 13.1 10.9 16.3









#59
2,3,4,6-Tetrachlorophenol
Concen: 47.662 ng
RT: 15.228 min Scan# 2064
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	232	Resp:	255565
Ion	Ratio	Lower	Upper
232	100		
131	36.6	26.5	39.7
130	2.2	2.1	3.1
166	27.4	19.4	29.0



#60
Diethylphthalate
Concen: 37.063 ng
RT: 15.416 min Scan# 2096
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:	149	Resp:	837819
Ion	Ratio	Lower	Upper
149	100		
177	23.5	19.2	28.8
150	12.3	9.9	14.9









#61
4-Chlorophenyl-phenylether
Concen: 38.693 ng
RT: 15.639 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 449079
Ion Ratio Lower Upper
204 100
206 32.3 26.0 39.0
141 53.1 44.2 66.2



#62
4-Nitroaniline
Concen: 40.920 ng
RT: 15.663 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:138 Resp: 208119
Ion Ratio Lower Upper
138 100
92 48.1 50.9 90.9#
108 88.0 154.7 194.7#









#63
Azobenzene
Concen: 31.225 ng
RT: 15.934 min Scan# 2184
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 739161




Ion	Ratio	Lower	Upper
77	100		
182	30.9	5.5	45.5
105	20.2	0.0	36.5
51	26.5	10.3	50.3




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.345 min Scan# 2424
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion: 188 Resp: 611793

Ion	Ratio	Lower	Upper
188	100		
94	7.3	5.0	7.6
80	8.0	5.9	8.9









#65
4,6-Dinitro-2-methylphenol
Concen: 51.141 ng
RT: 15.734 min Scan# 2150
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 162923
Ion Ratio Lower Upper
198 100
51 27.9 10.3 50.3
105 34.9 13.6 53.6



#66
n-Nitrosodiphenylamine
Concen: 36.177 ng
RT: 15.851 min Scan# 2170
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:169 Resp: 754179
Ion Ratio Lower Upper
169 100
168 64.9 55.2 82.8
167 35.3 36.6 54.8#









#67
4-Bromophenyl-phenylether
Concen: 42.705 ng
RT: 16.534 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 314742
Ion Ratio Lower Upper
248 100
250 98.7 77.5 116.3
141 58.8 50.1 75.1



#68
Hexachlorobenzene
Concen: 44.171 ng
RT: 16.657 min Scan# 2307
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:284 Resp: 387622
Ion Ratio Lower Upper
284 100
142 25.3 20.8 31.2
249 30.0 23.7 35.5









#69
Atrazine
Concen: 42.982 ng
RT: 16.804 min Scan# 2332
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:200 Resp: 256067
Ion Ratio Lower Upper
200 100
173 24.0 0.7 40.7
215 49.3 29.0 69.0



#70
Pentachlorophenol
Concen: 42.342 ng
RT: 16.998 min Scan# 2365
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:266 Resp: 186360
Ion Ratio Lower Upper
266 100
268 68.0 49.4 74.0
264 70.4 53.7 80.5









#71
Phenanthrene
Concen: 35.805 ng
RT: 17.392 min Scan# 2432
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 1364695
Ion Ratio Lower Upper
178 100
176 19.1 15.4 23.2
179 15.3 12.3 18.5



#72
Anthracene
Concen: 36.257 ng
RT: 17.481 min Scan# 2447
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:178 Resp: 1364490
Ion Ratio Lower Upper
178 100
176 18.8 15.2 22.8
179 15.9 12.6 19.0









#73
 Carbazole
 Concen: 35.055 ng
 RT: 17.745 min Scan# 2492
 Delta R.T. 0.000 min
 Lab File: BP005736.D
 Acq: 07 Jun 2021 11:03




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:167 Resp: 1298392
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.0
 139 13.1 10.4 15.6



#74
 Di-n-butylphthalate
 Concen: 36.906 ng
 RT: 18.292 min Scan# 2585
 Delta R.T. 0.000 min
 Lab File: BP005736.D
 Acq: 07 Jun 2021 11:03

Tgt Ion:149 Resp: 1463846
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.6 11.4
 104 4.9 8.9 13.3#









#75
Fluoranthene
Concen: 37.888 ng
RT: 19.345 min Scan# 2764
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:202 Resp: 1733518




Ion	Ratio	Lower	Upper
202	100		
101	8.7	0.0	26.1
203	17.7	0.0	37.2




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.416 min Scan# 3116
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:240 Resp: 725668

Ion	Ratio	Lower	Upper
240	100		
120	8.1	4.2	6.2#
236	25.4	20.2	30.2









#77
Benzidine
Concen: 49.275 ng
RT: 19.516 min Scan# 2793
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 679073
Ion Ratio Lower Upper
184 100
185 14.5 11.1 16.7
183 12.0 9.8 14.8



#78
Pyrene
Concen: 35.087 ng
RT: 19.698 min Scan# 2824
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:202 Resp: 1796625
Ion Ratio Lower Upper
202 100
200 21.1 16.9 25.3
203 17.8 13.8 20.6









#79
Terphenyl-d14
Concen: 74.721 ng
RT: 19.886 min Scan# 2856
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:244 Resp: 2889667
Ion Ratio Lower Upper
244 100
212 7.1 5.8 8.6
122 8.4 5.2 7.8#



#80
Butylbenzylphthalate
Concen: 35.799 ng
RT: 20.557 min Scan# 2970
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:149 Resp: 717666
Ion Ratio Lower Upper
149 100
91 65.1 60.2 90.4
206 23.5 23.9 35.9#









#81
Benzo(a)anthracene
Concen: 36.453 ng
RT: 21.398 min Scan# 3113
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 1941353
Ion Ratio Lower Upper
228 100
226 27.1 21.2 31.8
229 19.9 15.6 23.4



#82
3,3'-Dichlorobenzidine
Concen: 42.177 ng
RT: 21.327 min Scan# 3101
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:252 Resp: 725526
Ion Ratio Lower Upper
252 100
254 65.6 49.9 74.9
126 8.1 5.5 8.3









#83
Chrysene
Concen: 36.042 ng
RT: 21.451 min Scan# 3122
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 1858689
Ion Ratio Lower Upper
228 100
226 29.9 24.8 37.2
229 19.7 15.4 23.0



#84
Bis(2-ethylhexyl)phthalate
Concen: 35.410 ng
RT: 21.316 min Scan# 3099
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:149 Resp: 1024263
Ion Ratio Lower Upper
149 100
167 29.2 23.8 35.8
279 6.7 5.7 8.5









#85
Di-n-octyl phthalate
Concen: 38.723 ng
RT: 22.245 min Scan# 3257
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:149 Resp: 1882502




Ion	Ratio	Lower	Upper
149	100		
167	1.7	1.5	2.3
43	6.8	5.1	7.7




#86
Perylene-d12
Concen: 20.000 ng
RT: 23.851 min Scan# 3530
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:264 Resp: 877209

Ion	Ratio	Lower	Upper
264	100		
260	24.1	20.6	30.8
265	21.9	19.0	28.6









#87
Indeno(1,2,3-cd)pyrene
Concen: 36.877 ng
RT: 26.427 min Scan# 3968
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:276 Resp: 2752216
Ion Ratio Lower Upper
276 100
138 17.3 8.4 12.6#
277 25.7 20.6 31.0



#88
Benzo(b)fluoranthene
Concen: 35.643 ng
RT: 23.110 min Scan# 3404
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:252 Resp: 2268533
Ion Ratio Lower Upper
252 100
253 21.8 18.2 27.2
125 6.8 3.4 5.0#









#89
Benzo(k)fluoranthene
Concen: 33.976 ng
RT: 23.157 min Scan# 3412
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 2049115
Ion Ratio Lower Upper
252 100
253 22.0 17.2 25.8
125 7.3 3.8 5.6#



#90
Benzo(a)pyrene
Concen: 35.622 ng
RT: 23.745 min Scan# 3512
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:252 Resp: 2062563
Ion Ratio Lower Upper
252 100
253 22.2 17.1 25.7
125 8.3 3.8 5.6#









#91
Dibenzo(a,h)anthracene
Concen: 36.751 ng
RT: 26.439 min Scan# 3970
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 2380332
Ion Ratio Lower Upper
278 100
139 11.0 5.0 7.6#
279 24.2 18.6 28.0



#92
Benzo(g,h,i)perylene
Concen: 37.077 ng
RT: 27.215 min Scan# 4102
Delta R.T. 0.000 min
Lab File: BP005736.D
Acq: 07 Jun 2021 11:03

Tgt Ion:276 Resp: 2316250
Ion Ratio Lower Upper
276 100
277 24.1 19.8 29.8
138 15.1 7.7 11.5#

